

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** POWE02
Lab Code: ACE **SDG NO.:** Q3341
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 09/18/2025 09/18/2025

Client Sample No.: CCAL01 **Date Analyzed:** 10/15/2025
Lab Sample No.: PSTDCCC050 **Data File :** PD090596.D **Time Analyzed:** 09:46

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
4,4'-DDD	6.700	6.601	6.801	55.940	50.000	11.9
4,4'-DDE	6.190	6.092	6.292	56.340	50.000	12.7
4,4'-DDT	7.016	6.917	7.117	48.920	50.000	-2.2
Aldrin	5.264	5.166	5.366	56.580	50.000	13.2
Decachlorobiphenyl	9.065	8.968	9.168	51.590	50.000	3.2
Dieldrin	6.341	6.243	6.443	55.600	50.000	11.2
Tetrachloro-m-xylene	3.545	3.446	3.646	53.290	50.000	6.6

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** POWE02
Lab Code: ACE **SDG NO.:** Q3341
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 09/18/2025 09/18/2025

Client Sample No.: CCAL01 **Date Analyzed:** 10/15/2025
Lab Sample No.: PSTDCCC050 **Data File :** PD090596.D **Time Analyzed:** 09:46

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	5.920	5.822	6.022	57.880	50.000	15.8
4,4'-DDE	5.366	5.268	5.468	59.270	50.000	18.5
4,4'-DDT	6.173	6.076	6.276	51.930	50.000	3.9
Aldrin	4.360	4.262	4.462	59.500	50.000	19.0
Decachlorobiphenyl	8.061	7.964	8.164	51.230	50.000	2.5
Dieldrin	5.503	5.406	5.606	58.370	50.000	16.7
Tetrachloro-m-xylene	2.875	2.776	2.976	57.100	50.000	14.2