

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** POWE02
Lab Code: ACE **SDG NO.:** Q3341
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 09/18/2025 09/18/2025

Client Sample No.: CCAL02 **Date Analyzed:** 10/15/2025
Lab Sample No.: PSTDCCC050 **Data File :** PD090608.D **Time Analyzed:** 16:33

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
4,4'-DDD	6.701	6.601	6.801	55.820	50.000	11.6
4,4'-DDE	6.191	6.092	6.292	56.030	50.000	12.1
4,4'-DDT	7.017	6.917	7.117	50.140	50.000	0.3
Aldrin	5.265	5.166	5.366	56.740	50.000	13.5
Decachlorobiphenyl	9.067	8.968	9.168	53.480	50.000	7.0
Dieldrin	6.342	6.243	6.443	54.920	50.000	9.8
Tetrachloro-m-xylene	3.546	3.446	3.646	41.970	50.000	-16.1

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** POWE02
Lab Code: ACE **SDG NO.:** Q3341
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 09/18/2025 09/18/2025

Client Sample No.: CCAL02 **Date Analyzed:** 10/15/2025
Lab Sample No.: PSTDCCC050 **Data File :** PD090608.D **Time Analyzed:** 16:33

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	5.921	5.822	6.022	54.380	50.000	8.8
4,4'-DDE	5.366	5.268	5.468	56.790	50.000	13.6
4,4'-DDT	6.174	6.076	6.276	49.510	50.000	-1.0
Aldrin	4.361	4.262	4.462	57.840	50.000	15.7
Decachlorobiphenyl	8.062	7.964	8.164	46.740	50.000	-6.5
Dieldrin	5.504	5.406	5.606	55.940	50.000	11.9
Tetrachloro-m-xylene	2.875	2.776	2.976	44.360	50.000	-11.3