

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** POWE02
Lab Code: ACE **SDG NO.:** Q3341
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 09/18/2025 09/18/2025

Client Sample No.: CCAL04 **Date Analyzed:** 10/15/2025
Lab Sample No.: PSTDCCC050 **Data File :** PD090622.D **Time Analyzed:** 20:25

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	6.701	6.601	6.801	58.300	50.000	16.6
4,4'-DDE	6.190	6.092	6.292	57.160	50.000	14.3
4,4'-DDT	7.016	6.917	7.117	51.290	50.000	2.6
Aldrin	5.264	5.166	5.366	57.760	50.000	15.5
Decachlorobiphenyl	9.066	8.968	9.168	55.170	50.000	10.3
Dieldrin	6.341	6.243	6.443	56.490	50.000	13.0
Tetrachloro-m-xylene	3.545	3.446	3.646	42.480	50.000	-15.0

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** POWE02
Lab Code: ACE **SDG NO.:** Q3341
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 09/18/2025 09/18/2025

Client Sample No.: CCAL04 **Date Analyzed:** 10/15/2025
Lab Sample No.: PSTDCCC050 **Data File :** PD090622.D **Time Analyzed:** 20:25

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	5.920	5.822	6.022	55.720	50.000	11.4
4,4'-DDE	5.366	5.268	5.468	57.640	50.000	15.3
4,4'-DDT	6.174	6.076	6.276	51.020	50.000	2.0
Aldrin	4.360	4.262	4.462	58.410	50.000	16.8
Decachlorobiphenyl	8.062	7.964	8.164	49.650	50.000	-0.7
Dieldrin	5.504	5.406	5.606	56.850	50.000	13.7
Tetrachloro-m-xylene	2.875	2.776	2.976	44.510	50.000	-11.0