

284 Sheffield Street, Mountainside, New Jersey 07092, Phone: 908 789 8900,

Test:

PESTICIDE Group1

Fax: 908 789 8922

Report of Analysis

Client: Kleinfelder
Project: Webster School
Client Sample ID: C-1MSD
Lab Sample ID: Q3341-01MSD

Analytical Method: 8081B

Sample Wt/Vol: 30.07 g Prep Method: SW3541B

 or
 Date Collected: 10/13/25

 chool
 Date Received: 10/14/25

 SDG No.:
 Q3341

 MSD
 Matrix: SOIL

 % Solid: 78.9

g Final Vol: 10000 uL 541B Prep Date 10/15/25

CAS Number	r Parameter	Conc.	Qua. DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS								
309-00-2	Aldrin	20.6	1	0.15	2.10	ug/kg	10/15/25 19:17	PB170103
60-57-1	Dieldrin	20.3	1	0.18	2.10	ug/kg	10/15/25 19:17	PB170103
72-55-9	4,4-DDE	20.1	1	0.18	2.10	ug/kg	10/15/25 19:17	PB170103
72-54-8	4,4-DDD	20.0	1	0.19	2.10	ug/kg	10/15/25 19:17	PB170103
50-29-3	4,4-DDT	17.7	1	0.18	2.10	ug/kg	10/15/25 19:17	PB170103
SURROGATES								
2051-24-3	Decachlorobiphenyl	15.3		20 - 144	77%	SPK: 20		
877-09-8	Tetrachloro-m-xylene	16.6		19 - 148	83%	SPK: 20		

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit