

**Laboratory Control Sample/Laboratory Control Sample Duplicate Summary**

**SW-846**

|                 |                    |                           |                   |
|-----------------|--------------------|---------------------------|-------------------|
| <b>SDG No.:</b> | <u>Q3341</u>       | <b>Analytical Method:</b> | <u>8081B</u>      |
| <b>Client:</b>  | <u>Kleinfelder</u> | <b>Datafile :</b>         | <u>PD090599.D</u> |

| Lab Sample ID            | Parameter | Spike | Result | Units | Rec | RPD | Qual | RPD  |     | Limits |  | RPD |
|--------------------------|-----------|-------|--------|-------|-----|-----|------|------|-----|--------|--|-----|
|                          |           |       |        |       |     |     |      | Qual | Low | High   |  |     |
| PB170103BS<br>(Column 1) | Aldrin    | 16.66 | 17.6   | ug/kg | 106 |     |      |      | 82  | 124    |  |     |
|                          | Dieldrin  | 16.66 | 17.4   | ug/kg | 104 |     |      |      | 85  | 121    |  |     |
|                          | 4,4'-DDE  | 16.66 | 17.5   | ug/kg | 105 |     |      |      | 81  | 123    |  |     |
|                          | 4,4'-DDD  | 16.66 | 17.6   | ug/kg | 106 |     |      |      | 80  | 131    |  |     |
|                          | 4,4'-DDT  | 16.66 | 14.8   | ug/kg | 89  |     |      |      | 70  | 129    |  |     |
| PB170103BS<br>(Column 2) | Aldrin    | 16.66 | 18.4   | ug/kg | 110 |     |      |      | 82  | 124    |  |     |
|                          | Dieldrin  | 16.66 | 18.3   | ug/kg | 110 |     |      |      | 85  | 121    |  |     |
|                          | 4,4'-DDE  | 16.66 | 18.4   | ug/kg | 110 |     |      |      | 81  | 123    |  |     |
|                          | 4,4'-DDD  | 16.66 | 18.1   | ug/kg | 109 |     |      |      | 80  | 131    |  |     |
|                          | 4,4'-DDT  | 16.66 | 15.5   | ug/kg | 93  |     |      |      | 70  | 129    |  |     |