

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q3341

Client ID : SSTDCCC040

Date Analyzed: 10/15/2025

Lab File ID: BF143951.D

Time Analyzed: 14:33

Instrument ID: BNA_F

GC Column: DB-UI ID: 0.18 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	133330	6.881	482185	8.17	244444	9.92
UPPER LIMIT	266660	7.381	964370	8.669	488888	10.422
LOWER LIMIT	66665	6.381	241093	7.669	122222	9.422
EPA SAMPLE NO.						
01 LAYDOWN-YARD-1MS	181566	6.88	654257	8.17	308543	9.92
02 LAYDOWN-YARD-1MSD	162049	6.88	562605	8.17	257454	9.93
03 PB170106BL	127400	6.88	500628	8.16	280346	9.92
04 PB170106BS	118178	6.88	460406	8.16	251791	9.92
05 C-1	106992	6.88	421767	8.16	230177	9.92
06 C-2	102022	6.88	410764	8.16	225605	9.92
07 C-3	109785	6.88	433602	8.16	242042	9.92

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.