

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name:	<u>Alliance</u>	Contract:	<u>POWE02</u>
Lab Code:	<u>ACE</u>	SDG No.:	<u>Q3341</u>
Instrument ID:	<u>MSVOA_Y</u>	Calibration Date/Time:	<u>10/14/2025 11:25</u>
Lab File ID:	<u>VY023456.D</u>	Init. Calib. Date(s):	<u>10/07/2025 10/07/2025</u>
Heated Purge: (Y/N)	<u>Y</u>	Init. Calib. Time(s):	<u>09:17 12:32</u>
GC Column:	<u>RXI-624</u>	ID:	<u>0.25</u> (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
cis-1,2-Dichloroethene	0.567	0.543		-4.23	20
1,1,1-Trichloroethane	0.795	0.773		-2.77	20
Benzene	1.288	1.282		-0.47	20
Trichloroethene	0.377	0.371		-1.59	20
Toluene	0.836	0.857		2.51	20
Ethyl Benzene	1.760	1.806		2.61	20
m/p-Xylenes	0.706	0.733		3.82	20
o-Xylene	0.662	0.677		2.27	20
Isopropylbenzene	3.295	3.418		3.73	20
1,2-Dichloroethane-d4	0.418	0.388		-7.18	20
Dibromofluoromethane	0.307	0.308		0.33	20
Toluene-d8	1.144	1.166		1.92	20
4-Bromofluorobenzene	0.378	0.371		-1.85	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.