



Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY101425\
Data File : VY023455.D
Acq On : 14 Oct 2025 08:36
Operator : SY/MD
Sample : BFB
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100725S.M
Title : SW846 8260
Last Update : Wed Oct 08 05:12:50 2025



AutoFind: Scans 1752, 1753, 1754; Background Corrected with Scan 1744

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	17.2	24973	PASS
75	95	30	60	50.6	73587	PASS
95	95	100	100	100.0	145429	PASS
96	95	5	9	6.5	9381	PASS
173	174	0.00	2	0.9	1248	PASS
174	95	50	100	95.6	139029	PASS
175	174	5	9	7.5	10428	PASS
176	174	95	101	97.8	136035	PASS
177	176	5	9	6.5	8846	PASS