

Surrogate Summary

SDG No.: Q3341

Client: Kleinfelder

Analytical Method: SW8260D

Lab Sample ID	Client ID	Parameter	Spike	Result	Recovery (%)	Qual	Limits (%)	
							Low	High
Q3341-01	C-1	1,2-Dichloroethane-d4	50	63.4	127		63	155
		Dibromofluoromethane	50	55.4	111		70	134
		Toluene-d8	50	53.0	106		74	123
		4-Bromofluorobenzene	50	54.8	109		17	146
Q3341-02	C-2	1,2-Dichloroethane-d4	50	66.8	134		63	155
		Dibromofluoromethane	50	56.4	113		70	134
		Toluene-d8	50	53.4	107		74	123
		4-Bromofluorobenzene	50	55.8	112		17	146
Q3341-03	C-3	1,2-Dichloroethane-d4	50	66.5	133		63	155
		Dibromofluoromethane	50	56.6	113		70	134
		Toluene-d8	50	53.6	107		74	123
		4-Bromofluorobenzene	50	55.6	111		17	146
VY1014SBL01	VY1014SBL01	1,2-Dichloroethane-d4	50	59.3	119		63	155
		Dibromofluoromethane	50	54.2	108		70	134
		Toluene-d8	50	51.9	104		74	123
		4-Bromofluorobenzene	50	52.1	104		17	146
VY1014SBS01	VY1014SBS01	1,2-Dichloroethane-d4	50	46.2	92		63	155
		Dibromofluoromethane	50	48.9	98		70	134
		Toluene-d8	50	49.3	99		74	123
		4-Bromofluorobenzene	50	46.5	93		17	146
VY1014SBSD01	VY1014SBSD01	1,2-Dichloroethane-d4	50	46.2	92		63	155
		Dibromofluoromethane	50	49.1	98		70	134
		Toluene-d8	50	49.3	99		74	123
		4-Bromofluorobenzene	50	46.7	93		17	146