

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.:	<u>Q3341</u>	Analytical Method:	<u>SW8260D</u>
Client:	<u>Kleinfelder</u>	Datafile :	<u>VY023458.D</u>

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VY1014SBS01	cis-1,2-Dichloroethene	20	19.9	ug/Kg	100			82	123	
	1,1,1-Trichloroethane	20	20.6	ug/Kg	103			80	126	
	Benzene	20	20.7	ug/Kg	104			84	121	
	Trichloroethene	20	20.7	ug/Kg	104			83	122	
	Toluene	20	20.6	ug/Kg	103			83	122	
	Ethyl Benzene	20	19.9	ug/Kg	100			82	124	
	m/p-Xylenes	40	40.8	ug/Kg	102			83	124	
	o-Xylene	20	19.5	ug/Kg	98			83	123	
	Isopropylbenzene	20	19.9	ug/Kg	100			82	124	