

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.:	<u>Q3341</u>	Analytical Method:	<u>SW8260D</u>
Client:	<u>Kleinfelder</u>	Datafile :	<u>VY023459.D</u>

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VY1014SBSD01	cis-1,2-Dichloroethene	20	20.3	ug/Kg	102	2		82	123	20
	1,1,1-Trichloroethane	20	20.9	ug/Kg	104	1		80	126	20
	Benzene	20	20.8	ug/Kg	104	0		84	121	20
	Trichloroethene	20	20.3	ug/Kg	102	2		83	122	20
	Toluene	20	20.9	ug/Kg	104	1		83	122	20
	Ethyl Benzene	20	20.1	ug/Kg	101	1		82	124	20
	m/p-Xylenes	40	41.0	ug/Kg	103	1		83	124	20
	o-Xylene	20	20.0	ug/Kg	100	2		83	123	20
	Isopropylbenzene	20	20.2	ug/Kg	101	1		82	124	20