

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: POWE02
 Lab Code: ACE SDG NO.: Q3341
 Lab File ID: VY023456.D Date Analyzed: 10/14/2025
 Instrument ID: MSVOA_Y Time Analyzed: 11:25
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	571722	7.71	808604	8.62	715207	11.42
UPPER LIMIT	1143440	8.213	1617210	9.116	1430410	11.92
LOWER LIMIT	285861	7.213	404302	8.116	357604	10.92
EPA SAMPLE NO.						
C-1	550557	7.71	1073471	8.62	1115279	11.42
C-2	494390	7.71	985514	8.62	1038095	11.42
C-3	521903	7.71	1026356	8.62	1081652	11.42
VY1014SBL01	564646	7.71	1109577	8.62	1120800	11.42
VY1014SBS01	531866	7.71	775568	8.62	681796	11.42
VY1014SBSD01	517858	7.71	759118	8.62	662146	11.42

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.