

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name:	<u>Alliance</u>	Contract:	<u>POWE02</u>
Lab Code:	<u>ACE</u>	SDG NO.:	<u>Q3341</u>
Lab File ID:	<u>VY023456.D</u>	Date Analyzed:	<u>10/14/2025</u>
Instrument ID:	<u>MSVOA_Y</u>	Time Analyzed:	<u>11:25</u>
GC Column:	<u>RXI-624</u>	ID:	<u>0.25</u> (mm)
		Heated Purge:	(Y/N) <u>Y</u>

	IS4 AREA #	RT #				
12 HOUR STD	372874	13.346				
UPPER LIMIT	745748	13.846				
LOWER LIMIT	186437	12.846				
EPA SAMPLE NO.						
C-1	474493	13.35				
C-2	461613	13.35				
C-3	485422	13.35				
VY1014SBL01	479501	13.35				
VY1014SBS01	357890	13.35				
VY1014SBSD01	344738	13.35				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.