



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Initial and Continuing Calibration Verification

Client: Mars Chocolate North America, LLC

SDG No.: Q3344

Project: W.T.P. Final Effluent 2025

RunNo.: LB137539

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV1 Ammonia as N	mg/L	1	1	100	90-110	10/15/2025
Sample ID: CCV1 Ammonia as N	mg/L	0.99	1	99	90-110	10/15/2025
Sample ID: CCV2 Ammonia as N	mg/L	1	1	100	90-110	10/15/2025
Sample ID: CCV3 Ammonia as N	mg/L	1	1	100	90-110	10/15/2025

Initial and Continuing Calibration Verification

Client: Mars Chocolate North America, LLC

SDG No.: Q3344

Project: W.T.P. Final Effluent 2025

RunNo.: LB137541

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV COD	mg/L	50.173	50	100	95-105	09/08/2025
Sample ID: CCV1 COD	mg/L	48.150	50	96	95-105	10/15/2025
Sample ID: CCV2 COD	mg/L	51.184	50	102	95-105	10/15/2025

Initial and Continuing Calibration Verification

Client: Mars Chocolate North America, LLC

SDG No.: Q3344

Project: W.T.P. Final Effluent 2025

RunNo.: LB137597

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV Phosphorus, Total	mg/L	0.507	0.50	101	90-110	10/21/2025
Sample ID: CCV1 Phosphorus, Total	mg/L	0.502	0.50	100	90-110	10/21/2025
Sample ID: CCV2 Phosphorus, Total	mg/L	0.501	0.50	100	90-110	10/21/2025
Sample ID: CCV3 Phosphorus, Total	mg/L	0.505	0.50	101	90-110	10/21/2025