

LB137565


7470A

INSTRUMENT ID: CV1

Linear

μ Abs.:

4410E



A= 0.0000e+000

B= 2.2600e-004

C= 3.5038e-002

Rho= 0.9999218

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.028	0.028	-33	0.000	-33	0			
0.2	0.200	0.205	0.005	753	0.0 %	753	0			
2.5	2.500	2.503	0.003	10918	0.0 %	10918	0			
5.0	5.000	4.906	-0.094	21551	0.0 %	21551				
7.5	7.500	7.555	0.055	33275	0.0 %	33275				
10.0	10.000	10.004	0.004	44108	0.0 %	44108				

LB137565 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 16 Oct 2025 14:20:56

Type	Sample ID	Extended ID	Conc.	µ Abs	Units	Type	Std Conc	Method	Date
S	0.0 - 1	0.0	-	-33	PPB	Std	0.0000	7470A	16 Oct 2025 14:21:14
S	0.2 - 1	0.2	-	753	PPB	Std	0.2000	7470A	16 Oct 2025 14:23:31
S	2.5 - 1	2.5	-	10918	PPB	Std	2.5000	7470A	16 Oct 2025 14:25:48
S	5.0 - 1	5.0	-	21551	PPB	Std	5.0000	7470A	16 Oct 2025 14:28:05
S	7.5 - 1	7.5	-	33275	PPB	Std	7.5000	7470A	16 Oct 2025 14:30:24
S	10.0 - 1	10.0	-	44108	PPB	Std	10.0000	7470A	16 Oct 2025 14:32:43
U	ICV75 - 1	ICV75	3.6934	16187	PPB	SMPL	-7470A	7470A	16 Oct 2025 14:35:49
U	ICB75 - 1	ICB75	0.0000	-155	PPB	SMPL	-7470A	7470A	16 Oct 2025 14:38:04
U	CCV40 - 1	CCV40	4.5391	19929	PPB	SMPL	-7470A	7470A	16 Oct 2025 14:40:22
U	CCB40 - 1	CCB40	-0.0233	-258	PPB	SMPL	-7470A	7470A	16 Oct 2025 14:42:37
U	CRA - 1	CRA	0.1847	662	PPB	SMPL	-7470A	7470A	16 Oct 2025 14:44:56
U	HighStd - 1	HighStd	9.7733	43089	PPB	SMPL	-7470A	7470A	16 Oct 2025 14:47:17
U	ChkStd - 1	ChkStd	6.6850	29424	PPB	SMPL	-7470A	7470A	16 Oct 2025 14:49:33
U	PB170126BL - 1	PBW	-0.0291	-284	PPB	SMPL	-7470A	7470A	16 Oct 2025 14:52:00
U	PB170126BS - 1	LCSW	3.6611	16044	PPB	SMPL	-7470A	7470A	16 Oct 2025 15:00:24
U	Q3337-02 - 1	LAYDOWN-YARD-1	0.0029	-142	PPB	SMPL	-7470A	7470A	16 Oct 2025 15:02:39
U	Q3337-04 - 1	LAYDOWN-YARD-2	0.0418	30	PPB	SMPL	-7470A	7470A	16 Oct 2025 15:04:58
U	Q3342-02 - 1	CONC-ROLLOFF-WC-1	0.0445	42	PPB	SMPL	-7470A	7470A	16 Oct 2025 15:07:15
U	Q3343-04 - 1	SOIL-ROLLOFF-WC-1	0.0441	40	PPB	SMPL	-7470A	7470A	16 Oct 2025 15:09:32
U	Q3345-02 - 1	TANK-FARM-SOIL	0.0572	98	PPB	SMPL	-7470A	7470A	16 Oct 2025 15:11:48
U	CCV41 - 1	CCV41	4.5676	20055	PPB	SMPL	-7470A	7470A	16 Oct 2025 15:14:03
U	CCB41 - 1	CCB41	-0.0090	-195	PPB	SMPL	-7470A	7470A	16 Oct 2025 15:16:18
U	Q3345-04 - 1	2ND-STREET-SOIL	0.0210	-62	PPB	SMPL	-7470A	7470A	16 Oct 2025 15:18:37
U	Q3345-06 - 1	M-R-SOIL	0.0242	-48	PPB	SMPL	-7470A	7470A	16 Oct 2025 15:20:51
U	Q3346-01 - 1	Zeltmann Power Plant HSRG 7B	0.2986	1166	PPB	SMPL	-7470A	7470A	16 Oct 2025 15:23:07
U	Q3348-04 - 1	WCS-TP1	-0.0122	-209	PPB	SMPL	-7470A	7470A	16 Oct 2025 15:25:23
U	Q3348-08 - 1	WCS-TP2	0.0389	17	PPB	SMPL	-7470A	7470A	16 Oct 2025 15:27:40
U	Q3348-08DUP - 1	WCS-TP2DUP	0.0468	52	PPB	SMPL	-7470A	7470A	16 Oct 2025 15:29:56
U	Q3348-08MS - 1	WCS-TP2MS	3.6407	15954	PPB	SMPL	-7470A	7470A	16 Oct 2025 15:32:12
U	Q3348-08MSD - 1	WCS-TP2MSD	3.3196	14533	PPB	SMPL	-7470A	7470A	16 Oct 2025 15:34:29
U	Q3349-01 - 1	PURS Station	0.0027	-143	PPB	SMPL	-7470A	7470A	16 Oct 2025 15:36:48
U	Q3352-04 - 1	TP-3	0.0246	-46	PPB	SMPL	-7470A	7470A	16 Oct 2025 15:39:07
U	CCV42 - 1	CCV42	4.6885	20590	PPB	SMPL	-7470A	7470A	16 Oct 2025 15:44:14
U	CCB42 - 1	CCB42	-0.0298	-287	PPB	SMPL	-7470A	7470A	16 Oct 2025 15:46:29
U	Q3354-02 - 1	HT1403	0.0626	122	PPB	SMPL	-7470A	7470A	16 Oct 2025 15:48:50
U	Q3357-01 - 1	DRUM-1	0.0796	197	PPB	SMPL	-7470A	7470A	16 Oct 2025 15:51:06
U	Q3357-02 - 1	DRUM-2	0.0533	81	PPB	SMPL	-7470A	7470A	16 Oct 2025 15:53:21
U	Q3357-03 - 1	DRUM-3	0.0439	39	PPB	SMPL	-7470A	7470A	16 Oct 2025 15:55:37
U	Q3357-04 - 1	DRUM-4	0.0414	28	PPB	SMPL	-7470A	7470A	16 Oct 2025 15:57:53
U	PB170138BL - 1	PBW	0.0151	-88	PPB	SMPL	-7470A	7470A	16 Oct 2025 16:00:09
U	PB170138BS - 1	LCSW	3.2203	14094	PPB	SMPL	-7470A	7470A	16 Oct 2025 16:02:25
U	Q3356-01 - 1	AUD-25-0169-AUD-25-0170	3.5302	15465	PPB	SMPL	-7470A	7470A	16 Oct 2025 16:04:42
U	Q3356-01DUP - 1	AUD-25-0169-AUD-25-0170DUP	3.7607	16485	PPB	SMPL	-7470A	7470A	16 Oct 2025 16:07:10
U	Q3356-01MS - 1	AUD-25-0169-AUD-25-0170MS	5.9828	26317	PPB	SMPL	-7470A	7470A	16 Oct 2025 16:09:29
U	CCV43 - 1	CCV43	4.7355	20798	PPB	SMPL	-7470A	7470A	16 Oct 2025 16:17:41
U	CCB43 - 1	CCB43	-0.0212	-249	PPB	SMPL	-7470A	7470A	16 Oct 2025 16:19:56
U	Q3356-01MSD - 1	AUD-25-0169-AUD-25-0170MSD	6.0963	26819	PPB	SMPL	-7470A	7470A	16 Oct 2025 16:22:22
U	PB170081TB - 1		-0.0567	-406	PPB	SMPL	-7470A	7470A	16 Oct 2025 16:24:37
U	PB170116TB - 1		0.0400	22	PPB	SMPL	-7470A	7470A	16 Oct 2025 16:27:02
U	Q3348-08LX5 - 1		0.0384	15	PPB	SMPL	-7470A	7470A	16 Oct 2025 16:29:18
U	Q3348-08A - 1		3.5670	15628	PPB	SMPL	-7470A	7470A	16 Oct 2025 16:31:34
U	Q3356-01LX5 - 1		0.7399	3119	PPB	SMPL	-7470A	7470A	16 Oct 2025 16:33:51
U	Q3356-01A - 1		7.3045	32165	PPB	SMPL	-7470A	7470A	16 Oct 2025 16:36:17
U	CCV44 - 1	CCV44	4.6297	20330	PPB	SMPL	-7470A	7470A	16 Oct 2025 16:38:35
U	CCB44 - 1	CCB44	-0.0337	-304	PPB	SMPL	-7470A	7470A	16 Oct 2025 16:40:59