

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Core Environmental Consultants and Service

SDG No.: Q3346

Contract: CORE02

Lab Code: ACE

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB170112BS							
Arsenic	ug/L	4000	3540		88	80 - 120	P
Barium	ug/L	1000	892		89	80 - 120	P
Cadmium	ug/L	1000	832		83	80 - 120	P
Chromium	ug/L	2000	1780		89	80 - 120	P
Lead	ug/L	5000	4210		84	80 - 120	P
Selenium	ug/L	10000	9090		91	80 - 120	P
Silver	ug/L	380	337		89	80 - 120	P

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Core Environmental Consultants and Service

SDG No.: Q3346

Contract: CORE02

Lab Code: ACE

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB170126BS Mercury	ug/L	4.0	3.66		92	80 - 120	CV