

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: CORE02

Lab Code: ACE

SDG NO.: Q3346

Continuing Calib Date: 10/16/2025

Initial Calibration Date(s): 09/23/2025

09/23/2025

Continuing Calib Time: 10:29

Initial Calibration Time(s): 08:45

17:14

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW FROM TO		DIFF RT
Aroclor-1248-1 (1)	5.80	5.81	5.71	5.91	0.01
Aroclor-1248-2 (2)	6.07	6.08	5.98	6.18	0.01
Aroclor-1248-3 (3)	6.28	6.29	6.19	6.39	0.01
Aroclor-1248-4 (4)	6.68	6.68	6.58	6.78	0.00
Aroclor-1248-5 (5)	6.72	6.72	6.62	6.82	0.00
Tetrachloro-m-xylene	4.65	4.66	4.56	4.76	0.01
Decachlorobiphenyl	10.43	10.44	10.34	10.54	0.01

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: CORE02

Lab Code: ACE

SDG NO.: Q3346

Continuing Calib Date: 10/16/2025

Initial Calibration Date(s): 09/23/2025

09/23/2025

Continuing Calib Time: 10:29

Initial Calibration Time(s): 08:45

17:14

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Aroclor-1248-1 (1)	4.88	4.89	4.79	4.99	0.01
Aroclor-1248-2 (2)	5.11	5.11	5.01	5.21	0.00
Aroclor-1248-3 (3)	5.15	5.15	5.05	5.25	0.01
Aroclor-1248-4 (4)	5.32	5.33	5.23	5.43	0.01
Aroclor-1248-5 (5)	5.71	5.71	5.61	5.81	0.00
Tetrachloro-m-xylene	3.79	3.80	3.70	3.90	0.01
Decachlorobiphenyl	8.79	8.80	8.70	8.90	0.01