

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** CORE02
Lab Code: ACE **SDG NO.:** Q3346
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 09/23/2025 09/23/2025

Client Sample No.: CCAL01 **Date Analyzed:** 10/16/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP075814.D **Time Analyzed:** 09:56

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.805	5.708	5.908	524.060	500.000	4.8
Aroclor-1016-2	5.826	5.729	5.929	548.640	500.000	9.7
Aroclor-1016-3	5.890	5.792	5.992	517.900	500.000	3.6
Aroclor-1016-4	5.987	5.889	6.089	550.940	500.000	10.2
Aroclor-1016-5	6.279	6.181	6.381	564.660	500.000	12.9
Aroclor-1260-1	7.396	7.299	7.499	531.080	500.000	6.2
Aroclor-1260-2	7.649	7.551	7.751	521.710	500.000	4.3
Aroclor-1260-3	8.007	7.909	8.109	567.070	500.000	13.4
Aroclor-1260-4	8.236	8.137	8.337	576.380	500.000	15.3
Aroclor-1260-5	8.561	8.464	8.664	568.290	500.000	13.7
Decachlorobiphenyl	10.432	10.334	10.534	44.940	50.000	-10.1
Tetrachloro-m-xylene	4.653	4.555	4.755	50.300	50.000	0.6

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** CORE02
Lab Code: ACE **SDG NO.:** Q3346
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 09/23/2025 09/23/2025

Client Sample No.: CCAL01 **Date Analyzed:** 10/16/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP075814.D **Time Analyzed:** 09:56

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.885	4.793	4.993	530.840	500.000	6.2
Aroclor-1016-2	4.942	4.851	5.051	558.870	500.000	11.8
Aroclor-1016-3	5.062	4.971	5.171	510.110	500.000	2.0
Aroclor-1016-4	5.103	5.011	5.211	544.920	500.000	9.0
Aroclor-1016-5	5.317	5.225	5.425	535.260	500.000	7.1
Aroclor-1260-1	6.345	6.255	6.455	512.440	500.000	2.5
Aroclor-1260-2	6.532	6.441	6.641	517.060	500.000	3.4
Aroclor-1260-3	6.685	6.595	6.795	508.690	500.000	1.7
Aroclor-1260-4	7.154	7.065	7.265	473.290	500.000	-5.3
Aroclor-1260-5	7.393	7.304	7.504	492.900	500.000	-1.4
Decachlorobiphenyl	8.791	8.703	8.903	49.540	50.000	-0.9
Tetrachloro-m-xylene	3.789	3.696	3.896	48.430	50.000	-3.1