

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** CORE02
Lab Code: ACE **SDG NO.:** Q3346
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 09/23/2025 09/23/2025

Client Sample No.: CCAL02 **Date Analyzed:** 10/16/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP075829.D **Time Analyzed:** 15:33

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.806	5.708	5.908	516.640	500.000	3.3
Aroclor-1016-2	5.827	5.729	5.929	532.710	500.000	6.5
Aroclor-1016-3	5.890	5.792	5.992	504.020	500.000	0.8
Aroclor-1016-4	5.987	5.889	6.089	538.270	500.000	7.7
Aroclor-1016-5	6.279	6.181	6.381	542.030	500.000	8.4
Aroclor-1260-1	7.397	7.299	7.499	511.870	500.000	2.4
Aroclor-1260-2	7.650	7.551	7.751	502.310	500.000	0.5
Aroclor-1260-3	8.008	7.909	8.109	548.630	500.000	9.7
Aroclor-1260-4	8.237	8.137	8.337	559.350	500.000	11.9
Aroclor-1260-5	8.563	8.464	8.664	553.920	500.000	10.8
Decachlorobiphenyl	10.433	10.334	10.534	43.510	50.000	-13.0
Tetrachloro-m-xylene	4.653	4.555	4.755	48.900	50.000	-2.2

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** CORE02
Lab Code: ACE **SDG NO.:** Q3346
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 09/23/2025 09/23/2025

Client Sample No.: CCAL02 **Date Analyzed:** 10/16/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP075829.D **Time Analyzed:** 15:33

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.886	4.793	4.993	540.720	500.000	8.1
Aroclor-1016-2	4.944	4.851	5.051	559.970	500.000	12.0
Aroclor-1016-3	5.064	4.971	5.171	512.260	500.000	2.5
Aroclor-1016-4	5.105	5.011	5.211	547.200	500.000	9.4
Aroclor-1016-5	5.319	5.225	5.425	525.650	500.000	5.1
Aroclor-1260-1	6.347	6.255	6.455	491.540	500.000	-1.7
Aroclor-1260-2	6.534	6.441	6.641	504.780	500.000	1.0
Aroclor-1260-3	6.687	6.595	6.795	496.650	500.000	-0.7
Aroclor-1260-4	7.156	7.065	7.265	466.620	500.000	-6.7
Aroclor-1260-5	7.396	7.304	7.504	483.480	500.000	-3.3
Decachlorobiphenyl	8.793	8.703	8.903	49.870	50.000	-0.3
Tetrachloro-m-xylene	3.792	3.696	3.896	48.290	50.000	-3.4