



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q3346

Analytical Method: 8082A

Client: Core Environmental Consultants and Serv

Datafile : PP075820.D

Lab Sample ID	Parameter	Spike	Result	Units	Rec	RPD	Qual	RPD	Low	Limits	RPD
								Qual		High	
PB170119BS (Column 1)	AR1016	166.5	191	ug/kg	115				71	120	
	AR1260	166.5	184	ug/kg	111				65	130	
PB170119BS (Column 2)	AR1016	166.5	172	ug/kg	103				71	120	
	AR1260	166.5	171	ug/kg	103				65	130	