



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: Core Environmental Consultants and Services, Inc.

SDG No.: Q3349

Contract: CORE02

Lab Code: ACE

Initial Calibration Source:

Continuing Calibration Source:

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRI01	Arsenic	20.4	20.0	102	65 - 135	P	10/16/2025	12:09	LB137573
	Barium	98.6	100	99	65 - 135	P	10/16/2025	12:09	LB137573
	Cadmium	5.97	6.0	100	65 - 135	P	10/16/2025	12:09	LB137573
	Chromium	10.9	10.0	109	65 - 135	P	10/16/2025	12:09	LB137573
	Lead	10.5	12.0	88	65 - 135	P	10/16/2025	12:09	LB137573
	Selenium	21.5	20.0	107	65 - 135	P	10/16/2025	12:09	LB137573
	Silver	10.8	10.0	108	65 - 135	P	10/16/2025	12:09	LB137573
CRA	Mercury	0.18	0.2	92	70 - 130	CV	10/16/2025	14:44	LB137565
CRI01	Arsenic	21.5	20.0	107	65 - 135	P	10/17/2025	12:54	LB137588
	Barium	96.5	100	96	65 - 135	P	10/17/2025	12:54	LB137588
	Cadmium	5.61	6.0	94	65 - 135	P	10/17/2025	12:54	LB137588
	Chromium	10.3	10.0	103	65 - 135	P	10/17/2025	12:54	LB137588
	Lead	11.2	12.0	94	65 - 135	P	10/17/2025	12:54	LB137588
	Selenium	19.9	20.0	100	65 - 135	P	10/17/2025	12:54	LB137588
	Silver	10.7	10.0	107	65 - 135	P	10/17/2025	12:54	LB137588
CRI01	Arsenic	18.9	20.0	94	65 - 135	P	10/20/2025	12:54	LB137590
	Barium	101	100	101	65 - 135	P	10/20/2025	12:54	LB137590
	Cadmium	5.79	6.0	96	65 - 135	P	10/20/2025	12:54	LB137590
	Chromium	10.3	10.0	103	65 - 135	P	10/20/2025	12:54	LB137590
	Lead	11.8	12.0	98	65 - 135	P	10/20/2025	12:54	LB137590
	Selenium	22.3	20.0	112	65 - 135	P	10/20/2025	12:54	LB137590
	Silver	10.2	10.0	102	65 - 135	P	10/20/2025	12:54	LB137590