

Metals
- 3b -
PREPARATION BLANK SUMMARY

Client: Core Environmental Consultants and Services, **SDG No.:** Q3349

Instrument: CV1

Sample ID	Analyte	Result (ug/L)	Acceptance Limit	Conc Qual	CRQL ug/L	M	Analysis Date	Analysis Time	Run
PB170116TB		WATER		Batch Number:	PB170126		Prep Date:	10/16/2025	
	Mercury	0.76	<2	U	2.00	CV	10/16/2025	16:27	LB137565
Sample ID	Analyte	Result (ug/L)	Acceptance Limit	Conc Qual	CRQL ug/L	M	Analysis Date	Analysis Time	Run
PB170126BL		WATER		Batch Number:	PB170126		Prep Date:	10/16/2025	
	Mercury	0.076	<0.2	U	0.20	CV	10/16/2025	14:52	LB137565

Metals
- 3b -
PREPARATION BLANK SUMMARY

Client: Core Environmental Consultants and Services,

SDG No.: Q3349

Instrument: P4

Sample ID	Analyte	Result (ug/L)	Acceptance Limit	Conc Qual	CRQL ug/L	M	Analysis Date	Analysis Time	Run
PB170116TB		WATER		Batch Number:	PB170132		Prep Date:	10/16/2025	
	Arsenic	25.6	<50	U	100	P	10/20/2025	13:42	LB137590
	Barium	72.8	<250	U	500	P	10/20/2025	13:42	LB137590
	Cadmium	2.50	<15	U	30.0	P	10/20/2025	13:42	LB137590
	Chromium	15.9	<25	J	50.0	P	10/20/2025	13:42	LB137590
	Lead	11.5	<30	U	60.0	P	10/20/2025	13:42	LB137590
	Selenium	48.2	<50	U	100	P	10/20/2025	13:42	LB137590
	Silver	8.10	<25	U	50.0	P	10/20/2025	13:42	LB137590
Sample ID	Analyte	Result (ug/L)	Acceptance Limit	Conc Qual	CRQL ug/L	M	Analysis Date	Analysis Time	Run
PB170132BL		WATER		Batch Number:	PB170132		Prep Date:	10/16/2025	
	Arsenic	25.6	<50	U	100	P	10/16/2025	18:59	LB137573
	Barium	72.8	<250	U	500	P	10/16/2025	18:59	LB137573
	Cadmium	2.50	<15	U	30.0	P	10/16/2025	18:59	LB137573
	Chromium	10.6	<25	U	50.0	P	10/16/2025	18:59	LB137573
	Lead	11.5	<30	U	60.0	P	10/16/2025	18:59	LB137573
	Selenium	48.2	<50	U	100	P	10/16/2025	18:59	LB137573
	Silver	8.10	<25	U	50.0	P	10/16/2025	18:59	LB137573