

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Core Environmental Consultants and Service

SDG No.: Q3349

Contract: CORE02

Lab Code: ACE

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB170126BS Mercury	ug/L	4.0	3.66		92	80 - 120	CV

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Core Environmental Consultants and Service

SDG No.: Q3349

Contract: CORE02

Lab Code: ACE

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB170132BS							
Arsenic	ug/L	4000	3580		90	80 - 120	P
Barium	ug/L	1000	837		84	80 - 120	P
Cadmium	ug/L	1000	835		84	80 - 120	P
Chromium	ug/L	2000	1810		90	80 - 120	P
Lead	ug/L	5000	4240		85	80 - 120	P
Selenium	ug/L	10000	9160		92	80 - 120	P
Silver	ug/L	380	347		91	80 - 120	P