



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

IDENTIFICATION SUMMARY
FOR MULTICOMPONENT ANALYTES

SAMPLE NO.

PB170119BS

Lab Name: Alliance

Contract: CORE02

Lab Code: ACE

SDG NO.: Q3349

Lab Sample ID: PB170119BS

Date(s) Analyzed: 10/16/2025 10/16/2025

Instrument ID (1): ECD_P

Instrument ID (2): ECD_P

GC Column: (1): ZB-MR1 ID: 0.32 (mm)

GC Column: (2): ZB-MR2 ID: 0.32 (mm)

Data file PP075820.D

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	MEAN CONCENTRATION	%RPD
			FROM	TO			
Aroclor-1016	1	5.807	5.757	5.857	188	191	10.47
COLUMN 1	2	5.828	5.778	5.878	192		
	3	5.891	5.841	5.941	184		
	4	5.988	5.938	6.038	197		
	5	6.28	6.23	6.33	193		
	1	4.888	4.838	4.938	176		
COLUMN 2	2	4.945	4.895	4.995	181	172	
	3	5.066	5.016	5.116	155		
	4	5.105	5.055	5.155	172		
	5	5.319	5.269	5.369	175		
Aroclor-1260	1	7.398	7.348	7.448	191	184	7.32
COLUMN 1	2	7.651	7.601	7.701	186		
	3	8.009	7.959	8.059	173		
	4	8.236	8.186	8.286	191		
	5	8.564	8.514	8.614	177		
	1	6.347	6.297	6.397	185		
COLUMN 2	2	6.535	6.485	6.585	186	171	
	3	6.686	6.636	6.736	182		
	4	7.157	7.107	7.207	146		
	5	7.397	7.347	7.447	153		

IDENTIFICATION SUMMARY
FOR MULTICOMPONENT ANALYTES

SAMPLE NO.

WCS-TP2MS

Lab Name: Alliance

Contract: CORE02

Lab Code: ACE

SDG NO.: Q3349

Lab Sample ID: Q3348-05MS

Date(s) Analyzed: 10/16/2025 10/16/2025

Instrument ID (1): ECD_P

Instrument ID (2): ECD_P

GC Column: (1): ZB-MR1 ID: 0.32 (mm)

GC Column: (2): ZB-MR2 ID: 0.32 (mm)

Data file PP075824.D

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	MEAN CONCENTRATION	%RPD
			FROM	TO			
Aroclor-1016	1	5.803	5.753	5.853	227	248	30.43
COLUMN 1	2	5.825	5.775	5.875	249		
	3	5.887	5.837	5.937	240		
	4	5.984	5.934	6.034	264		
	5	6.277	6.227	6.327	259		
	1	4.885	4.835	4.935	281		
COLUMN 2	2	4.943	4.893	4.993	337	337	
	3	5.064	5.014	5.114	300		
	4	5.105	5.055	5.155	292		
	5	5.322	5.272	5.372	476		
Aroclor-1260	1	7.395	7.345	7.445	237	222	6.99
COLUMN 1	2	7.647	7.597	7.697	220		
	3	8.006	7.956	8.056	203		
	4	8.234	8.184	8.284	231		
	5	8.559	8.509	8.609	216		
	1	6.347	6.297	6.397	214		
COLUMN 2	2	6.533	6.483	6.583	235	207	
	3	6.686	6.636	6.736	212		
	4	7.155	7.105	7.205	184		
	5	7.395	7.345	7.445	191		

IDENTIFICATION SUMMARY
FOR MULTICOMPONENT ANALYTES

SAMPLE NO.

WCS-TP2MSD

Lab Name: Alliance

Contract: CORE02

Lab Code: ACE

SDG NO.: Q3349

Lab Sample ID: Q3348-05MSD

Date(s) Analyzed: 10/16/2025 10/16/2025

Instrument ID (1): ECD_P

Instrument ID (2): ECD_P

GC Column: (1): ZB-MR1 ID: 0.32 (mm)

GC Column: (2): ZB-MR2 ID: 0.32 (mm)

Data file PP075825.D

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	MEAN CONCENTRATION	%RPD
			FROM	TO			
Aroclor-1016	1	5.803	5.753	5.853	244	263	31.68
COLUMN 1	2	5.824	5.774	5.874	265		
	3	5.887	5.837	5.937	257		
	4	5.984	5.934	6.034	275		
	5	6.276	6.226	6.326	272		
	1	4.886	4.836	4.936	307		
COLUMN 2	2	4.945	4.895	4.995	360	362	
	3	5.064	5.014	5.114	307		
	4	5.104	5.054	5.154	318		
	5	5.321	5.271	5.371	521		
Aroclor-1260	1	7.394	7.344	7.444	244	232	7.14
COLUMN 1	2	7.647	7.597	7.697	228		
	3	8.005	7.955	8.055	212		
	4	8.233	8.183	8.283	250		
	5	8.559	8.509	8.609	226		
	1	6.347	6.297	6.397	220		
COLUMN 2	2	6.533	6.483	6.583	252	216	
	3	6.686	6.636	6.736	219		
	4	7.155	7.105	7.205	191		
	5	7.395	7.345	7.445	199		