

Surrogate Summary

SDG No.: Q3349

Client: Core Environmental Consultants and Ser

Analytical Method: 8082A

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Recovery(%)	Qual	Limits(%)	
								Low	High
I.BLK-PP075246.D	PIBLK-PP075246.D	Tetrachloro-m-xyl	1	20	19.9	100		60	140
		Decachlorobiphen	1	20	19.4	97		60	140
		Tetrachloro-m-xyl	2	20	17.9	90		60	140
		Decachlorobiphen	2	20	20.0	100		60	140
I.BLK-PP075818.D	PIBLK-PP075818.D	Tetrachloro-m-xyl	1	20	22.3	112		60	140
		Decachlorobiphen	1	20	20.7	104		60	140
		Tetrachloro-m-xyl	2	20	18.9	94		60	140
		Decachlorobiphen	2	20	22.0	110		60	140
PB170119BL	PB170119BL	Tetrachloro-m-xyl	1	20	23.3	116		32	144
		Decachlorobiphen	1	20	22.0	110		32	175
		Tetrachloro-m-xyl	2	20	20.9	104		32	144
		Decachlorobiphen	2	20	23.8	119		32	175
PB170119BS	PB170119BS	Tetrachloro-m-xyl	1	20	26.2	131		32	144
		Decachlorobiphen	1	20	23.8	119		32	175
		Tetrachloro-m-xyl	2	20	12.6	63		32	144
		Decachlorobiphen	2	20	25.5	128		32	175
Q3348-05MS	WCS-TP2MS	Tetrachloro-m-xyl	1	20	25.0	125		32	144
		Decachlorobiphen	1	20	17.2	86		32	175
		Tetrachloro-m-xyl	2	20	14.9	74		32	144
		Decachlorobiphen	2	20	18.1	90		32	175
Q3348-05MSD	WCS-TP2MSD	Tetrachloro-m-xyl	1	20	26.5	133		32	144
		Decachlorobiphen	1	20	18.6	93		32	175
		Tetrachloro-m-xyl	2	20	11.4	57		32	144
		Decachlorobiphen	2	20	19.3	97		32	175
Q3349-02	PURS Station	Tetrachloro-m-xyl	1	20	12.3	62		32	144
		Decachlorobiphen	1	20	17.0	85		32	175
		Tetrachloro-m-xyl	2	20	5.80	29	*	32	144
		Decachlorobiphen	2	20	7.20	36		32	175
I.BLK-PP075833.D	PIBLK-PP075833.D	Tetrachloro-m-xyl	1	20	21.4	107		60	140
		Decachlorobiphen	1	20	19.8	99		60	140
		Tetrachloro-m-xyl	2	20	18.8	94		60	140
		Decachlorobiphen	2	20	21.6	108		60	140