

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q3349

Client ID : SSTDCCC040

Date Analyzed: 10/17/2025

Lab File ID: BF143984.D

Time Analyzed: 14:43

Instrument ID: BNA_F

GC Column: DB-UI ID: 0.18 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	156623	6.887	566623	8.17	278942	9.92
UPPER LIMIT	313246	7.387	1133250	8.669	557884	10.422
LOWER LIMIT	78311.5	6.387	283312	7.669	139471	9.422
EPA SAMPLE NO.						
01 PB170149BL	124841	6.88	484003	8.16	266250	9.92
02 PB170149BS	119613	6.88	469641	8.16	253761	9.92
03 PB170116TB	124101	6.88	484858	8.16	277642	9.92
04 TP-3MS	84940	6.88	333517	8.16	183894	9.92
05 TP-3MSD	83393	6.88	325316	8.16	173895	9.92
06 PURS Station	87692	6.88	299344	8.16	127355 *	9.92

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.