

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: CORE02
 Lab Code: ACE SDG NO.: Q3349
 Lab File ID: VX048201.D Date Analyzed: 10/17/2025
 Instrument ID: MSVOA_X Time Analyzed: 09:55
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	144905	5.54	233738	6.75	208719	10.04
UPPER LIMIT	289810	6.044	467476	7.245	417438	10.537
LOWER LIMIT	72452.5	5.044	116869	6.245	104360	9.537
EPA SAMPLE NO.						
PURS Station	84578	5.54	169663	6.75	184791	10.04
VX1017WBL01	129436	5.54	259480	6.75	275267	10.04
VX1017WBS01	157010	5.54	248908	6.75	238012	10.04
VX1017WBSD01	149269	5.53	237242	6.75	228640	10.04

IS1 = Pentafluorobenzene

IS2 = 1,4-Difluorobenzene

IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.