

Laboratory Control Sample Summary

Client: ATG-GREENVILLE AEC

SDG No.: Q3351

Project: AEC-2025-0013 - CSC 6113 Taylorville

Run No.: LB137541

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB137541BS							
COD	mg/L	50	51.2		102	1	90-110	10/15/2025

Laboratory Control Sample Summary

Client: ATG-GREENVILLE AEC

SDG No.: Q3351

Project: AEC-2025-0013 - CSC 6113 Taylorville

Run No.: LB137553

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB137553BS							
TSS	mg/L	550	561		102	1	90-110	10/16/2025