

Laboratory Control Sample Summary

Client: Aramark Uniforms

SDG No.: Q3359

Project: Monthly 2025

Run No.: LB137553

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB137553BS							
TSS	mg/L	550	561		102	1	90-110	10/16/2025

Laboratory Control Sample Summary

Client: Aramark Uniforms

SDG No.: Q3359

Project: Monthly 2025

Run No.: LB137557

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB137557BS							
TPH	mg/L	20.0	17.0		85	1	78-114	10/16/2025

Laboratory Control Sample Summary

Client: Aramark Uniforms

SDG No.: Q3359

Project: Monthly 2025

Run No.: LB137557

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB137557BSD							
TPH	mg/L	20.0	17.1		86	1	78-114	10/16/2025

Laboratory Control Sample Summary

Client: Aramark Uniforms

SDG No.: Q3359

Project: Monthly 2025

Run No.: LB137577

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB137577BS							
BOD5	mg/L	198	186		94	1	84.6-115.4	10/17/2025