

DATA PACKAGE

VOLATILE ORGANICS

PROJECT NAME : WASTE WATER 2025

GARDEN STATE LABORATORIES, INC.

410 Hillside Avenue

Hillside, NJ - 07205

Phone No: 800-273-8901

ORDER ID : Q3361

ATTENTION : Sharon Ercoliani



Laboratory Certification ID # 20012



Table Of Contents for Q3361

1) Signature Page	3
2) Case Narrative	4
3) Qualifier Page	6
4) QA Checklist	7
5) VOCMS Group1 Data	8
6) VOCMS Group2 Data	13
7) Shipping Document	23
7.1) CHAIN OF CUSTODY	24
7.2) Lab Certificate	25
7.3) Internal COC	26

1
2
3
4
5
6
7

Cover Page

Order ID : Q3361

Project ID : Waste Water 2025

Client : Garden State Laboratories, Inc.

Lab Sample Number

Q3361-01
Q3361-02

Client Sample Number

VOA 251015111-01
Trip blank 251015076-04

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. Release of the data contained in this hard copy data package has been authorized by the laboratory manager or his designee, as verified by the following signature.

Signature :

APPROVED

By Nimisha Pandya, QA/QC Supervisor at 9:50 am, Oct 28, 2025

Date: 10/27/2025

NYDOH CERTIFICATION NO - 11376

NJDEP CERTIFICATION NO - 20012

CASE NARRATIVE

Garden State Laboratories, Inc.

Project Name: Waste Water 2025

Project # N/A

Order ID # Q3361

Test Name: VOCMS Group1,VOCMS Group2

A. Number of Samples and Date of Receipt:

2 Water samples were received on 10/16/2025.

B. Parameters

According to the Chain of Custody document, the following analyses were requested:

. This data package contains results for VOCMS Group1(624.1), VOCMS Group2(8260-D).

C. Analytical Techniques:

VOCMS Group1 : The analysis performed on instrument MSVOA_N were done using GC column Rxi-624SIL MS 30m, 0.25mm, 1.4 um, Cat. #13868.The analysis of VOCMS Group1 was based on method 624.1.

VOCMS Group2 : The analysis performed on instrument MSVOA_X were done using GC column DB-624UI 20m 0.18mm 1.0 um. Cat#121-1324UI. The analysis of VOCMS Group2 was based on method 8260D.

D. QA/ QC Samples:

The Holding Times were met for all analysis.

The Surrogate recoveries were met for all analysis.

The Internal Standards Areas were met for all analysis.

The Retention Times were met for all analysis.

The RPD were met for all analysis.

The Blank Spike met requirements for all compounds.

The Blank Spike Duplicate met requirements for all compounds.

The Blank analysis did not indicate the presence of lab contamination.

The Initial Calibration met the requirements.

The Continuous Calibration met the requirements except following

VOCMS Group2 : The Continuous Calibration File ID VX048252.D met the requirements except for 4-Methyl-2-Pentanone is failing marginally low therefore no corrective action taken.

The Tuning criteria met requirements.

VOCMS Group2 : Sample VOA 251015111-01 was diluted due to high concentration.

E. Additional Comments:

VOCMS Group1 : “As per method, MS/MSD is required to be performed with the sample analysis. However, Lab did not receive sufficient volume to perform the MS/MSD therefore MS/MSD were not performed for this project. However, Lab has performed LCS/LCSD instead.”

VOCMS Group2 : Samples for MS/MSD for VOC analysis were not provided with this set of samples. The Blank Spike Duplicate is reported with the data.

F. Manual Integration Comments:

Please refer to the Manual integration Report included with the Run Logs for information on the manual integrations performed.

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

APPROVED

By Nimisha Pandya, QA/QC Supervisor at 9:50 am, Oct 28, 2025

Signature_____

DATA REPORTING QUALIFIERS- ORGANIC

For reporting results, the following “Results Qualifiers” are used:

Value	If the result is a value greater than or equal to the detection limit, report the value
U	Indicates the compound was analyzed for but was not detected. Report the minimum detection limit for the sample with the U, i.e. “10 U”. This is not necessarily the instrument detection limit attainable for this particular sample based on any concentration or dilution that may have been required.
ND	Indicates the analyte was analyzed for, but not detected
J	Indicates an estimated value. This flag is used: (1) When estimating a concentration for a tentatively identified compound (library search hits, where a 1:1 response is assumed.) (2) When the mass spectral data indicated the identification, however the result was less than the specified detection limit greater than zero. If the detection limit was 10ug/L and a concentration of 3 ug/L was calculated report as 3 J. This flag is used when similar situation arise on any organic parameter i.e. Pest, PCB and others.
B	Indicates the analyte was found in the blank as well as the sample report as “12 B”.
E	Indicates the analyte ‘s concentration exceeds the calibrated range of the instrument for that specific analysis.
D	This flag identifies all compounds identified in an analysis at a secondary dilution factor.
P	This flag is used for Pesticide/PCB target analyte when there is >25% difference for detected concentrations between the two GC columns. The lower of the two values is reported on Form 1 and flagged with a “P”.
N	This flag indicates presumptive evidence of a compound. This is only used for tentatively identified compounds (TICs), where the identification is based on a mass spectral library search. It applies to all TIC results. For generic characterization of a TIC, such as chlorinated hydrocarbon, the flag is not used.
A	This flag indicates that a Tentatively Identified Compound is a suspected aldol-condensation product.
Q	Indicates the LCS did not meet the control limits requirements

APPENDIX A

QA REVIEW GENERAL DOCUMENTATION

Project #: Q3361

Completed

For thorough review, the report must have the following:

GENERAL:

Are all original paperwork present (chain of custody, record of communication,airbill, sample management lab chronicle, login page)

✓

Check chain-of-custody for proper relinquish/return of samples

✓

Is the chain of custody signed and complete

✓

Check internal chain-of-custody for proper relinquish/return of samples /sample extracts

✓

Collect information for each project id from server. Were all requirements followed

✓

COVER PAGE:

Do numbers of samples correspond to the number of samples in the Chain of Custody on login page

✓

Do lab numbers and client Ids on cover page agree with the Chain of Custody

✓

CHAIN OF CUSTODY:

Do requested analyses on Chain of Custody agree with form I results

✓

Do requested analyses on Chain of Custody agree with the log-in page

✓

Were the correct method log-in for analysis according to the Analytical Request and Chain of Custody

✓

Were the samples received within hold time

✓

Were any problems found with the samples at arrival recorded in the Sample Management Laboratory Chronicle

✓

ANALYTICAL:

Was method requirement followed?

✓

Was client requirement followed?

✓

Does the case narrative summarize all QC failure?

✓

All runlogs and manual integration are reviewed for requirements

✓

All manual calculations and /or hand notations verified

✓

QA Review Signature: SOHIL JODHANI

Date: 10/27/2025

Hit Summary Sheet
624.1

SDG No.: Q3361

Client: Garden State Laboratories, Inc.

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	RDL	Units
-----------	-----------	--------	-----------	---------------	---	-----	-----	-------

Client ID:

0

Total Voc :

Total Concentration:

A

B

C

D



SAMPLE DATA

Report of Analysis

Client: Garden State Laboratories, Inc.
Project: Waste Water 2025
Client Sample ID: VOA 251015111-01
Lab Sample ID: Q3361-01
Analytical Method: E624.1
Sample Wt/Vol: 5 mL

Level : LOW
Final Vol: 5000 uL

Date Collected: 10/15/25
Date Received: 10/16/25
SDG No.: Q3361
Matrix: Water
% Solid: 0
Test: VOCMS Group1

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS									
107-02-8	Acrolein	6.60	U	1	6.60	25.0	ug/L	10/17/25 12:34	VN101725
107-13-1	Acrylonitrile	2.80	U	1	2.80	25.0	ug/L	10/17/25 12:34	VN101725
SURROGATES									
17060-07-0	1,2-Dichloroethane-d4	30.9			91 - 110	103%	SPK: 30		
2037-26-5	Toluene-d8	30.8			91 - 112	103%	SPK: 30		
460-00-4	4-Bromofluorobenzene	29.6			63 - 112	99%	SPK: 30		
INTERNAL STANDARDS									
		Area Count							
74-97-5	Bromochloromethane	56300							
540-36-3	1,4-Difluorobenzene	286000							
3114-55-4	Chlorobenzene-d5	253000							

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client: Garden State Laboratories, Inc.
Project: Waste Water 2025
Client Sample ID: Trip blank 251015076-04
Lab Sample ID: Q3361-02
Analytical Method: E624.1
Sample Wt/Vol: 5 mL

Level : LOW
Final Vol: 5000 uL

Date Collected: 10/15/25
Date Received: 10/16/25
SDG No.: Q3361
Matrix: Water
% Solid: 0
Test: VOCMS Group1

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS									
107-02-8	Acrolein	6.60	U	1	6.60	25.0	ug/L	10/17/25 11:32	VN101725
107-13-1	Acrylonitrile	2.80	U	1	2.80	25.0	ug/L	10/17/25 11:32	VN101725
SURROGATES									
17060-07-0	1,2-Dichloroethane-d4	32.3			91 - 110	108%	SPK: 30		
2037-26-5	Toluene-d8	31.2			91 - 112	104%	SPK: 30		
460-00-4	4-Bromofluorobenzene	28.9			63 - 112	96%	SPK: 30		
INTERNAL STANDARDS									
		Area Count							
74-97-5	Bromochloromethane	35300							
540-36-3	1,4-Difluorobenzene	198000							
3114-55-4	Chlorobenzene-d5	171000							

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

LAB CHRONICLE

OrderID:	Q3361	OrderDate:	10/16/2025 9:52:00 AM
Client:	Garden State Laboratories, Inc.	Project:	Waste Water 2025
Contact:	Sharon Ercoliani	Location:	VOA Lab

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q3361-01	VOA 251015111-01	Water	VOCMS Group1	624.1	10/15/25		10/17/25	10/16/25
Q3361-02	Trip blank 251015076-04	Water	VOCMS Group1	624.1	10/15/25		10/17/25	10/16/25

Hit Summary Sheet
SW-846

SDG No.: Q3361
Client: Garden State Laboratories, Inc.

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	RDL	Units
Client ID: VOA 251015111-01								
Q3361-01	VOA 251015111-01	Water	Vinyl Chloride	0.37	J	0.26	1.00	ug/L
Q3361-01	VOA 251015111-01	Water	Acetone	1700	E	1.50	5.00	ug/L
Q3361-01	VOA 251015111-01	Water	Methyl tert-butyl Ether	2.70		0.16	1.00	ug/L
Q3361-01	VOA 251015111-01	Water	Methyl Acetate	8.10		0.27	1.00	ug/L
Q3361-01	VOA 251015111-01	Water	Methylene Chloride	1.60		0.28	1.00	ug/L
Q3361-01	VOA 251015111-01	Water	2-Butanone	2400	E	0.98	5.00	ug/L
Q3361-01	VOA 251015111-01	Water	Benzene	5.70		0.15	1.00	ug/L
Q3361-01	VOA 251015111-01	Water	1,2-Dichloroethane	2.00		0.22	1.00	ug/L
Q3361-01	VOA 251015111-01	Water	4-Methyl-2-Pentanone	13.6		0.68	5.00	ug/L
Q3361-01	VOA 251015111-01	Water	Toluene	19.8		0.14	1.00	ug/L
Q3361-01	VOA 251015111-01	Water	Chlorobenzene	2.10		0.12	1.00	ug/L
Q3361-01	VOA 251015111-01	Water	Ethyl Benzene	12.0		0.13	1.00	ug/L
Q3361-01	VOA 251015111-01	Water	m/p-Xylenes	11.9		0.24	2.00	ug/L
Q3361-01	VOA 251015111-01	Water	o-Xylene	6.90		0.12	1.00	ug/L
Q3361-01	VOA 251015111-01	Water	Styrene	0.99	J	0.15	1.00	ug/L
Q3361-01	VOA 251015111-01	Water	Isopropylbenzene	1.70		0.12	1.00	ug/L
Q3361-01	VOA 251015111-01	Water	1,4-Dichlorobenzene	3.40		0.19	1.00	ug/L
Q3361-01	VOA 251015111-01	Water	1,2-Dichlorobenzene	0.27	J	0.16	1.00	ug/L
			Total Voc :			4190		
Q3361-01	VOA 251015111-01	Water	(+)-2-Bornanone	* 56.2	J	0	0	ug/L
Q3361-01	VOA 251015111-01	Water	Eucalyptol	* 11.0	J	0	0	ug/L
Q3361-01	VOA 251015111-01	Water	Fenchone	* 33.5	J	0	0	ug/L
Q3361-01	VOA 251015111-01	Water	Cyclohexene,1-(2-methylpropy	* 29.6	J	0	0	ug/L
Q3361-01	VOA 251015111-01	Water	3-Hexanone, 2-methyl-	* 14.7	J	0	0	ug/L
Q3361-01	VOA 251015111-01	Water	Ethyl Acetate	* 24.5	J	0.31	1.00	ug/L
Q3361-01	VOA 251015111-01	Water	Isopropyl Acetate	* 2.70	J	0.32	1.00	ug/L
Q3361-01	VOA 251015111-01	Water	Tetrahydrofuran	* 800	J	0.99	5.00	ug/L
Q3361-01	VOA 251015111-01	Water	Tert butyl alcohol	* 7600	J	5.50	25.0	ug/L
Q3361-01	VOA 251015111-01	Water	Diethyl Ether	* 6.10	J	0.31	1.00	ug/L
Q3361-01	VOA 251015111-01	Water	n-propylbenzene	* 0.54	J	0.13	1.00	ug/L
Q3361-01	VOA 251015111-01	Water	1,3,5-Trimethylbenzene	* 1.00	J	0.15	1.00	ug/L
Q3361-01	VOA 251015111-01	Water	1,2,4-Trimethylbenzene	* 4.00	J	0.14	1.00	ug/L
Q3361-01	VOA 251015111-01	Water	p-Isopropyltoluene	* 9.70	J	0.13	1.00	ug/L
Q3361-01	VOA 251015111-01	Water	Naphthalene	* 33.0	J	0.20	1.00	ug/L
			Total Tics :			8630		
			Total Concentration:			12800		

Hit Summary Sheet
SW-846

SDG No.: Q3361

Client: Garden State Laboratories, Inc.

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	RDL	Units
Client ID: VOA 251015111-01DL								
Q3361-01DL	VOA 251015111-01	Water	Acetone	1300	D	15.1	50.0	ug/L
Q3361-01DL	VOA 251015111-01	Water	2-Butanone	1500	D	9.80	50.0	ug/L
Q3361-01DL	VOA 251015111-01	Water	Benzene	4.60	JD	1.50	10.0	ug/L
Q3361-01DL	VOA 251015111-01	Water	Toluene	12.8	D	1.40	10.0	ug/L
Q3361-01DL	VOA 251015111-01	Water	Ethyl Benzene	7.70	JD	1.30	10.0	ug/L
Q3361-01DL	VOA 251015111-01	Water	m/p-Xylenes	6.90	JD	2.40	20.0	ug/L
Q3361-01DL	VOA 251015111-01	Water	o-Xylene	4.40	JD	1.20	10.0	ug/L
Total Voc :				2840				
Total Concentration:				2840				



SAMPLE DATA

Report of Analysis

Client: Garden State Laboratories, Inc.
Project: Waste Water 2025
Client Sample ID: VOA 251015111-01
Lab Sample ID: Q3361-01
Analytical Method: 8260D
Sample Wt/Vol: 5 mL

Level : LOW
Final Vol: 5000 uL

Date Collected: 10/15/25
Date Received: 10/16/25
SDG No.: Q3361
Matrix: Water
% Solid: 0
Test: VOCMS Group2

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS									
75-71-8	Dichlorodifluoromethane	0.22	U	1	0.22	1.00	ug/L	10/17/25 18:37	VX101725
74-87-3	Chloromethane	0.32	U	1	0.32	1.00	ug/L	10/17/25 18:37	VX101725
75-01-4	Vinyl Chloride	0.37	J	1	0.26	1.00	ug/L	10/17/25 18:37	VX101725
74-83-9	Bromomethane	1.40	U	1	1.40	5.00	ug/L	10/17/25 18:37	VX101725
75-00-3	Chloroethane	0.47	U	1	0.47	1.00	ug/L	10/17/25 18:37	VX101725
75-69-4	Trichlorofluoromethane	0.33	U	1	0.33	1.00	ug/L	10/17/25 18:37	VX101725
76-13-1	1,1,2-Trichlorotrifluoroethane	0.25	U	1	0.25	1.00	ug/L	10/17/25 18:37	VX101725
75-35-4	1,1-Dichloroethene	0.23	U	1	0.23	1.00	ug/L	10/17/25 18:37	VX101725
67-64-1	Acetone	1700	E	1	1.50	5.00	ug/L	10/17/25 18:37	VX101725
75-15-0	Carbon Disulfide	0.21	U	1	0.21	1.00	ug/L	10/17/25 18:37	VX101725
1634-04-4	Methyl tert-butyl Ether	2.70		1	0.16	1.00	ug/L	10/17/25 18:37	VX101725
79-20-9	Methyl Acetate	8.10		1	0.27	1.00	ug/L	10/17/25 18:37	VX101725
75-09-2	Methylene Chloride	1.60		1	0.28	1.00	ug/L	10/17/25 18:37	VX101725
156-60-5	trans-1,2-Dichloroethene	0.23	U	1	0.23	1.00	ug/L	10/17/25 18:37	VX101725
75-34-3	1,1-Dichloroethane	0.23	U	1	0.23	1.00	ug/L	10/17/25 18:37	VX101725
110-82-7	Cyclohexane	1.50	U	1	1.50	5.00	ug/L	10/17/25 18:37	VX101725
78-93-3	2-Butanone	2400	E	1	0.98	5.00	ug/L	10/17/25 18:37	VX101725
56-23-5	Carbon Tetrachloride	0.25	U	1	0.25	1.00	ug/L	10/17/25 18:37	VX101725
156-59-2	cis-1,2-Dichloroethene	0.19	U	1	0.19	1.00	ug/L	10/17/25 18:37	VX101725
74-97-5	Bromochloromethane	0.22	U	1	0.22	1.00	ug/L	10/17/25 18:37	VX101725
67-66-3	Chloroform	0.25	U	1	0.25	1.00	ug/L	10/17/25 18:37	VX101725
71-55-6	1,1,1-Trichloroethane	0.20	U	1	0.20	1.00	ug/L	10/17/25 18:37	VX101725
108-87-2	Methylcyclohexane	0.16	U	1	0.16	1.00	ug/L	10/17/25 18:37	VX101725
71-43-2	Benzene	5.70		1	0.15	1.00	ug/L	10/17/25 18:37	VX101725
107-06-2	1,2-Dichloroethane	2.00		1	0.22	1.00	ug/L	10/17/25 18:37	VX101725
79-01-6	Trichloroethene	0.090	U	1	0.090	1.00	ug/L	10/17/25 18:37	VX101725
78-87-5	1,2-Dichloropropane	0.20	U	1	0.20	1.00	ug/L	10/17/25 18:37	VX101725
75-27-4	Bromodichloromethane	0.22	U	1	0.22	1.00	ug/L	10/17/25 18:37	VX101725
108-10-1	4-Methyl-2-Pentanone	13.6		1	0.68	5.00	ug/L	10/17/25 18:37	VX101725
108-88-3	Toluene	19.8		1	0.14	1.00	ug/L	10/17/25 18:37	VX101725
10061-02-6	t-1,3-Dichloropropene	0.17	U	1	0.17	1.00	ug/L	10/17/25 18:37	VX101725
10061-01-5	cis-1,3-Dichloropropene	0.16	U	1	0.16	1.00	ug/L	10/17/25 18:37	VX101725
79-00-5	1,1,2-Trichloroethane	0.21	U	1	0.21	1.00	ug/L	10/17/25 18:37	VX101725
591-78-6	2-Hexanone	0.89	U	1	0.89	5.00	ug/L	10/17/25 18:37	VX101725
124-48-1	Dibromochloromethane	0.18	U	1	0.18	1.00	ug/L	10/17/25 18:37	VX101725
106-93-4	1,2-Dibromoethane	0.15	U	1	0.15	1.00	ug/L	10/17/25 18:37	VX101725
127-18-4	Tetrachloroethene	0.23	U	1	0.23	1.00	ug/L	10/17/25 18:37	VX101725
108-90-7	Chlorobenzene	2.10		1	0.12	1.00	ug/L	10/17/25 18:37	VX101725
100-41-4	Ethyl Benzene	12.0		1	0.13	1.00	ug/L	10/17/25 18:37	VX101725
179601-23-1	m/p-Xylenes	11.9		1	0.24	2.00	ug/L	10/17/25 18:37	VX101725
95-47-6	o-Xylene	6.90		1	0.12	1.00	ug/L	10/17/25 18:37	VX101725
100-42-5	Styrene	0.99	J	1	0.15	1.00	ug/L	10/17/25 18:37	VX101725
75-25-2	Bromoform	0.19	U	1	0.19	1.00	ug/L	10/17/25 18:37	VX101725

Report of Analysis

Client: Garden State Laboratories, Inc.
Project: Waste Water 2025
Client Sample ID: VOA 251015111-01
Lab Sample ID: Q3361-01
Analytical Method: 8260D
Sample Wt/Vol: 5 mL

Level: LOW
Final Vol: 5000 uL

Date Collected: 10/15/25
Date Received: 10/16/25
SDG No.: Q3361
Matrix: Water
% Solid: 0
Test: VOCMS Group2

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
98-82-8	Isopropylbenzene	1.70		1	0.12	1.00	ug/L	10/17/25 18:37	VX101725
79-34-5	1,1,2,2-Tetrachloroethane	0.26	U	1	0.26	1.00	ug/L	10/17/25 18:37	VX101725
541-73-1	1,3-Dichlorobenzene	0.16	U	1	0.16	1.00	ug/L	10/17/25 18:37	VX101725
106-46-7	1,4-Dichlorobenzene	3.40		1	0.19	1.00	ug/L	10/17/25 18:37	VX101725
95-50-1	1,2-Dichlorobenzene	0.27	J	1	0.16	1.00	ug/L	10/17/25 18:37	VX101725
96-12-8	1,2-Dibromo-3-Chloropropane	0.53	U	1	0.53	1.00	ug/L	10/17/25 18:37	VX101725
120-82-1	1,2,4-Trichlorobenzene	0.20	U	1	0.20	1.00	ug/L	10/17/25 18:37	VX101725
87-61-6	1,2,3-Trichlorobenzene	0.20	U	1	0.20	1.00	ug/L	10/17/25 18:37	VX101725
SURROGATES									
17060-07-0	1,2-Dichloroethane-d4	50.0			74 - 125	100%	SPK: 50		
1868-53-7	Dibromofluoromethane	50.0			75 - 124	100%	SPK: 50		
2037-26-5	Toluene-d8	48.0			86 - 113	96%	SPK: 50		
460-00-4	4-Bromofluorobenzene	58.8			77 - 121	118%	SPK: 50		
INTERNAL STANDARDS									
363-72-4	Pentafluorobenzene	77400							
540-36-3	1,4-Difluorobenzene	149000							
3114-55-4	Chlorobenzene-d5	160000							
3855-82-1	1,4-Dichlorobenzene-d4	85700							
TENTATIVE IDENTIFIED COMPOUNDS									
60-29-7	Diethyl Ether	6.10	J			2.14	ug/L		
75-65-0	Tert butyl alcohol	7600	J			2.93	ug/L		
141-78-6	Ethyl Acetate	24.5	J			4.70	ug/L		
109-99-9	Tetrahydrofuran	800	J			4.98	ug/L		
108-21-4	Isopropyl Acetate	2.70	J			6.33	ug/L		
007379-12-6	3-Hexanone, 2-methyl-	14.7	J			9.36	ug/L		
103-65-1	n-propylbenzene	0.54	J			11.3	ug/L		
108-67-8	1,3,5-Trimethylbenzene	1.00	J			11.4	ug/L		
95-63-6	1,2,4-Trimethylbenzene	4.00	J			11.7	ug/L		
99-87-6	p-Isopropyltoluene	9.70	J			12.0	ug/L		
000470-82-6	Eucalyptol	11.0	J			12.1	ug/L		
001195-79-5	Fenchone	33.5	J			12.9	ug/L		
000464-49-3	(+)-2-Bornanone	56.2	J			13.5	ug/L		
003983-03-7	Cyclohexene,1-(2-methylpropyl)-	29.6	J			13.6	ug/L		
91-20-3	Naphthalene	33.0	J			13.8	ug/L		

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client: Garden State Laboratories, Inc.
Project: Waste Water 2025
Client Sample ID: VOA 251015111-01DL
Lab Sample ID: Q3361-01DL
Analytical Method: 8260D
Sample Wt/Vol: 5 mL

Level : LOW
Final Vol: 5000 uL

Date Collected: 10/15/25
Date Received: 10/16/25
SDG No.: Q3361
Matrix: Water
% Solid: 0
Test: VOCMS Group2

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS									
75-71-8	Dichlorodifluoromethane	2.20	UD	10	2.20	10.0	ug/L	10/21/25 17:07	VX102125
74-87-3	Chloromethane	3.20	UD	10	3.20	10.0	ug/L	10/21/25 17:07	VX102125
75-01-4	Vinyl Chloride	2.60	UD	10	2.60	10.0	ug/L	10/21/25 17:07	VX102125
74-83-9	Bromomethane	14.4	UD	10	14.4	50.0	ug/L	10/21/25 17:07	VX102125
75-00-3	Chloroethane	4.70	UD	10	4.70	10.0	ug/L	10/21/25 17:07	VX102125
75-69-4	Trichlorofluoromethane	3.30	UD	10	3.30	10.0	ug/L	10/21/25 17:07	VX102125
76-13-1	1,1,2-Trichlorotrifluoroethane	2.50	UD	10	2.50	10.0	ug/L	10/21/25 17:07	VX102125
75-35-4	1,1-Dichloroethene	2.30	UD	10	2.30	10.0	ug/L	10/21/25 17:07	VX102125
67-64-1	Acetone	1300	D	10	15.1	50.0	ug/L	10/21/25 17:07	VX102125
75-15-0	Carbon Disulfide	2.10	UD	10	2.10	10.0	ug/L	10/21/25 17:07	VX102125
1634-04-4	Methyl tert-butyl Ether	1.60	UD	10	1.60	10.0	ug/L	10/21/25 17:07	VX102125
79-20-9	Methyl Acetate	2.70	UD	10	2.70	10.0	ug/L	10/21/25 17:07	VX102125
75-09-2	Methylene Chloride	2.80	UD	10	2.80	10.0	ug/L	10/21/25 17:07	VX102125
156-60-5	trans-1,2-Dichloroethene	2.30	UD	10	2.30	10.0	ug/L	10/21/25 17:07	VX102125
75-34-3	1,1-Dichloroethane	2.30	UD	10	2.30	10.0	ug/L	10/21/25 17:07	VX102125
110-82-7	Cyclohexane	14.5	UD	10	14.5	50.0	ug/L	10/21/25 17:07	VX102125
78-93-3	2-Butanone	1500	D	10	9.80	50.0	ug/L	10/21/25 17:07	VX102125
56-23-5	Carbon Tetrachloride	2.50	UD	10	2.50	10.0	ug/L	10/21/25 17:07	VX102125
156-59-2	cis-1,2-Dichloroethene	1.90	UD	10	1.90	10.0	ug/L	10/21/25 17:07	VX102125
74-97-5	Bromochloromethane	2.20	UD	10	2.20	10.0	ug/L	10/21/25 17:07	VX102125
67-66-3	Chloroform	2.50	UD	10	2.50	10.0	ug/L	10/21/25 17:07	VX102125
71-55-6	1,1,1-Trichloroethane	2.00	UD	10	2.00	10.0	ug/L	10/21/25 17:07	VX102125
108-87-2	Methylcyclohexane	1.60	UD	10	1.60	10.0	ug/L	10/21/25 17:07	VX102125
71-43-2	Benzene	4.60	JD	10	1.50	10.0	ug/L	10/21/25 17:07	VX102125
107-06-2	1,2-Dichloroethane	2.20	UD	10	2.20	10.0	ug/L	10/21/25 17:07	VX102125
79-01-6	Trichloroethene	0.93	UD	10	0.93	10.0	ug/L	10/21/25 17:07	VX102125
78-87-5	1,2-Dichloropropane	2.00	UD	10	2.00	10.0	ug/L	10/21/25 17:07	VX102125
75-27-4	Bromodichloromethane	2.20	UD	10	2.20	10.0	ug/L	10/21/25 17:07	VX102125
108-10-1	4-Methyl-2-Pentanone	6.80	UD	10	6.80	50.0	ug/L	10/21/25 17:07	VX102125
108-88-3	Toluene	12.8	D	10	1.40	10.0	ug/L	10/21/25 17:07	VX102125
10061-02-6	t-1,3-Dichloropropene	1.70	UD	10	1.70	10.0	ug/L	10/21/25 17:07	VX102125
10061-01-5	cis-1,3-Dichloropropene	1.60	UD	10	1.60	10.0	ug/L	10/21/25 17:07	VX102125
79-00-5	1,1,2-Trichloroethane	2.10	UD	10	2.10	10.0	ug/L	10/21/25 17:07	VX102125
591-78-6	2-Hexanone	8.90	UD	10	8.90	50.0	ug/L	10/21/25 17:07	VX102125
124-48-1	Dibromochloromethane	1.80	UD	10	1.80	10.0	ug/L	10/21/25 17:07	VX102125
106-93-4	1,2-Dibromoethane	1.50	UD	10	1.50	10.0	ug/L	10/21/25 17:07	VX102125
127-18-4	Tetrachloroethene	2.30	UD	10	2.30	10.0	ug/L	10/21/25 17:07	VX102125
108-90-7	Chlorobenzene	1.20	UD	10	1.20	10.0	ug/L	10/21/25 17:07	VX102125
100-41-4	Ethyl Benzene	7.70	JD	10	1.30	10.0	ug/L	10/21/25 17:07	VX102125
179601-23-1	m/p-Xylenes	6.90	JD	10	2.40	20.0	ug/L	10/21/25 17:07	VX102125
95-47-6	o-Xylene	4.40	JD	10	1.20	10.0	ug/L	10/21/25 17:07	VX102125
100-42-5	Styrene	1.50	UD	10	1.50	10.0	ug/L	10/21/25 17:07	VX102125
75-25-2	Bromoform	1.90	UD	10	1.90	10.0	ug/L	10/21/25 17:07	VX102125

Report of Analysis

Client: Garden State Laboratories, Inc.
Project: Waste Water 2025
Client Sample ID: VOA 251015111-01DL
Lab Sample ID: Q3361-01DL
Analytical Method: 8260D
Sample Wt/Vol: 5 mL

Level : LOW
Final Vol: 5000 uL

Date Collected: 10/15/25
Date Received: 10/16/25
SDG No.: Q3361
Matrix: Water
% Solid: 0
Test: VOCMS Group2

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
98-82-8	Isopropylbenzene	1.20	UD	10	1.20	10.0	ug/L	10/21/25 17:07	VX102125
79-34-5	1,1,2,2-Tetrachloroethane	2.60	UD	10	2.60	10.0	ug/L	10/21/25 17:07	VX102125
541-73-1	1,3-Dichlorobenzene	1.60	UD	10	1.60	10.0	ug/L	10/21/25 17:07	VX102125
106-46-7	1,4-Dichlorobenzene	1.90	UD	10	1.90	10.0	ug/L	10/21/25 17:07	VX102125
95-50-1	1,2-Dichlorobenzene	1.60	UD	10	1.60	10.0	ug/L	10/21/25 17:07	VX102125
96-12-8	1,2-Dibromo-3-Chloropropane	5.30	UD	10	5.30	10.0	ug/L	10/21/25 17:07	VX102125
120-82-1	1,2,4-Trichlorobenzene	2.00	UD	10	2.00	10.0	ug/L	10/21/25 17:07	VX102125
87-61-6	1,2,3-Trichlorobenzene	2.00	UD	10	2.00	10.0	ug/L	10/21/25 17:07	VX102125
SURROGATES									
17060-07-0	1,2-Dichloroethane-d4	55.0			74 - 125	110%	SPK: 50		
1868-53-7	Dibromofluoromethane	48.4			75 - 124	97%	SPK: 50		
2037-26-5	Toluene-d8	47.9			86 - 113	96%	SPK: 50		
460-00-4	4-Bromofluorobenzene	57.9			77 - 121	116%	SPK: 50		
INTERNAL STANDARDS									
363-72-4	Pentafluorobenzene	106000							
540-36-3	1,4-Difluorobenzene	201000							
3114-55-4	Chlorobenzene-d5	219000							
3855-82-1	1,4-Dichlorobenzene-d4	121000							

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client: Garden State Laboratories, Inc.
Project: Waste Water 2025
Client Sample ID: Trip blank 251015076-04
Lab Sample ID: Q3361-02
Analytical Method: 8260D
Sample Wt/Vol: 5 mL

Level : LOW
Final Vol: 5000 uL

Date Collected: 10/15/25
Date Received: 10/16/25
SDG No.: Q3361
Matrix: Water
% Solid: 0
Test: VOCMS Group2

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS									
75-71-8	Dichlorodifluoromethane	0.22	U	1	0.22	1.00	ug/L	10/17/25 14:51	VX101725
74-87-3	Chloromethane	0.32	U	1	0.32	1.00	ug/L	10/17/25 14:51	VX101725
75-01-4	Vinyl Chloride	0.26	U	1	0.26	1.00	ug/L	10/17/25 14:51	VX101725
74-83-9	Bromomethane	1.40	U	1	1.40	5.00	ug/L	10/17/25 14:51	VX101725
75-00-3	Chloroethane	0.47	U	1	0.47	1.00	ug/L	10/17/25 14:51	VX101725
75-69-4	Trichlorofluoromethane	0.33	U	1	0.33	1.00	ug/L	10/17/25 14:51	VX101725
76-13-1	1,1,2-Trichlorotrifluoroethane	0.25	U	1	0.25	1.00	ug/L	10/17/25 14:51	VX101725
75-35-4	1,1-Dichloroethene	0.23	U	1	0.23	1.00	ug/L	10/17/25 14:51	VX101725
67-64-1	Acetone	1.50	U	1	1.50	5.00	ug/L	10/17/25 14:51	VX101725
75-15-0	Carbon Disulfide	0.21	U	1	0.21	1.00	ug/L	10/17/25 14:51	VX101725
1634-04-4	Methyl tert-butyl Ether	0.16	U	1	0.16	1.00	ug/L	10/17/25 14:51	VX101725
79-20-9	Methyl Acetate	0.27	U	1	0.27	1.00	ug/L	10/17/25 14:51	VX101725
75-09-2	Methylene Chloride	0.28	U	1	0.28	1.00	ug/L	10/17/25 14:51	VX101725
156-60-5	trans-1,2-Dichloroethene	0.23	U	1	0.23	1.00	ug/L	10/17/25 14:51	VX101725
75-34-3	1,1-Dichloroethane	0.23	U	1	0.23	1.00	ug/L	10/17/25 14:51	VX101725
110-82-7	Cyclohexane	1.50	U	1	1.50	5.00	ug/L	10/17/25 14:51	VX101725
78-93-3	2-Butanone	0.98	U	1	0.98	5.00	ug/L	10/17/25 14:51	VX101725
56-23-5	Carbon Tetrachloride	0.25	U	1	0.25	1.00	ug/L	10/17/25 14:51	VX101725
156-59-2	cis-1,2-Dichloroethene	0.19	U	1	0.19	1.00	ug/L	10/17/25 14:51	VX101725
74-97-5	Bromochloromethane	0.22	U	1	0.22	1.00	ug/L	10/17/25 14:51	VX101725
67-66-3	Chloroform	0.25	U	1	0.25	1.00	ug/L	10/17/25 14:51	VX101725
71-55-6	1,1,1-Trichloroethane	0.20	U	1	0.20	1.00	ug/L	10/17/25 14:51	VX101725
108-87-2	Methylcyclohexane	0.16	U	1	0.16	1.00	ug/L	10/17/25 14:51	VX101725
71-43-2	Benzene	0.15	U	1	0.15	1.00	ug/L	10/17/25 14:51	VX101725
107-06-2	1,2-Dichloroethane	0.22	U	1	0.22	1.00	ug/L	10/17/25 14:51	VX101725
79-01-6	Trichloroethene	0.090	U	1	0.090	1.00	ug/L	10/17/25 14:51	VX101725
78-87-5	1,2-Dichloropropane	0.20	U	1	0.20	1.00	ug/L	10/17/25 14:51	VX101725
75-27-4	Bromodichloromethane	0.22	U	1	0.22	1.00	ug/L	10/17/25 14:51	VX101725
108-10-1	4-Methyl-2-Pentanone	0.68	U	1	0.68	5.00	ug/L	10/17/25 14:51	VX101725
108-88-3	Toluene	0.14	U	1	0.14	1.00	ug/L	10/17/25 14:51	VX101725
10061-02-6	t-1,3-Dichloropropene	0.17	U	1	0.17	1.00	ug/L	10/17/25 14:51	VX101725
10061-01-5	cis-1,3-Dichloropropene	0.16	U	1	0.16	1.00	ug/L	10/17/25 14:51	VX101725
79-00-5	1,1,2-Trichloroethane	0.21	U	1	0.21	1.00	ug/L	10/17/25 14:51	VX101725
591-78-6	2-Hexanone	0.89	U	1	0.89	5.00	ug/L	10/17/25 14:51	VX101725
124-48-1	Dibromochloromethane	0.18	U	1	0.18	1.00	ug/L	10/17/25 14:51	VX101725
106-93-4	1,2-Dibromoethane	0.15	U	1	0.15	1.00	ug/L	10/17/25 14:51	VX101725
127-18-4	Tetrachloroethene	0.23	U	1	0.23	1.00	ug/L	10/17/25 14:51	VX101725
108-90-7	Chlorobenzene	0.12	U	1	0.12	1.00	ug/L	10/17/25 14:51	VX101725
100-41-4	Ethyl Benzene	0.13	U	1	0.13	1.00	ug/L	10/17/25 14:51	VX101725
179601-23-1	m/p-Xylenes	0.24	U	1	0.24	2.00	ug/L	10/17/25 14:51	VX101725
95-47-6	o-Xylene	0.12	U	1	0.12	1.00	ug/L	10/17/25 14:51	VX101725
100-42-5	Styrene	0.15	U	1	0.15	1.00	ug/L	10/17/25 14:51	VX101725
75-25-2	Bromoform	0.19	U	1	0.19	1.00	ug/L	10/17/25 14:51	VX101725

Report of Analysis

Client: Garden State Laboratories, Inc.
Project: Waste Water 2025
Client Sample ID: Trip blank 251015076-04
Lab Sample ID: Q3361-02
Analytical Method: 8260D
Sample Wt/Vol: 5 mL

Level : LOW
Final Vol: 5000 uL

Date Collected: 10/15/25
Date Received: 10/16/25
SDG No.: Q3361
Matrix: Water
% Solid: 0
Test: VOCMS Group2

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
98-82-8	Isopropylbenzene	0.12	U	1	0.12	1.00	ug/L	10/17/25 14:51	VX101725
79-34-5	1,1,2,2-Tetrachloroethane	0.26	U	1	0.26	1.00	ug/L	10/17/25 14:51	VX101725
541-73-1	1,3-Dichlorobenzene	0.16	U	1	0.16	1.00	ug/L	10/17/25 14:51	VX101725
106-46-7	1,4-Dichlorobenzene	0.19	U	1	0.19	1.00	ug/L	10/17/25 14:51	VX101725
95-50-1	1,2-Dichlorobenzene	0.16	U	1	0.16	1.00	ug/L	10/17/25 14:51	VX101725
96-12-8	1,2-Dibromo-3-Chloropropane	0.53	U	1	0.53	1.00	ug/L	10/17/25 14:51	VX101725
120-82-1	1,2,4-Trichlorobenzene	0.20	U	1	0.20	1.00	ug/L	10/17/25 14:51	VX101725
87-61-6	1,2,3-Trichlorobenzene	0.20	U	1	0.20	1.00	ug/L	10/17/25 14:51	VX101725
SURROGATES									
17060-07-0	1,2-Dichloroethane-d4	57.9			74 - 125	116%	SPK: 50		
1868-53-7	Dibromofluoromethane	50.4			75 - 124	101%	SPK: 50		
2037-26-5	Toluene-d8	47.6			86 - 113	95%	SPK: 50		
460-00-4	4-Bromofluorobenzene	57.9			77 - 121	116%	SPK: 50		
INTERNAL STANDARDS									
		Area Count							
363-72-4	Pentafluorobenzene	105000							
540-36-3	1,4-Difluorobenzene	219000							
3114-55-4	Chlorobenzene-d5	233000							
3855-82-1	1,4-Dichlorobenzene-d4	119000							

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

LAB CHRONICLE

OrderID:	Q3361	OrderDate:	10/16/2025 9:52:00 AM
Client:	Garden State Laboratories, Inc.	Project:	Waste Water 2025
Contact:	Sharon Ercoliani	Location:	VOA Lab

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q3361-01	VOA 251015111-01	Water			10/15/25			10/16/25
			VOCMS Group1	624.1				
			VOCMS Group2	8260-Low			10/17/25 10/17/25	
Q3361-01DL	VOA 251015111-01DL	Water			10/15/25			10/16/25
			VOCMS Group2	8260-Low			10/21/25	
Q3361-02	Trip blank 251015076-04	Water			10/15/25			10/16/25
			VOCMS Group1	624.1			10/17/25	
			VOCMS Group2	8260-Low			10/17/25	



SHIPPING DOCUMENTS

CHAIN OF CUSTODY RECORD - PRESS HARD AND PRINT CLEARLY - USE BALL POINT PEN

Garden State Laboratories, Inc.

Main Lab - 410 Hillside Avenue, Hillside NJ 07205 - NJDEP Lab Cert. #20044
Jersey Shore Lab - 54 Main Street, Waretown NJ 08758 - NJDEP Lab Cert. #15037

Tel. 800-273-8901/908-688-8900 Fax 908-688-8966 www.gslabs.com info@gslabs.com

Office and Drop off Locations

North Jersey Office: 225 Sparta Avenue, Sparta, NJ 07871 Tel. 973-729-1827

West Jersey Office: 2050 Route 31 North, Glen Gardner, NJ 08826 Tel. 908-537-7414

CLIENT INFORMATION (REPORT TO BE SENT TO)

Name: Garden State Laboratories, Inc. Contact/Authorized by: Robert Szot
Mailing Address: 410 Hillside Ave. Phone: 908-688-8900 EXT 129
City/State/Zip: Hillside, NJ. 07205 Email: rszot@gslabs.com

SAMPLE INFORMATION

SAMPLE TYPE: WASTE WATER

SAMPLE LOCATION:

Grab Comp	SAMPLE ID	SAMPLE COLLECTION				ANALYSIS REQUIRED (Print Legibly)	CONTAINER INFORMATION			
		Date	Time	AM	PM		No.	Type*	Size	Pres.*
X	VOA 251015111-01	10/15/25	9:20	X		EPA 8260 TCL LIST + Acrolien & Acrylonitrile	3	V	40mL	A
X	Trip blank 251015076-04					EPA 8260 TCL LIST + Acrolien & Acrylonitrile	2	V	40mL	A

*Container Type: P = Plastic G = Glass A = Amber Glass I = Sterile Thio V = Vial Other/Specify:

*Preservation Code: A = Non Preserved B = Sulfuric Acid C = Sodium Hydroxide D = Nitric Acid
E = Hydrochloric Acid F = Zinc Acetate G = Sodium Thiosulfate H = Ascorbic Acid I = Cooled Other/Specify:

☒ SUBCONTRACTED WORK

TURNAROUND TIME: ☒ Standard ☐ Rush (IF RUSH REQUESTED) Rush Due by:

REPORT FORMAT: ☒ Standard Report ☐ Other/Specify:

☐ Standard Report + E2 PWS ID#:

SEND TO: Chem Tech

DATE/TIME:

METHOD OF SHIPMENT:

Deliver

PAYMENT INFORMATION

☐ Sampling/Pick-up Fee: \$ ☐ Composite Fee: \$ ☐ Rush Fee: \$ Amount Due: \$

Payment Method: ☐ Credit Card Type: ☐ Check # ☐ Other: See Quote

Note:

VOA UNPRESERVED DUE TO EFFERVESCENSE - 3 DAY TAT PER JORDAN HEDVAT

SAMPLE CUSTODY EXCHANGES MUST BE DOCUMENTED BELOW EACH TIME SAMPLES CHANGE POSSESSION

PLEASE PRINT YOUR NAME LEGIBLY, USE FULL LEGAL SIGNATURE, DATE AND TIME

Sampled by (PRINT):	Signature:	Date/Time:
Client/Client's Representative (PRINT):	Signature:	Date/Time:
1. Received/Relinquished by (PRINT): Megan Hawanich	Signature: <i>Megan Hawanich</i>	Date/Time: 10/15/25 14:55
2. Received/Relinquished by (PRINT):	Signature: <i>Cassandra Pena</i>	Date/Time: 10/16/25 9:40

The liability of Garden State Laboratories, Inc. for services rendered shall in no event exceed the amount of the invoice.
Main Lab Certified by NJ Dept. of Health, NJDEP, TN, NY Dept. of Health #11550 and PADEP #68-03680

Q3361

FOR SAMPLE RECEIVING USE ONLY

DATE/TIME/TEMP. REC'D AT LAB:

Page of

GSL CLIENT #

MICRO #

CHEM. #

SAMPLE REC'D BY:

☒ GSL FIELD SAMPLER/PICK-UP

☐ PICK-UP AT DROP OFF LOCATION

☐ DELIVERED BY CLIENT

2.3

Laboratory Certification

Certified By	License No.
Connecticut	PH-0830
DOD ELAP (ANAB)	L2219
Maine	2024021
Maryland	296
New Hampshire	255425
New Jersey	20012
New York	11376
Pennsylvania	68-00548
Soil Permit	525-24-234-08441
Texas	TX-C25-00189
Virginia	460312

LOGIN REPORT/SAMPLE TRANSFER

Order ID : Q3361	GARD04	Order Date : 10/16/2025 9:52:00 AM	Project Mgr :
Client Name : Garden State Laboratories,]		Project Name : Waste Water 2025	Report Type : Level 1
Client Contact : Sharon Ercoliani		Receive DateTime : 10/16/2025 9:40:00 AM	EDD Type : EXCEL NOCLEANUP
Invoice Name : Garden State Laboratories,]		Purchase Order :	Hard Copy Date :
Invoice Contact : Sharon Ercoliani			Date Signoff :

LAB ID	CLIENT ID	MATRIX	SAMPLE DATE	SAMPLE TIME	TEST	TEST GROUP	METHOD	FAX DATE	DUE DATES
Q3361-01	VOA 251015111-01	Water	10/15/2025	09:20					
					VOCMS Group1		624.1	10 Bus. Days	
					VOCMS Group2		8260-Low	10 Bus. Days	
Q3361-02	Trip blank 251015076-04	Water	10/15/2025	09:20					
					VOCMS Group1		624.1	10 Bus. Days	
					VOCMS Group2		8260-Low	10 Bus. Days	

Relinquished By :

Date / Time : 10/16/25 1050

Received By :

Date / Time :

Storage Area : VOA Refridgerator Room