

VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name:	<u>Alliance</u>	Contract:	<u>GARD04</u>
Lab Code:	<u>ACE</u>	SDG No.:	<u>Q3361</u>
Instrument ID:	<u>MSVOA_X</u>	Calibration Date(s):	<u>10/20/2025</u> <u>10/20/2025</u>
Heated Purge: (Y/N)	<u>N</u>	Calibration Time(s):	<u>10:59</u> <u>13:49</u>
GC Column:	<u>DB-624UI</u>	ID:	<u>0.18</u> (mm)

LAB FILE ID:		RRF001 = VX048229.D		RRF020 = VX048231.D		RRF050 = VX048232.D		
		RRF100 = VX048233.D		RRF150 = VX048234.D		RRF005 = VX048236.D		
COMPOUND	RRF001	RRF020	RRF050	RRF100	RRF150	RRF005	RRF	% RSD
Dichlorodifluoromethane	0.522	0.513	0.483	0.540	0.504	0.522	0.514	3.7
Chloromethane	0.506	0.468	0.449	0.476	0.471	0.508	0.480	4.8
Vinyl Chloride	0.447	0.437	0.416	0.462	0.445	0.458	0.444	3.7
Bromomethane		0.323	0.282	0.280	0.276	0.346	0.302	10.4
Chloroethane	0.233	0.277	0.247	0.282	0.267	0.261	0.261	7.2
Trichlorofluoromethane	0.776	0.874	0.807	0.907	0.869	0.880	0.852	5.8
1,1,2-Trichlorotrifluoroethane	0.512	0.540	0.501	0.583	0.549	0.553	0.540	5.5
1,1-Dichloroethene	0.515	0.548	0.485	0.558	0.531	0.514	0.525	5
Acetone	0.300	0.250	0.202	0.221	0.206	0.262	0.240	15.7
Carbon Disulfide	2.034	1.612	1.442	1.617	1.532	1.689	1.654	12.4
Methyl tert-butyl Ether	1.894	1.909	1.704	1.817	1.764	1.657	1.791	5.7
Methyl Acetate	0.514	0.551	0.466	0.514	0.504	0.464	0.502	6.6
Methylene Chloride	0.555	0.651	0.552	0.597	0.567	0.595	0.586	6.4
trans-1,2-Dichloroethene	0.587	0.585	0.535	0.592	0.558	0.603	0.577	4.4
1,1-Dichloroethane	1.116	1.085	0.977	1.055	1.019	1.054	1.051	4.6
Cyclohexane		0.894	0.821	0.926	0.866	0.843	0.870	4.8
2-Butanone	0.311	0.329	0.281	0.301	0.290	0.315	0.304	5.7
Carbon Tetrachloride	0.650	0.594	0.553	0.604	0.567	0.622	0.599	5.9
cis-1,2-Dichloroethene	0.682	0.700	0.626	0.695	0.661	0.673	0.673	4
Bromochloromethane	0.388	0.404	0.441	0.462	0.452	0.469	0.436	7.5
Chloroform	1.147	1.208	1.046	1.127	1.067	1.143	1.123	5.2
1,1,1-Trichloroethane	1.049	1.085	0.973	1.055	1.007	1.041	1.035	3.8
Methylcyclohexane	0.551	0.548	0.540	0.626	0.582	0.590	0.573	5.7
Benzene	1.493	1.480	1.301	1.426	1.347	1.405	1.409	5.3
1,2-Dichloroethane	0.529	0.566	0.500	0.528	0.496	0.499	0.520	5.2
Trichloroethene	0.400	0.373	0.348	0.380	0.363	0.366	0.372	4.7
1,2-Dichloropropane	0.309	0.362	0.321	0.347	0.333	0.337	0.335	5.5
Bromodichloromethane	0.570	0.592	0.538	0.573	0.538	0.567	0.563	3.8
4-Methyl-2-Pentanone	0.366	0.422	0.377	0.384	0.362	0.374	0.381	5.7
Toluene	0.881	0.944	0.862	0.910	0.849	0.925	0.895	4.2

* Compounds with required minimum RRF and maximum %RSD values.
All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: <u>Alliance</u>	Contract: <u>GARD04</u>
Lab Code: <u>ACE</u>	SDG No.: <u>Q3361</u>
Instrument ID: <u>MSVOA_X</u>	Calibration Date(s): <u>10/20/2025</u> <u>10/20/2025</u>
Heated Purge: (Y/N) <u>N</u>	Calibration Time(s): <u>10:59</u> <u>13:49</u>
GC Column: <u>DB-624UI</u> ID: <u>0.18</u> (mm)	

LAB FILE ID:		RRF001 = VX048229.D		RRF020 = VX048231.D		RRF050 = VX048232.D		
		RRF100 = VX048233.D		RRF150 = VX048234.D		RRF005 = VX048236.D		
COMPOUND	RRF001	RRF020	RRF050	RRF100	RRF150	RRF005	RRF	% RSD
t-1,3-Dichloropropene	0.477	0.582	0.545	0.574	0.553	0.526	0.543	7
cis-1,3-Dichloropropene	0.582	0.604	0.572	0.615	0.585	0.556	0.586	3.7
1,1,2-Trichloroethane	0.318	0.357	0.321	0.333	0.313	0.355	0.333	5.8
2-Hexanone	0.231	0.298	0.267	0.271	0.256	0.271	0.266	8.3
Dibromochloromethane	0.482	0.476	0.425	0.442	0.418	0.441	0.447	5.9
1,2-Dibromoethane	0.346	0.393	0.342	0.359	0.336	0.353	0.355	5.7
Tetrachloroethene	0.418	0.378	0.328	0.370	0.361	0.385	0.373	7.9
Chlorobenzene	1.149	1.159	1.045	1.148	1.102	1.145	1.125	3.9
Ethyl Benzene	1.905	1.954	1.825	2.008	1.922	1.876	1.915	3.3
m/p-Xylenes	0.647	0.750	0.687	0.750	0.707	0.727	0.712	5.6
o-Xylene	0.652	0.712	0.664	0.723	0.686	0.684	0.687	3.9
Styrene	1.054	1.236	1.143	1.247	1.173	1.165	1.170	6
Bromoform	0.338	0.350	0.308	0.335	0.326	0.325	0.330	4.3
Isopropylbenzene	3.618	3.656	3.430	3.875	3.754	3.579	3.652	4.2
1,1,2,2-Tetrachloroethane	0.988	1.065	0.946	1.018	0.992	0.974	0.997	4.1
1,3-Dichlorobenzene	1.863	1.742	1.550	1.747	1.707	1.750	1.726	5.9
1,4-Dichlorobenzene	2.143	1.803	1.577	1.724	1.710	1.772	1.788	10.6
1,2-Dichlorobenzene	1.803	1.631	1.494	1.619	1.588	1.590	1.621	6.3
1,2-Dibromo-3-Chloropropane	0.167	0.219	0.195	0.224	0.223	0.187	0.202	11.5
1,2,4-Trichlorobenzene	1.441	0.975	0.959	1.114	1.094	0.994	1.096	16.5
1,2,3-Trichlorobenzene	1.154	0.942	0.909	1.050	1.033	0.947	1.006	9
1,2-Dichloroethane-d4		0.709	0.682	0.690	0.698	0.725	0.701	2.4
Dibromofluoromethane		0.345	0.330	0.346	0.348	0.338	0.341	2.2
Toluene-d8		1.185	1.185	1.196	1.195	1.221	1.196	1.2
4-Bromofluorobenzene		0.445	0.444	0.446	0.441	0.500	0.455	5.5

* Compounds with required minimum RRF and maximum %RSD values.
All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.