

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QC Batch ID # VX101725

Review By	John Carlone	Review On	10/20/2025 8:29:16 AM
Supervise By	Maresh Dadoda	Supervise On	10/20/2025 3:54:51 PM
SubDirectory	VX101725	HP Acquire Method	HP Processing Method 82X101725W.M

STD. NAME	STD REF.#
Tune/Reschk	VP135829
Initial Calibration Stds	VP135861,VP135862,VP135863,VP135864,VP135865,VP135866
CCC	VP135830
Internal Standard/PEM	
ICV/I.BLK	VP135867
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VX048197.D	17 Oct 2025 08:17		JC/MD	Ok
2	VSTDIC001	VSTDIC001	VX048198.D	17 Oct 2025 08:45	pH#Lot#V12668	JC/MD	Ok,M
3	VSTDIC005	VSTDIC005	VX048199.D	17 Oct 2025 09:14		JC/MD	Ok,M
4	VSTDIC020	VSTDIC020	VX048200.D	17 Oct 2025 09:35		JC/MD	Ok,M
5	VSTDIC050	VSTDIC050	VX048201.D	17 Oct 2025 09:55	Comp. #58 on Linear Regression	JC/MD	Ok,M
6	VSTDIC100	VSTDIC100	VX048202.D	17 Oct 2025 10:16		JC/MD	Ok,M
7	VSTDIC150	VSTDIC150	VX048203.D	17 Oct 2025 10:36		JC/MD	Ok,M
8	IBLK	IBLK	VX048204.D	17 Oct 2025 10:57		JC/MD	Ok
9	VSTDICV050	ICVVX101725	VX048205.D	17 Oct 2025 12:05	ICV Failed for comp. #13 for DOD	JC/MD	Ok,M
10	VX1017MBL01	VX1017MBL01	VX048206.D	17 Oct 2025 12:54		JC/MD	Ok
11	VX1017WBL01	VX1017WBL01	VX048207.D	17 Oct 2025 13:15		JC/MD	Ok
12	Q3311-12DL	MW-214DDL	VX048208.D	17 Oct 2025 13:43	vial B pH<2	JC/MD	Ok
13	VX1017WBS01	VX1017WBS01	VX048209.D	17 Oct 2025 14:03		JC/MD	Ok,M
14	VX1017WBSD01	VX1017WBSD01	VX048210.D	17 Oct 2025 14:30		JC/MD	Ok,M
15	Q3361-02	Trip blank 251015076-0	VX048211.D	17 Oct 2025 14:51	vial B pH#6.0 TB	JC/MD	Ok
16	Q3362-02	Trip blank 251015076-0	VX048212.D	17 Oct 2025 15:11	vial B pH#6.0 TB	JC/MD	Ok
17	PB170117TB	PB170117TB	VX048213.D	17 Oct 2025 15:32		JC/MD	Ok

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QC Batch ID # VX101725

Review By	John Carlone	Review On	10/20/2025 8:29:16 AM
Supervise By	Maresh Dadoda	Supervise On	10/20/2025 3:54:51 PM
SubDirectory	VX101725	HP Acquire Method	HP Processing Method 82X101725W.M
STD. NAME	STD REF.#		
Tune/Reschk	VP135829		
Initial Calibration Stds	VP135861,VP135862,VP135863,VP135864,VP135865,VP135866		
CCC	VP135830		
Internal Standard/PEM			
ICV/I.BLK	VP135867		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

18	Q3348-04	WCS-TP1	VX048214.D	17 Oct 2025 15:52	vial B pH#5.0	JC/MD	Ok,M
19	Q3352-04	TP-3	VX048215.D	17 Oct 2025 16:13	vial A pH#5.0	JC/MD	Ok
20	Q3354-02	HT1403	VX048216.D	17 Oct 2025 16:34	vial A pH#5.0	JC/MD	Ok
21	Q3357-01	DRUM-1	VX048217.D	17 Oct 2025 16:54	vial A pH#5.0	JC/MD	Ok
22	Q3357-02	DRUM-2	VX048218.D	17 Oct 2025 17:15	vial A pH#5.0	JC/MD	Ok
23	Q3357-03	DRUM-3	VX048219.D	17 Oct 2025 17:36	vial A pH#5.0	JC/MD	Ok
24	Q3357-04	DRUM-4	VX048220.D	17 Oct 2025 17:56	vial A pH#5.0	JC/MD	Ok
25	Q3362-01	VOA 251015085-01	VX048221.D	17 Oct 2025 18:17	vial C pH#6.0 Surrogate Fail	JC/MD	Ok
26	Q3361-01	VOA 251015111-01	VX048222.D	17 Oct 2025 18:37	vial C pH#6.0 Need 10X	JC/MD	Dilution
27	IBLK	IBLK	VX048223.D	17 Oct 2025 18:58		JC/MD	Ok
28	Q3349-01	PURS Station	VX048224.D	17 Oct 2025 19:19	vial A pH#5.0 Oil sample	JC/MD	Ok
29	VSTDCCC050	VSTDCCC050EC	VX048225.D	17 Oct 2025 19:39		JC/MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QC Batch ID # VX102025

Review By	John Carlone	Review On	10/21/2025 8:15:04 AM
Supervise By	Semsettin Yesilyurt	Supervise On	10/21/2025 10:44:02 AM
SubDirectory	VX102025	HP Acquire Method	HP Processing Method 82X102025W.M

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP135849 VP135851,VP135852,VP135853,VP135854,VP135855,VP135856
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP135850

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VX048226.D	20 Oct 2025 08:15		JC/MD	Ok
2	VSTDCCC050	VSTDCCC050	VX048227.D	20 Oct 2025 10:13		JC/MD	Ok
3	IBLK	IBLK	VX048228.D	20 Oct 2025 10:38		JC/MD	Not Ok
4	VSTDICC001	VSTDICC001	VX048229.D	20 Oct 2025 10:59	LR-58,94	JC/MD	Ok,M
5	VSTDICC005	VSTDICC005	VX048230.D	20 Oct 2025 11:20		JC/MD	Not Ok
6	VSTDICC020	VSTDICC020	VX048231.D	20 Oct 2025 11:40		JC/MD	Ok,M
7	VSTDICCC050	VSTDICCC050	VX048232.D	20 Oct 2025 12:01		JC/MD	Ok,M
8	VSTDICC100	VSTDICC100	VX048233.D	20 Oct 2025 12:22		JC/MD	Ok,M
9	VSTDICC150	VSTDICC150	VX048234.D	20 Oct 2025 12:43		JC/MD	Ok,M
10	IBLK	IBLK	VX048235.D	20 Oct 2025 13:04		JC/MD	Ok
11	VSTDICC005	VSTDICC005	VX048236.D	20 Oct 2025 13:49		JC/MD	Ok,M
12	VSTDICV050	ICVVX102025	VX048237.D	20 Oct 2025 14:28	pH#Lot#V12668	JC/MD	Ok,M
13	VX1020MBL01	VX1020MBL01	VX048238.D	20 Oct 2025 14:56		JC/MD	Ok
14	VX1020WBL01	VX1020WBL01	VX048239.D	20 Oct 2025 15:23		JC/MD	Ok
15	VX1020WBS01	VX1020WBS01	VX048240.D	20 Oct 2025 15:50		JC/MD	Ok,M
16	Q3371-03	VPB184-HYD-2025101	VX048241.D	20 Oct 2025 16:18	vial A pH<2 Surrogate Fail	JC/MD	ReRun
17	Q3371-06	VPB184-HYD-2025101	VX048242.D	20 Oct 2025 16:38	vial A pH<2 Surrogate Fail	JC/MD	ReRun
18	Q3371-01	BP-VPB-184-TB-20251	VX048243.D	20 Oct 2025 16:59	vial A pH<2 TB; Surrogate Fail	JC/MD	ReRun

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QC Batch ID # VX102025

Review By	John Carlone	Review On	10/21/2025 8:15:04 AM
Supervise By	Semsettin Yesilyurt	Supervise On	10/21/2025 10:44:02 AM
SubDirectory	VX102025	HP Acquire Method	HP Processing Method 82X102025W.M
STD. NAME	STD REF.#		
Tune/Reschk	VP135849		
Initial Calibration Stds	VP135851,VP135852,VP135853,VP135854,VP135855,VP135856		
CCC	VP135850		
Internal Standard/PEM			
ICV/I.BLK			
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

19	Q3371-02	BP-VPB-184-EB-20251	VX048244.D	20 Oct 2025 17:20	vial A pH<2 EB; Surrogate Fail; Hit of com.#16	JC/MD	ReRun
20	Q3371-04	BP-VPB-184-GW-915-9	VX048245.D	20 Oct 2025 17:41	vial A pH<2	JC/MD	Ok
21	Q3371-05	BP-VPB-184-GW-920-9	VX048246.D	20 Oct 2025 18:02	vial A pH<2	JC/MD	Ok
22	VX1020WBSD01	VX1020WBSD01	VX048247.D	20 Oct 2025 18:22		JC/MD	Ok,M
23	VX1020MBS01	VX1020MBS01	VX048248.D	20 Oct 2025 18:43		JC/MD	Ok,M
24	Q3390-01	4174	VX048249.D	20 Oct 2025 19:04	cloth material	JC/MD	Ok
25	VSTDCCC050	VSTDCCC050EC	VX048250.D	20 Oct 2025 19:25		JC/MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QC Batch ID # VX102125

Review By	John Carlone	Review On	10/23/2025 1:52:10 PM
Supervise By	Maresh Dadoda	Supervise On	10/24/2025 2:32:28 PM
SubDirectory	VX102125	HP Acquire Method	HP Processing Method 82X102025W.M

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP135868
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP135869,VP135870

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VX048251.D	21 Oct 2025 08:12		JC/MD	Ok
2	VSTDCCC050	VSTDCCC050	VX048252.D	21 Oct 2025 08:43	pH#Lot#V12668	JC/MD	Ok,M
3	VX1021MBL01	VX1021MBL01	VX048253.D	21 Oct 2025 09:12		JC/MD	Ok
4	VX1021WBL01	VX1021WBL01	VX048254.D	21 Oct 2025 09:40		JC/MD	Ok
5	VX1021WBS01	VX1021WBS01	VX048255.D	21 Oct 2025 10:06		JC/MD	Ok,M
6	VX1021WBSD01	VX1021WBSD01	VX048256.D	21 Oct 2025 10:33		JC/MD	Ok,M
7	Q3371-03	VPB184-HYD-2025101	VX048257.D	21 Oct 2025 10:54	vial B pH<2	JC/MD	Ok
8	Q3371-01	BP-VPB-184-TB-20251	VX048258.D	21 Oct 2025 11:15	vial B pH<2 TB	JC/MD	Ok
9	Q3371-02	BP-VPB-184-EB-20251	VX048259.D	21 Oct 2025 11:35	vial B pH<2 EB	JC/MD	Ok
10	Q3371-06	VPB184-HYD-2025101	VX048260.D	21 Oct 2025 11:56	vial B pH<2	JC/MD	Ok
11	Q3413-02	BP-VPB-184-TB-20251	VX048261.D	21 Oct 2025 12:17	vial A pH<2 TB	JC/MD	Ok
12	PB170148TB	PB170148TB	VX048262.D	21 Oct 2025 12:37		JC/MD	Ok
13	Q3369-04	MH-22	VX048263.D	21 Oct 2025 12:58	vial A pH#5.0	JC/MD	Ok
14	Q3369-08	MH-21	VX048264.D	21 Oct 2025 13:19	vial A pH#5.0	JC/MD	Ok
15	Q3374-04	TP-4-WC	VX048265.D	21 Oct 2025 13:40	vial A pH#5.0	JC/MD	Ok,M
16	Q3374-08	TP-5-WC	VX048266.D	21 Oct 2025 14:00	vial A pH#5.0	JC/MD	Ok
17	Q3381-02	VNJ-222	VX048267.D	21 Oct 2025 14:21	vial A pH#5.0	JC/MD	Ok
18	Q3381-04	VNJ-245	VX048268.D	21 Oct 2025 14:42	vial A pH#5.0	JC/MD	Ok

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QC Batch ID # VX102125

Review By	John Carlone	Review On	10/23/2025 1:52:10 PM
Supervise By	Maresh Dadoda	Supervise On	10/24/2025 2:32:28 PM
SubDirectory	VX102125	HP Acquire Method	HP Processing Method 82X102025W.M
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	VP135868		
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP135869,VP135870		

19	Q3381-06	RBR200059	VX048269.D	21 Oct 2025 15:02	vial A pH#5.0	JC/MD	Ok
20	Q3381-08	VNJ-255	VX048270.D	21 Oct 2025 15:23	vial A pH#5.0	JC/MD	Ok
21	Q3392-04	MH-20	VX048271.D	21 Oct 2025 15:44	vial A pH#5.0	JC/MD	Ok
22	Q3392-08	MH-19	VX048272.D	21 Oct 2025 16:04	vial A pH#5.0	JC/MD	Ok
23	IBLK	IBLK	VX048273.D	21 Oct 2025 16:25		JC/MD	Ok
24	Q3400-01	FRAC-TANK-0760450	VX048274.D	21 Oct 2025 16:46	vial A+B pH<2	JC/MD	Ok
25	Q3361-01DL	VOA 251015111-01DL	VX048275.D	21 Oct 2025 17:07	vial B pH#6.0	JC/MD	Ok
26	IBLK	IBLK	VX048276.D	21 Oct 2025 17:27		JC/MD	Ok
27	Q3376-07	PR132-FB01-20251016	VX048277.D	21 Oct 2025 17:48	vial A pH<2	JC/MD	Ok
28	VSTDCCC050	VSTDCCC050EC	VX048278.D	21 Oct 2025 18:09		JC/MD	Ok,M

M : Manual Integration