

Surrogate Summary

SDG No.: **Q3361**

Client: **Garden State Laboratories, Inc.**

Analytical Method: **SW8260-Low**

Lab Sample ID	Client ID	Parameter	Spike	Result	Recovery (%)	Qual	Limits (%)	
							Low	High
Q3361-01	VOA 251015111-01	1,2-Dichloroethane-d4	50	50.0	100		74	125
		Dibromofluoromethane	50	50.0	100		75	124
		Toluene-d8	50	48.0	96		86	113
		4-Bromofluorobenzene	50	58.8	118		77	121
Q3361-01DL	VOA 251015111-01DL	1,2-Dichloroethane-d4	50	55.0	110		74	125
		Dibromofluoromethane	50	48.4	97		75	124
		Toluene-d8	50	47.9	96		86	113
		4-Bromofluorobenzene	50	57.9	116		77	121
Q3361-02	Trip blank 251015076-04	1,2-Dichloroethane-d4	50	57.9	116		74	125
		Dibromofluoromethane	50	50.4	101		75	124
		Toluene-d8	50	47.6	95		86	113
		4-Bromofluorobenzene	50	57.9	116		77	121
VX1017WBL01	VX1017WBL01	1,2-Dichloroethane-d4	50	53.1	106		74	125
		Dibromofluoromethane	50	51.2	102		75	124
		Toluene-d8	50	50.0	100		86	113
		4-Bromofluorobenzene	50	59.1	118		77	121
VX1017WBS01	VX1017WBS01	1,2-Dichloroethane-d4	50	43.7	87		74	125
		Dibromofluoromethane	50	51.2	102		75	124
		Toluene-d8	50	48.1	96		86	113
		4-Bromofluorobenzene	50	47.9	96		77	121
VX1017WBSD0	VX1017WBSD01	1,2-Dichloroethane-d4	50	45.5	91		74	125
		Dibromofluoromethane	50	50.6	101		75	124
		Toluene-d8	50	47.7	95		86	113
		4-Bromofluorobenzene	50	48.3	97		77	121
VX1021WBL01	VX1021WBL01	1,2-Dichloroethane-d4	50	52.5	105		74	125
		Dibromofluoromethane	50	47.1	94		75	124
		Toluene-d8	50	46.4	93		86	113
		4-Bromofluorobenzene	50	52.6	105		77	121
VX1021WBS01	VX1021WBS01	1,2-Dichloroethane-d4	50	49.2	98		74	125
		Dibromofluoromethane	50	49.3	99		75	124
		Toluene-d8	50	47.8	96		86	113
		4-Bromofluorobenzene	50	49.7	99		77	121