

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.:	<u>Q3361</u>	Analytical Method:	<u>SW8260-Low</u>
Client:	<u>Garden State Laboratories, Inc.</u>	Datafile :	<u>VX048210.D</u>

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VX1017WBSD01	Dichlorodifluoromethane	20	17.5	ug/L	88	4		69	116	19
	Chloromethane	20	17.5	ug/L	88	6		65	116	21
	Vinyl chloride	20	18.5	ug/L	93	1		65	117	19
	Bromomethane	20	18.5	ug/L	93	3		58	125	20
	Chloroethane	20	17.3	ug/L	86	5		56	128	20
	Trichlorofluoromethane	20	17.8	ug/L	89	0		73	115	16
	1,1,2-Trichlorotrifluoroethane	20	18.7	ug/L	94	4		80	112	15
	1,1-Dichloroethene	20	18.5	ug/L	93	3		74	110	20
	Acetone	100	83.7	ug/L	84	2		60	125	20
	Carbon disulfide	20	18.5	ug/L	93	1		64	112	20
	Methyl tert-butyl Ether	20	18.8	ug/L	94	2		78	114	20
	Methyl Acetate	20	19.6	ug/L	98	3		67	125	20
	Methylene Chloride	20	18.3	ug/L	92	2		72	114	20
	trans-1,2-Dichloroethene	20	18.4	ug/L	92	9		75	108	16
	1,1-Dichloroethane	20	17.7	ug/L	89	2		78	112	20
	Cyclohexane	20	20.3	ug/L	102	2		75	110	20
	2-Butanone	100	96.7	ug/L	97	2		65	122	26
	Carbon Tetrachloride	20	19.7	ug/L	99	1		77	113	15
	cis-1,2-Dichloroethene	20	18.2	ug/L	91	0		77	110	20
	Bromochloromethane	20	18.8	ug/L	94	1		70	124	20
	Chloroform	20	18.5	ug/L	93	2		79	113	20
	1,1,1-Trichloroethane	20	19.2	ug/L	96	3		80	108	20
	Methylcyclohexane	20	19.9	ug/L	100	3		72	115	20
	Benzene	20	19.2	ug/L	96	2		82	109	15
	1,2-Dichloroethane	20	20.5	ug/L	103	3		80	115	20
	Trichloroethene	20	18.5	ug/L	93	5		77	113	15
	1,2-Dichloropropane	20	19.7	ug/L	99	2		83	111	16
	Bromodichloromethane	20	19.5	ug/L	98	1		83	110	16
	4-Methyl-2-Pentanone	100	110	ug/L	110	0		74	118	25
	Toluene	20	20.0	ug/L	100	1		82	110	16
	t-1,3-Dichloropropene	20	20.5	ug/L	103	6		79	110	20
	cis-1,3-Dichloropropene	20	19.8	ug/L	99	1		82	110	16
	1,1,2-Trichloroethane	20	20.5	ug/L	103	0		83	112	20
	2-Hexanone	100	110	ug/L	110	10		73	117	25
	Dibromochloromethane	20	21.2	ug/L	106	3		82	110	20
	1,2-Dibromoethane	20	21.1	ug/L	106	2		81	110	20
	Tetrachloroethene	20	18.9	ug/L	95	0		67	123	15
	Chlorobenzene	20	19.3	ug/L	97	0		82	109	15
	Ethyl Benzene	20	18.2	ug/L	91	6		83	109	16
	m/p-Xylenes	40	37.2	ug/L	93	6		82	110	15
	o-Xylene	20	18.5	ug/L	93	7		83	109	20
	Styrene	20	19.0	ug/L	95	5		80	111	17
	Bromoform	20	18.8	ug/L	94	1		79	109	20
	Isopropylbenzene	20	19.9	ug/L	100	1		83	112	29
	1,1,2,2-Tetrachloroethane	20	20.1	ug/L	101	4		76	118	20
	1,3-Dichlorobenzene	20	19.4	ug/L	97	2		82	108	20
	1,4-Dichlorobenzene	20	18.9	ug/L	95	1		82	107	15
	1,2-Dichlorobenzene	20	19.3	ug/L	97	2		82	109	20

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q3361

Analytical Method: SW8260-Low

Client: Garden State Laboratories, Inc.

Datafile : VX048210.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits	RPD
									High	
VX1017WBSD01	1,2-Dibromo-3-Chloropropane	20	20.3	ug/L	102	7		68	112	20
	1,2,4-Trichlorobenzene	20	19.3	ug/L	97	1		75	113	29
	1,2,3-Trichlorobenzene	20	20.5	ug/L	103	6		76	114	29