

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: GARD04
 Lab Code: ACE SDG NO.: Q3361
 Lab File ID: VX048252.D Date Analyzed: 10/21/2025
 Instrument ID: MSVOA_X Time Analyzed: 08:43
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	144122	5.53	227103	6.74	197016	10.04
UPPER LIMIT	288244	6.031	454206	7.238	394032	10.537
LOWER LIMIT	72061	5.031	113552	6.238	98508	9.537
EPA SAMPLE NO.						
VOA 251015111-01DL	105623	5.54	200756	6.75	218533	10.04
VX1021WBL01	121915	5.53	240535	6.75	247603	10.04
VX1021WBS01	141199	5.54	230109	6.74	212342	10.04

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.