

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: GARD04
 Lab Code: ACE SDG NO.: Q3361
 Lab File ID: VX048252.D Date Analyzed: 10/21/2025
 Instrument ID: MSVOA_X Time Analyzed: 08:43
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	96991	12.00€				
UPPER LIMIT	193982	12.50€				
LOWER LIMIT	48495.5	11.50€				
EPA SAMPLE NO.						
VOA 251015111-01DL	120634	12.01				
VX1021WBL01	126551	12.01				
VX1021WBS01	110789	12.01				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.