



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Hit Summary Sheet
SW-846

SDG No.: Q3362
Client: Garden State Laboratories, Inc.

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	RDL	Units
-----------	-----------	--------	-----------	---------------	---	-----	-----	-------

Client ID: 0

Total Voc :
Total Concentration:

- A
- B
- C
- D



SAMPLE DATA

A

B

C

D

Report of Analysis

Client: Garden State Laboratories, Inc.
Project: Waste Water 2025
Client Sample ID: VOA 251015085-01
Lab Sample ID: Q3362-01
Analytical Method: E624.1
Sample Wt/Vol: 5 mL

Level : LOW
Final Vol: 5000 uL

Date Collected: 10/15/25
Date Received: 10/16/25
SDG No.: Q3362
Matrix: Water
% Solid: 0
Test: VOCMS Group1

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS									
107-02-8	Acrolein	6.60	U	1	6.60	25.0	ug/L	10/17/25 13:36	VN101725
107-13-1	Acrylonitrile	2.80	U	1	2.80	25.0	ug/L	10/17/25 13:36	VN101725
SURROGATES									
17060-07-0	1,2-Dichloroethane-d4	31.4			91 - 110	105%	SPK: 30		
2037-26-5	Toluene-d8	30.3			91 - 112	101%	SPK: 30		
460-00-4	4-Bromofluorobenzene	29.0			63 - 112	97%	SPK: 30		
INTERNAL STANDARDS									
		Area Count							
74-97-5	Bromochloromethane	60000							
540-36-3	1,4-Difluorobenzene	311000							
3114-55-4	Chlorobenzene-d5	282000							

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client: Garden State Laboratories, Inc.
Project: Waste Water 2025
Client Sample ID: Trip blank 251015076-03
Lab Sample ID: Q3362-02
Analytical Method: E624.1
Sample Wt/Vol: 5 mL

Level : LOW
Final Vol: 5000 uL

Date Collected: 10/15/25
Date Received: 10/16/25
SDG No.: Q3362
Matrix: Water
% Solid: 0
Test: VOCMS Group1

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS									
107-02-8	Acrolein	6.60	U	1	6.60	25.0	ug/L	10/17/25 11:53	VN101725
107-13-1	Acrylonitrile	2.80	U	1	2.80	25.0	ug/L	10/17/25 11:53	VN101725
SURROGATES									
17060-07-0	1,2-Dichloroethane-d4	32.4			91 - 110	108%	SPK: 30		
2037-26-5	Toluene-d8	30.6			91 - 112	102%	SPK: 30		
460-00-4	4-Bromofluorobenzene	28.9			63 - 112	96%	SPK: 30		
INTERNAL STANDARDS									
		Area Count							
74-97-5	Bromochloromethane	43900							
540-36-3	1,4-Difluorobenzene	238000							
3114-55-4	Chlorobenzene-d5	213000							

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

LAB CHRONICLE

OrderID:	Q3362	OrderDate:	10/16/2025 9:55:00 AM
Client:	Garden State Laboratories, Inc.	Project:	Waste Water 2025
Contact:	Sharon Ercoliani	Location:	VOA Lab

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q3362-01	VOA 251015085-01	Water	VOCMS Group1	624.1	10/15/25		10/17/25	10/16/25
Q3362-02	Trip blank 251015076-03	Water	VOCMS Group1	624.1	10/15/25		10/17/25	10/16/25

Hit Summary Sheet SW-846

SDG No.: Q3362

Client: Garden State Laboratories, Inc.

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	RDL	Units
Client ID:	VOA 251015085-01							
Q3362-01	VOA 251015085-01 Water		Dichlorodifluoromethane	0.31	J	0.22	1.00	ug/L
Q3362-01	VOA 251015085-01 Water		Chloroethane	0.66	J	0.47	1.00	ug/L
Q3362-01	VOA 251015085-01 Water		Methyl tert-butyl Ether	0.71	J	0.16	1.00	ug/L
Q3362-01	VOA 251015085-01 Water		Benzene	1.70		0.15	1.00	ug/L
Q3362-01	VOA 251015085-01 Water		Toluene	0.44	J	0.14	1.00	ug/L
Q3362-01	VOA 251015085-01 Water		Chlorobenzene	16.2		0.12	1.00	ug/L
Q3362-01	VOA 251015085-01 Water		m/p-Xylenes	3.00		0.24	2.00	ug/L
Q3362-01	VOA 251015085-01 Water		o-Xylene	2.40		0.12	1.00	ug/L
Q3362-01	VOA 251015085-01 Water		Isopropylbenzene	0.57	J	0.12	1.00	ug/L
Q3362-01	VOA 251015085-01 Water		1,4-Dichlorobenzene	7.00		0.19	1.00	ug/L
Q3362-01	VOA 251015085-01 Water		1,2-Dichlorobenzene	0.55	J	0.16	1.00	ug/L
			Total Voc :			33.5		
Q3362-01	VOA 251015085-01 Water		Tetrahydrofuran	* 170	J	0.99	5.00	ug/L
Q3362-01	VOA 251015085-01 Water		Tert butyl alcohol	* 370	J	5.50	25.0	ug/L
Q3362-01	VOA 251015085-01 Water		Diethyl Ether	* 15.8	J	0.31	1.00	ug/L
Q3362-01	VOA 251015085-01 Water		2-Chlorotoluene	* 0.51	J	0.14	1.00	ug/L
Q3362-01	VOA 251015085-01 Water		1,3,5-Trimethylbenzene	* 0.24	J	0.15	1.00	ug/L
Q3362-01	VOA 251015085-01 Water		1,2,4-Trimethylbenzene	* 1.30	J	0.14	1.00	ug/L
Q3362-01	VOA 251015085-01 Water		Naphthalene	* 3.60	J	0.20	1.00	ug/L
Q3362-01	VOA 251015085-01 Water		1,4-Dioxane	* 210	J	6.90	100	ug/L
			Total Tics :			771		
			Total Concentration:			805		



SAMPLE DATA

A

B

C

D

Report of Analysis

Client: Garden State Laboratories, Inc.
Project: Waste Water 2025
Client Sample ID: VOA 251015085-01
Lab Sample ID: Q3362-01
Analytical Method: 8260D
Sample Wt/Vol: 5 mL

Level : LOW
Final Vol: 5000 uL

Date Collected: 10/15/25
Date Received: 10/16/25
SDG No.: Q3362
Matrix: Water
% Solid: 0
Test: VOCMS Group2

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS									
75-71-8	Dichlorodifluoromethane	0.31	J	1	0.22	1.00	ug/L	10/17/25 18:17	VX101725
74-87-3	Chloromethane	0.32	U	1	0.32	1.00	ug/L	10/17/25 18:17	VX101725
75-01-4	Vinyl Chloride	0.26	U	1	0.26	1.00	ug/L	10/17/25 18:17	VX101725
74-83-9	Bromomethane	1.40	U	1	1.40	5.00	ug/L	10/17/25 18:17	VX101725
75-00-3	Chloroethane	0.66	J	1	0.47	1.00	ug/L	10/17/25 18:17	VX101725
75-69-4	Trichlorofluoromethane	0.33	U	1	0.33	1.00	ug/L	10/17/25 18:17	VX101725
76-13-1	1,1,2-Trichlorotrifluoroethane	0.25	U	1	0.25	1.00	ug/L	10/17/25 18:17	VX101725
75-35-4	1,1-Dichloroethene	0.23	U	1	0.23	1.00	ug/L	10/17/25 18:17	VX101725
67-64-1	Acetone	1.50	U	1	1.50	5.00	ug/L	10/17/25 18:17	VX101725
75-15-0	Carbon Disulfide	0.21	U	1	0.21	1.00	ug/L	10/17/25 18:17	VX101725
1634-04-4	Methyl tert-butyl Ether	0.71	J	1	0.16	1.00	ug/L	10/17/25 18:17	VX101725
79-20-9	Methyl Acetate	0.27	U	1	0.27	1.00	ug/L	10/17/25 18:17	VX101725
75-09-2	Methylene Chloride	0.28	U	1	0.28	1.00	ug/L	10/17/25 18:17	VX101725
156-60-5	trans-1,2-Dichloroethene	0.23	U	1	0.23	1.00	ug/L	10/17/25 18:17	VX101725
75-34-3	1,1-Dichloroethane	0.23	U	1	0.23	1.00	ug/L	10/17/25 18:17	VX101725
110-82-7	Cyclohexane	1.50	U	1	1.50	5.00	ug/L	10/17/25 18:17	VX101725
78-93-3	2-Butanone	0.98	U	1	0.98	5.00	ug/L	10/17/25 18:17	VX101725
56-23-5	Carbon Tetrachloride	0.25	U	1	0.25	1.00	ug/L	10/17/25 18:17	VX101725
156-59-2	cis-1,2-Dichloroethene	0.19	U	1	0.19	1.00	ug/L	10/17/25 18:17	VX101725
74-97-5	Bromochloromethane	0.22	U	1	0.22	1.00	ug/L	10/17/25 18:17	VX101725
67-66-3	Chloroform	0.25	U	1	0.25	1.00	ug/L	10/17/25 18:17	VX101725
71-55-6	1,1,1-Trichloroethane	0.20	U	1	0.20	1.00	ug/L	10/17/25 18:17	VX101725
108-87-2	Methylcyclohexane	0.16	U	1	0.16	1.00	ug/L	10/17/25 18:17	VX101725
71-43-2	Benzene	1.70		1	0.15	1.00	ug/L	10/17/25 18:17	VX101725
107-06-2	1,2-Dichloroethane	0.22	U	1	0.22	1.00	ug/L	10/17/25 18:17	VX101725
79-01-6	Trichloroethene	0.090	U	1	0.090	1.00	ug/L	10/17/25 18:17	VX101725
78-87-5	1,2-Dichloropropane	0.20	U	1	0.20	1.00	ug/L	10/17/25 18:17	VX101725
75-27-4	Bromodichloromethane	0.22	U	1	0.22	1.00	ug/L	10/17/25 18:17	VX101725
108-10-1	4-Methyl-2-Pentanone	0.68	U	1	0.68	5.00	ug/L	10/17/25 18:17	VX101725
108-88-3	Toluene	0.44	J	1	0.14	1.00	ug/L	10/17/25 18:17	VX101725
10061-02-6	t-1,3-Dichloropropene	0.17	U	1	0.17	1.00	ug/L	10/17/25 18:17	VX101725
10061-01-5	cis-1,3-Dichloropropene	0.16	U	1	0.16	1.00	ug/L	10/17/25 18:17	VX101725
79-00-5	1,1,2-Trichloroethane	0.21	U	1	0.21	1.00	ug/L	10/17/25 18:17	VX101725
591-78-6	2-Hexanone	0.89	U	1	0.89	5.00	ug/L	10/17/25 18:17	VX101725
124-48-1	Dibromochloromethane	0.18	U	1	0.18	1.00	ug/L	10/17/25 18:17	VX101725
106-93-4	1,2-Dibromoethane	0.15	U	1	0.15	1.00	ug/L	10/17/25 18:17	VX101725
127-18-4	Tetrachloroethene	0.23	U	1	0.23	1.00	ug/L	10/17/25 18:17	VX101725
108-90-7	Chlorobenzene	16.2		1	0.12	1.00	ug/L	10/17/25 18:17	VX101725
100-41-4	Ethyl Benzene	0.13	U	1	0.13	1.00	ug/L	10/17/25 18:17	VX101725
179601-23-1	m/p-Xylenes	3.00		1	0.24	2.00	ug/L	10/17/25 18:17	VX101725
95-47-6	o-Xylene	2.40		1	0.12	1.00	ug/L	10/17/25 18:17	VX101725
100-42-5	Styrene	0.15	U	1	0.15	1.00	ug/L	10/17/25 18:17	VX101725
75-25-2	Bromoform	0.19	U	1	0.19	1.00	ug/L	10/17/25 18:17	VX101725

Report of Analysis

Client: Garden State Laboratories, Inc.
Project: Waste Water 2025
Client Sample ID: VOA 251015085-01
Lab Sample ID: Q3362-01
Analytical Method: 8260D
Sample Wt/Vol: 5 mL

Level : LOW
Final Vol: 5000 uL

Date Collected: 10/15/25
Date Received: 10/16/25
SDG No.: Q3362
Matrix: Water
% Solid: 0
Test: VOCMS Group2

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
98-82-8	Isopropylbenzene	0.57	J	1	0.12	1.00	ug/L	10/17/25 18:17	VX101725
79-34-5	1,1,2,2-Tetrachloroethane	0.26	U	1	0.26	1.00	ug/L	10/17/25 18:17	VX101725
541-73-1	1,3-Dichlorobenzene	0.16	U	1	0.16	1.00	ug/L	10/17/25 18:17	VX101725
106-46-7	1,4-Dichlorobenzene	7.00		1	0.19	1.00	ug/L	10/17/25 18:17	VX101725
95-50-1	1,2-Dichlorobenzene	0.55	J	1	0.16	1.00	ug/L	10/17/25 18:17	VX101725
96-12-8	1,2-Dibromo-3-Chloropropane	0.53	U	1	0.53	1.00	ug/L	10/17/25 18:17	VX101725
120-82-1	1,2,4-Trichlorobenzene	0.20	U	1	0.20	1.00	ug/L	10/17/25 18:17	VX101725
87-61-6	1,2,3-Trichlorobenzene	0.20	U	1	0.20	1.00	ug/L	10/17/25 18:17	VX101725
SURROGATES									
17060-07-0	1,2-Dichloroethane-d4	53.5			74 - 125	107%	SPK: 50		
1868-53-7	Dibromofluoromethane	51.8			75 - 124	104%	SPK: 50		
2037-26-5	Toluene-d8	48.2			86 - 113	96%	SPK: 50		
460-00-4	4-Bromofluorobenzene	61.0	*		77 - 121	122%	SPK: 50		
INTERNAL STANDARDS									
		Area Count							
363-72-4	Pentafluorobenzene	97800							
540-36-3	1,4-Difluorobenzene	190000							
3114-55-4	Chlorobenzene-d5	206000							
3855-82-1	1,4-Dichlorobenzene-d4	112000							
TENTATIVE IDENTIFIED COMPOUNDS									
60-29-7	Diethyl Ether	15.8	J			2.14	ug/L		
75-65-0	Tert butyl alcohol	370	J			2.94	ug/L		
109-99-9	Tetrahydrofuran	170	J			5.00	ug/L		
123-91-1	1,4-Dioxane	210	J			7.65	ug/L		
95-49-8	2-Chlorotoluene	0.51	J			11.3	ug/L		
108-67-8	1,3,5-Trimethylbenzene	0.24	J			11.4	ug/L		
95-63-6	1,2,4-Trimethylbenzene	1.30	J			11.7	ug/L		
91-20-3	Naphthalene	3.60	J			13.8	ug/L		

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client: Garden State Laboratories, Inc.
Project: Waste Water 2025
Client Sample ID: Trip blank 251015076-03
Lab Sample ID: Q3362-02
Analytical Method: 8260D
Sample Wt/Vol: 5 mL

Level : LOW
Final Vol: 5000 uL

Date Collected: 10/15/25
Date Received: 10/16/25
SDG No.: Q3362
Matrix: Water
% Solid: 0
Test: VOCMS Group2

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS									
75-71-8	Dichlorodifluoromethane	0.22	U	1	0.22	1.00	ug/L	10/17/25 15:11	VX101725
74-87-3	Chloromethane	0.32	U	1	0.32	1.00	ug/L	10/17/25 15:11	VX101725
75-01-4	Vinyl Chloride	0.26	U	1	0.26	1.00	ug/L	10/17/25 15:11	VX101725
74-83-9	Bromomethane	1.40	U	1	1.40	5.00	ug/L	10/17/25 15:11	VX101725
75-00-3	Chloroethane	0.47	U	1	0.47	1.00	ug/L	10/17/25 15:11	VX101725
75-69-4	Trichlorofluoromethane	0.33	U	1	0.33	1.00	ug/L	10/17/25 15:11	VX101725
76-13-1	1,1,2-Trichlorotrifluoroethane	0.25	U	1	0.25	1.00	ug/L	10/17/25 15:11	VX101725
75-35-4	1,1-Dichloroethene	0.23	U	1	0.23	1.00	ug/L	10/17/25 15:11	VX101725
67-64-1	Acetone	1.50	U	1	1.50	5.00	ug/L	10/17/25 15:11	VX101725
75-15-0	Carbon Disulfide	0.21	U	1	0.21	1.00	ug/L	10/17/25 15:11	VX101725
1634-04-4	Methyl tert-butyl Ether	0.16	U	1	0.16	1.00	ug/L	10/17/25 15:11	VX101725
79-20-9	Methyl Acetate	0.27	U	1	0.27	1.00	ug/L	10/17/25 15:11	VX101725
75-09-2	Methylene Chloride	0.28	U	1	0.28	1.00	ug/L	10/17/25 15:11	VX101725
156-60-5	trans-1,2-Dichloroethene	0.23	U	1	0.23	1.00	ug/L	10/17/25 15:11	VX101725
75-34-3	1,1-Dichloroethane	0.23	U	1	0.23	1.00	ug/L	10/17/25 15:11	VX101725
110-82-7	Cyclohexane	1.50	U	1	1.50	5.00	ug/L	10/17/25 15:11	VX101725
78-93-3	2-Butanone	0.98	U	1	0.98	5.00	ug/L	10/17/25 15:11	VX101725
56-23-5	Carbon Tetrachloride	0.25	U	1	0.25	1.00	ug/L	10/17/25 15:11	VX101725
156-59-2	cis-1,2-Dichloroethene	0.19	U	1	0.19	1.00	ug/L	10/17/25 15:11	VX101725
74-97-5	Bromochloromethane	0.22	U	1	0.22	1.00	ug/L	10/17/25 15:11	VX101725
67-66-3	Chloroform	0.25	U	1	0.25	1.00	ug/L	10/17/25 15:11	VX101725
71-55-6	1,1,1-Trichloroethane	0.20	U	1	0.20	1.00	ug/L	10/17/25 15:11	VX101725
108-87-2	Methylcyclohexane	0.16	U	1	0.16	1.00	ug/L	10/17/25 15:11	VX101725
71-43-2	Benzene	0.15	U	1	0.15	1.00	ug/L	10/17/25 15:11	VX101725
107-06-2	1,2-Dichloroethane	0.22	U	1	0.22	1.00	ug/L	10/17/25 15:11	VX101725
79-01-6	Trichloroethene	0.090	U	1	0.090	1.00	ug/L	10/17/25 15:11	VX101725
78-87-5	1,2-Dichloropropane	0.20	U	1	0.20	1.00	ug/L	10/17/25 15:11	VX101725
75-27-4	Bromodichloromethane	0.22	U	1	0.22	1.00	ug/L	10/17/25 15:11	VX101725
108-10-1	4-Methyl-2-Pentanone	0.68	U	1	0.68	5.00	ug/L	10/17/25 15:11	VX101725
108-88-3	Toluene	0.14	U	1	0.14	1.00	ug/L	10/17/25 15:11	VX101725
10061-02-6	t-1,3-Dichloropropene	0.17	U	1	0.17	1.00	ug/L	10/17/25 15:11	VX101725
10061-01-5	cis-1,3-Dichloropropene	0.16	U	1	0.16	1.00	ug/L	10/17/25 15:11	VX101725
79-00-5	1,1,2-Trichloroethane	0.21	U	1	0.21	1.00	ug/L	10/17/25 15:11	VX101725
591-78-6	2-Hexanone	0.89	U	1	0.89	5.00	ug/L	10/17/25 15:11	VX101725
124-48-1	Dibromochloromethane	0.18	U	1	0.18	1.00	ug/L	10/17/25 15:11	VX101725
106-93-4	1,2-Dibromoethane	0.15	U	1	0.15	1.00	ug/L	10/17/25 15:11	VX101725
127-18-4	Tetrachloroethene	0.23	U	1	0.23	1.00	ug/L	10/17/25 15:11	VX101725
108-90-7	Chlorobenzene	0.12	U	1	0.12	1.00	ug/L	10/17/25 15:11	VX101725
100-41-4	Ethyl Benzene	0.13	U	1	0.13	1.00	ug/L	10/17/25 15:11	VX101725
179601-23-1	m/p-Xylenes	0.24	U	1	0.24	2.00	ug/L	10/17/25 15:11	VX101725
95-47-6	o-Xylene	0.12	U	1	0.12	1.00	ug/L	10/17/25 15:11	VX101725
100-42-5	Styrene	0.15	U	1	0.15	1.00	ug/L	10/17/25 15:11	VX101725
75-25-2	Bromoform	0.19	U	1	0.19	1.00	ug/L	10/17/25 15:11	VX101725

Report of Analysis

Client: Garden State Laboratories, Inc.
Project: Waste Water 2025
Client Sample ID: Trip blank 251015076-03
Lab Sample ID: Q3362-02
Analytical Method: 8260D
Sample Wt/Vol: 5 mL

Level : LOW
Final Vol: 5000 uL

Date Collected: 10/15/25
Date Received: 10/16/25
SDG No.: Q3362
Matrix: Water
% Solid: 0
Test: VOCMS Group2

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
98-82-8	Isopropylbenzene	0.12	U	1	0.12	1.00	ug/L	10/17/25 15:11	VX101725
79-34-5	1,1,2,2-Tetrachloroethane	0.26	U	1	0.26	1.00	ug/L	10/17/25 15:11	VX101725
541-73-1	1,3-Dichlorobenzene	0.16	U	1	0.16	1.00	ug/L	10/17/25 15:11	VX101725
106-46-7	1,4-Dichlorobenzene	0.19	U	1	0.19	1.00	ug/L	10/17/25 15:11	VX101725
95-50-1	1,2-Dichlorobenzene	0.16	U	1	0.16	1.00	ug/L	10/17/25 15:11	VX101725
96-12-8	1,2-Dibromo-3-Chloropropane	0.53	U	1	0.53	1.00	ug/L	10/17/25 15:11	VX101725
120-82-1	1,2,4-Trichlorobenzene	0.20	U	1	0.20	1.00	ug/L	10/17/25 15:11	VX101725
87-61-6	1,2,3-Trichlorobenzene	0.20	U	1	0.20	1.00	ug/L	10/17/25 15:11	VX101725
SURROGATES									
17060-07-0	1,2-Dichloroethane-d4	54.1			74 - 125	108%	SPK: 50		
1868-53-7	Dibromofluoromethane	58.5			75 - 124	117%	SPK: 50		
2037-26-5	Toluene-d8	47.6			86 - 113	95%	SPK: 50		
460-00-4	4-Bromofluorobenzene	58.3			77 - 121	117%	SPK: 50		
INTERNAL STANDARDS									
363-72-4	Pentafluorobenzene	101000							
540-36-3	1,4-Difluorobenzene	201000							
3114-55-4	Chlorobenzene-d5	215000							
3855-82-1	1,4-Dichlorobenzene-d4	114000							

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

LAB CHRONICLE

OrderID:	Q3362	OrderDate:	10/16/2025 9:55:00 AM
Client:	Garden State Laboratories, Inc.	Project:	Waste Water 2025
Contact:	Sharon Ercoliani	Location:	VOA Lab

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q3362-01	VOA 251015085-01	Water			10/15/25			10/16/25
			VOCMS Group1	624.1				
			VOCMS Group2	8260-Low				
Q3362-02	Trip blank 251015076-03	Water			10/15/25			10/16/25
			VOCMS Group1	624.1				
			VOCMS Group2	8260-Low				