

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name:	<u>Alliance</u>	Contract:	<u>GARD04</u>
Lab Code:	<u>ACE</u>	SDG No.:	<u>Q3362</u>
Instrument ID:	<u>MSVOA_N</u>	Calibration Date/Time:	<u>10/17/2025 09:29</u>
Lab File ID:	<u>VN088055.D</u>	Init. Calib. Date(s):	<u>10/08/2025 10/08/2025</u>
Heated Purge: (Y/N)	<u>N</u>	Init. Calib. Time(s):	<u>09:34 11:12</u>
GC Column:	<u>RXI-624</u>	ID:	<u>0.25</u> (mm)

COMPOUND	RRF	RRF020	MIN RRF	%D	MAX%D
Acrolein	0.397	0.365		-8.06	
Acrylonitrile	1.238	1.060		-14.38	
1,2-Dichloroethane-d4	2.336	2.445	0.01	4.67	
Toluene-d8	1.337	1.360	0.01	1.72	
4-Bromofluorobenzene	0.495	0.493	0.2	-0.4	

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.