

## DATA REPORTING QUALIFIERS- ORGANIC

For reporting results, the following “ Results Qualifiers” are used:

|       |                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                           |
|-------|-------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|
| Value | If the result is a value greater than or equal to the detection limit, report the value                                                                                                                                                                                                                                                                                                                                                                                                                                                   |
| U     | Indicates the compound was analyzed for but was not detected. Report the minimum detection limit for the sample with the U, i.e. “10 U”. This is not necessarily the instrument detection limit attainable for this particular sample based on any concentration or dilution that may have been required.                                                                                                                                                                                                                                 |
| ND    | Indicates the analyte was analyzed for, but not detected                                                                                                                                                                                                                                                                                                                                                                                                                                                                                  |
| J     | Indicates an estimated value. This flag is used:<br>(1) When estimating a concentration for a tentatively identified compound (library search hits, where a 1:1 response is assumed.)<br>(2) When the mass spectral data indicated the identification, however the result was less than the specified detection limit greater than zero. If the detection limit was 10ug/L and a concentration of 3 ug/L was calculated report as 3 J. This flag is used when similar situation arise on any organic parameter i.e. Pest, PCB and others. |
| B     | Indicates the analyte was found in the blank as well as the sample report as “12 B”.                                                                                                                                                                                                                                                                                                                                                                                                                                                      |
| E     | Indicates the analyte ‘s concentration exceeds the calibrated range of the instrument for that specific analysis.                                                                                                                                                                                                                                                                                                                                                                                                                         |
| D     | This flag identifies all compounds identified in an analysis at a secondary dilution factor.                                                                                                                                                                                                                                                                                                                                                                                                                                              |
| P     | This flag is used for Pesticide/PCB target analyte when there is >25% difference for detected concentrations between the two GC columns. The lower of the two values is reported on Form 1 and flagged with a “P”.                                                                                                                                                                                                                                                                                                                        |
| N     | This flag indicates presumptive evidence of a compound. This is only used for tentatively identified compounds (TICs), where the identification is based on a mass spectral library search. It applies to all TIC results. For generic characterization of a TIC, such as chlorinated hydrocarbon, the flag is not used.                                                                                                                                                                                                                  |
| A     | This flag indicates that a Tentatively Identified Compound is a suspected aldol-condensation product.                                                                                                                                                                                                                                                                                                                                                                                                                                     |
| Q     | Indicates the LCS did not meet the control limits requirements                                                                                                                                                                                                                                                                                                                                                                                                                                                                            |

## LAB CHRONICLE

|                                                |                                         |
|------------------------------------------------|-----------------------------------------|
| <b>OrderID:</b> Q3362                          | <b>OrderDate:</b> 10/16/2025 9:55:00 AM |
| <b>Client:</b> Garden State Laboratories, Inc. | <b>Project:</b> Waste Water 2025        |
| <b>Contact:</b> Sharon Ercoliani               | <b>Location:</b> VOA Lab                |

| LabID    | ClientID                   | Matrix | Test         | Method | Sample Date | Prep Date | Anal Date | Received |
|----------|----------------------------|--------|--------------|--------|-------------|-----------|-----------|----------|
| Q3362-01 | VOA 251015085-01           | Water  | VOCMS Group1 | 624.1  | 10/15/25    |           | 10/17/25  | 10/16/25 |
| Q3362-02 | Trip blank<br>251015076-03 | Water  | VOCMS Group1 | 624.1  | 10/15/25    |           | 10/17/25  | 10/16/25 |



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,  
Fax : 908 789 8922

**Hit Summary Sheet**  
**SW-846**

**SDG No.:** Q3362

**Client:** Garden State Laboratories, Inc.

| Sample ID | Client ID | Matrix | Parameter | Concentration | C | MDL | RDL | Units |
|-----------|-----------|--------|-----------|---------------|---|-----|-----|-------|
|-----------|-----------|--------|-----------|---------------|---|-----|-----|-------|

**Client ID:**

0

**Total Voc :**

**Total Concentration:**



# QC SUMMARY

## Surrogate Summary

SDG No.: Q3362

Client: Garden State Laboratories, Inc.

Analytical Method: SW624.1

| Lab Sample ID | Client ID               | Parameter             | Spike | Result | Recovery (%) | Qual | Limits (%) |      |
|---------------|-------------------------|-----------------------|-------|--------|--------------|------|------------|------|
|               |                         |                       |       |        |              |      | Low        | High |
| Q3362-01      | VOA 251015085-01        | 1,2-Dichloroethane-d4 | 30    | 31.4   | 105          |      | 91         | 110  |
|               |                         | Toluene-d8            | 30    | 30.3   | 101          |      | 91         | 112  |
|               |                         | 4-Bromofluorobenzene  | 30    | 29.0   | 97           |      | 63         | 112  |
| Q3362-02      | Trip blank 251015076-03 | 1,2-Dichloroethane-d4 | 30    | 32.4   | 108          |      | 91         | 110  |
|               |                         | Toluene-d8            | 30    | 30.6   | 102          |      | 91         | 112  |
|               |                         | 4-Bromofluorobenzene  | 30    | 28.9   | 96           |      | 63         | 112  |
| VN1017WBL01   | VN1017WBL01             | 1,2-Dichloroethane-d4 | 30    | 32.4   | 108          |      | 91         | 110  |
|               |                         | Toluene-d8            | 30    | 31.5   | 105          |      | 91         | 112  |
|               |                         | 4-Bromofluorobenzene  | 30    | 28.4   | 95           |      | 63         | 112  |
| VN1017WBS01   | VN1017WBS01             | 1,2-Dichloroethane-d4 | 30    | 32.0   | 107          |      | 91         | 110  |
|               |                         | Toluene-d8            | 30    | 30.9   | 103          |      | 91         | 112  |
|               |                         | 4-Bromofluorobenzene  | 30    | 29.9   | 100          |      | 63         | 112  |
| VN1017WBSD0   | VN1017WBSD01            | 1,2-Dichloroethane-d4 | 30    | 31.5   | 105          |      | 91         | 110  |
|               |                         | Toluene-d8            | 30    | 30.6   | 102          |      | 91         | 112  |
|               |                         | 4-Bromofluorobenzene  | 30    | 29.9   | 100          |      | 63         | 112  |

**Laboratory Control Sample/Laboratory Control Sample Duplicate Summary**

**SW-846**

|                 |                                        |                           |                   |
|-----------------|----------------------------------------|---------------------------|-------------------|
| <b>SDG No.:</b> | <u>Q3362</u>                           | <b>Analytical Method:</b> | <u>SW624.1</u>    |
| <b>Client:</b>  | <u>Garden State Laboratories, Inc.</u> | <b>Datafile :</b>         | <u>VN088056.D</u> |

| Lab Sample ID | Parameter     | Spike | Result | Unit | Rec | RPD | Qual | Low | Limits | RPD |
|---------------|---------------|-------|--------|------|-----|-----|------|-----|--------|-----|
|               |               |       |        |      |     |     |      |     | High   |     |
| VN1017WBS01   | Acrolein      | 100   | 84.0   | ug/L | 84  |     |      | 60  | 140    |     |
|               | Acrylonitrile | 100   | 91.3   | ug/L | 91  |     |      | 60  | 140    |     |

**Laboratory Control Sample/Laboratory Control Sample Duplicate Summary**

**SW-846**

|                 |                                               |                           |                          |
|-----------------|-----------------------------------------------|---------------------------|--------------------------|
| <b>SDG No.:</b> | <u><b>Q3362</b></u>                           | <b>Analytical Method:</b> | <u><b>SW624.1</b></u>    |
| <b>Client:</b>  | <u><b>Garden State Laboratories, Inc.</b></u> | <b>Datafile :</b>         | <u><b>VN088057.D</b></u> |

| Lab Sample ID | Parameter     | Spike | Result | Unit | Rec | RPD | Qual | Low | Limits | RPD |
|---------------|---------------|-------|--------|------|-----|-----|------|-----|--------|-----|
|               |               |       |        |      |     |     |      |     | High   |     |
| VN1017WBSD01  | Acrolein      | 100   | 91.8   | ug/L | 92  | 9   |      | 60  | 140    | 20  |
|               | Acrylonitrile | 100   | 90.6   | ug/L | 91  | 0   |      | 60  | 140    | 20  |

VOLATILE METHOD BLANK SUMMARY

Client ID

VN1017WBL01

Lab Name: Alliance

Contract: GARD04

Lab Code: ACE

SDG NO.: Q3362

Lab File ID: VN088058.D

Lab Sample ID: VN1017WBL01

Date Analyzed: 10/17/2025

Time Analyzed: 10:56

GC Column: RXI-624 ID: 0.25 (mm)

Heated Purge: (Y/N) N

Instrument ID: MSVOA\_N

THIS METHOD BLANK APPLIES TO THE FOLLOWING SAMPLES, MS AND MSD:

| EPA<br>SAMPLE NO.       | LAB<br>SAMPLE ID | LAB<br>FILE ID | DATE<br>ANALYZED |
|-------------------------|------------------|----------------|------------------|
| VN1017WBS01             | VN1017WBS01      | VN088056.D     | 10/17/2025       |
| VN1017WBSD01            | VN1017WBSD01     | VN088057.D     | 10/17/2025       |
| Trip blank 251015076-03 | Q3362-02         | VN088060.D     | 10/17/2025       |
| VOA 251015085-01        | Q3362-01         | VN088065.D     | 10/17/2025       |

COMMENTS: \_\_\_\_\_  
\_\_\_\_\_





284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,  
Fax : 908 789 8922

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK  
BROMOFLUOROBENZENE (BFB)

Lab Name: Alliance

Contract: GARD04

Lab Code: ACE

SDG NO.: Q3362

Lab File ID: VN088029.D

BFB Injection Date: 10/08/2025

Instrument ID: MSVOA\_N

BFB Injection Time: 09:00

GC Column: RXI-624 ID: 0.25 (mm)

Heated Purge: Y/N N

| m/e | ION ABUNDANCE CRITERIA             | % RELATIVE ABUNDANCE |
|-----|------------------------------------|----------------------|
| 50  | 15.0 - 40.0% of mass 95            | 20                   |
| 75  | 30.0 - 60.0% of mass 95            | 52.6                 |
| 95  | Base Peak, 100% relative abundance | 100                  |
| 96  | 5.0 - 9.0% of mass 95              | 6.4                  |
| 173 | Less than 2.0% of mass 174         | 0.5 ( 0.7 ) 1        |
| 174 | 50.0 - 100.0% of mass 95           | 72.1                 |
| 175 | 5.0 - 9.0% of mass 174             | 5.6 ( 7.8 ) 1        |
| 176 | 95.0 - 101.0% of mass 174          | 69.5 ( 96.4 ) 1      |
| 177 | 5.0 - 9.0% of mass 176             | 4.7 ( 6.8 ) 2        |

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

| CLIENT ID | LAB SAMPLE ID | LAB FILE ID | DATE ANALYZED | TIME ANALYZED |
|-----------|---------------|-------------|---------------|---------------|
| VSTDIC005 | VSTDIC005     | VN088030.D  | 10/08/2025    | 09:34         |
| VSTDIC020 | VSTDIC020     | VN088031.D  | 10/08/2025    | 10:08         |
| VSTDIC050 | VSTDIC050     | VN088032.D  | 10/08/2025    | 10:30         |
| VSTDIC100 | VSTDIC100     | VN088033.D  | 10/08/2025    | 10:51         |
| VSTDIC150 | VSTDIC150     | VN088034.D  | 10/08/2025    | 11:12         |



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,  
Fax : 908 789 8922

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK  
BROMOFLUOROBENZENE (BFB)

Lab Name: Alliance

Contract: GARD04

Lab Code: ACE

SDG NO.: Q3362

Lab File ID: VN088054.D

BFB Injection Date: 10/17/2025

Instrument ID: MSVOA\_N

BFB Injection Time: 08:56

GC Column: RXI-624 ID: 0.25 (mm)

Heated Purge: Y/N N

| m/e | ION ABUNDANCE CRITERIA             | % RELATIVE ABUNDANCE |
|-----|------------------------------------|----------------------|
| 50  | 15.0 - 40.0% of mass 95            | 21.6                 |
| 75  | 30.0 - 60.0% of mass 95            | 53.9                 |
| 95  | Base Peak, 100% relative abundance | 100                  |
| 96  | 5.0 - 9.0% of mass 95              | 7.2                  |
| 173 | Less than 2.0% of mass 174         | 0.8 ( 1.2 ) 1        |
| 174 | 50.0 - 100.0% of mass 95           | 67.1                 |
| 175 | 5.0 - 9.0% of mass 174             | 5 ( 7.5 ) 1          |
| 176 | 95.0 - 101.0% of mass 174          | 64.1 ( 95.6 ) 1      |
| 177 | 5.0 - 9.0% of mass 176             | 4.3 ( 6.8 ) 2        |

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

| CLIENT ID               | LAB SAMPLE ID | LAB FILE ID | DATE ANALYZED | TIME ANALYZED |
|-------------------------|---------------|-------------|---------------|---------------|
| VSTDCCC020              | VSTDCCC020    | VN088055.D  | 10/17/2025    | 09:29         |
| VN1017WBS01             | VN1017WBS01   | VN088056.D  | 10/17/2025    | 10:03         |
| VN1017WBSD01            | VN1017WBSD01  | VN088057.D  | 10/17/2025    | 10:36         |
| VN1017WBL01             | VN1017WBL01   | VN088058.D  | 10/17/2025    | 10:56         |
| Trip blank 251015076-03 | Q3362-02      | VN088060.D  | 10/17/2025    | 11:53         |
| VOA 251015085-01        | Q3362-01      | VN088065.D  | 10/17/2025    | 13:36         |

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: GARD04  
 Lab Code: ACE SDG NO.: Q3362  
 Lab File ID: VN088055.D Date Analyzed: 10/17/2025  
 Instrument ID: MSVOA\_N Time Analyzed: 09:29  
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

|                         | IS1<br>AREA # | RT #  | IS2<br>AREA # | RT #  | IS3<br>AREA # | RT #   |
|-------------------------|---------------|-------|---------------|-------|---------------|--------|
| 12 HOUR STD             | 68741         | 7.79  | 379209        | 9.08  | 339769        | 11.84  |
| UPPER LIMIT             | 137482        | 8.294 | 758418        | 9.576 | 679538        | 12.341 |
| LOWER LIMIT             | 34370.5       | 7.294 | 189605        | 8.576 | 169885        | 11.341 |
| EPA SAMPLE NO.          |               |       |               |       |               |        |
| VOA 251015085-01        | 59993         | 7.80  | 311111        | 9.08  | 282237        | 11.85  |
| Trip blank 251015076-03 | 43937         | 7.79  | 238376        | 9.08  | 212709        | 11.84  |
| VN1017WBL01             | 56868         | 7.79  | 316193        | 9.08  | 274787        | 11.84  |
| VN1017WBS01             | 61010         | 7.79  | 345518        | 9.08  | 303706        | 11.84  |
| VN1017WBSD01            | 65413         | 7.80  | 359810        | 9.08  | 319803        | 11.84  |

IS1 = Bromochloromethane  
 IS2 = 1,4-Difluorobenzene  
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area  
 AREA LOWER LIMIT = -50% of internal standard area  
 RT UPPER LIMIT = +0.50 minutes of internal standard RT  
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

# Column used to flag values outside QC limits with an asterisk.  
 \* Values outside of QC limits.

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

|                |                   |                |                   |
|----------------|-------------------|----------------|-------------------|
| Lab Name:      | <u>Alliance</u>   | Contract:      | <u>GARD04</u>     |
| Lab Code:      | <u>ACE</u>        | SDG NO.:       | <u>Q3362</u>      |
| Lab File ID:   | <u>VN088055.D</u> | Date Analyzed: | <u>10/17/2025</u> |
| Instrument ID: | <u>MSVOA_N</u>    | Time Analyzed: | <u>09:29</u>      |
| GC Column:     | <u>RXI-624</u>    | ID:            | <u>0.25</u> (mm)  |
|                |                   | Heated Purge:  | (Y/N) <u>N</u>    |

|                         | IS4<br>AREA # | RT # |  |  |  |  |
|-------------------------|---------------|------|--|--|--|--|
| 12 HOUR STD             | 0             | 0    |  |  |  |  |
| UPPER LIMIT             | 0             |      |  |  |  |  |
| LOWER LIMIT             | 0             |      |  |  |  |  |
| EPA SAMPLE NO.          |               |      |  |  |  |  |
| VOA 251015085-01        | 0             | 0.00 |  |  |  |  |
| Trip blank 251015076-03 | 0             | 0.00 |  |  |  |  |
| VN1017WBL01             | 0             | 0.00 |  |  |  |  |
| VN1017WBS01             | 0             | 0.00 |  |  |  |  |
| VN1017WBSD01            | 0             | 0.00 |  |  |  |  |

IS4 =

AREA UPPER LIMIT = +100% of internal standard area  
 AREA LOWER LIMIT = -50% of internal standard area  
 RT UPPER LIMIT = +0.50 minutes of internal standard RT  
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

# Column used to flag values outside QC limits with an asterisk.  
 \* Values outside of QC limits.



# SAMPLE DATA



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,  
Fax : 908 789 8922

## Report of Analysis

Client: Garden State Laboratories, Inc.

Project: Waste Water 2025

Client Sample ID: VOA 251015085-01

Lab Sample ID: Q3362-01

Analytical Method: E624.1

Sample Wt/Vol: 5 mL

Level : LOW

Final Vol: 5000 uL

Date Collected: 10/15/25

Date Received: 10/16/25

SDG No.: Q3362

Matrix: Water

% Solid: 0

Test: VOCMS Group1

| CAS Number                | Parameter             | Conc.             | Qua. | DF | MDL      | LOQ / CRQL | Units   | Date Ana.      | BatchID  |
|---------------------------|-----------------------|-------------------|------|----|----------|------------|---------|----------------|----------|
| <b>TARGETS</b>            |                       |                   |      |    |          |            |         |                |          |
| 107-02-8                  | Acrolein              | 6.60              | U    | 1  | 6.60     | 25.0       | ug/L    | 10/17/25 13:36 | VN101725 |
| 107-13-1                  | Acrylonitrile         | 2.80              | U    | 1  | 2.80     | 25.0       | ug/L    | 10/17/25 13:36 | VN101725 |
| <b>SURROGATES</b>         |                       |                   |      |    |          |            |         |                |          |
| 17060-07-0                | 1,2-Dichloroethane-d4 | 31.4              |      |    | 91 - 110 | 105%       | SPK: 30 |                |          |
| 2037-26-5                 | Toluene-d8            | 30.3              |      |    | 91 - 112 | 101%       | SPK: 30 |                |          |
| 460-00-4                  | 4-Bromofluorobenzene  | 29.0              |      |    | 63 - 112 | 97%        | SPK: 30 |                |          |
| <b>INTERNAL STANDARDS</b> |                       |                   |      |    |          |            |         |                |          |
|                           |                       | <b>Area Count</b> |      |    |          |            |         |                |          |
| 74-97-5                   | Bromochloromethane    | 60000             |      |    |          |            |         |                |          |
| 540-36-3                  | 1,4-Difluorobenzene   | 311000            |      |    |          |            |         |                |          |
| 3114-55-4                 | Chlorobenzene-d5      | 282000            |      |    |          |            |         |                |          |

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

\* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN101725\  
 Data File : VN088065.D  
 Acq On : 17 Oct 2025 13:36  
 Operator : JC\MD  
 Sample : Q3362-01  
 Misc : 5.0mL/MSVOA\_N/WATER  
 ALS Vial : 12 Sample Multiplier: 1

Instrument :  
 MSVOA\_N  
 ClientSampleId :  
 VOA 251015085-01

Manual Integrations  
 APPROVED

Reviewed By :John Carlone 10/20/2025  
 Supervised By :Mahesh Dadoda 10/20/2025

Quant Time: Oct 18 01:29:23 2025  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\624N100825W.M  
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS  
 QLast Update : Thu Oct 09 02:36:32 2025  
 Response via : Initial Calibration

| Compound                    | R.T.           | QIon | Response            | Conc    | Units  | Dev(Min) |
|-----------------------------|----------------|------|---------------------|---------|--------|----------|
| -----                       |                |      |                     |         |        |          |
| Internal Standards          |                |      |                     |         |        |          |
| 1) Bromochloromethane       | 7.800          | 128  | 59993               | 30.000  | ug/l   | 0.00     |
| 28) 1,4-Difluorobenzene     | 9.082          | 114  | 311111              | 30.000  | ug/l   | 0.00     |
| 57) Chlorobenzene-d5        | 11.847         | 117  | 282237              | 30.000  | ug/l   | 0.00     |
|                             |                |      |                     |         |        |          |
| System Monitoring Compounds |                |      |                     |         |        |          |
| 27) 1,2-Dichloroethane-d4   | 8.559          | 65   | 146632              | 31.388  | ug/l   | 0.00     |
| Spiked Amount 30.000        | Range 91 - 110 |      | Recovery = 104.633% |         |        |          |
| 60) 4-Bromofluorobenzene    | 12.829         | 95   | 134961              | 29.006  | ug/l   | 0.00     |
| Spiked Amount 30.000        | Range 63 - 112 |      | Recovery = 96.700%  |         |        |          |
| 63) Toluene-d8              | 10.547         | 98   | 381294              | 30.322  | ug/l   | 0.00     |
| Spiked Amount 30.000        | Range 91 - 112 |      | Recovery = 101.067% |         |        |          |
|                             |                |      |                     |         |        |          |
| Target Compounds            |                |      |                     |         |        | Qvalue   |
| 2) Dichlorodifluoromethane  | 2.148          | 85   | 1180m               | 0.343   | ug/l   |          |
| 6) Chloroethane             | 3.148          | 64   | 2599                | 0.970   | ug/l   | 91       |
| 8) Diethyl Ether            | 3.959          | 74   | 44491               | 17.826  | ug/l   | 93       |
| 23) tert-Butyl Alcohol      | 5.512          | 59   | 246221              | 248.079 | ug/l # | 100      |
| 24) Methyl tert-Butyl Ether | 5.800          | 73   | 8441m               | 0.620   | ug/l   |          |
| 34) Benzene                 | 8.583          | 78   | 32137               | 2.173   | ug/l # | 30       |
| 45) 1,4-Dioxane             | 9.682          | 88   | 15497               | 225.643 | ug/l # | 88       |
| 62) Toluene                 | 10.606         | 91   | 8529                | 0.533   | ug/l   | 98       |
| 64) Chlorobenzene           | 11.871         | 112  | 203043              | 19.856  | ug/l   | 98       |
| 66) Ethyl Benzene           | 11.941         | 91   | 12189               | 0.700   | ug/l   | 94       |
| 67) m/p-Xylenes             | 12.047         | 106  | 28348               | 4.224   | ug/l   | 96       |
| 68) o-Xylene                | 12.371         | 106  | 20472               | 3.064   | ug/l   | 92       |
| 70) Isopropylbenzene        | 12.670         | 105  | 17470               | 1.067   | ug/l   | 99       |
| 74) n-propylbenzene         | 13.023         | 91   | 5647                | 0.286   | ug/l   | 98       |
| 75) 2-Chlorotoluene         | 13.100         | 91   | 8603                | 0.712   | ug/l   | 99       |
| 76) 1,3,5-Trimethylbenzene  | 13.159         | 105  | 5140                | 0.370   | ug/l   | 51       |
| 80) 1,2,4-Trimethylbenzene  | 13.465         | 105  | 28157               | 2.022   | ug/l   | 96       |
| 84) 1,4-Dichlorobenzene     | 13.794         | 146  | 66054               | 8.506   | ug/l   | 99       |
| 87) 1,2-Dichlorobenzene     | 14.088         | 146  | 5312                | 0.717   | ug/l   | 95       |
| 91) Naphthalene             | 15.611         | 128  | 109504              | 6.780   | ug/l   | 99       |
| -----                       |                |      |                     |         |        |          |

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN101725\  
Data File : VN088065.D  
Acq On : 17 Oct 2025 13:36  
Operator : JC\MD  
Sample : Q3362-01  
Misc : 5.0mL/MSVOA\_N/WATER  
ALS Vial : 12 Sample Multiplier: 1

Instrument :  
MSVOA\_N  
ClientSampleId :  
VOA 251015085-01


Manual Integrations  
APPROVED

Reviewed By : John Carlone 10/20/2025  
Supervised By : Mahesh Dadoda 10/20/2025

Quant Time: Oct 18 01:29:23 2025  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\624N100825W.M  
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS  
QLast Update : Thu Oct 09 02:36:32 2025  
Response via : Initial Calibration











#1  
Bromochloromethane  
Concen: 30.000 ug/l  
RT: 7.800 min Scan# 994  
Delta R.T. 0.006 min  
Lab File: VN088065.D  
Acq: 17 Oct 2025 13:36

Instrument :  
MSVOA\_N  
ClientSampleId :  
VOA 251015085-01

| Tgt Ion | Ratio | Lower | Upper |
|---------|-------|-------|-------|
| 128     | 100   |       |       |
| 49      | 181.9 | 0.0   | 439.8 |
| 130     | 121.8 | 0.0   | 325.8 |




Manual Integrations  
APPROVED


Reviewed By :John Carlone 10/20/2025  
Supervised By :Mahesh Dadoda 10/20/2025



#2  
Dichlorodifluoromethane  
Concen: 0.343 ug/l m  
RT: 2.148 min Scan# 33  
Delta R.T. 0.006 min  
Lab File: VN088065.D  
Acq: 17 Oct 2025 13:36


| Tgt Ion | Ratio | Lower | Upper |
|---------|-------|-------|-------|
| 85      | 100   |       |       |
| 87      | 21.8  | 17.9  | 53.8  |





#6  
Chloroethane  
Concen: 0.970 ug/l  
RT: 3.148 min Scan# 201  
Delta R.T. 0.012 min  
Lab File: VN088065.D  
Acq: 17 Oct 2025 13:36




Instrument :  
MSVOA\_N  
ClientSampleId :  
VOA 251015085-01



Tgt Ion: 64 Resp: 2599  
Ion Ratio Lower Upper  
64 100  
66 38.9 26.8 40.2




Manual Integrations  
APPROVED


Reviewed By :John Carlone 10/20/2025  
Supervised By :Mahesh Dadoda 10/20/2025



#8  
Diethyl Ether  
Concen: 17.826 ug/l  
RT: 3.959 min Scan# 341  
Delta R.T. 0.000 min  
Lab File: VN088065.D  
Acq: 17 Oct 2025 13:36


Tgt Ion: 74 Resp: 44491  
Ion Ratio Lower Upper  
74 100  
45 109.5 20.4 183.8





#23  
tert-Butyl Alcohol  
Concen: 248.079 ug/l  
RT: 5.512 min Scan# 608  
Delta R.T. -0.018 min  
Lab File: VN088065.D  
Acq: 17 Oct 2025 13:36




Instrument :  
MSVOA\_N  
ClientSampleId :  
VOA 251015085-01



Tgt Ion: 59 Resp: 24622  
Ion Ratio Lower Upper  
59 100  
52 0.1 0.0 0.0




Manual Integrations  
APPROVED


Reviewed By :John Carlone 10/20/2025  
Supervised By :Mahesh Dadoda 10/20/2025



#24  
Methyl tert-Butyl Ether  
Concen: 0.620 ug/l m  
RT: 5.800 min Scan# 654  
Delta R.T. 0.012 min  
Lab File: VN088065.D  
Acq: 17 Oct 2025 13:36

Tgt Ion: 73 Resp: 8441  
Ion Ratio Lower Upper  
73 100  
43 11.4 17.0 25.4





#27

1,2-Dichloroethane-d4

Concen: 31.388 ug/l

RT: 8.559 min Scan# 1123

Delta R.T. -0.006 min

Lab File: VN088065.D

Acq: 17 Oct 2025 13:36

Instrument :

MSVOA\_N

ClientSampleId :

VOA 251015085-01

Tgt Ion: 65 Resp: 146632

Ion Ratio Lower Upper

65 100





67 52.7 41.9 62.9

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/20/2025

Supervised By :Mahesh Dadoda 10/20/2025



#28

1,4-Difluorobenzene

Concen: 30.000 ug/l

RT: 9.082 min Scan# 1212

Delta R.T. 0.000 min

Lab File: VN088065.D

Acq: 17 Oct 2025 13:36




Tgt Ion:114 Resp: 311111


Ion Ratio Lower Upper

114 100

63 22.2 16.6 25.0


88 16.5 12.6 18.8





#34  
Benzene  
Concen: 2.173 ug/l  
RT: 8.583 min Scan# 1127  
Delta R.T. -0.006 min  
Lab File: VN088065.D  
Acq: 17 Oct 2025 13:36




Instrument :  
MSVOA\_N  
ClientSampleId :  
VOA 251015085-01



Tgt Ion: 78 Resp: 3213  
Ion Ratio Lower Upper  
78 100  
77 25.5 18.9 28.3  
51 87.4 15.6 23.4




Manual Integrations  
APPROVED


Reviewed By :John Carlone 10/20/2025  
Supervised By :Mahesh Dadoda 10/20/2025



#45  
1,4-Dioxane  
Concen: 225.643 ug/l  
RT: 9.682 min Scan# 1314  
Delta R.T. 0.006 min  
Lab File: VN088065.D  
Acq: 17 Oct 2025 13:36

Tgt Ion: 88 Resp: 15497  
Ion Ratio Lower Upper  
88 100  
43 32.7 18.6 27.8  
58 63.5 57.4 86.2









#57  
Chlorobenzene-d5  
Concen: 30.000 ug/l  
RT: 11.847 min Scan# 1682  
Delta R.T. 0.000 min  
Lab File: VN088065.D  
Acq: 17 Oct 2025 13:36

Instrument :  
MSVOA\_N  
ClientSampleId :  
VOA 251015085-01

Tgt Ion: 117 Resp: 28223  
Ion Ratio Lower Upper  
117 100  
82 57.9 45.2 67.8  
119 32.0 25.9 38.9




Manual Integrations  
APPROVED


Reviewed By :John Carlone 10/20/2025  
Supervised By :Mahesh Dadoda 10/20/2025



#60  
4-Bromofluorobenzene  
Concen: 29.006 ug/l  
RT: 12.829 min Scan# 1849  
Delta R.T. 0.000 min  
Lab File: VN088065.D  
Acq: 17 Oct 2025 13:36


Tgt Ion: 95 Resp: 134961  
Ion Ratio Lower Upper  
95 100  
174 67.3 57.4 86.0  
176 64.5 57.0 85.4





#62  
Toluene  
Concen: 0.533 ug/l  
RT: 10.606 min Scan# 1472  
Delta R.T. -0.006 min  
Lab File: VN088065.D  
Acq: 17 Oct 2025 13:36




Instrument :  
MSVOA\_N  
ClientSampleId :  
VOA 251015085-01




Tgt Ion: 91 Resp: 8529  
Ion Ratio Lower Upper  
91 100  
92 56.1 45.8 68.8

Manual Integrations  
APPROVED



Reviewed By :John Carlone 10/20/2025  
Supervised By :Mahesh Dadoda 10/20/2025




#63  
Toluene-d8  
Concen: 30.322 ug/l  
RT: 10.547 min Scan# 1461  
Delta R.T. 0.000 min  
Lab File: VN088065.D  
Acq: 17 Oct 2025 13:36



Tgt Ion: 98 Resp: 381294  
Ion Ratio Lower Upper  
98 100  
100 65.0 53.1 79.7









#64  
Chlorobenzene  
Concen: 19.856 ug/l  
RT: 11.871 min Scan# 1686  
Delta R.T. 0.000 min  
Lab File: VN088065.D  
Acq: 17 Oct 2025 13:36

Instrument :  
MSVOA\_N  
ClientSampleId :  
VOA 251015085-01

Tgt Ion: 112 Resp: 20304  
Ion Ratio Lower Upper  
112 100  
114 31.1 25.7 38.5




Manual Integrations  
APPROVED

Reviewed By :John Carlone 10/20/2025  
Supervised By :Mahesh Dadoda 10/20/2025




#66  
Ethyl Benzene  
Concen: 0.700 ug/l  
RT: 11.941 min Scan# 1698  
Delta R.T. 0.000 min  
Lab File: VN088065.D  
Acq: 17 Oct 2025 13:36

Tgt Ion: 91 Resp: 12189  
Ion Ratio Lower Upper  
91 100  
106 33.3 24.0 36.0











#67  
m/p-Xylenes  
Concen: 4.224 ug/l  
RT: 12.047 min Scan# 1716  
Delta R.T. 0.000 min  
Lab File: VN088065.D  
Acq: 17 Oct 2025 13:36

Instrument :  
MSVOA\_N  
ClientSampleId :  
VOA 251015085-01

Tgt Ion:106 Resp: 2834  
Ion Ratio Lower Upper  
106 100  
91 195.7 161.4 242.0




Manual Integrations  
APPROVED


Reviewed By :John Carlone 10/20/2025  
Supervised By :Mahesh Dadoda 10/20/2025



#68  
o-Xylene  
Concen: 3.064 ug/l  
RT: 12.371 min Scan# 1771  
Delta R.T. -0.006 min  
Lab File: VN088065.D  
Acq: 17 Oct 2025 13:36


Tgt Ion:106 Resp: 20472  
Ion Ratio Lower Upper  
106 100  
91 223.5 105.2 315.6





#70  
Isopropylbenzene  
Concen: 1.067 ug/l  
RT: 12.670 min Scan# 1823  
Delta R.T. -0.006 min  
Lab File: VN088065.D  
Acq: 17 Oct 2025 13:36




Instrument :  
MSVOA\_N  
ClientSampleId :  
VOA 251015085-01



Tgt Ion: 105 Resp: 17470  
Ion Ratio Lower Upper  
105 100  
120 27.7 19.1 35.5




Manual Integrations  
APPROVED


Reviewed By :John Carlone 10/20/2025  
Supervised By :Mahesh Dadoda 10/20/2025



#74  
n-propylbenzene  
Concen: 0.286 ug/l  
RT: 13.023 min Scan# 1882  
Delta R.T. 0.006 min  
Lab File: VN088065.D  
Acq: 17 Oct 2025 13:36

Tgt Ion: 91 Resp: 5647  
Ion Ratio Lower Upper  
91 100  
120 24.1 0.0 46.2









#75  
2-Chlorotoluene  
Concen: 0.712 ug/l  
RT: 13.100 min Scan# 1895  
Delta R.T. -0.006 min  
Lab File: VN088065.D  
Acq: 17 Oct 2025 13:36

Instrument :  
MSVOA\_N  
ClientSampleId :  
VOA 251015085-01

Tgt Ion: 91 Resp: 860  
Ion Ratio Lower Upper  
91 100  
126 32.7 0.0 66.6




Manual Integrations  
APPROVED


Reviewed By :John Carlone 10/20/2025  
Supervised By :Mahesh Dadoda 10/20/2025



#76  
1,3,5-Trimethylbenzene  
Concen: 0.370 ug/l  
RT: 13.159 min Scan# 1905  
Delta R.T. 0.006 min  
Lab File: VN088065.D  
Acq: 17 Oct 2025 13:36

Tgt Ion:105 Resp: 5140  
Ion Ratio Lower Upper  
105 100  
120 16.4 0.0 100.6









#80  
1,2,4-Trimethylbenzene  
Concen: 2.022 ug/l  
RT: 13.465 min Scan# 1956  
Delta R.T. 0.006 min  
Lab File: VN088065.D  
Acq: 17 Oct 2025 13:36

Instrument : MSVOA\_N  
ClientSampleId : VOA 251015085-01

Tgt Ion:105 Resp: 2815  
Ion Ratio Lower Upper  
105 100  
120 44.5 0.0 93.8




Manual Integrations  
APPROVED


Reviewed By :John Carlone 10/20/2025  
Supervised By :Mahesh Dadoda 10/20/2025



#84  
1,4-Dichlorobenzene  
Concen: 8.506 ug/l  
RT: 13.794 min Scan# 2013  
Delta R.T. 0.006 min  
Lab File: VN088065.D  
Acq: 17 Oct 2025 13:36

Tgt Ion:146 Resp: 66054  
Ion Ratio Lower Upper  
146 100  
111 41.5 20.5 61.6  
148 62.6 31.9 95.9





#87

1,2-Dichlorobenzene

Concen: 0.717 ug/l

RT: 14.088 min Scan# 2063

Delta R.T. 0.000 min

Lab File: VN088065.D

Acq: 17 Oct 2025 13:36

Instrument :

MSVOA\_N

ClientSampleId :





VOA 251015085-01

|           |       |       |      |
|-----------|-------|-------|------|
| Tgt Ion:  | 146   | Resp: | 531  |
| Ion Ratio | Lower | Upper |      |
| 146       | 100   |       |      |
| 111       | 50.3  | 22.3  | 66.8 |
| 148       | 64.7  | 31.4  | 94.3 |

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/20/2025  
Supervised By :Mahesh Dadoda 10/20/2025



#91

Naphthalene

Concen: 6.780 ug/l




RT: 15.611 min Scan# 2322

Delta R.T. -0.006 min

Lab File: VN088065.D

Acq: 17 Oct 2025 13:36

|           |       |       |        |
|-----------|-------|-------|--------|
| Tgt Ion:  | 128   | Resp: | 109504 |
| Ion Ratio | Lower | Upper |        |
| 128       | 100   |       |        |
| 127       | 13.5  | 10.3  | 15.5   |
| 129       | 10.9  | 8.6   | 13.0   |





284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,  
Fax : 908 789 8922

## Report of Analysis

Client: Garden State Laboratories, Inc.  
Project: Waste Water 2025  
Client Sample ID: Trip blank 251015076-03  
Lab Sample ID: Q3362-02  
Analytical Method: E624.1  
Sample Wt/Vol: 5 mL

Level : LOW  
Final Vol: 5000 uL

Date Collected: 10/15/25  
Date Received: 10/16/25  
SDG No.: Q3362  
Matrix: Water  
% Solid: 0  
Test: VOCMS Group1

| CAS Number                | Parameter             | Conc.             | Qua. | DF | MDL      | LOQ / CRQL | Units   | Date Ana.      | BatchID  |
|---------------------------|-----------------------|-------------------|------|----|----------|------------|---------|----------------|----------|
| <b>TARGETS</b>            |                       |                   |      |    |          |            |         |                |          |
| 107-02-8                  | Acrolein              | 6.60              | U    | 1  | 6.60     | 25.0       | ug/L    | 10/17/25 11:53 | VN101725 |
| 107-13-1                  | Acrylonitrile         | 2.80              | U    | 1  | 2.80     | 25.0       | ug/L    | 10/17/25 11:53 | VN101725 |
| <b>SURROGATES</b>         |                       |                   |      |    |          |            |         |                |          |
| 17060-07-0                | 1,2-Dichloroethane-d4 | 32.4              |      |    | 91 - 110 | 108%       | SPK: 30 |                |          |
| 2037-26-5                 | Toluene-d8            | 30.6              |      |    | 91 - 112 | 102%       | SPK: 30 |                |          |
| 460-00-4                  | 4-Bromofluorobenzene  | 28.9              |      |    | 63 - 112 | 96%        | SPK: 30 |                |          |
| <b>INTERNAL STANDARDS</b> |                       |                   |      |    |          |            |         |                |          |
|                           |                       | <b>Area Count</b> |      |    |          |            |         |                |          |
| 74-97-5                   | Bromochloromethane    | 43900             |      |    |          |            |         |                |          |
| 540-36-3                  | 1,4-Difluorobenzene   | 238000            |      |    |          |            |         |                |          |
| 3114-55-4                 | Chlorobenzene-d5      | 213000            |      |    |          |            |         |                |          |

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

\* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN101725\  
 Data File : VN088060.D  
 Acq On : 17 Oct 2025 11:53  
 Operator : JC\MD  
 Sample : Q3362-02  
 Misc : 5.0mL/MSVOA\_N/WATER  
 ALS Vial : 7 Sample Multiplier: 1

Instrument :  
 MSVOA\_N  
 ClientSampleId :  
 Trip blank 251015076-03

Quant Time: Oct 18 01:27:36 2025  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\624N100825W.M  
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS  
 QLast Update : Thu Oct 09 02:36:32 2025  
 Response via : Initial Calibration

| Compound                    | R.T.   | QIon  | Response | Conc     | Units | Dev(Min) |
|-----------------------------|--------|-------|----------|----------|-------|----------|
| -----                       |        |       |          |          |       |          |
| Internal Standards          |        |       |          |          |       |          |
| 1) Bromochloromethane       | 7.794  | 128   | 43937    | 30.000   | ug/l  | 0.00     |
| 28) 1,4-Difluorobenzene     | 9.082  | 114   | 238376   | 30.000   | ug/l  | 0.00     |
| 57) Chlorobenzene-d5        | 11.841 | 117   | 212709   | 30.000   | ug/l  | 0.00     |
| System Monitoring Compounds |        |       |          |          |       |          |
| 27) 1,2-Dichloroethane-d4   | 8.559  | 65    | 110886   | 32.410   | ug/l  | 0.00     |
| Spiked Amount               | 30.000 | Range | 91 - 110 | Recovery | =     | 108.033% |
| 60) 4-Bromofluorobenzene    | 12.829 | 95    | 101235   | 28.870   | ug/l  | 0.00     |
| Spiked Amount               | 30.000 | Range | 63 - 112 | Recovery | =     | 96.233%  |
| 63) Toluene-d8              | 10.547 | 98    | 289573   | 30.555   | ug/l  | 0.00     |
| Spiked Amount               | 30.000 | Range | 91 - 112 | Recovery | =     | 101.867% |


| Target Compounds | Qvalue |
|------------------|--------|
| -----            |        |

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN101725\  
Data File : VN088060.D  
Acq On : 17 Oct 2025 11:53  
Operator : JC\MD  
Sample : Q3362-02  
Misc : 5.0mL/MSVOA\_N/WATER  
ALS Vial : 7 Sample Multiplier: 1

Instrument :  
MSVOA\_N  
ClientSampleId :  
Trip blank 251015076-03

Quant Time: Oct 18 01:27:36 2025  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\624N100825W.M  
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS  
QLast Update : Thu Oct 09 02:36:32 2025  
Response via : Initial Calibration











#1  
Bromochloromethane  
Concen: 30.000 ug/l  
RT: 7.794 min Scan# 993  
Delta R.T. 0.000 min  
Lab File: VN088060.D  
Acq: 17 Oct 2025 11:53




Instrument :  
MSVOA\_N  
ClientSampleId :  
Trip blank 251015076-03


| Tgt Ion | Ratio | Lower | Upper |
|---------|-------|-------|-------|
| 128     | 100   |       |       |
| 49      | 188.7 | 0.0   | 439.8 |
| 130     | 127.8 | 0.0   | 325.8 |



#27  
1,2-Dichloroethane-d4  
Concen: 32.410 ug/l  
RT: 8.559 min Scan# 1123  
Delta R.T. -0.006 min  
Lab File: VN088060.D  
Acq: 17 Oct 2025 11:53

| Tgt Ion | Ratio | Lower | Upper |
|---------|-------|-------|-------|
| 65      | 100   |       |       |
| 67      | 53.2  | 41.9  | 62.9  |





#28

1,4-Difluorobenzene

Concen: 30.000 ug/l

RT: 9.082 min Scan# 11

Delta R.T. 0.000 min

Lab File: VN088060.D

Acq: 17 Oct 2025 11:53

Instrument :

MSVOA\_N

ClientSampleId :

Trip blank 251015076-03





Tgt Ion:114 Resp: 238376

Ion Ratio Lower Upper

114 100

63 21.8 16.6 25.0

88 17.2 12.6 18.8



#57

Chlorobenzene-d5

Concen: 30.000 ug/l

RT: 11.841 min Scan# 1681

Delta R.T. -0.006 min

Lab File: VN088060.D

Acq: 17 Oct 2025 11:53




Tgt Ion:117 Resp: 212709


Ion Ratio Lower Upper

117 100

82 58.0 45.2 67.8

119 31.2 25.9 38.9









#60  
4-Bromofluorobenzene  
Concen: 28.870 ug/l  
RT: 12.829 min Scan# 1849  
Delta R.T. 0.000 min  
Lab File: VN088060.D  
Acq: 17 Oct 2025 11:53




Instrument :  
MSVOA\_N  
ClientSampleId :  
Trip blank 251015076-03

Tgt Ion: 95 Resp: 101235  
Ion Ratio Lower Upper  
95 100  
174 66.3 57.4 86.0  
176 62.7 57.0 85.4



#63  
Toluene-d8  
Concen: 30.555 ug/l  
RT: 10.547 min Scan# 1461  
Delta R.T. 0.000 min  
Lab File: VN088060.D  
Acq: 17 Oct 2025 11:53

Tgt Ion: 98 Resp: 289573  
Ion Ratio Lower Upper  
98 100  
100 64.7 53.1 79.7





# CALIBRATION SUMMARY



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,  
Fax : 908 789 8922

# VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: Alliance Contract: GARD04  
 Lab Code: ACE SDG No.: Q3362  
 Instrument ID: MSVOA\_N Calibration Date(s): 10/08/2025 10/08/2025  
 Heated Purge: (Y/N) N Calibration Time(s): 09:34 11:12  
 GC Column: RXI-624 ID: 0.25 (mm)

|                       |        |                     |        |                     |        |                     |       |       |  |
|-----------------------|--------|---------------------|--------|---------------------|--------|---------------------|-------|-------|--|
| LAB FILE ID:          |        | RRF005 = VN088030.D |        | RRF020 = VN088031.D |        | RRF050 = VN088032.D |       |       |  |
|                       |        | RRF100 = VN088033.D |        | RRF150 = VN088034.D |        | RRF =               |       |       |  |
| COMPOUND              | RRF005 | RRF020              | RRF050 | RRF100              | RRF150 | RRF                 | RRF   | % RSD |  |
| Acrolein              | 0.452  | 0.353               | 0.360  | 0.411               | 0.408  |                     | 0.397 | 10.3  |  |
| Acrylonitrile         | 1.269  | 1.221               | 1.203  | 1.302               | 1.193  |                     | 1.238 | 3.7   |  |
| 1,2-Dichloroethane-d4 | 2.303  | 2.319               | 2.360  | 2.378               | 2.319  |                     | 2.336 | 1.4   |  |
| Toluene-d8            | 1.344  | 1.343               | 1.333  | 1.329               | 1.334  |                     | 1.337 | 0.5   |  |
| 4-Bromofluorobenzene  | 0.488  | 0.497               | 0.493  | 0.507               | 0.489  |                     | 0.495 | 1.5   |  |

\* Compounds with required minimum RRF and maximum %RSD values.  
 All other compounds must meet a minimum RRF of 0.010.  
 RRF of 1,4-Dioxane = Value should be divide by 1000.

Method Path : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\  
 Method File : 624N100825W.M  
 Title : METHOD 624 VOLATILE ORGANIC ANALYSIS  
 Last Update : Thu Oct 09 02:36:32 2025  
 Response Via : Initial Calibration

## Calibration Files

5 =VN088030.D 20 =VN088031.D 50 =VN088032.D 100 =VN088033.D 150 =VN088034.D

| Compound |                     | 5              | 20    | 50    | 100   | 150   | Avg   | %RSD  |
|----------|---------------------|----------------|-------|-------|-------|-------|-------|-------|
| -----    |                     |                |       |       |       |       |       |       |
| 1) I     | Bromochloromethane  | -----ISTD----- |       |       |       |       |       |       |
| 2) M     | Dichlorodifluoro... | 1.490          | 1.788 | 1.711 | 1.867 | 1.736 | 1.718 | 8.19  |
| 3) M     | Chloromethane       | 1.918          | 1.942 | 1.878 | 2.057 | 1.945 | 1.948 | 3.42  |
| 4) M     | Vinyl Chloride      | 2.020          | 2.044 | 1.970 | 2.142 | 1.981 | 2.031 | 3.38  |
| 5) M     | Bromomethane        | 1.037          | 1.041 | 1.067 | 1.211 | 1.105 | 1.092 | 6.58  |
| 6) M     | Chloroethane        | 1.430          | 1.330 | 1.281 | 1.393 | 1.265 | 1.340 | 5.28  |
| 7) M     | Trichlorofluorom... | 2.820          | 2.737 | 2.698 | 2.902 | 2.655 | 2.762 | 3.59  |
| 8) T     | Diethyl Ether       | 1.353          | 1.224 | 1.198 | 1.288 | 1.177 | 1.248 | 5.75  |
| 9)       | 1,1,2-Trichlorot... | 1.510          | 1.460 | 1.453 | 1.558 | 1.412 | 1.479 | 3.81  |
| 10) M    | 1,1-Dichloroethene  | 1.710          | 1.559 | 1.529 | 1.593 | 1.496 | 1.577 | 5.22  |
| 11)      | Methyl Iodide       | 0.781          | 1.201 | 1.541 | 1.911 | 1.834 | 1.454 | 32.23 |
| 12)      | Methyl Acetate      | 2.679          | 2.533 | 2.442 | 2.697 | 2.471 | 2.564 | 4.60  |
| 13) M    | Acrolein            | 0.452          | 0.353 | 0.360 | 0.411 | 0.408 | 0.397 | 10.34 |
| 14) M    | Acrylonitrile       | 1.269          | 1.221 | 1.203 | 1.302 | 1.193 | 1.237 | 3.72  |
| 15) M    | Acetone             | 0.419          | 0.365 | 0.346 | 0.362 | 0.320 | 0.362 | 10.07 |
| 16) M    | Carbon Disulfide    | 5.111          | 4.866 | 4.796 | 5.172 | 4.779 | 4.945 | 3.72  |
| 17)      | Allyl chloride      | 3.248          | 2.939 | 2.812 | 3.033 | 2.850 | 2.976 | 5.85  |
| 18) M    | Methylene Chloride  | 2.188          | 1.946 | 1.946 | 2.049 | 1.895 | 2.005 | 5.82  |
| 19) M    | trans-1,2-Dichlo... | 1.892          | 1.826 | 1.733 | 1.891 | 1.776 | 1.824 | 3.84  |
| 20) T    | Diisopropyl ether   | 6.682          | 6.730 | 6.581 | 7.169 | 6.720 | 6.776 | 3.35  |
| 21) M    | 1,1-Dichloroethane  | 3.634          | 3.511 | 3.377 | 3.682 | 3.433 | 3.528 | 3.67  |
| 22) M    | cis-1,2-Dichloro... | 2.393          | 2.264 | 2.205 | 2.381 | 2.172 | 2.283 | 4.40  |
| 23) M    | tert-Butyl Alcohol  | 0.549          | 0.494 | 0.481 | 0.511 | 0.446 | 0.496 | 7.68  |
| 24) M    | Methyl tert-Buty... | 7.020          | 6.709 | 6.609 | 7.119 | 6.586 | 6.809 | 3.60  |
| 25) M    | Chloroform          | 3.927          | 3.583 | 3.492 | 3.728 | 3.472 | 3.640 | 5.20  |
| 26)      | Cyclohexane         | 2.967          | 2.964 | 2.826 | 3.030 | 2.827 | 2.923 | 3.13  |
| 27) s    | 1,2-Dichloroetha... | 2.303          | 2.319 | 2.360 | 2.378 | 2.319 | 2.336 | 1.36  |
|          |                     |                |       |       |       |       |       |       |
| 28) I    | 1,4-Difluorobenzene | -----ISTD----- |       |       |       |       |       |       |
| 29)      | 1,1-Dichloropropene | 0.443          | 0.421 | 0.413 | 0.452 | 0.422 | 0.430 | 3.84  |
| 30) M    | 2-Butanone          | 0.330          | 0.315 | 0.305 | 0.333 | 0.303 | 0.317 | 4.43  |
| 31)      | 2,2-Dichloropropane | 0.613          | 0.550 | 0.528 | 0.565 | 0.514 | 0.554 | 6.95  |
| 32) M    | 1,1,1-Trichloroe... | 0.564          | 0.539 | 0.513 | 0.557 | 0.514 | 0.537 | 4.41  |
| 33) M    | Carbon Tetrachlo... | 0.457          | 0.445 | 0.435 | 0.475 | 0.439 | 0.450 | 3.60  |
| 34) M    | Benzene             | 1.491          | 1.402 | 1.366 | 1.489 | 1.384 | 1.426 | 4.17  |
| 35)      | Methacrylonitrile   | 0.349          | 0.333 | 0.326 | 0.343 | 0.316 | 0.334 | 3.89  |
| 36) M    | 1,2-Dichloroethane  | 0.517          | 0.490 | 0.473 | 0.514 | 0.482 | 0.495 | 3.97  |
| 37) M    | Trichloroethene     | 0.358          | 0.334 | 0.322 | 0.351 | 0.324 | 0.338 | 4.79  |
| 38)      | Methylcyclohexane   | 0.560          | 0.522 | 0.513 | 0.555 | 0.506 | 0.531 | 4.67  |
| 39) M    | 1,2-Dichloropropane | 0.379          | 0.350 | 0.337 | 0.375 | 0.347 | 0.358 | 5.08  |
| 40)      | Dibromomethane      | 0.270          | 0.258 | 0.251 | 0.271 | 0.251 | 0.260 | 3.85  |
| 41) M    | Bromodichloromet... | 0.571          | 0.517 | 0.502 | 0.556 | 0.512 | 0.532 | 5.61  |
| 42) M    | Vinyl Acetate       | 1.073          | 1.006 | 1.017 | 1.195 | 1.111 | 1.080 | 7.13  |
| 43)      | Ethyl Acetate       | 0.627          | 0.609 | 0.570 | 0.653 | 0.590 | 0.610 | 5.26  |
| 44)      | Isopropyl Acetate   | 0.988          | 0.978 | 0.970 | 1.061 | 0.982 | 0.996 | 3.72  |
| 45) T    | 1,4-Dioxane         | 0.007          | 0.007 | 0.007 | 0.007 | 0.006 | 0.007 | 8.01  |
| 46)      | Methyl methacrylate | 0.453          | 0.385 | 0.430 | 0.480 | 0.443 | 0.439 | 7.98  |
| 47)      | n-amyl Acetate      | 0.377          | 0.459 | 0.444 | 0.504 | 0.576 | 0.472 | 15.62 |
| 48) M    | t-1,3-Dichloropr... | 0.587          | 0.570 | 0.571 | 0.637 | 0.590 | 0.591 | 4.63  |
| 49) T    | cis-1,3-Dichloro... | 0.628          | 0.604 | 0.589 | 0.655 | 0.607 | 0.617 | 4.12  |
| 50) M    | 1,1,2-Trichloroe... | 0.366          | 0.338 | 0.331 | 0.357 | 0.332 | 0.345 | 4.53  |
| 51)      | Ethyl methacrylate  | 0.512          | 0.552 | 0.583 | 0.659 | 0.611 | 0.583 | 9.57  |
| 52)      | 1,3-Dichloropropane | 0.612          | 0.593 | 0.576 | 0.642 | 0.585 | 0.601 | 4.36  |
| 53) M    | Dibromochloromet... | 0.415          | 0.400 | 0.398 | 0.443 | 0.405 | 0.412 | 4.56  |
| 54) M    | 1,2-Dibromoethane   | 0.381          | 0.359 | 0.352 | 0.391 | 0.360 | 0.369 | 4.55  |
| 55) M    | 2-Chloroethyl vi... | 0.290          | 0.319 | 0.313 | 0.338 | 0.314 | 0.315 | 5.39  |
| 56) M    | Bromoform           | 0.269          | 0.265 | 0.273 | 0.306 | 0.281 | 0.279 | 5.81  |

Method Path : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\  
 Method File : 624N100825W.M

|       |                     | -----ISTD----- |       |       |       |       |       |       |
|-------|---------------------|----------------|-------|-------|-------|-------|-------|-------|
| 57) I | Chlorobenzene-d5    |                |       |       |       |       |       |       |
| 58) M | 4-Methyl-2-Penta... | 0.679          | 0.671 | 0.657 | 0.702 | 0.646 | 0.671 | 3.21  |
| 59) M | 2-Hexanone          | 0.360          | 0.437 | 0.456 | 0.500 | 0.463 | 0.443 | 11.67 |
| 60) S | 4-Bromofluoroben... | 0.488          | 0.497 | 0.493 | 0.507 | 0.489 | 0.495 | 1.54  |
| 61) M | Tetrachloroethene   | 0.323          | 0.306 | 0.293 | 0.309 | 0.287 | 0.304 | 4.62  |
| 62) M | Toluene             | 1.796          | 1.700 | 1.645 | 1.730 | 1.638 | 1.702 | 3.81  |
| 63) S | Toluene-d8          | 1.344          | 1.343 | 1.333 | 1.329 | 1.334 | 1.337 | 0.50  |
| 64) M | Chlorobenzene       | 1.118          | 1.084 | 1.059 | 1.124 | 1.049 | 1.087 | 3.07  |
| 65)   | 1,1,1,2-Tetrachl... | 0.382          | 0.382 | 0.369 | 0.383 | 0.357 | 0.375 | 2.99  |
| 66) M | Ethyl Benzene       | 1.902          | 1.839 | 1.800 | 1.919 | 1.792 | 1.850 | 3.13  |
| 67) M | m/p-Xylenes         | 0.729          | 0.714 | 0.693 | 0.739 | 0.693 | 0.713 | 2.89  |
| 68) M | o-Xylene            | 0.746          | 0.710 | 0.685 | 0.732 | 0.677 | 0.710 | 4.17  |
| 69) M | Styrene             | 1.155          | 1.133 | 1.178 | 1.245 | 1.162 | 1.175 | 3.63  |
| 70)   | Isopropylbenzene    | 1.791          | 1.721 | 1.707 | 1.804 | 1.679 | 1.740 | 3.14  |
| 71) M | 1,1,2,2-Tetrachl... | 0.620          | 0.583 | 0.576 | 0.612 | 0.560 | 0.590 | 4.22  |
| 72)   | 1,2,3-Trichlorop... | 0.580          | 0.517 | 0.547 | 0.571 | 0.524 | 0.548 | 5.03  |
| 73)   | Bromobenzene        | 0.431          | 0.428 | 0.422 | 0.439 | 0.407 | 0.425 | 2.88  |
| 74)   | n-propylbenzene     | 2.089          | 2.082 | 2.067 | 2.224 | 2.043 | 2.101 | 3.37  |
| 75)   | 2-Chlorotoluene     | 1.299          | 1.273 | 1.252 | 1.352 | 1.248 | 1.285 | 3.31  |
| 76)   | 1,3,5-Trimethylb... | 1.521          | 1.464 | 1.453 | 1.548 | 1.405 | 1.478 | 3.84  |
| 77)   | t-1,4-Dichloro-2... | 0.202          | 0.196 | 0.214 | 0.248 | 0.232 | 0.218 | 9.81  |
| 78)   | 4-Chlorotoluene     | 1.254          | 1.296 | 1.295 | 1.393 | 1.288 | 1.305 | 3.96  |
| 79)   | tert-butylbenzene   | 1.295          | 1.287 | 1.268 | 1.342 | 1.215 | 1.282 | 3.59  |
| 80)   | 1,2,4-Trimethylb... | 1.505          | 1.473 | 1.453 | 1.563 | 1.406 | 1.480 | 3.95  |
| 81)   | sec-Butylbenzene    | 1.842          | 1.802 | 1.770 | 1.884 | 1.712 | 1.802 | 3.65  |
| 82)   | p-Isopropyltoluene  | 1.590          | 1.540 | 1.522 | 1.603 | 1.437 | 1.538 | 4.29  |
| 83) M | 1,3-Dichlorobenzene | 0.810          | 0.804 | 0.798 | 0.835 | 0.755 | 0.800 | 3.61  |
| 84) M | 1,4-Dichlorobenzene | 0.855          | 0.833 | 0.802 | 0.848 | 0.790 | 0.825 | 3.48  |
| 85)   | n-Butylbenzene      | 1.435          | 1.433 | 1.420 | 1.506 | 1.362 | 1.431 | 3.58  |
| 86) T | Hexachloroethane    | 0.326          | 0.303 | 0.304 | 0.317 | 0.291 | 0.308 | 4.32  |
| 87) M | 1,2-Dichlorobenzene | 0.821          | 0.786 | 0.766 | 0.819 | 0.744 | 0.787 | 4.26  |
| 88)   | 1,2-Dibromo-3-Ch... | 0.134          | 0.140 | 0.133 | 0.141 | 0.129 | 0.136 | 3.68  |
| 89)   | 1,2,4-Trichlorob... | 0.467          | 0.470 | 0.455 | 0.493 | 0.460 | 0.469 | 3.10  |
| 90)   | Hexachlorobutadiene | 0.177          | 0.160 | 0.162 | 0.166 | 0.152 | 0.163 | 5.60  |
| 91) M | Naphthalene         | 1.691          | 1.709 | 1.674 | 1.820 | 1.690 | 1.717 | 3.44  |
| 92)   | 1,2,3-Trichlorob... | 0.467          | 0.470 | 0.455 | 0.493 | 0.460 | 0.469 | 3.10  |

(#) = Out of Range

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN100825\  
 Data File : VN088030.D  
 Acq On : 08 Oct 2025 09:34  
 Operator : JC\MD  
 Sample : VSTDICC005  
 Misc : 5.0mL/MSVOA\_N/WATER  
 ALS Vial : 2 Sample Multiplier: 1

## Instrument :

MSVOA\_N

## ClientSampleId :

VSTDICC005

## Manual Integrations

## APPROVED

Reviewed By :John  
 Carlone

10/09/2025

Supervised By :Mahesh  
 Dadoda

10/10/2025

Quant Time: Oct 09 02:00:45 2025  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\624N100825W.M  
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS  
 QLast Update : Thu Oct 09 01:59:51 2025  
 Response via : Initial Calibration

| Compound                     | R.T.           | QIon | Response   | Conc     | Units  | Dev(Min) |
|------------------------------|----------------|------|------------|----------|--------|----------|
| -----                        |                |      |            |          |        |          |
| Internal Standards           |                |      |            |          |        |          |
| 1) Bromochloromethane        | 7.800          | 128  | 70500      | 30.000   | ug/l   | 0.00     |
| 28) 1,4-Difluorobenzene      | 9.083          | 114  | 388491     | 30.000   | ug/l   | 0.00     |
| 57) Chlorobenzene-d5         | 11.847         | 117  | 345679     | 30.000   | ug/l   | 0.00     |
|                              |                |      |            |          |        |          |
| System Monitoring Compounds  |                |      |            |          |        |          |
| 27) 1,2-Dichloroethane-d4    | 8.565          | 65   | 162371     | 29.577   | ug/l   | 0.00     |
| Spiked Amount 30.000         | Range 91 - 110 |      | Recovery = | 98.600%  |        |          |
| 60) 4-Bromofluorobenzene     | 12.829         | 95   | 168531     | 29.574   | ug/l   | 0.00     |
| Spiked Amount 30.000         | Range 63 - 112 |      | Recovery = | 98.567%  |        |          |
| 63) Toluene-d8               | 10.547         | 98   | 464668     | 30.170   | ug/l   | 0.00     |
| Spiked Amount 30.000         | Range 91 - 112 |      | Recovery = | 100.567% |        |          |
|                              |                |      |            |          |        |          |
| Target Compounds             |                |      |            |          |        | Qvalue   |
| 2) Dichlorodifluoromethane   | 2.142          | 85   | 17512      | 4.337    | ug/l   | 91       |
| 3) Chloromethane             | 2.383          | 50   | 22534      | 4.922    | ug/l   | 89       |
| 4) Vinyl Chloride            | 2.536          | 62   | 23730      | 4.971    | ug/l   | 100      |
| 5) Bromomethane              | 2.977          | 94   | 12182      | 4.746    | ug/l   | 89       |
| 6) Chloroethane              | 3.136          | 64   | 16801      | 5.336    | ug/l   | 99       |
| 7) Trichlorofluoromethane    | 3.506          | 101  | 33130      | 5.104    | ug/l   | 93       |
| 8) Diethyl Ether             | 3.965          | 74   | 15893      | 5.419    | ug/l   | 56       |
| 9) 1,1,2-Trichlorotrifluo... | 4.365          | 101  | 17744      | 5.106    | ug/l   | 91       |
| 10) 1,1-Dichloroethene       | 4.336          | 96   | 20091      | 5.420    | ug/l   | 87       |
| 11) Methyl Iodide            | 4.571          | 142  | 9178m      | 2.686    | ug/l   |          |
| 12) Methyl Acetate           | 5.018          | 43   | 31482m     | 5.370    | ug/l   |          |
| 13) Acrolein                 | 4.171          | 56   | 26567      | 28.489   | ug/l # | 1        |
| 14) Acrylonitrile            | 5.706          | 53   | 74537      | 25.631   | ug/l # | 89       |
| 15) Acetone                  | 4.418          | 58   | 24635m     | 28.837   | ug/l   |          |
| 16) Carbon Disulfide         | 4.700          | 76   | 60049      | 5.168    | ug/l   | 97       |
| 17) Allyl chloride           | 5.018          | 41   | 38166m     | 5.456    | ug/l   |          |
| 18) Methylene Chloride       | 5.271          | 84   | 25710m     | 5.457    | ug/l   |          |
| 19) trans-1,2-Dichloroethene | 5.777          | 96   | 22235      | 5.188    | ug/l # | 91       |
| 20) Diisopropyl ether        | 6.659          | 45   | 78511      | 4.930    | ug/l   | 98       |
| 21) 1,1-Dichloroethane       | 6.553          | 63   | 42704      | 5.151    | ug/l   | 96       |
| 22) cis-1,2-Dichloroethene   | 7.471          | 96   | 28117      | 5.241    | ug/l   | 97       |
| 23) tert-Butyl Alcohol       | 5.524          | 59   | 32274      | 27.671   | ug/l # | 100      |
| 24) Methyl tert-Butyl Ether  | 5.789          | 73   | 82482      | 5.155    | ug/l   | 96       |
| 25) Chloroform               | 7.947          | 83   | 46137      | 5.393    | ug/l   | 92       |
| 26) Cyclohexane              | 8.241          | 56   | 34857      | 5.075    | ug/l # | 98       |
| 29) 1,1-Dichloropropene      | 8.353          | 75   | 28652      | 5.145    | ug/l   | 95       |
| 30) 2-Butanone               | 7.471          | 43   | 106757     | 25.994   | ug/l   | 100      |
| 31) 2,2-Dichloropropane      | 7.471          | 77   | 39713m     | 5.527    | ug/l   |          |
| 32) 1,1,1-Trichloroethane    | 8.147          | 97   | 36518      | 5.248    | ug/l   | 98       |
| 33) Carbon Tetrachloride     | 8.347          | 117  | 29564      | 5.072    | ug/l   | 93       |
| 34) Benzene                  | 8.583          | 78   | 96524      | 5.226    | ug/l # | 81       |
| 35) Methacrylonitrile        | 7.771          | 41   | 22614      | 5.236    | ug/l   | 95       |
| 36) 1,2-Dichloroethane       | 8.659          | 62   | 33473      | 5.220    | ug/l   | 98       |
| 37) Trichloroethene          | 9.335          | 130  | 23203      | 5.302    | ug/l   | 98       |
| 38) Methylcyclohexane        | 9.583          | 83   | 36254      | 5.272    | ug/l   | 97       |
| 39) 1,2-Dichloropropane      | 9.606          | 63   | 24526      | 5.296    | ug/l   | 97       |
| 40) Dibromomethane           | 9.694          | 93   | 17510      | 5.194    | ug/l   | 96       |
| 41) Bromodichloromethane     | 9.871          | 83   | 36971      | 5.370    | ug/l # | 97       |
| 42) Vinyl Acetate            | 6.594          | 43   | 347500     | 24.837   | ug/l   | 99       |



Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN100825\  
 Data File : VN088030.D  
 Acq On : 08 Oct 2025 09:34  
 Operator : JC\MD  
 Sample : VSTDICC005  
 Misc : 5.0mL/MSVOA\_N/WATER  
 ALS Vial : 2 Sample Multiplier: 1

## Instrument :

MSVOA\_N

## ClientSampleId :

VSTDICC005

## Manual Integrations

## APPROVED

Reviewed By :John  
 Carlone

10/09/2025

Supervised By :Mahesh  
 Dadoda

10/10/2025

Quant Time: Oct 09 02:00:45 2025  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\624N100825W.M  
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS  
 QLast Update : Thu Oct 09 01:59:51 2025  
 Response via : Initial Calibration

| Compound                      | R.T.   | QIon | Response | Conc   | Units | Dev(Min) |     |
|-------------------------------|--------|------|----------|--------|-------|----------|-----|
| 43) Ethyl Acetate             | 7.553  | 43   | 40587    | 5.141  | ug/l  | #        | 71  |
| 44) Isopropyl Acetate         | 8.677  | 43   | 63965    | 4.961  | ug/l  |          | 97  |
| 45) 1,4-Dioxane               | 9.683  | 88   | 8538     | 99.555 | ug/l  | #        | 91  |
| 46) Methyl methacrylate       | 9.665  | 41   | 29344    | 5.167  | ug/l  |          | 87  |
| 47) n-amyl Acetate            | 12.965 | 43   | 24408m   | 4.308  | ug/l  |          |     |
| 48) t-1,3-Dichloropropene     | 10.818 | 75   | 38023    | 4.966  | ug/l  |          | 93  |
| 49) cis-1,3-Dichloropropene   | 10.294 | 75   | 40685    | 5.094  | ug/l  |          | 90  |
| 50) 1,1,2-Trichloroethane     | 10.994 | 97   | 23673    | 5.302  | ug/l  | #        | 91  |
| 51) Ethyl methacrylate        | 10.859 | 69   | 33154    | 4.388  | ug/l  |          | 94  |
| 52) 1,3-Dichloropropane       | 11.141 | 76   | 39602    | 5.084  | ug/l  |          | 98  |
| 53) Dibromochloromethane      | 11.335 | 129  | 26900    | 5.040  | ug/l  |          | 100 |
| 54) 1,2-Dibromoethane         | 11.453 | 107  | 24696    | 5.174  | ug/l  |          | 97  |
| 55) 2-Chloroethyl vinyl ether | 10.141 | 63   | 93982    | 23.044 | ug/l  |          | 98  |
| 56) Bromoform                 | 12.559 | 173  | 17392    | 4.820  | ug/l  |          | 92  |
| 58) 4-Methyl-2-Pentanone      | 10.430 | 43   | 195537   | 25.289 | ug/l  |          | 99  |
| 59) 2-Hexanone                | 11.182 | 43   | 103702   | 20.314 | ug/l  |          | 99  |
| 61) Tetrachloroethene         | 11.088 | 164  | 18609    | 5.316  | ug/l  |          | 94  |
| 62) Toluene                   | 10.612 | 91   | 103446   | 5.276  | ug/l  |          | 98  |
| 64) Chlorobenzene             | 11.871 | 112  | 64403    | 5.142  | ug/l  |          | 92  |
| 65) 1,1,1,2-Tetrachloroethane | 11.941 | 131  | 21997    | 5.096  | ug/l  |          | 93  |
| 66) Ethyl Benzene             | 11.941 | 91   | 109599   | 5.140  | ug/l  |          | 92  |
| 67) m/p-Xylenes               | 12.053 | 106  | 83946    | 10.213 | ug/l  |          | 100 |
| 68) o-Xylene                  | 12.376 | 106  | 42988    | 5.254  | ug/l  |          | 98  |
| 69) Styrene                   | 12.394 | 104  | 66556    | 4.917  | ug/l  |          | 99  |
| 70) Isopropylbenzene          | 12.676 | 105  | 103205   | 5.146  | ug/l  |          | 97  |
| 71) 1,1,2,2-Tetrachloroethane | 12.918 | 83   | 35698    | 5.250  | ug/l  |          | 100 |
| 72) 1,2,3-Trichloropropane    | 12.976 | 75   | 33421m   | 5.064  | ug/l  |          |     |
| 73) Bromobenzene              | 12.959 | 156  | 24839    | 5.068  | ug/l  |          | 99  |
| 74) n-propylbenzene           | 13.018 | 91   | 120356   | 4.971  | ug/l  |          | 100 |
| 75) 2-Chlorotoluene           | 13.106 | 91   | 74825    | 5.055  | ug/l  |          | 98  |
| 76) 1,3,5-Trimethylbenzene    | 13.153 | 105  | 87631    | 5.144  | ug/l  |          | 97  |
| 77) t-1,4-Dichloro-2-butene   | 12.723 | 75   | 11665m   | 4.776  | ug/l  |          |     |
| 78) 4-Chlorotoluene           | 13.206 | 91   | 72267    | 4.805  | ug/l  |          | 100 |
| 79) tert-butylbenzene         | 13.418 | 119  | 74615    | 5.053  | ug/l  |          | 97  |
| 80) 1,2,4-Trimethylbenzene    | 13.459 | 105  | 86692    | 5.084  | ug/l  |          | 99  |
| 81) sec-Butylbenzene          | 13.594 | 105  | 106110   | 5.111  | ug/l  |          | 100 |
| 82) p-Isopropyltoluene        | 13.712 | 119  | 91611    | 5.168  | ug/l  |          | 98  |
| 83) 1,3-Dichlorobenzene       | 13.712 | 146  | 46688    | 5.062  | ug/l  |          | 97  |
| 84) 1,4-Dichlorobenzene       | 13.794 | 146  | 49265    | 5.180  | ug/l  |          | 99  |
| 85) n-Butylbenzene            | 14.035 | 91   | 82683    | 5.013  | ug/l  |          | 99  |
| 86) Hexachloroethane          | 14.306 | 117  | 18772    | 5.286  | ug/l  |          | 91  |
| 87) 1,2-Dichlorobenzene       | 14.088 | 146  | 47327    | 5.217  | ug/l  |          | 98  |
| 88) 1,2-Dibromo-3-Chloropr... | 14.694 | 75   | 7710     | 4.938  | ug/l  |          | 95  |
| 89) 1,2,4-Trichlorobenzene    | 15.817 | 180  | 26899    | 4.978  | ug/l  |          | 92  |
| 90) Hexachlorobutadiene       | 15.470 | 225  | 10187    | 5.417  | ug/l  |          | 96  |
| 91) Naphthalene               | 15.617 | 128  | 97405    | 4.924  | ug/l  |          | 98  |
| 92) 1,2,3-Trichlorobenzene    | 15.817 | 180  | 26899    | 4.978  | ug/l  |          | 92  |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN100825\  
Data File : VN088030.D  
Acq On : 08 Oct 2025 09:34  
Operator : JC\MD  
Sample : VSTDIC005  
Misc : 5.0mL/MSVOA\_N/WATER  
ALS Vial : 2 Sample Multiplier: 1

## Instrument :

MSVOA\_N

## ClientSampleId :

VSTDIC005

## Manual Integrations

## APPROVED

Reviewed By :John  
Carlone


10/09/2025

Supervised By :Mahesh

Dadoda

10/10/2025

Quant Time: Oct 09 02:00:45 2025  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\624N100825W.M  
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS  
QLast Update : Thu Oct 09 01:59:51 2025  
Response via : Initial Calibration



Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN100825\  
 Data File : VN088031.D  
 Acq On : 08 Oct 2025 10:08  
 Operator : JC\MD  
 Sample : VSTDICCC020  
 Misc : 5.0mL/MSVOA\_N/WATER  
 ALS Vial : 3 Sample Multiplier: 1

Instrument :  
 MSVOA\_N  
 ClientSampleId :  
 VSTDICCC020

Manual Integrations  
 APPROVED

Reviewed By : John Carlone 10/09/2025  
 Supervised By : Mahesh Dadoda 10/10/2025

Quant Time: Oct 09 02:01:45 2025  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\624N100825W.M  
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS  
 QLast Update : Thu Oct 09 01:59:51 2025  
 Response via : Initial Calibration

| Compound                     | R.T.           | QIon | Response   | Conc     | Units  | Dev(Min) |
|------------------------------|----------------|------|------------|----------|--------|----------|
| -----                        |                |      |            |          |        |          |
| Internal Standards           |                |      |            |          |        |          |
| 1) Bromochloromethane        | 7.794          | 128  | 71387      | 30.000   | ug/l   | 0.00     |
| 28) 1,4-Difluorobenzene      | 9.082          | 114  | 401663     | 30.000   | ug/l   | 0.00     |
| 57) Chlorobenzene-d5         | 11.847         | 117  | 355387     | 30.000   | ug/l   | 0.00     |
|                              |                |      |            |          |        |          |
| System Monitoring Compounds  |                |      |            |          |        |          |
| 27) 1,2-Dichloroethane-d4    | 8.565          | 65   | 165568     | 29.785   | ug/l   | 0.00     |
| Spiked Amount 30.000         | Range 91 - 110 |      | Recovery = | 99.267%  |        |          |
| 60) 4-Bromofluorobenzene     | 12.829         | 95   | 176557     | 30.136   | ug/l   | 0.00     |
| Spiked Amount 30.000         | Range 63 - 112 |      | Recovery = | 100.467% |        |          |
| 63) Toluene-d8               | 10.547         | 98   | 477319     | 30.145   | ug/l   | 0.00     |
| Spiked Amount 30.000         | Range 91 - 112 |      | Recovery = | 100.500% |        |          |
|                              |                |      |            |          |        |          |
| Target Compounds             |                |      |            |          |        | Qvalue   |
| 2) Dichlorodifluoromethane   | 2.142          | 85   | 85073      | 20.806   | ug/l   | 100      |
| 3) Chloromethane             | 2.383          | 50   | 92431      | 19.939   | ug/l   | 100      |
| 4) Vinyl Chloride            | 2.536          | 62   | 97272      | 20.124   | ug/l   | 100      |
| 5) Bromomethane              | 2.983          | 94   | 49554      | 19.065   | ug/l   | 100      |
| 6) Chloroethane              | 3.136          | 64   | 63320      | 19.859   | ug/l   | 98       |
| 7) Trichlorofluoromethane    | 3.506          | 101  | 130253     | 19.817   | ug/l   | 100      |
| 8) Diethyl Ether             | 3.959          | 74   | 58264      | 19.618   | ug/l   | 100      |
| 9) 1,1,2-Trichlorotrifluo... | 4.365          | 101  | 69507      | 19.754   | ug/l   | 100      |
| 10) 1,1-Dichloroethene       | 4.330          | 96   | 74211      | 19.771   | ug/l   | 100      |
| 11) Methyl Iodide            | 4.577          | 142  | 57156      | 16.522   | ug/l   | 100      |
| 12) Methyl Acetate           | 5.018          | 43   | 120527     | 20.303   | ug/l   | 100      |
| 13) Acrolein                 | 4.177          | 56   | 83909      | 88.861   | ug/l # | 4        |
| 14) Acrylonitrile            | 5.706          | 53   | 290536     | 98.664   | ug/l   | 100      |
| 15) Acetone                  | 4.424          | 58   | 86871      | 100.424  | ug/l   | 100      |
| 16) Carbon Disulfide         | 4.694          | 76   | 231559     | 19.680   | ug/l   | 100      |
| 17) Allyl chloride           | 5.012          | 41   | 139881     | 19.750   | ug/l   | 100      |
| 18) Methylene Chloride       | 5.259          | 84   | 92593      | 19.409   | ug/l   | 100      |
| 19) trans-1,2-Dichloroethene | 5.777          | 96   | 86884      | 20.021   | ug/l   | 100      |
| 20) Diisopropyl ether        | 6.659          | 45   | 320294     | 19.864   | ug/l   | 100      |
| 21) 1,1-Dichloroethane       | 6.553          | 63   | 167086     | 19.905   | ug/l   | 100      |
| 22) cis-1,2-Dichloroethene   | 7.471          | 96   | 107728     | 19.831   | ug/l   | 100      |
| 23) tert-Butyl Alcohol       | 5.530          | 59   | 117569     | 99.549   | ug/l # | 100      |
| 24) Methyl tert-Butyl Ether  | 5.788          | 73   | 319297     | 19.708   | ug/l   | 100      |
| 25) Chloroform               | 7.953          | 83   | 170511     | 19.684   | ug/l   | 100      |
| 26) Cyclohexane              | 8.241          | 56   | 141068     | 20.283   | ug/l   | 100      |
| 29) 1,1-Dichloropropene      | 8.353          | 75   | 112668     | 19.569   | ug/l   | 100      |
| 30) 2-Butanone               | 7.477          | 43   | 422066     | 99.396   | ug/l   | 100      |
| 31) 2,2-Dichloropropane      | 7.477          | 77   | 147354     | 19.837   | ug/l   | 100      |
| 32) 1,1,1-Trichloroethane    | 8.153          | 97   | 144223     | 20.047   | ug/l   | 100      |
| 33) Carbon Tetrachloride     | 8.347          | 117  | 119094     | 19.764   | ug/l   | 100      |
| 34) Benzene                  | 8.588          | 78   | 375303     | 19.653   | ug/l   | 100      |
| 35) Methacrylonitrile        | 7.759          | 41   | 89107      | 19.955   | ug/l   | 100      |
| 36) 1,2-Dichloroethane       | 8.653          | 62   | 131101     | 19.775   | ug/l   | 100      |
| 37) Trichloroethene          | 9.335          | 130  | 89526      | 19.787   | ug/l   | 100      |
| 38) Methylcyclohexane        | 9.582          | 83   | 139683     | 19.646   | ug/l   | 100      |
| 39) 1,2-Dichloropropane      | 9.600          | 63   | 93775      | 19.584   | ug/l   | 100      |
| 40) Dibromomethane           | 9.688          | 93   | 69093      | 19.823   | ug/l   | 100      |
| 41) Bromodichloromethane     | 9.871          | 83   | 138534     | 19.460   | ug/l   | 100      |
| 42) Vinyl Acetate            | 6.588          | 43   | 1347005    | 93.117   | ug/l   | 100      |

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN100825\  
 Data File : VN088031.D  
 Acq On : 08 Oct 2025 10:08  
 Operator : JC\MD  
 Sample : VSTDICCC020  
 Misc : 5.0mL/MSVOA\_N/WATER  
 ALS Vial : 3 Sample Multiplier: 1

Instrument :  
 MSVOA\_N  
 ClientSampleId :  
 VSTDICCC020

Manual Integrations  
 APPROVED

Reviewed By : John Carlone 10/09/2025  
 Supervised By : Mahesh Dadoda 10/10/2025

Quant Time: Oct 09 02:01:45 2025  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\624N100825W.M  
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS  
 QLast Update : Thu Oct 09 01:59:51 2025  
 Response via : Initial Calibration

| Compound                      | R.T.   | QIon | Response | Conc    | Units | Dev(Min) |
|-------------------------------|--------|------|----------|---------|-------|----------|
| 43) Ethyl Acetate             | 7.553  | 43   | 163041   | 19.974  | ug/l  | 100      |
| 44) Isopropyl Acetate         | 8.677  | 43   | 261811   | 19.638  | ug/l  | 100      |
| 45) 1,4-Dioxane               | 9.676  | 88   | 38180    | 430.590 | ug/l  | 100      |
| 46) Methyl methacrylate       | 9.665  | 41   | 103153   | 17.569  | ug/l  | 100      |
| 47) n-amyl Acetate            | 12.570 | 43   | 122978m  | 20.993  | ug/l  |          |
| 48) t-1,3-Dichloropropene     | 10.818 | 75   | 152693   | 19.289  | ug/l  | 100      |
| 49) cis-1,3-Dichloropropene   | 10.294 | 75   | 161768   | 19.591  | ug/l  | 100      |
| 50) 1,1,2-Trichloroethane     | 10.994 | 97   | 90592    | 19.626  | ug/l  | 100      |
| 51) Ethyl methacrylate        | 10.859 | 69   | 147903   | 18.933  | ug/l  | 100      |
| 52) 1,3-Dichloropropane       | 11.141 | 76   | 158729   | 19.710  | ug/l  | 100      |
| 53) Dibromochloromethane      | 11.335 | 129  | 106989   | 19.388  | ug/l  | 100      |
| 54) 1,2-Dibromoethane         | 11.447 | 107  | 96091    | 19.472  | ug/l  | 100      |
| 55) 2-Chloroethyl vinyl ether | 10.141 | 63   | 427409   | 101.363 | ug/l  | 100      |
| 56) Bromoform                 | 12.559 | 173  | 71081    | 19.053  | ug/l  | 100      |
| 58) 4-Methyl-2-Pentanone      | 10.429 | 43   | 794535   | 99.952  | ug/l  | 100      |
| 59) 2-Hexanone                | 11.182 | 43   | 517498   | 98.602  | ug/l  | 100      |
| 61) Tetrachloroethene         | 11.082 | 164  | 72535    | 20.155  | ug/l  | 100      |
| 62) Toluene                   | 10.612 | 91   | 402666   | 19.975  | ug/l  | 100      |
| 64) Chlorobenzene             | 11.870 | 112  | 256863   | 19.949  | ug/l  | 100      |
| 65) 1,1,1,2-Tetrachloroethane | 11.941 | 131  | 90536    | 20.403  | ug/l  | 100      |
| 66) Ethyl Benzene             | 11.941 | 91   | 435618   | 19.873  | ug/l  | 100      |
| 67) m/p-Xylenes               | 12.047 | 106  | 338147   | 40.016  | ug/l  | 100      |
| 68) o-Xylene                  | 12.376 | 106  | 168249   | 20.001  | ug/l  | 100      |
| 69) Styrene                   | 12.394 | 104  | 268372   | 19.286  | ug/l  | 100      |
| 70) Isopropylbenzene          | 12.676 | 105  | 407637   | 19.771  | ug/l  | 100      |
| 71) 1,1,2,2-Tetrachloroethane | 12.917 | 83   | 138052   | 19.750  | ug/l  | 100      |
| 72) 1,2,3-Trichloropropane    | 12.976 | 75   | 122603m  | 18.069  | ug/l  |          |
| 73) Bromobenzene              | 12.959 | 156  | 101455   | 20.133  | ug/l  | 100      |
| 74) n-propylbenzene           | 13.017 | 91   | 493301   | 19.820  | ug/l  | 100      |
| 75) 2-Chlorotoluene           | 13.106 | 91   | 301590   | 19.817  | ug/l  | 100      |
| 76) 1,3,5-Trimethylbenzene    | 13.153 | 105  | 346960   | 19.811  | ug/l  | 100      |
| 77) t-1,4-Dichloro-2-butene   | 12.723 | 75   | 46433    | 18.492  | ug/l  | 100      |
| 78) 4-Chlorotoluene           | 13.200 | 91   | 307157   | 19.866  | ug/l  | 100      |
| 79) tert-butylbenzene         | 13.417 | 119  | 304884   | 20.082  | ug/l  | 100      |
| 80) 1,2,4-Trimethylbenzene    | 13.459 | 105  | 349065   | 19.910  | ug/l  | 100      |
| 81) sec-Butylbenzene          | 13.594 | 105  | 426975   | 20.002  | ug/l  | 100      |
| 82) p-Isopropyltoluene        | 13.706 | 119  | 364945   | 20.025  | ug/l  | 100      |
| 83) 1,3-Dichlorobenzene       | 13.711 | 146  | 190476   | 20.090  | ug/l  | 100      |
| 84) 1,4-Dichlorobenzene       | 13.788 | 146  | 197354   | 20.183  | ug/l  | 100      |
| 85) n-Butylbenzene            | 14.035 | 91   | 339471   | 20.022  | ug/l  | 100      |
| 86) Hexachloroethane          | 14.311 | 117  | 71806    | 19.668  | ug/l  | 100      |
| 87) 1,2-Dichlorobenzene       | 14.088 | 146  | 186272   | 19.974  | ug/l  | 100      |
| 88) 1,2-Dibromo-3-Chloropr... | 14.700 | 75   | 33189    | 20.674  | ug/l  | 100      |
| 89) 1,2,4-Trichlorobenzene    | 15.811 | 180  | 111303   | 20.037  | ug/l  | 100      |
| 90) Hexachlorobutadiene       | 15.476 | 225  | 37802    | 19.553  | ug/l  | 100      |
| 91) Naphthalene               | 15.617 | 128  | 404900   | 19.910  | ug/l  | 100      |
| 92) 1,2,3-Trichlorobenzene    | 15.811 | 180  | 111303   | 20.037  | ug/l  | 100      |

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN100825\  
 Data File : VN088031.D  
 Acq On : 08 Oct 2025 10:08  
 Operator : JC\MD  
 Sample : VSTDICCC020  
 Misc : 5.0mL/MSVOA\_N/WATER  
 ALS Vial : 3 Sample Multiplier: 1

Instrument :  
 MSVOA\_N  
 ClientSampleId :  
 VSTDICCC020

### Manual Integrations APPROVED

Reviewed By : John Carlone 10/09/2025  
 Supervised By : Mahesh Dadoda 10/10/2025

Quant Time: Oct 09 02:01:45 2025  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\624N100825W.M  
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS  
 QLast Update : Thu Oct 09 01:59:51 2025  
 Response via : Initial Calibration





Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN100825\  
 Data File : VN088032.D  
 Acq On : 08 Oct 2025 10:30  
 Operator : JC\MD  
 Sample : VSTDICC050  
 Misc : 5.0mL/MSVOA\_N/WATER  
 ALS Vial : 4 Sample Multiplier: 1

Instrument :  
 MSVOA\_N  
 ClientSampleId :  
 VSTDICC050

Manual Integrations  
 APPROVED

Reviewed By : John Carlone 10/09/2025  
 Supervised By : Mahesh Dadoda 10/10/2025

Quant Time: Oct 09 02:02:43 2025  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\624N100825W.M  
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS  
 QLast Update : Thu Oct 09 01:59:51 2025  
 Response via : Initial Calibration

| Compound                     | R.T.           | QIon | Response            | Conc    | Units  | Dev(Min) |
|------------------------------|----------------|------|---------------------|---------|--------|----------|
| -----                        |                |      |                     |         |        |          |
| Internal Standards           |                |      |                     |         |        |          |
| 1) Bromochloromethane        | 7.800          | 128  | 71521               | 30.000  | ug/l   | 0.00     |
| 28) 1,4-Difluorobenzene      | 9.082          | 114  | 407063              | 30.000  | ug/l   | 0.00     |
| 57) Chlorobenzene-d5         | 11.841         | 117  | 363868              | 30.000  | ug/l   | 0.00     |
|                              |                |      |                     |         |        |          |
| System Monitoring Compounds  |                |      |                     |         |        |          |
| 27) 1,2-Dichloroethane-d4    | 8.559          | 65   | 168820              | 30.313  | ug/l   | 0.00     |
| Spiked Amount 30.000         | Range 91 - 110 |      | Recovery = 101.033% |         |        |          |
| 60) 4-Bromofluorobenzene     | 12.829         | 95   | 179340              | 29.897  | ug/l   | 0.00     |
| Spiked Amount 30.000         | Range 63 - 112 |      | Recovery = 99.667%  |         |        |          |
| 63) Toluene-d8               | 10.547         | 98   | 485008              | 29.917  | ug/l   | 0.00     |
| Spiked Amount 30.000         | Range 91 - 112 |      | Recovery = 99.733%  |         |        |          |
|                              |                |      |                     |         |        |          |
| Target Compounds             |                |      |                     |         |        |          |
|                              |                |      |                     |         | Qvalue |          |
| 2) Dichlorodifluoromethane   | 2.142          | 85   | 203904              | 49.775  | ug/l   | 95       |
| 3) Chloromethane             | 2.383          | 50   | 223881              | 48.205  | ug/l   | 98       |
| 4) Vinyl Chloride            | 2.536          | 62   | 234864              | 48.499  | ug/l   | 99       |
| 5) Bromomethane              | 2.971          | 94   | 127206              | 48.849  | ug/l   | 96       |
| 6) Chloroethane              | 3.136          | 64   | 152751              | 47.817  | ug/l   | 97       |
| 7) Trichlorofluoromethane    | 3.512          | 101  | 321596              | 48.835  | ug/l   | 97       |
| 8) Diethyl Ether             | 3.965          | 74   | 142839              | 48.005  | ug/l   | 100      |
| 9) 1,1,2-Trichlorotrifluo... | 4.359          | 101  | 173195              | 49.131  | ug/l   | 97       |
| 10) 1,1-Dichloroethene       | 4.336          | 96   | 182248              | 48.464  | ug/l   | 98       |
| 11) Methyl Iodide            | 4.577          | 142  | 183665              | 52.992  | ug/l   | 97       |
| 12) Methyl Acetate           | 5.018          | 43   | 291092              | 48.943  | ug/l   | 91       |
| 13) Acrolein                 | 4.177          | 56   | 214422              | 226.650 | ug/l # | 6        |
| 14) Acrylonitrile            | 5.706          | 53   | 717034              | 243.043 | ug/l   | 99       |
| 15) Acetone                  | 4.424          | 58   | 206350              | 238.097 | ug/l   | 90       |
| 16) Carbon Disulfide         | 4.700          | 76   | 571725              | 48.499  | ug/l   | 97       |
| 17) Allyl chloride           | 5.012          | 41   | 335231              | 47.243  | ug/l   | 98       |
| 18) Methylene Chloride       | 5.259          | 84   | 231926              | 48.524  | ug/l   | 96       |
| 19) trans-1,2-Dichloroethene | 5.771          | 96   | 206627              | 47.525  | ug/l   | 90       |
| 20) Diisopropyl ether        | 6.659          | 45   | 784464              | 48.559  | ug/l   | 98       |
| 21) 1,1-Dichloroethane       | 6.553          | 63   | 402552              | 47.866  | ug/l   | 97       |
| 22) cis-1,2-Dichloroethene   | 7.471          | 96   | 262837              | 48.293  | ug/l   | 99       |
| 23) tert-Butyl Alcohol       | 5.530          | 59   | 286613              | 242.230 | ug/l # | 100      |
| 24) Methyl tert-Butyl Ether  | 5.789          | 73   | 787846              | 48.537  | ug/l   | 98       |
| 25) Chloroform               | 7.947          | 83   | 416216              | 47.959  | ug/l   | 100      |
| 26) Cyclohexane              | 8.241          | 56   | 336848              | 48.343  | ug/l # | 95       |
| 29) 1,1-Dichloropropene      | 8.353          | 75   | 280084              | 48.002  | ug/l   | 99       |
| 30) 2-Butanone               | 7.471          | 43   | 1034181             | 240.318 | ug/l   | 99       |
| 31) 2,2-Dichloropropane      | 7.471          | 77   | 357946              | 47.547  | ug/l   | 100      |
| 32) 1,1,1-Trichloroethane    | 8.153          | 97   | 348014              | 47.733  | ug/l   | 99       |
| 33) Carbon Tetrachloride     | 8.341          | 117  | 295178              | 48.335  | ug/l   | 97       |
| 34) Benzene                  | 8.588          | 78   | 926605              | 47.877  | ug/l   | 99       |
| 35) Methacrylonitrile        | 7.765          | 41   | 221496              | 48.945  | ug/l   | 94       |
| 36) 1,2-Dichloroethane       | 8.653          | 62   | 320898              | 47.761  | ug/l   | 99       |
| 37) Trichloroethene          | 9.329          | 130  | 218229              | 47.593  | ug/l   | 100      |
| 38) Methylcyclohexane        | 9.582          | 83   | 347822              | 48.271  | ug/l   | 99       |
| 39) 1,2-Dichloropropane      | 9.600          | 63   | 228797              | 47.147  | ug/l   | 94       |
| 40) Dibromomethane           | 9.688          | 93   | 170079              | 48.148  | ug/l   | 98       |
| 41) Bromodichloromethane     | 9.871          | 83   | 340823              | 47.241  | ug/l   | 99       |
| 42) Vinyl Acetate            | 6.588          | 43   | 3449178             | 235.274 | ug/l   | 99       |

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN100825\  
 Data File : VN088032.D  
 Acq On : 08 Oct 2025 10:30  
 Operator : JC\MD  
 Sample : VSTDICC050  
 Misc : 5.0mL/MSVOA\_N/WATER  
 ALS Vial : 4 Sample Multiplier: 1

Instrument :  
 MSVOA\_N  
 ClientSampleId :  
 VSTDICC050

Manual Integrations  
 APPROVED

Reviewed By : John Carlone 10/09/2025  
 Supervised By : Mahesh Dadoda 10/10/2025

Quant Time: Oct 09 02:02:43 2025  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\624N100825W.M  
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS  
 QLast Update : Thu Oct 09 01:59:51 2025  
 Response via : Initial Calibration


| Compound                      | R.T.   | QIon | Response | Conc    | Units | Dev(Min) |
|-------------------------------|--------|------|----------|---------|-------|----------|
| 43) Ethyl Acetate             | 7.547  | 43   | 386755   | 46.753  | ug/l  | 95       |
| 44) Isopropyl Acetate         | 8.671  | 43   | 658172   | 48.713  | ug/l  | 98       |
| 45) 1,4-Dioxane               | 9.682  | 88   | 88246    | 982.026 | ug/l  | 98       |
| 46) Methyl methacrylate       | 9.665  | 41   | 292033   | 49.080  | ug/l  | 96       |
| 47) n-amyl Acetate            | 12.506 | 43   | 301319m  | 50.755  | ug/l  |          |
| 48) t-1,3-Dichloropropene     | 10.818 | 75   | 387494   | 48.302  | ug/l  | 97       |
| 49) cis-1,3-Dichloropropene   | 10.294 | 75   | 399785   | 47.774  | ug/l  | 94       |
| 50) 1,1,2-Trichloroethane     | 10.994 | 97   | 224582   | 48.009  | ug/l  | 95       |
| 51) Ethyl methacrylate        | 10.853 | 69   | 395806   | 49.995  | ug/l  | 99       |
| 52) 1,3-Dichloropropane       | 11.141 | 76   | 390914   | 47.897  | ug/l  | 99       |
| 53) Dibromochloromethane      | 11.335 | 129  | 269681   | 48.222  | ug/l  | 100      |
| 54) 1,2-Dibromoethane         | 11.447 | 107  | 238540   | 47.697  | ug/l  | 99       |
| 55) 2-Chloroethyl vinyl ether | 10.141 | 63   | 1061685  | 248.447 | ug/l  | 100      |
| 56) Bromoform                 | 12.559 | 173  | 185069   | 48.948  | ug/l  | 99       |
| 58) 4-Methyl-2-Pentanone      | 10.424 | 43   | 1992861  | 244.858 | ug/l  | 99       |
| 59) 2-Hexanone                | 11.176 | 43   | 1382348  | 257.249 | ug/l  | 99       |
| 61) Tetrachloroethene         | 11.082 | 164  | 177933   | 48.289  | ug/l  | 95       |
| 62) Toluene                   | 10.612 | 91   | 997746   | 48.341  | ug/l  | 99       |
| 64) Chlorobenzene             | 11.871 | 112  | 642529   | 48.738  | ug/l  | 98       |
| 65) 1,1,1,2-Tetrachloroethane | 11.941 | 131  | 223784   | 49.257  | ug/l  | 97       |
| 66) Ethyl Benzene             | 11.941 | 91   | 1091616  | 48.639  | ug/l  | 97       |
| 67) m/p-Xylenes               | 12.047 | 106  | 840749   | 97.174  | ug/l  | 99       |
| 68) o-Xylene                  | 12.376 | 106  | 415634   | 48.257  | ug/l  | 96       |
| 69) Styrene                   | 12.388 | 104  | 714533   | 50.151  | ug/l  | 98       |
| 70) Isopropylbenzene          | 12.676 | 105  | 1034903  | 49.025  | ug/l  | 100      |
| 71) 1,1,2,2-Tetrachloroethane | 12.918 | 83   | 349263   | 48.801  | ug/l  | 98       |
| 72) 1,2,3-Trichloropropane    | 12.970 | 75   | 331890m  | 47.774  | ug/l  |          |
| 73) Bromobenzene              | 12.959 | 156  | 255670   | 49.554  | ug/l  | 98       |
| 74) n-propylbenzene           | 13.012 | 91   | 1253446  | 49.187  | ug/l  | 100      |
| 75) 2-Chlorotoluene           | 13.100 | 91   | 759413   | 48.738  | ug/l  | 99       |
| 76) 1,3,5-Trimethylbenzene    | 13.153 | 105  | 881209   | 49.143  | ug/l  | 99       |
| 77) t-1,4-Dichloro-2-butene   | 12.718 | 75   | 129603   | 50.413  | ug/l  | 93       |
| 78) 4-Chlorotoluene           | 13.200 | 91   | 785069   | 49.593  | ug/l  | 98       |
| 79) tert-butylbenzene         | 13.412 | 119  | 769068   | 49.475  | ug/l  | 99       |
| 80) 1,2,4-Trimethylbenzene    | 13.459 | 105  | 881036   | 49.081  | ug/l  | 100      |
| 81) sec-Butylbenzene          | 13.594 | 105  | 1073333  | 49.110  | ug/l  | 100      |
| 82) p-Isopropyltoluene        | 13.706 | 119  | 922870   | 49.458  | ug/l  | 99       |
| 83) 1,3-Dichlorobenzene       | 13.712 | 146  | 483694   | 49.826  | ug/l  | 99       |
| 84) 1,4-Dichlorobenzene       | 13.788 | 146  | 486086   | 48.552  | ug/l  | 99       |
| 85) n-Butylbenzene            | 14.035 | 91   | 861074   | 49.601  | ug/l  | 99       |
| 86) Hexachloroethane          | 14.312 | 117  | 184332   | 49.313  | ug/l  | 96       |
| 87) 1,2-Dichlorobenzene       | 14.082 | 146  | 464249   | 48.621  | ug/l  | 100      |
| 88) 1,2-Dibromo-3-Chloropr... | 14.700 | 75   | 80611    | 49.043  | ug/l  | 98       |
| 89) 1,2,4-Trichlorobenzene    | 15.811 | 180  | 275876   | 48.506  | ug/l  | 98       |
| 90) Hexachlorobutadiene       | 15.470 | 225  | 98031    | 49.524  | ug/l  | 97       |
| 91) Naphthalene               | 15.611 | 128  | 1015294  | 48.760  | ug/l  | 99       |
| 92) 1,2,3-Trichlorobenzene    | 15.811 | 180  | 275876   | 48.506  | ug/l  | 98       |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

**Instrument :**  
MSVOA\_N  
**ClientSampleId :**  
VSTDICC050

## Manual Integrations APPROVED

Reviewed By :John Carlone 10/09/2025  
Supervised By :Mahesh Dadoda 10/10/2025





Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN100825\  
 Data File : VN088033.D  
 Acq On : 08 Oct 2025 10:51  
 Operator : JC\MD  
 Sample : VSTDICC100  
 Misc : 5.0mL/MSVOA\_N/WATER  
 ALS Vial : 5 Sample Multiplier: 1

Instrument :  
 MSVOA\_N  
 ClientSampleId :  
 VSTDICC100

Manual Integrations  
 APPROVED

Reviewed By : John Carlone 10/09/2025  
 Supervised By : Mahesh Dadoda 10/10/2025

Quant Time: Oct 09 02:03:50 2025  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\624N100825W.M  
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS  
 QLast Update : Thu Oct 09 01:59:51 2025  
 Response via : Initial Calibration

| Compound                     | R.T.           | QIon | Response            | Conc    | Units  | Dev(Min) |
|------------------------------|----------------|------|---------------------|---------|--------|----------|
| -----                        |                |      |                     |         |        |          |
| Internal Standards           |                |      |                     |         |        |          |
| 1) Bromochloromethane        | 7.794          | 128  | 67628               | 30.000  | ug/l   | 0.00     |
| 28) 1,4-Difluorobenzene      | 9.083          | 114  | 378736              | 30.000  | ug/l   | 0.00     |
| 57) Chlorobenzene-d5         | 11.847         | 117  | 351656              | 30.000  | ug/l   | 0.00     |
|                              |                |      |                     |         |        |          |
| System Monitoring Compounds  |                |      |                     |         |        |          |
| 27) 1,2-Dichloroethane-d4    | 8.559          | 65   | 160840              | 30.543  | ug/l   | 0.00     |
| Spiked Amount 30.000         | Range 91 - 110 |      | Recovery = 101.800% |         |        |          |
| 60) 4-Bromofluorobenzene     | 12.829         | 95   | 178151              | 30.730  | ug/l   | 0.00     |
| Spiked Amount 30.000         | Range 63 - 112 |      | Recovery = 102.433% |         |        |          |
| 63) Toluene-d8               | 10.547         | 98   | 467474              | 29.837  | ug/l   | 0.00     |
| Spiked Amount 30.000         | Range 91 - 112 |      | Recovery = 99.467%  |         |        |          |
|                              |                |      |                     |         |        |          |
| Target Compounds             |                |      |                     |         |        | Qvalue   |
| 2) Dichlorodifluoromethane   | 2.142          | 85   | 420916              | 108.665 | ug/l   | 95       |
| 3) Chloromethane             | 2.383          | 50   | 463730              | 105.596 | ug/l   | 98       |
| 4) Vinyl Chloride            | 2.536          | 62   | 482888              | 105.455 | ug/l   | 100      |
| 5) Bromomethane              | 2.959          | 94   | 273063              | 110.897 | ug/l   | 98       |
| 6) Chloroethane              | 3.130          | 64   | 314009              | 103.956 | ug/l   | 100      |
| 7) Trichlorofluoromethane    | 3.501          | 101  | 654258              | 105.071 | ug/l   | 97       |
| 8) Diethyl Ether             | 3.959          | 74   | 290416              | 103.221 | ug/l   | 99       |
| 9) 1,1,2-Trichlorotrifluo... | 4.365          | 101  | 351149              | 105.345 | ug/l   | 97       |
| 10) 1,1-Dichloroethene       | 4.336          | 96   | 359076              | 100.983 | ug/l   | 99       |
| 11) Methyl Iodide            | 4.577          | 142  | 430900              | 131.483 | ug/l   | 94       |
| 12) Methyl Acetate           | 5.012          | 43   | 608023m             | 108.115 | ug/l   |          |
| 13) Acrolein                 | 4.177          | 56   | 463612              | 518.261 | ug/l # | 7        |
| 14) Acrylonitrile            | 5.712          | 53   | 1467014             | 525.879 | ug/l   | 99       |
| 15) Acetone                  | 4.424          | 58   | 407619              | 497.406 | ug/l   | 93       |
| 16) Carbon Disulfide         | 4.700          | 76   | 1166017             | 104.606 | ug/l   | 100      |
| 17) Allyl chloride           | 5.012          | 41   | 683636              | 101.888 | ug/l   | 96       |
| 18) Methylene Chloride       | 5.259          | 84   | 461958              | 102.216 | ug/l   | 95       |
| 19) trans-1,2-Dichloroethene | 5.771          | 96   | 426250              | 103.682 | ug/l   | 93       |
| 20) Diisopropyl ether        | 6.659          | 45   | 1616050             | 105.793 | ug/l   | 98       |
| 21) 1,1-Dichloroethane       | 6.553          | 63   | 830085              | 104.385 | ug/l   | 95       |
| 22) cis-1,2-Dichloroethene   | 7.471          | 96   | 536722              | 104.292 | ug/l   | 99       |
| 23) tert-Butyl Alcohol       | 5.530          | 59   | 576246              | 515.047 | ug/l # | 100      |
| 24) Methyl tert-Butyl Ether  | 5.789          | 73   | 1604709             | 104.553 | ug/l   | 98       |
| 25) Chloroform               | 7.947          | 83   | 840390              | 102.409 | ug/l   | 100      |
| 26) Cyclohexane              | 8.236          | 56   | 682990              | 103.662 | ug/l # | 95       |
| 29) 1,1-Dichloropropene      | 8.353          | 75   | 570703              | 105.125 | ug/l   | 98       |
| 30) 2-Butanone               | 7.471          | 43   | 2104280             | 525.556 | ug/l   | 99       |
| 31) 2,2-Dichloropropane      | 7.471          | 77   | 713241              | 101.828 | ug/l   | 100      |
| 32) 1,1,1-Trichloroethane    | 8.147          | 97   | 703319              | 103.680 | ug/l   | 99       |
| 33) Carbon Tetrachloride     | 8.341          | 117  | 599874              | 105.575 | ug/l   | 98       |
| 34) Benzene                  | 8.588          | 78   | 1880026             | 104.406 | ug/l   | 98       |
| 35) Methacrylonitrile        | 7.765          | 41   | 432524              | 102.726 | ug/l   | 94       |
| 36) 1,2-Dichloroethane       | 8.653          | 62   | 649488              | 103.896 | ug/l   | 99       |
| 37) Trichloroethene          | 9.335          | 130  | 443090              | 103.860 | ug/l   | 100      |
| 38) Methylcyclohexane        | 9.582          | 83   | 700603              | 104.502 | ug/l   | 99       |
| 39) 1,2-Dichloropropane      | 9.600          | 63   | 473093              | 104.780 | ug/l   | 99       |
| 40) Dibromomethane           | 9.688          | 93   | 342480              | 104.205 | ug/l   | 97       |
| 41) Bromodichloromethane     | 9.871          | 83   | 701393              | 104.491 | ug/l   | 99       |
| 42) Vinyl Acetate            | 6.589          | 43   | 7544223             | 553.092 | ug/l   | 97       |

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN100825\  
 Data File : VN088033.D  
 Acq On : 08 Oct 2025 10:51  
 Operator : JC\MD  
 Sample : VSTDICC100  
 Misc : 5.0mL/MSVOA\_N/WATER  
 ALS Vial : 5 Sample Multiplier: 1

Instrument :  
 MSVOA\_N  
 ClientSampleId :  
 VSTDICC100

Manual Integrations  
 APPROVED

Reviewed By : John Carlone 10/09/2025  
 Supervised By : Mahesh Dadoda 10/10/2025

Quant Time: Oct 09 02:03:50 2025  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\624N100825W.M  
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS  
 QLast Update : Thu Oct 09 01:59:51 2025  
 Response via : Initial Calibration

| Compound                      | R.T.   | QIon | Response | Conc     | Units  | Dev(Min) |
|-------------------------------|--------|------|----------|----------|--------|----------|
| 43) Ethyl Acetate             | 7.547  | 43   | 824040   | 107.065  | ug/l   | 96       |
| 44) Isopropyl Acetate         | 8.671  | 43   | 1339571  | 106.561  | ug/l   | 98       |
| 45) 1,4-Dioxane               | 9.682  | 88   | 178569   | 2135.793 | ug/l   | 96       |
| 46) Methyl methacrylate       | 9.665  | 41   | 606493   | 109.554  | ug/l   | 96       |
| 47) n-amyl Acetate            | 12.488 | 43   | 635857   | 115.116  | ug/l # | 64       |
| 48) t-1,3-Dichloropropene     | 10.818 | 75   | 804743   | 107.816  | ug/l   | 100      |
| 49) cis-1,3-Dichloropropene   | 10.294 | 75   | 826516   | 106.155  | ug/l   | 95       |
| 50) 1,1,2-Trichloroethane     | 10.994 | 97   | 450541   | 103.515  | ug/l   | 99       |
| 51) Ethyl methacrylate        | 10.853 | 69   | 831363   | 112.866  | ug/l   | 98       |
| 52) 1,3-Dichloropropane       | 11.141 | 76   | 810597   | 106.748  | ug/l   | 98       |
| 53) Dibromochloromethane      | 11.335 | 129  | 559787   | 107.583  | ug/l   | 98       |
| 54) 1,2-Dibromoethane         | 11.447 | 107  | 493699   | 106.100  | ug/l   | 100      |
| 55) 2-Chloroethyl vinyl ether | 10.141 | 63   | 2132450  | 536.342  | ug/l   | 100      |
| 56) Bromoform                 | 12.559 | 173  | 386003   | 109.728  | ug/l   | 98       |
| 58) 4-Methyl-2-Pentanone      | 10.424 | 43   | 4116685  | 523.371  | ug/l   | 99       |
| 59) 2-Hexanone                | 11.177 | 43   | 2928211  | 563.852  | ug/l   | 99       |
| 61) Tetrachloroethene         | 11.082 | 164  | 362506   | 101.796  | ug/l   | 93       |
| 62) Toluene                   | 10.612 | 91   | 2027659  | 101.652  | ug/l   | 98       |
| 64) Chlorobenzene             | 11.871 | 112  | 1317139  | 103.379  | ug/l   | 98       |
| 65) 1,1,1,2-Tetrachloroethane | 11.941 | 131  | 448505   | 102.148  | ug/l   | 98       |
| 66) Ethyl Benzene             | 11.941 | 91   | 2249065  | 103.692  | ug/l   | 97       |
| 67) m/p-Xylenes               | 12.047 | 106  | 1731517  | 207.080  | ug/l   | 99       |
| 68) o-Xylene                  | 12.376 | 106  | 858288   | 103.111  | ug/l   | 97       |
| 69) Styrene                   | 12.388 | 104  | 1459636  | 106.005  | ug/l   | 98       |
| 70) Isopropylbenzene          | 12.676 | 105  | 2115128  | 103.677  | ug/l   | 99       |
| 71) 1,1,2,2-Tetrachloroethane | 12.918 | 83   | 717005   | 103.664  | ug/l   | 99       |
| 72) 1,2,3-Trichloropropane    | 12.970 | 75   | 669076m  | 99.654   | ug/l   |          |
| 73) Bromobenzene              | 12.959 | 156  | 514996   | 103.282  | ug/l   | 95       |
| 74) n-propylbenzene           | 13.012 | 91   | 2606776  | 105.847  | ug/l   | 99       |
| 75) 2-Chlorotoluene           | 13.100 | 91   | 1584486  | 105.221  | ug/l   | 99       |
| 76) 1,3,5-Trimethylbenzene    | 13.153 | 105  | 1814925  | 104.730  | ug/l   | 98       |
| 77) t-1,4-Dichloro-2-butene   | 12.718 | 75   | 290843   | 117.060  | ug/l   | 89       |
| 78) 4-Chlorotoluene           | 13.200 | 91   | 1632331  | 106.695  | ug/l   | 98       |
| 79) tert-butylbenzene         | 13.418 | 119  | 1573625  | 104.749  | ug/l   | 99       |
| 80) 1,2,4-Trimethylbenzene    | 13.459 | 105  | 1831829  | 105.591  | ug/l   | 98       |
| 81) sec-Butylbenzene          | 13.594 | 105  | 2207832  | 104.528  | ug/l   | 99       |
| 82) p-Isopropyltoluene        | 13.706 | 119  | 1878975  | 104.194  | ug/l   | 100      |
| 83) 1,3-Dichlorobenzene       | 13.712 | 146  | 978389   | 104.286  | ug/l   | 100      |
| 84) 1,4-Dichlorobenzene       | 13.788 | 146  | 993969   | 102.730  | ug/l   | 98       |
| 85) n-Butylbenzene            | 14.035 | 91   | 1765582  | 105.237  | ug/l   | 99       |
| 86) Hexachloroethane          | 14.312 | 117  | 371189   | 102.750  | ug/l   | 98       |
| 87) 1,2-Dichlorobenzene       | 14.082 | 146  | 959910   | 104.023  | ug/l   | 99       |
| 88) 1,2-Dibromo-3-Chloropr... | 14.694 | 75   | 165583   | 104.237  | ug/l   | 95       |
| 89) 1,2,4-Trichlorobenzene    | 15.812 | 180  | 577645   | 105.091  | ug/l   | 98       |
| 90) Hexachlorobutadiene       | 15.476 | 225  | 194567   | 101.706  | ug/l   | 98       |
| 91) Naphthalene               | 15.611 | 128  | 2133295  | 106.010  | ug/l   | 100      |
| 92) 1,2,3-Trichlorobenzene    | 15.812 | 180  | 577645   | 105.091  | ug/l   | 98       |

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN100825\  
Data File : VN088033.D  
Acq On : 08 Oct 2025 10:51  
Operator : JC\MD  
Sample : VSTDICC100  
Misc : 5.0mL/MSVOA\_N/WATER  
ALS Vial : 5 Sample Multiplier: 1

**Instrument :**  
MSVOA\_N  
**ClientSampleId :**  
VSTDICC100

## Manual Integrations APPROVED

Reviewed By :John Carlone 10/09/2025  
Supervised By :Mahesh Dadoda 10/10/2025

Quant Time: Oct 09 02:03:50 2025  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\624N100825W.M  
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS  
QLast Update : Thu Oct 09 01:59:51 2025  
Response via : Initial Calibration



Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN100825\  
 Data File : VN088034.D  
 Acq On : 08 Oct 2025 11:12  
 Operator : JC\MD  
 Sample : VSTDICC150  
 Misc : 5.0mL/MSVOA\_N/WATER  
 ALS Vial : 6 Sample Multiplier: 1

## Instrument :

MSVOA\_N

## ClientSampleId :

VSTDICC150

## Manual Integrations

## APPROVED

Reviewed By :John Carlone 10/09/2025

Supervised By :Mahesh Dadoda 10/10/2025

Quant Time: Oct 09 02:04:52 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\624N100825W.M

Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS

QLast Update : Thu Oct 09 01:59:51 2025

Response via : Initial Calibration

| Compound                     | R.T.           | QIon | Response   | Conc    | Units  | Dev(Min) |
|------------------------------|----------------|------|------------|---------|--------|----------|
| -----                        |                |      |            |         |        |          |
| Internal Standards           |                |      |            |         |        |          |
| 1) Bromochloromethane        | 7.800          | 128  | 70768      | 30.000  | ug/l   | 0.00     |
| 28) 1,4-Difluorobenzene      | 9.083          | 114  | 396399     | 30.000  | ug/l   | 0.00     |
| 57) Chlorobenzene-d5         | 11.841         | 117  | 360347     | 30.000  | ug/l   | 0.00     |
|                              |                |      |            |         |        |          |
| System Monitoring Compounds  |                |      |            |         |        |          |
| 27) 1,2-Dichloroethane-d4    | 8.565          | 65   | 164118     | 29.782  | ug/l   | 0.00     |
| Spiked Amount 30.000         | Range 91 - 110 |      | Recovery = | 99.267% |        |          |
| 60) 4-Bromofluorobenzene     | 12.829         | 95   | 176216     | 29.663  | ug/l   | 0.00     |
| Spiked Amount 30.000         | Range 63 - 112 |      | Recovery = | 98.867% |        |          |
| 63) Toluene-d8               | 10.547         | 98   | 480539     | 29.931  | ug/l   | 0.00     |
| Spiked Amount 30.000         | Range 91 - 112 |      | Recovery = | 99.767% |        |          |
|                              |                |      |            |         |        |          |
| Target Compounds             |                |      |            |         |        | Qvalue   |
| 2) Dichlorodifluoromethane   | 2.142          | 85   | 614191     | 151.526 | ug/l   | 95       |
| 3) Chloromethane             | 2.383          | 50   | 688323     | 149.783 | ug/l   | 98       |
| 4) Vinyl Chloride            | 2.536          | 62   | 700818     | 146.257 | ug/l   | 100      |
| 5) Bromomethane              | 2.954          | 94   | 390993     | 151.745 | ug/l   | 96       |
| 6) Chloroethane              | 3.124          | 64   | 447592     | 141.605 | ug/l   | 95       |
| 7) Trichlorofluoromethane    | 3.506          | 101  | 939283     | 144.151 | ug/l   | 100      |
| 8) Diethyl Ether             | 3.959          | 74   | 416480     | 141.459 | ug/l   | 100      |
| 9) 1,1,2-Trichlorotrifluo... | 4.359          | 101  | 499641     | 143.242 | ug/l   | 97       |
| 10) 1,1-Dichloroethene       | 4.330          | 96   | 529290     | 142.248 | ug/l   | 98       |
| 11) Methyl Iodide            | 4.577          | 142  | 649112     | 189.280 | ug/l   | 96       |
| 12) Methyl Acetate           | 5.018          | 43   | 874170     | 148.543 | ug/l   | 90       |
| 13) Acrolein                 | 4.171          | 56   | 722228     | 771.539 | ug/l # | 6        |
| 14) Acrylonitrile            | 5.706          | 53   | 2111000    | 723.151 | ug/l   | 100      |
| 15) Acetone                  | 4.424          | 58   | 565771     | 659.761 | ug/l   | 92       |
| 16) Carbon Disulfide         | 4.700          | 76   | 1690921    | 144.966 | ug/l   | 100      |
| 17) Allyl chloride           | 5.006          | 41   | 1008388    | 143.620 | ug/l   | 98       |
| 18) Methylene Chloride       | 5.259          | 84   | 670686     | 141.817 | ug/l   | 97       |
| 19) trans-1,2-Dichloroethene | 5.771          | 96   | 628517     | 146.099 | ug/l   | 89       |
| 20) Diisopropyl ether        | 6.659          | 45   | 2377730    | 148.749 | ug/l   | 97       |
| 21) 1,1-Dichloroethane       | 6.547          | 63   | 1214899    | 145.997 | ug/l   | 96       |
| 22) cis-1,2-Dichloroethene   | 7.471          | 96   | 768626     | 142.727 | ug/l   | 98       |
| 23) tert-Butyl Alcohol       | 5.536          | 59   | 789077     | 673.981 | ug/l # | 100      |
| 24) Methyl tert-Butyl Ether  | 5.783          | 73   | 2330339    | 145.094 | ug/l   | 99       |
| 25) Chloroform               | 7.947          | 83   | 1228670    | 143.082 | ug/l   | 97       |
| 26) Cyclohexane              | 8.241          | 56   | 1000427    | 145.104 | ug/l # | 96       |
| 29) 1,1-Dichloropropene      | 8.353          | 75   | 836265     | 147.179 | ug/l   | 97       |
| 30) 2-Butanone               | 7.471          | 43   | 2998106    | 715.429 | ug/l   | 98       |
| 31) 2,2-Dichloropropane      | 7.471          | 77   | 1019106    | 139.013 | ug/l   | 100      |
| 32) 1,1,1-Trichloroethane    | 8.153          | 97   | 1018706    | 143.482 | ug/l   | 99       |
| 33) Carbon Tetrachloride     | 8.341          | 117  | 869642     | 146.233 | ug/l   | 97       |
| 34) Benzene                  | 8.588          | 78   | 2743909    | 145.591 | ug/l   | 98       |
| 35) Methacrylonitrile        | 7.765          | 41   | 627210     | 142.327 | ug/l   | 97       |
| 36) 1,2-Dichloroethane       | 8.653          | 62   | 954996     | 145.960 | ug/l   | 99       |
| 37) Trichloroethene          | 9.335          | 130  | 642810     | 143.961 | ug/l   | 100      |
| 38) Methylcyclohexane        | 9.582          | 83   | 1002978    | 142.937 | ug/l   | 98       |
| 39) 1,2-Dichloropropane      | 9.600          | 63   | 688269     | 145.644 | ug/l   | 99       |
| 40) Dibromomethane           | 9.688          | 93   | 497967     | 144.763 | ug/l   | 98       |
| 41) Bromodichloromethane     | 9.871          | 83   | 1015193    | 144.501 | ug/l   | 98       |
| 42) Vinyl Acetate            | 6.589          | 43   | 11007914   | 771.067 | ug/l   | 97       |

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN100825\  
 Data File : VN088034.D  
 Acq On : 08 Oct 2025 11:12  
 Operator : JC\MD  
 Sample : VSTDICC150  
 Misc : 5.0mL/MSVOA\_N/WATER  
 ALS Vial : 6 Sample Multiplier: 1

## Instrument :

MSVOA\_N

## ClientSampleId :

VSTDICC150

## Manual Integrations

## APPROVED

Reviewed By :John Carlone 10/09/2025

Supervised By :Mahesh Dadoda 10/10/2025

Quant Time: Oct 09 02:04:52 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\624N100825W.M

Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS

QLast Update : Thu Oct 09 01:59:51 2025

Response via : Initial Calibration

| Compound                      | R.T.   | QIon | Response | Conc     | Units  | Dev(Min) |
|-------------------------------|--------|------|----------|----------|--------|----------|
| 43) Ethyl Acetate             | 7.547  | 43   | 1168944  | 145.110  | ug/l   | 94       |
| 44) Isopropyl Acetate         | 8.677  | 43   | 1946213  | 147.920  | ug/l   | 98       |
| 45) 1,4-Dioxane               | 9.682  | 88   | 230506   | 2634.144 | ug/l   | 98       |
| 46) Methyl methacrylate       | 9.665  | 41   | 878594   | 151.633  | ug/l   | 98       |
| 47) n-amyl Acetate            | 12.482 | 43   | 1141223  | 197.401  | ug/l # | 64       |
| 48) t-1,3-Dichloropropene     | 10.818 | 75   | 1169571  | 149.712  | ug/l   | 100      |
| 49) cis-1,3-Dichloropropene   | 10.294 | 75   | 1203477  | 147.684  | ug/l   | 96       |
| 50) 1,1,2-Trichloroethane     | 10.994 | 97   | 657933   | 144.430  | ug/l   | 97       |
| 51) Ethyl methacrylate        | 10.853 | 69   | 1210974  | 157.077  | ug/l   | 99       |
| 52) 1,3-Dichloropropane       | 11.141 | 76   | 1159064  | 145.836  | ug/l   | 100      |
| 53) Dibromochloromethane      | 11.341 | 129  | 802448   | 147.348  | ug/l   | 100      |
| 54) 1,2-Dibromoethane         | 11.447 | 107  | 713476   | 146.499  | ug/l   | 100      |
| 55) 2-Chloroethyl vinyl ether | 10.141 | 63   | 3115151  | 748.593  | ug/l   | 99       |
| 56) Bromoform                 | 12.559 | 173  | 556224   | 151.071  | ug/l   | 99       |
| 58) 4-Methyl-2-Pentanone      | 10.424 | 43   | 5819803  | 722.051  | ug/l   | 99       |
| 59) 2-Hexanone                | 11.177 | 43   | 4169658  | 783.538  | ug/l   | 100      |
| 61) Tetrachloroethene         | 11.082 | 164  | 517435   | 141.798  | ug/l   | 94       |
| 62) Toluene                   | 10.612 | 91   | 2951946  | 144.419  | ug/l   | 99       |
| 64) Chlorobenzene             | 11.871 | 112  | 1890882  | 144.832  | ug/l   | 99       |
| 65) 1,1,1,2-Tetrachloroethane | 11.941 | 131  | 643792   | 143.089  | ug/l   | 98       |
| 66) Ethyl Benzene             | 11.941 | 91   | 3229064  | 145.284  | ug/l   | 98       |
| 67) m/p-Xylenes               | 12.047 | 106  | 2496336  | 291.347  | ug/l   | 99       |
| 68) o-Xylene                  | 12.376 | 106  | 1219294  | 142.948  | ug/l   | 95       |
| 69) Styrene                   | 12.388 | 104  | 2093655  | 148.383  | ug/l   | 99       |
| 70) Isopropylbenzene          | 12.671 | 105  | 3025763  | 144.736  | ug/l   | 99       |
| 71) 1,1,2,2-Tetrachloroethane | 12.918 | 83   | 1009711  | 142.462  | ug/l   | 99       |
| 72) 1,2,3-Trichloropropane    | 12.971 | 75   | 944811m  | 137.329  | ug/l   |          |
| 73) Bromobenzene              | 12.959 | 156  | 732655   | 143.390  | ug/l   | 95       |
| 74) n-propylbenzene           | 13.012 | 91   | 3681361  | 145.874  | ug/l   | 99       |
| 75) 2-Chlorotoluene           | 13.106 | 91   | 2247982  | 145.681  | ug/l   | 99       |
| 76) 1,3,5-Trimethylbenzene    | 13.153 | 105  | 2531712  | 142.568  | ug/l   | 99       |
| 77) t-1,4-Dichloro-2-butene   | 12.718 | 75   | 417387   | 163.941  | ug/l   | 90       |
| 78) 4-Chlorotoluene           | 13.200 | 91   | 2320550  | 148.022  | ug/l   | 98       |
| 79) tert-butylbenzene         | 13.418 | 119  | 2189957  | 142.259  | ug/l   | 99       |
| 80) 1,2,4-Trimethylbenzene    | 13.459 | 105  | 2534010  | 142.543  | ug/l   | 98       |
| 81) sec-Butylbenzene          | 13.594 | 105  | 3085212  | 142.543  | ug/l   | 99       |
| 82) p-Isopropyltoluene        | 13.706 | 119  | 2589136  | 140.112  | ug/l   | 100      |
| 83) 1,3-Dichlorobenzene       | 13.712 | 146  | 1360760  | 141.545  | ug/l   | 99       |
| 84) 1,4-Dichlorobenzene       | 13.788 | 146  | 1422602  | 143.484  | ug/l   | 99       |
| 85) n-Butylbenzene            | 14.035 | 91   | 2454577  | 142.775  | ug/l   | 99       |
| 86) Hexachloroethane          | 14.312 | 117  | 525061   | 141.838  | ug/l   | 97       |
| 87) 1,2-Dichlorobenzene       | 14.082 | 146  | 1340645  | 141.778  | ug/l   | 99       |
| 88) 1,2-Dibromo-3-Chloropr... | 14.694 | 75   | 233320   | 143.337  | ug/l   | 96       |
| 89) 1,2,4-Trichlorobenzene    | 15.812 | 180  | 829215   | 147.221  | ug/l   | 99       |
| 90) Hexachlorobutadiene       | 15.476 | 225  | 273878   | 139.711  | ug/l   | 96       |
| 91) Naphthalene               | 15.612 | 128  | 3044921  | 147.662  | ug/l   | 100      |
| 92) 1,2,3-Trichlorobenzene    | 15.812 | 180  | 829215   | 147.221  | ug/l   | 99       |

(#) = qualifier out of range (m) = manual integration (+) = signals summed



Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN100825\  
 Data File : VN088034.D  
 Acq On : 08 Oct 2025 11:12  
 Operator : JC\MD  
 Sample : VSTDIC150  
 Misc : 5.0mL/MSVOA\_N/WATER  
 ALS Vial : 6 Sample Multiplier: 1

## Instrument :

MSVOA\_N

## Client Sample Id :

VSTDIC150

## Manual Integrations

## APPROVED

Reviewed By : John Carlone 10/09/2025

Supervised By : Mahesh Dadoda 10/10/2025


Quant Time: Oct 09 02:04:52 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\624N100825W.M

Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS

QLast Update : Thu Oct 09 01:59:51 2025

Response via : Initial Calibration



Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN100825\  
 Data File : VN088036.D  
 Acq On : 08 Oct 2025 12:52  
 Operator : JC\MD  
 Sample : VSTDICV020  
 Misc : 5.0mL/MSVOA\_N/WATER  
 ALS Vial : 8 Sample Multiplier: 1

Instrument :  
 MSVOA\_N  
 ClientSampleId :  
 ICVVN100825

Manual Integrations  
 APPROVED

Reviewed By : John Carlone 10/09/2025  
 Supervised By : Mahesh Dadoda 10/10/2025

Quant Time: Oct 09 02:40:57 2025  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\624N100825W.M  
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS  
 QLast Update : Thu Oct 09 02:36:32 2025  
 Response via : Initial Calibration

| Compound                     | R.T.           | QIon | Response   | Conc     | Units  | Dev(Min) |
|------------------------------|----------------|------|------------|----------|--------|----------|
| -----                        |                |      |            |          |        |          |
| Internal Standards           |                |      |            |          |        |          |
| 1) Bromochloromethane        | 7.794          | 128  | 71678      | 30.000   | ug/l   | 0.00     |
| 28) 1,4-Difluorobenzene      | 9.083          | 114  | 399537     | 30.000   | ug/l   | 0.00     |
| 57) Chlorobenzene-d5         | 11.847         | 117  | 349709     | 30.000   | ug/l   | 0.00     |
|                              |                |      |            |          |        |          |
| System Monitoring Compounds  |                |      |            |          |        |          |
| 27) 1,2-Dichloroethane-d4    | 8.559          | 65   | 167135     | 29.945   | ug/l   | 0.00     |
| Spiked Amount 30.000         | Range 91 - 110 |      | Recovery = | 99.800%  |        |          |
| 60) 4-Bromofluorobenzene     | 12.829         | 95   | 172458     | 29.914   | ug/l   | 0.00     |
| Spiked Amount 30.000         | Range 63 - 112 |      | Recovery = | 99.700%  |        |          |
| 63) Toluene-d8               | 10.547         | 98   | 468600     | 30.075   | ug/l   | 0.00     |
| Spiked Amount 30.000         | Range 91 - 112 |      | Recovery = | 100.267% |        |          |
|                              |                |      |            |          |        |          |
| Target Compounds             |                |      |            |          |        | Qvalue   |
| 2) Dichlorodifluoromethane   | 2.142          | 85   | 82744      | 20.154   | ug/l   | 95       |
| 3) Chloromethane             | 2.383          | 50   | 93360      | 20.058   | ug/l   | 98       |
| 4) Vinyl Chloride            | 2.536          | 62   | 95282      | 19.632   | ug/l   | 96       |
| 5) Bromomethane              | 2.977          | 94   | 57202      | 21.918   | ug/l   | 99       |
| 6) Chloroethane              | 3.136          | 64   | 60278      | 18.828   | ug/l   | 97       |
| 7) Trichlorofluoromethane    | 3.512          | 101  | 127893     | 19.378   | ug/l   | 99       |
| 8) Diethyl Ether             | 3.965          | 74   | 55257      | 18.530   | ug/l   | 100      |
| 9) 1,1,2-Trichlorotrifluo... | 4.359          | 101  | 69948      | 19.799   | ug/l   | 97       |
| 10) 1,1-Dichloroethene       | 4.336          | 96   | 73588      | 19.526   | ug/l   | 98       |
| 11) Methyl Iodide            | 4.583          | 142  | 59897      | 17.244   | ug/l   | 93       |
| 12) Methyl Acetate           | 5.012          | 43   | 112275     | 18.325   | ug/l # | 88       |
| 13) Acrolein                 | 4.171          | 56   | 93925      | 99.064   | ug/l # | 7        |
| 14) Acrylonitrile            | 5.706          | 53   | 277894     | 93.988   | ug/l   | 100      |
| 15) Acetone                  | 4.430          | 58   | 77464m     | 89.462   | ug/l   |          |
| 16) Carbon Disulfide         | 4.706          | 76   | 233680     | 19.779   | ug/l   | 97       |
| 17) Allyl chloride           | 5.006          | 41   | 134223     | 18.874   | ug/l   | 95       |
| 18) Methylene Chloride       | 5.259          | 84   | 91812      | 19.167   | ug/l   | 94       |
| 19) trans-1,2-Dichloroethene | 5.777          | 96   | 84343      | 19.357   | ug/l   | 94       |
| 20) Diisopropyl ether        | 6.659          | 45   | 308391     | 19.048   | ug/l # | 97       |
| 21) 1,1-Dichloroethane       | 6.547          | 63   | 160073     | 18.992   | ug/l   | 97       |
| 22) cis-1,2-Dichloroethene   | 7.471          | 96   | 103897     | 19.048   | ug/l   | 98       |
| 23) tert-Butyl Alcohol       | 5.518          | 59   | 111186     | 93.763   | ug/l # | 100      |
| 24) Methyl tert-Butyl Ether  | 5.783          | 73   | 305027     | 18.751   | ug/l   | 100      |
| 25) Chloroform               | 7.953          | 83   | 166302     | 19.120   | ug/l   | 94       |
| 26) Cyclohexane              | 8.241          | 56   | 135705     | 19.433   | ug/l # | 97       |
| 29) 1,1-Dichloropropene      | 8.353          | 75   | 111713     | 19.507   | ug/l   | 98       |
| 30) 2-Butanone               | 7.471          | 43   | 383183     | 90.720   | ug/l   | 100      |
| 31) 2,2-Dichloropropane      | 7.477          | 77   | 136899     | 18.552   | ug/l   | 98       |
| 32) 1,1,1-Trichloroethane    | 8.153          | 97   | 138868     | 19.406   | ug/l   | 100      |
| 33) Carbon Tetrachloride     | 8.341          | 117  | 117258     | 19.562   | ug/l   | 93       |
| 34) Benzene                  | 8.583          | 78   | 362455     | 19.081   | ug/l   | 97       |
| 35) Methacrylonitrile        | 7.759          | 41   | 88280      | 19.875   | ug/l   | 95       |
| 36) 1,2-Dichloroethane       | 8.653          | 62   | 127793     | 19.378   | ug/l   | 100      |
| 37) Trichloroethene          | 9.336          | 130  | 86434      | 19.205   | ug/l   | 95       |
| 38) Methylcyclohexane        | 9.583          | 83   | 139333     | 19.701   | ug/l   | 99       |
| 39) 1,2-Dichloropropane      | 9.600          | 63   | 88633      | 18.608   | ug/l   | 98       |
| 40) Dibromomethane           | 9.688          | 93   | 65564      | 18.910   | ug/l   | 97       |
| 41) Bromodichloromethane     | 9.871          | 83   | 135709     | 19.165   | ug/l   | 98       |
| 42) Vinyl Acetate            | 6.589          | 43   | 1445002    | 100.422  | ug/l   | 98       |

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN100825\  
 Data File : VN088036.D  
 Acq On : 08 Oct 2025 12:52  
 Operator : JC\MD  
 Sample : VSTDICV020  
 Misc : 5.0mL/MSVOA\_N/WATER  
 ALS Vial : 8 Sample Multiplier: 1

Instrument :  
 MSVOA\_N  
 ClientSampleId :  
 ICVVN100825

Manual Integrations  
 APPROVED

Reviewed By : John Carlone 10/09/2025  
 Supervised By : Mahesh Dadoda 10/10/2025

Quant Time: Oct 09 02:40:57 2025  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\624N100825W.M  
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS  
 QLast Update : Thu Oct 09 02:36:32 2025  
 Response via : Initial Calibration

| Compound                      | R.T.   | QIon | Response | Conc    | Units | Dev(Min) |
|-------------------------------|--------|------|----------|---------|-------|----------|
| 43) Ethyl Acetate             | 7.547  | 43   | 156744   | 19.305  | ug/l  | 96       |
| 44) Isopropyl Acetate         | 8.671  | 43   | 246443   | 18.583  | ug/l  | 93       |
| 45) 1,4-Dioxane               | 9.677  | 88   | 36239    | 410.874 | ug/l  | 92       |
| 46) Methyl methacrylate       | 9.665  | 41   | 109997   | 18.835  | ug/l  | 98       |
| 47) n-amyl Acetate            | 12.618 | 43   | 128465m  | 20.438  | ug/l  |          |
| 48) t-1,3-Dichloropropene     | 10.818 | 75   | 145368   | 18.462  | ug/l  | 96       |
| 49) cis-1,3-Dichloropropene   | 10.288 | 75   | 153293   | 18.663  | ug/l  | 95       |
| 50) 1,1,2-Trichloroethane     | 10.994 | 97   | 85732    | 18.672  | ug/l  | 97       |
| 51) Ethyl methacrylate        | 10.859 | 69   | 146337   | 18.832  | ug/l  | 97       |
| 52) 1,3-Dichloropropane       | 11.141 | 76   | 152014   | 18.977  | ug/l  | 100      |
| 53) Dibromochloromethane      | 11.335 | 129  | 103530   | 18.861  | ug/l  | 99       |
| 54) 1,2-Dibromoethane         | 11.447 | 107  | 91960    | 18.734  | ug/l  | 100      |
| 55) 2-Chloroethyl vinyl ether | 10.141 | 63   | 416502   | 99.302  | ug/l  | 100      |
| 56) Bromoform                 | 12.559 | 173  | 68875    | 18.560  | ug/l  | 99       |
| 58) 4-Methyl-2-Pentanone      | 10.424 | 43   | 758466   | 96.964  | ug/l  | 99       |
| 59) 2-Hexanone                | 11.177 | 43   | 483068   | 93.537  | ug/l  | 99       |
| 61) Tetrachloroethene         | 11.082 | 164  | 71033    | 20.058  | ug/l  | 97       |
| 62) Toluene                   | 10.606 | 91   | 386375   | 19.478  | ug/l  | 98       |
| 64) Chlorobenzene             | 11.871 | 112  | 244551   | 19.301  | ug/l  | 99       |
| 65) 1,1,1,2-Tetrachloroethane | 11.941 | 131  | 85260    | 19.526  | ug/l  | 97       |
| 66) Ethyl Benzene             | 11.941 | 91   | 429071   | 19.892  | ug/l  | 97       |
| 67) m/p-Xylenes               | 12.047 | 106  | 324024   | 38.967  | ug/l  | 99       |
| 68) o-Xylene                  | 12.377 | 106  | 165214   | 19.959  | ug/l  | 99       |
| 69) Styrene                   | 12.388 | 104  | 265810   | 19.412  | ug/l  | 97       |
| 70) Isopropylbenzene          | 12.676 | 105  | 398248   | 19.630  | ug/l  | 100      |
| 71) 1,1,2,2-Tetrachloroethane | 12.918 | 83   | 131239   | 19.080  | ug/l  | 99       |
| 72) 1,2,3-Trichloropropane    | 12.971 | 75   | 127895m  | 20.021  | ug/l  |          |
| 73) Bromobenzene              | 12.959 | 156  | 95991    | 19.358  | ug/l  | 99       |
| 74) n-propylbenzene           | 13.012 | 91   | 477849   | 19.511  | ug/l  | 100      |
| 75) 2-Chlorotoluene           | 13.106 | 91   | 293367   | 19.590  | ug/l  | 99       |
| 76) 1,3,5-Trimethylbenzene    | 13.153 | 105  | 338338   | 19.632  | ug/l  | 99       |
| 77) t-1,4-Dichloro-2-butene   | 12.718 | 75   | 45124    | 17.725  | ug/l  | 92       |
| 78) 4-Chlorotoluene           | 13.200 | 91   | 304451   | 20.011  | ug/l  | 96       |
| 79) tert-butylbenzene         | 13.418 | 119  | 300216   | 20.095  | ug/l  | 99       |
| 80) 1,2,4-Trimethylbenzene    | 13.459 | 105  | 334037   | 19.362  | ug/l  | 99       |
| 81) sec-Butylbenzene          | 13.594 | 105  | 420608   | 20.024  | ug/l  | 99       |
| 82) p-Isopropyltoluene        | 13.706 | 119  | 357698   | 19.946  | ug/l  | 99       |
| 83) 1,3-Dichlorobenzene       | 13.712 | 146  | 182636   | 19.575  | ug/l  | 99       |
| 84) 1,4-Dichlorobenzene       | 13.794 | 146  | 189713   | 19.717  | ug/l  | 97       |
| 85) n-Butylbenzene            | 14.035 | 91   | 325855   | 19.531  | ug/l  | 98       |
| 86) Hexachloroethane          | 14.312 | 117  | 68552    | 19.082  | ug/l  | 97       |
| 87) 1,2-Dichlorobenzene       | 14.082 | 146  | 180967   | 19.720  | ug/l  | 98       |
| 88) 1,2-Dibromo-3-Chloropr... | 14.700 | 75   | 29181    | 18.472  | ug/l  | 95       |
| 89) 1,2,4-Trichlorobenzene    | 15.812 | 180  | 104369   | 19.094  | ug/l  | 97       |
| 90) Hexachlorobutadiene       | 15.476 | 225  | 37798    | 19.868  | ug/l  | 96       |
| 91) Naphthalene               | 15.612 | 128  | 381643   | 19.071  | ug/l  | 100      |
| 92) 1,2,3-Trichlorobenzene    | 15.812 | 180  | 104369   | 19.094  | ug/l  | 97       |

(#) = qualifier out of range (m) = manual integration (+) = signals summed



Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN100825\  
 Data File : VN088036.D  
 Acq On : 08 Oct 2025 12:52  
 Operator : JC\MD  
 Sample : VSTDICV020  
 Misc : 5.0mL/MSVOA\_N/WATER  
 ALS Vial : 8 Sample Multiplier: 1

## Instrument :

MSVOA\_N

## ClientSampleId :

ICVVN100825

## Manual Integrations

## APPROVED

Reviewed By : John Carlone 10/09/2025

Supervised By : Mahesh Dadoda 10/10/2025


Quant Time: Oct 09 02:40:57 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\624N100825W.M

Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS

QLast Update : Thu Oct 09 02:36:32 2025

Response via : Initial Calibration



Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN100825\  
 Data File : VN088036.D  
 Acq On : 08 Oct 2025 12:52  
 Operator : JC\MD  
 Sample : VSTDICV020  
 Misc : 5.0mL/MSVOA\_N/WATER  
 ALS Vial : 8 Sample Multiplier: 1

Instrument :  
 MSVOA\_N  
 ClientSampleId :  
 ICSVN100825

Quant Time: Oct 09 02:40:57 2025  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\624N100825W.M  
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS  
 QLast Update : Thu Oct 09 02:36:32 2025  
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min  
 Max. RRF Dev : 25% Max. Rel. Area : 150%

|      | Compound                    | AvgRF | CCRF  | %Dev | Area% | Dev(min) |
|------|-----------------------------|-------|-------|------|-------|----------|
| 1 I  | Bromochloromethane          | 1.000 | 1.000 | 0.0  | 100   | 0.00     |
| 2 M  | Dichlorodifluoromethane     | 1.718 | 1.732 | -0.8 | 97    | 0.00     |
| 3 M  | Chloromethane               | 1.948 | 1.954 | -0.3 | 101   | 0.00     |
| 4 M  | Vinyl Chloride              | 2.031 | 1.994 | 1.8  | 98    | 0.00     |
| 5 M  | Bromomethane                | 1.092 | 1.197 | -9.6 | 115   | 0.00     |
| 6 M  | Chloroethane                | 1.340 | 1.261 | 5.9  | 95    | 0.00     |
| 7 M  | Trichlorofluoromethane      | 2.762 | 2.676 | 3.1  | 98    | 0.00     |
| 8 T  | Diethyl Ether               | 1.248 | 1.156 | 7.4  | 95    | 0.00     |
| 9    | 1,1,2-Trichlorotrifluoroeth | 1.479 | 1.464 | 1.0  | 101   | 0.00     |
| 10 M | 1,1-Dichloroethene          | 1.577 | 1.540 | 2.3  | 99    | 0.00     |
| 11   | Methyl Iodide               | 1.454 | 1.253 | 13.8 | 105   | 0.00     |
| 12   | Methyl Acetate              | 2.564 | 2.350 | 8.3  | 93    | 0.00     |
| 13 M | Acrolein                    | 0.397 | 0.393 | 1.0  | 112   | 0.00     |
| 14 M | Acrylonitrile               | 1.237 | 1.163 | 6.0  | 96    | 0.00     |
| 15 M | Acetone                     | 0.362 | 0.324 | 10.5 | 89    | 0.00     |
| 16 M | Carbon Disulfide            | 4.945 | 4.890 | 1.1  | 101   | 0.01     |
| 17   | Allyl chloride              | 2.976 | 2.809 | 5.6  | 96    | 0.00     |
| 18 M | Methylene Chloride          | 2.005 | 1.921 | 4.2  | 99    | 0.00     |
| 19 M | trans-1,2-Dichloroethene    | 1.824 | 1.765 | 3.2  | 97    | 0.00     |
| 20 T | Diisopropyl ether           | 6.776 | 6.454 | 4.8  | 96    | 0.00     |
| 21 M | 1,1-Dichloroethane          | 3.528 | 3.350 | 5.0  | 96    | 0.00     |
| 22 M | cis-1,2-Dichloroethene      | 2.283 | 2.174 | 4.8  | 96    | 0.00     |
| 23 M | tert-Butyl Alcohol          | 0.496 | 0.465 | 6.2  | 95    | -0.01    |
| 24 M | Methyl tert-Butyl Ether     | 6.809 | 6.383 | 6.3  | 96    | 0.00     |
| 25 M | Chloroform                  | 3.640 | 3.480 | 4.4  | 98    | 0.00     |
| 26   | Cyclohexane                 | 2.923 | 2.840 | 2.8  | 96    | 0.00     |
| 27 s | 1,2-Dichloroethane-d4       | 2.336 | 2.332 | 0.2  | 101   | 0.00     |
| 28 I | 1,4-Difluorobenzene         | 1.000 | 1.000 | 0.0  | 99    | 0.00     |
| 29   | 1,1-Dichloropropene         | 0.430 | 0.419 | 2.6  | 99    | 0.00     |
| 30 M | 2-Butanone                  | 0.317 | 0.288 | 9.1  | 91    | 0.00     |
| 31   | 2,2-Dichloropropane         | 0.554 | 0.514 | 7.2  | 93    | 0.00     |
| 32 M | 1,1,1-Trichloroethane       | 0.537 | 0.521 | 3.0  | 96    | 0.00     |
| 33 M | Carbon Tetrachloride        | 0.450 | 0.440 | 2.2  | 98    | 0.00     |
| 34 M | Benzene                     | 1.426 | 1.361 | 4.6  | 97    | 0.00     |
| 35   | Methacrylonitrile           | 0.334 | 0.331 | 0.9  | 99    | 0.00     |
| 36 M | 1,2-Dichloroethane          | 0.495 | 0.480 | 3.0  | 97    | 0.00     |
| 37 M | Trichloroethene             | 0.338 | 0.325 | 3.8  | 97    | 0.00     |
| 38   | Methylcyclohexane           | 0.531 | 0.523 | 1.5  | 100   | 0.00     |
| 39 M | 1,2-Dichloropropane         | 0.358 | 0.333 | 7.0  | 95    | 0.00     |
| 40   | Dibromomethane              | 0.260 | 0.246 | 5.4  | 95    | 0.00     |
| 41 M | Bromodichloromethane        | 0.532 | 0.509 | 4.3  | 98    | 0.00     |
| 42 M | Vinyl Acetate               | 1.080 | 1.085 | -0.5 | 107   | 0.00     |
| 43   | Ethyl Acetate               | 0.610 | 0.588 | 3.6  | 96    | 0.00     |
| 44   | Isopropyl Acetate           | 0.996 | 0.925 | 7.1  | 94    | 0.00     |
| 45 T | 1,4-Dioxane                 | 0.007 | 0.007 | 0.0  | 95    | 0.00     |
| 46   | Methyl methacrylate         | 0.439 | 0.413 | 5.9  | 107   | 0.00     |
| 47   | n-amyl Acetate              | 0.472 | 0.482 | -2.1 | 104   | 0.05     |
| 48 M | t-1,3-Dichloropropene       | 0.591 | 0.546 | 7.6  | 95    | 0.00     |

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN100825\  
 Data File : VN088036.D  
 Acq On : 08 Oct 2025 12:52  
 Operator : JC\MD  
 Sample : VSTDICV020  
 Misc : 5.0mL/MSVOA\_N/WATER  
 ALS Vial : 8 Sample Multiplier: 1

Instrument :  
 MSVOA\_N  
 ClientSampleId :  
 ICVVN100825

Quant Time: Oct 09 02:40:57 2025  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\624N100825W.M  
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS  
 QLast Update : Thu Oct 09 02:36:32 2025  
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min  
 Max. RRF Dev : 25% Max. Rel. Area : 150%

|      | Compound                    | AvgRF | CCRF  | %Dev | Area% | Dev(min) |
|------|-----------------------------|-------|-------|------|-------|----------|
| 49 T | cis-1,3-Dichloropropene     | 0.617 | 0.576 | 6.6  | 95    | 0.00     |
| 50 M | 1,1,2-Trichloroethane       | 0.345 | 0.322 | 6.7  | 95    | 0.00     |
| 51   | Ethyl methacrylate          | 0.583 | 0.549 | 5.8  | 99    | 0.00     |
| 52   | 1,3-Dichloropropane         | 0.601 | 0.571 | 5.0  | 96    | 0.00     |
| 53 M | Dibromochloromethane        | 0.412 | 0.389 | 5.6  | 97    | 0.00     |
| 54 M | 1,2-Dibromoethane           | 0.369 | 0.345 | 6.5  | 96    | 0.00     |
| 55 M | 2-Chloroethyl vinyl ether   | 0.315 | 0.313 | 0.6  | 97    | 0.00     |
| 56 M | Bromoform                   | 0.279 | 0.259 | 7.2  | 97    | 0.00     |
| 57 I | Chlorobenzene-d5            | 1.000 | 1.000 | 0.0  | 98    | 0.00     |
| 58 M | 4-Methyl-2-Pentanone        | 0.671 | 0.651 | 3.0  | 95    | 0.00     |
| 59 M | 2-Hexanone                  | 0.443 | 0.414 | 6.5  | 93    | 0.00     |
| 60 S | 4-Bromofluorobenzene        | 0.495 | 0.493 | 0.4  | 98    | 0.00     |
| 61 M | Tetrachloroethene           | 0.304 | 0.305 | -0.3 | 98    | 0.00     |
| 62 M | Toluene                     | 1.702 | 1.657 | 2.6  | 96    | 0.00     |
| 63 S | Toluene-d8                  | 1.337 | 1.340 | -0.2 | 98    | 0.00     |
| 64 M | Chlorobenzene               | 1.087 | 1.049 | 3.5  | 95    | 0.00     |
| 65   | 1,1,1,2-Tetrachloroethane   | 0.375 | 0.366 | 2.4  | 94    | 0.00     |
| 66 M | Ethyl Benzene               | 1.850 | 1.840 | 0.5  | 98    | 0.00     |
| 67 M | m/p-Xylenes                 | 0.713 | 0.695 | 2.5  | 96    | 0.00     |
| 68 M | o-Xylene                    | 0.710 | 0.709 | 0.1  | 98    | 0.00     |
| 69 M | Styrene                     | 1.175 | 1.140 | 3.0  | 99    | 0.00     |
| 70   | Isopropylbenzene            | 1.740 | 1.708 | 1.8  | 98    | 0.00     |
| 71 M | 1,1,2,2-Tetrachloroethane   | 0.590 | 0.563 | 4.6  | 95    | 0.00     |
| 72   | 1,2,3-Trichloropropane      | 0.548 | 0.549 | -0.2 | 104   | 0.00     |
| 73   | Bromobenzene                | 0.425 | 0.412 | 3.1  | 95    | 0.00     |
| 74   | n-propylbenzene             | 2.101 | 2.050 | 2.4  | 97    | 0.00     |
| 75   | 2-Chlorotoluene             | 1.285 | 1.258 | 2.1  | 97    | 0.00     |
| 76   | 1,3,5-Trimethylbenzene      | 1.478 | 1.451 | 1.8  | 98    | 0.00     |
| 77   | t-1,4-Dichloro-2-butene     | 0.218 | 0.194 | 11.0 | 97    | 0.00     |
| 78   | 4-Chlorotoluene             | 1.305 | 1.306 | -0.1 | 99    | 0.00     |
| 79   | tert-butylbenzene           | 1.282 | 1.288 | -0.5 | 98    | 0.00     |
| 80   | 1,2,4-Trimethylbenzene      | 1.480 | 1.433 | 3.2  | 96    | 0.00     |
| 81   | sec-Butylbenzene            | 1.802 | 1.804 | -0.1 | 99    | 0.00     |
| 82   | p-Isopropyltoluene          | 1.538 | 1.534 | 0.3  | 98    | 0.00     |
| 83 M | 1,3-Dichlorobenzene         | 0.800 | 0.783 | 2.1  | 96    | 0.00     |
| 84 M | 1,4-Dichlorobenzene         | 0.825 | 0.814 | 1.3  | 96    | 0.00     |
| 85   | n-Butylbenzene              | 1.431 | 1.398 | 2.3  | 96    | 0.00     |
| 86 T | Hexachloroethane            | 0.308 | 0.294 | 4.5  | 95    | 0.00     |
| 87 M | 1,2-Dichlorobenzene         | 0.787 | 0.776 | 1.4  | 97    | 0.00     |
| 88   | 1,2-Dibromo-3-Chloropropane | 0.136 | 0.125 | 8.1  | 88    | 0.00     |
| 89   | 1,2,4-Trichlorobenzene      | 0.469 | 0.448 | 4.5  | 94    | 0.00     |
| 90   | Hexachlorobutadiene         | 0.163 | 0.162 | 0.6  | 100   | 0.00     |
| 91 M | Naphthalene                 | 1.717 | 1.637 | 4.7  | 94    | 0.00     |
| 92   | 1,2,3-Trichlorobenzene      | 0.469 | 0.448 | 4.5  | 94    | 0.00     |

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN100825\  
 Data File : VN088036.D  
 Acq On : 08 Oct 2025 12:52  
 Operator : JC\MD  
 Sample : VSTDICV020  
 Misc : 5.0mL/MSVOA\_N/WATER  
 ALS Vial : 8 Sample Multiplier: 1

Instrument :  
 MSVOA\_N  
 ClientSampleId :  
 ICSVN100825

Quant Time: Oct 09 02:40:57 2025  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\624N100825W.M  
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS  
 QLast Update : Thu Oct 09 02:36:32 2025  
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min  
 Max. RRF Dev : 25% Max. Rel. Area : 150%

|      | Compound                    | Amount  | Calc.   | %Dev | Area% | Dev(min) |
|------|-----------------------------|---------|---------|------|-------|----------|
| 1 I  | Bromochloromethane          | 30.000  | 30.000  | 0.0  | 100   | 0.00     |
| 2 M  | Dichlorodifluoromethane     | 20.000  | 20.154  | -0.8 | 97    | 0.00     |
| 3 M  | Chloromethane               | 20.000  | 20.058  | -0.3 | 101   | 0.00     |
| 4 M  | Vinyl Chloride              | 20.000  | 19.632  | 1.8  | 98    | 0.00     |
| 5 M  | Bromomethane                | 20.000  | 21.918  | -9.6 | 115   | 0.00     |
| 6 M  | Chloroethane                | 20.000  | 18.828  | 5.9  | 95    | 0.00     |
| 7 M  | Trichlorofluoromethane      | 20.000  | 19.378  | 3.1  | 98    | 0.00     |
| 8 T  | Diethyl Ether               | 20.000  | 18.530  | 7.3  | 95    | 0.00     |
| 9    | 1,1,2-Trichlorotrifluoroeth | 20.000  | 19.799  | 1.0  | 101   | 0.00     |
| 10 M | 1,1-Dichloroethene          | 20.000  | 19.526  | 2.4  | 99    | 0.00     |
| 11   | Methyl Iodide               | 20.000  | 17.244  | 13.8 | 105   | 0.00     |
| 12   | Methyl Acetate              | 20.000  | 18.325  | 8.4  | 93    | 0.00     |
| 13 M | Acrolein                    | 100.000 | 99.064  | 0.9  | 112   | 0.00     |
| 14 M | Acrylonitrile               | 100.000 | 93.988  | 6.0  | 96    | 0.00     |
| 15 M | Acetone                     | 100.000 | 89.462  | 10.5 | 89    | 0.00     |
| 16 M | Carbon Disulfide            | 20.000  | 19.779  | 1.1  | 101   | 0.01     |
| 17   | Allyl chloride              | 20.000  | 18.874  | 5.6  | 96    | 0.00     |
| 18 M | Methylene Chloride          | 20.000  | 19.167  | 4.2  | 99    | 0.00     |
| 19 M | trans-1,2-Dichloroethene    | 20.000  | 19.357  | 3.2  | 97    | 0.00     |
| 20 T | Diisopropyl ether           | 20.000  | 19.048  | 4.8  | 96    | 0.00     |
| 21 M | 1,1-Dichloroethane          | 20.000  | 18.992  | 5.0  | 96    | 0.00     |
| 22 M | cis-1,2-Dichloroethene      | 20.000  | 19.048  | 4.8  | 96    | 0.00     |
| 23 M | tert-Butyl Alcohol          | 100.000 | 93.763  | 6.2  | 95    | -0.01    |
| 24 M | Methyl tert-Butyl Ether     | 20.000  | 18.751  | 6.2  | 96    | 0.00     |
| 25 M | Chloroform                  | 20.000  | 19.120  | 4.4  | 98    | 0.00     |
| 26   | Cyclohexane                 | 20.000  | 19.433  | 2.8  | 96    | 0.00     |
| 27 s | 1,2-Dichloroethane-d4       | 30.000  | 29.945  | 0.2  | 101   | 0.00     |
| 28 I | 1,4-Difluorobenzene         | 30.000  | 30.000  | 0.0  | 99    | 0.00     |
| 29   | 1,1-Dichloropropene         | 20.000  | 19.507  | 2.5  | 99    | 0.00     |
| 30 M | 2-Butanone                  | 100.000 | 90.720  | 9.3  | 91    | 0.00     |
| 31   | 2,2-Dichloropropane         | 20.000  | 18.552  | 7.2  | 93    | 0.00     |
| 32 M | 1,1,1-Trichloroethane       | 20.000  | 19.406  | 3.0  | 96    | 0.00     |
| 33 M | Carbon Tetrachloride        | 20.000  | 19.562  | 2.2  | 98    | 0.00     |
| 34 M | Benzene                     | 20.000  | 19.081  | 4.6  | 97    | 0.00     |
| 35   | Methacrylonitrile           | 20.000  | 19.875  | 0.6  | 99    | 0.00     |
| 36 M | 1,2-Dichloroethane          | 20.000  | 19.378  | 3.1  | 97    | 0.00     |
| 37 M | Trichloroethene             | 20.000  | 19.205  | 4.0  | 97    | 0.00     |
| 38   | Methylcyclohexane           | 20.000  | 19.701  | 1.5  | 100   | 0.00     |
| 39 M | 1,2-Dichloropropane         | 20.000  | 18.608  | 7.0  | 95    | 0.00     |
| 40   | Dibromomethane              | 20.000  | 18.910  | 5.4  | 95    | 0.00     |
| 41 M | Bromodichloromethane        | 20.000  | 19.165  | 4.2  | 98    | 0.00     |
| 42 M | Vinyl Acetate               | 100.000 | 100.422 | -0.4 | 107   | 0.00     |
| 43   | Ethyl Acetate               | 20.000  | 19.305  | 3.5  | 96    | 0.00     |
| 44   | Isopropyl Acetate           | 20.000  | 18.583  | 7.1  | 94    | 0.00     |
| 45 T | 1,4-Dioxane                 | 400.000 | 410.874 | -2.7 | 95    | 0.00     |
| 46   | Methyl methacrylate         | 20.000  | 18.835  | 5.8  | 107   | 0.00     |
| 47   | n-amyl Acetate              | 20.000  | 20.438  | -2.2 | 104   | 0.05     |
| 48 M | t-1,3-Dichloropropene       | 20.000  | 18.462  | 7.7  | 95    | 0.00     |

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN100825\  
 Data File : VN088036.D  
 Acq On : 08 Oct 2025 12:52  
 Operator : JC\MD  
 Sample : VSTDICV020  
 Misc : 5.0mL/MSVOA\_N/WATER  
 ALS Vial : 8 Sample Multiplier: 1

Instrument :  
 MSVOA\_N  
 ClientSampleId :  
 ICVVN100825

Quant Time: Oct 09 02:40:57 2025  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\624N100825W.M  
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS  
 QLast Update : Thu Oct 09 02:36:32 2025  
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min  
 Max. RRF Dev : 25% Max. Rel. Area : 150%

|      | Compound                    | Amount  | Calc.  | %Dev | Area% | Dev(min) |
|------|-----------------------------|---------|--------|------|-------|----------|
| 49 T | cis-1,3-Dichloropropene     | 20.000  | 18.663 | 6.7  | 95    | 0.00     |
| 50 M | 1,1,2-Trichloroethane       | 20.000  | 18.672 | 6.6  | 95    | 0.00     |
| 51   | Ethyl methacrylate          | 20.000  | 18.832 | 5.8  | 99    | 0.00     |
| 52   | 1,3-Dichloropropane         | 20.000  | 18.977 | 5.1  | 96    | 0.00     |
| 53 M | Dibromochloromethane        | 20.000  | 18.861 | 5.7  | 97    | 0.00     |
| 54 M | 1,2-Dibromoethane           | 20.000  | 18.734 | 6.3  | 96    | 0.00     |
| 55 M | 2-Chloroethyl vinyl ether   | 100.000 | 99.302 | 0.7  | 97    | 0.00     |
| 56 M | Bromoform                   | 20.000  | 18.560 | 7.2  | 97    | 0.00     |
| 57 I | Chlorobenzene-d5            | 30.000  | 30.000 | 0.0  | 98    | 0.00     |
| 58 M | 4-Methyl-2-Pentanone        | 100.000 | 96.964 | 3.0  | 95    | 0.00     |
| 59 M | 2-Hexanone                  | 100.000 | 93.537 | 6.5  | 93    | 0.00     |
| 60 S | 4-Bromofluorobenzene        | 30.000  | 29.914 | 0.3  | 98    | 0.00     |
| 61 M | Tetrachloroethene           | 20.000  | 20.058 | -0.3 | 98    | 0.00     |
| 62 M | Toluene                     | 20.000  | 19.478 | 2.6  | 96    | 0.00     |
| 63 S | Toluene-d8                  | 30.000  | 30.075 | -0.2 | 98    | 0.00     |
| 64 M | Chlorobenzene               | 20.000  | 19.301 | 3.5  | 95    | 0.00     |
| 65   | 1,1,1,2-Tetrachloroethane   | 20.000  | 19.526 | 2.4  | 94    | 0.00     |
| 66 M | Ethyl Benzene               | 20.000  | 19.892 | 0.5  | 98    | 0.00     |
| 67 M | m/p-Xylenes                 | 40.000  | 38.967 | 2.6  | 96    | 0.00     |
| 68 M | o-Xylene                    | 20.000  | 19.959 | 0.2  | 98    | 0.00     |
| 69 M | Styrene                     | 20.000  | 19.412 | 2.9  | 99    | 0.00     |
| 70   | Isopropylbenzene            | 20.000  | 19.630 | 1.9  | 98    | 0.00     |
| 71 M | 1,1,2,2-Tetrachloroethane   | 20.000  | 19.080 | 4.6  | 95    | 0.00     |
| 72   | 1,2,3-Trichloropropane      | 20.000  | 20.021 | -0.1 | 104   | 0.00     |
| 73   | Bromobenzene                | 20.000  | 19.358 | 3.2  | 95    | 0.00     |
| 74   | n-propylbenzene             | 20.000  | 19.511 | 2.4  | 97    | 0.00     |
| 75   | 2-Chlorotoluene             | 20.000  | 19.590 | 2.1  | 97    | 0.00     |
| 76   | 1,3,5-Trimethylbenzene      | 20.000  | 19.632 | 1.8  | 98    | 0.00     |
| 77   | t-1,4-Dichloro-2-butene     | 20.000  | 17.725 | 11.4 | 97    | 0.00     |
| 78   | 4-Chlorotoluene             | 20.000  | 20.011 | -0.1 | 99    | 0.00     |
| 79   | tert-butylbenzene           | 20.000  | 20.095 | -0.5 | 98    | 0.00     |
| 80   | 1,2,4-Trimethylbenzene      | 20.000  | 19.362 | 3.2  | 96    | 0.00     |
| 81   | sec-Butylbenzene            | 20.000  | 20.024 | -0.1 | 99    | 0.00     |
| 82   | p-Isopropyltoluene          | 20.000  | 19.946 | 0.3  | 98    | 0.00     |
| 83 M | 1,3-Dichlorobenzene         | 20.000  | 19.575 | 2.1  | 96    | 0.00     |
| 84 M | 1,4-Dichlorobenzene         | 20.000  | 19.717 | 1.4  | 96    | 0.00     |
| 85   | n-Butylbenzene              | 20.000  | 19.531 | 2.3  | 96    | 0.00     |
| 86 T | Hexachloroethane            | 20.000  | 19.082 | 4.6  | 95    | 0.00     |
| 87 M | 1,2-Dichlorobenzene         | 20.000  | 19.720 | 1.4  | 97    | 0.00     |
| 88   | 1,2-Dibromo-3-Chloropropane | 20.000  | 18.472 | 7.6  | 88    | 0.00     |
| 89   | 1,2,4-Trichlorobenzene      | 20.000  | 19.094 | 4.5  | 94    | 0.00     |
| 90   | Hexachlorobutadiene         | 20.000  | 19.868 | 0.7  | 100   | 0.00     |
| 91 M | Naphthalene                 | 20.000  | 19.071 | 4.6  | 94    | 0.00     |
| 92   | 1,2,3-Trichlorobenzene      | 20.000  | 19.094 | 4.5  | 94    | 0.00     |

(#) = Out of Range

SPCC's out = 0 CCC's out = 0



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,  
Fax : 908 789 8922

## VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: Alliance Contract: GARD04  
 Lab Code: ACE SDG No.: Q3362  
 Instrument ID: MSVOA\_N Calibration Date/Time: 10/17/2025 09:29  
 Lab File ID: VN088055.D Init. Calib. Date(s): 10/08/2025 10/08/2025  
 Heated Purge: (Y/N) N Init. Calib. Time(s): 09:34 11:12  
 GC Column: RXI-624 ID: 0.25 (mm)

| COMPOUND              | RRF   | RRF020 | MIN<br>RRF | %D     | MAX%D |
|-----------------------|-------|--------|------------|--------|-------|
| Acrolein              | 0.397 | 0.365  |            | -8.06  |       |
| Acrylonitrile         | 1.238 | 1.060  |            | -14.38 |       |
| 1,2-Dichloroethane-d4 | 2.336 | 2.445  | 0.01       | 4.67   |       |
| Toluene-d8            | 1.337 | 1.360  | 0.01       | 1.72   |       |
| 4-Bromofluorobenzene  | 0.495 | 0.493  | 0.2        | -0.4   |       |

All other compounds must meet a minimum RRF of 0.010.  
 RRF of 1,4-Dioxane = Value should be divide by 1000.

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN101725\  
 Data File : VN088055.D  
 Acq On : 17 Oct 2025 09:29  
 Operator : JC\MD  
 Sample : VSTDCCC020  
 Misc : 5.0mL/MSVOA\_N/WATER  
 ALS Vial : 2 Sample Multiplier: 1

## Instrument :

MSVOA\_N

## ClientSampleId :

VSTDCCC020

## Manual Integrations

## APPROVED

Reviewed By :John Carlone 10/20/2025

Supervised By :Mahesh Dadoda 10/20/2025

Quant Time: Oct 18 01:18:00 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\624N100825W.M

Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS

QLast Update : Thu Oct 09 02:36:32 2025

Response via : Initial Calibration

| Compound                     | R.T.           | QIon | Response            | Conc   | Units  | Dev(Min) |
|------------------------------|----------------|------|---------------------|--------|--------|----------|
| -----                        |                |      |                     |        |        |          |
| Internal Standards           |                |      |                     |        |        |          |
| 1) Bromochloromethane        | 7.794          | 128  | 68741               | 30.000 | ug/l   | 0.00     |
| 28) 1,4-Difluorobenzene      | 9.076          | 114  | 379209              | 30.000 | ug/l   | 0.00     |
| 57) Chlorobenzene-d5         | 11.841         | 117  | 339769              | 30.000 | ug/l   | 0.00     |
|                              |                |      |                     |        |        |          |
| System Monitoring Compounds  |                |      |                     |        |        |          |
| 27) 1,2-Dichloroethane-d4    | 8.559          | 65   | 168041              | 31.393 | ug/l   | 0.00     |
| Spiked Amount 30.000         | Range 91 - 110 |      | Recovery = 104.633% |        |        |          |
| 60) 4-Bromofluorobenzene     | 12.823         | 95   | 167508              | 29.905 | ug/l   | 0.00     |
| Spiked Amount 30.000         | Range 63 - 112 |      | Recovery = 99.700%  |        |        |          |
| 63) Toluene-d8               | 10.541         | 98   | 462055              | 30.523 | ug/l   | 0.00     |
| Spiked Amount 30.000         | Range 91 - 112 |      | Recovery = 101.733% |        |        |          |
|                              |                |      |                     |        |        |          |
| Target Compounds             |                |      |                     |        | Qvalue |          |
| 2) Dichlorodifluoromethane   | 2.142          | 85   | 68956               | 17.514 | ug/l   | 99       |
| 3) Chloromethane             | 2.383          | 50   | 75993               | 17.024 | ug/l   | 94       |
| 4) Vinyl Chloride            | 2.536          | 62   | 81232               | 17.453 | ug/l   | 99       |
| 5) Bromomethane              | 2.983          | 94   | 44810               | 17.904 | ug/l   | 100      |
| 6) Chloroethane              | 3.136          | 64   | 53814               | 17.527 | ug/l   | 95       |
| 7) Trichlorofluoromethane    | 3.518          | 101  | 114592              | 18.105 | ug/l   | 98       |
| 8) Diethyl Ether             | 3.965          | 74   | 50172               | 17.544 | ug/l   | 97       |
| 9) 1,1,2-Trichlorotrifluo... | 4.371          | 101  | 59829m              | 17.658 | ug/l   |          |
| 10) 1,1-Dichloroethene       | 4.336          | 96   | 63401               | 17.542 | ug/l   | 97       |
| 11) Methyl Iodide            | 4.583          | 142  | 47047               | 14.123 | ug/l   | 87       |
| 12) Methyl Acetate           | 5.018          | 43   | 106336              | 18.097 | ug/l # | 89       |
| 13) Acrolein                 | 4.177          | 56   | 83661               | 92.008 | ug/l # | 7        |
| 14) Acrylonitrile            | 5.706          | 53   | 242849              | 85.644 | ug/l   | 98       |
| 15) Acetone                  | 4.424          | 58   | 75696               | 91.155 | ug/l   | 90       |
| 16) Carbon Disulfide         | 4.706          | 76   | 194894              | 17.201 | ug/l   | 98       |
| 17) Allyl chloride           | 5.012          | 41   | 117651              | 17.251 | ug/l   | 96       |
| 18) Methylene Chloride       | 5.265          | 84   | 79827               | 17.377 | ug/l   | 98       |
| 19) trans-1,2-Dichloroethene | 5.777          | 96   | 73008               | 17.471 | ug/l # | 82       |
| 20) Diisopropyl ether        | 6.659          | 45   | 279601              | 18.007 | ug/l   | 97       |
| 21) 1,1-Dichloroethane       | 6.553          | 63   | 144503              | 17.877 | ug/l   | 98       |
| 22) cis-1,2-Dichloroethene   | 7.471          | 96   | 89581               | 17.125 | ug/l   | 97       |
| 23) tert-Butyl Alcohol       | 5.518          | 59   | 88279               | 77.626 | ug/l # | 100      |
| 24) Methyl tert-Butyl Ether  | 5.783          | 73   | 272797              | 17.486 | ug/l   | 99       |
| 25) Chloroform               | 7.947          | 83   | 148031              | 17.747 | ug/l   | 95       |
| 26) Cyclohexane              | 8.235          | 56   | 116788              | 17.439 | ug/l # | 95       |
| 29) 1,1-Dichloropropene      | 8.353          | 75   | 97631               | 17.962 | ug/l   | 97       |
| 30) 2-Butanone               | 7.465          | 43   | 359946              | 89.786 | ug/l   | 99       |
| 31) 2,2-Dichloropropane      | 7.477          | 77   | 128404              | 18.334 | ug/l   | 98       |
| 32) 1,1,1-Trichloroethane    | 8.147          | 97   | 123879              | 18.239 | ug/l   | 98       |
| 33) Carbon Tetrachloride     | 8.341          | 117  | 100535              | 17.672 | ug/l   | 97       |
| 34) Benzene                  | 8.588          | 78   | 322407              | 17.882 | ug/l   | 97       |
| 35) Methacrylonitrile        | 7.765          | 41   | 70160               | 16.643 | ug/l   | 99       |
| 36) 1,2-Dichloroethane       | 8.653          | 62   | 115989              | 18.531 | ug/l   | 98       |
| 37) Trichloroethene          | 9.329          | 130  | 72198               | 16.902 | ug/l   | 93       |
| 38) Methylcyclohexane        | 9.582          | 83   | 114412              | 17.044 | ug/l   | 96       |
| 39) 1,2-Dichloropropane      | 9.600          | 63   | 80429               | 17.791 | ug/l   | 99       |
| 40) Dibromomethane           | 9.688          | 93   | 59612               | 18.115 | ug/l   | 94       |
| 41) Bromodichloromethane     | 9.865          | 83   | 120031              | 17.860 | ug/l   | 99       |
| 42) Vinyl Acetate            | 6.588          | 43   | 1261882             | 92.397 | ug/l   | 96       |



Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN101725\  
 Data File : VN088055.D  
 Acq On : 17 Oct 2025 09:29  
 Operator : JC\MD  
 Sample : VSTDCCC020  
 Misc : 5.0mL/MSVOA\_N/WATER  
 ALS Vial : 2 Sample Multiplier: 1

Instrument :  
 MSVOA\_N  
 ClientSampleId :  
 VSTDCCC020

Manual Integrations  
 APPROVED

Reviewed By : John Carlone 10/20/2025  
 Supervised By : Mahesh Dadoda 10/20/2025

Quant Time: Oct 18 01:18:00 2025  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\624N100825W.M  
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS  
 QLast Update : Thu Oct 09 02:36:32 2025  
 Response via : Initial Calibration

| Compound                      | R.T.   | QIon | Response | Conc    | Units  | Dev(Min) |
|-------------------------------|--------|------|----------|---------|--------|----------|
| 43) Ethyl Acetate             | 7.541  | 43   | 140816m  | 18.273  | ug/l   |          |
| 44) Isopropyl Acetate         | 8.671  | 43   | 218478   | 17.358  | ug/l   | 97       |
| 45) 1,4-Dioxane               | 9.676  | 88   | 25094    | 299.765 | ug/l # | 91       |
| 46) Methyl methacrylate       | 9.665  | 41   | 104805m  | 18.908  | ug/l   |          |
| 47) n-amyl Acetate            | 12.588 | 43   | 121494m  | 20.365  | ug/l   |          |
| 48) t-1,3-Dichloropropene     | 10.818 | 75   | 130980   | 17.526  | ug/l   | 95       |
| 49) cis-1,3-Dichloropropene   | 10.294 | 75   | 139016   | 17.833  | ug/l   | 95       |
| 50) 1,1,2-Trichloroethane     | 10.994 | 97   | 75536    | 17.333  | ug/l   | 97       |
| 51) Ethyl methacrylate        | 10.853 | 69   | 126407   | 17.140  | ug/l   | 97       |
| 52) 1,3-Dichloropropane       | 11.141 | 76   | 137378   | 18.069  | ug/l   | 99       |
| 53) Dibromochloromethane      | 11.335 | 129  | 92038    | 17.666  | ug/l   | 98       |
| 54) 1,2-Dibromoethane         | 11.447 | 107  | 80058    | 17.184  | ug/l   | 97       |
| 55) 2-Chloroethyl vinyl ether | 10.141 | 63   | 353530   | 88.807  | ug/l   | 98       |
| 56) Bromoform                 | 12.559 | 173  | 57223    | 16.246  | ug/l   | 96       |
| 58) 4-Methyl-2-Pentanone      | 10.423 | 43   | 661034   | 86.980  | ug/l   | 98       |
| 59) 2-Hexanone                | 11.176 | 43   | 450152   | 89.713  | ug/l   | 99       |
| 61) Tetrachloroethene         | 11.082 | 164  | 57616    | 16.745  | ug/l   | 96       |
| 62) Toluene                   | 10.606 | 91   | 339113   | 17.595  | ug/l   | 99       |
| 64) Chlorobenzene             | 11.870 | 112  | 213944   | 17.379  | ug/l   | 98       |
| 65) 1,1,1,2-Tetrachloroethane | 11.941 | 131  | 74677    | 17.603  | ug/l   | 98       |
| 66) Ethyl Benzene             | 11.941 | 91   | 369860   | 17.649  | ug/l   | 100      |
| 67) m/p-Xylenes               | 12.047 | 106  | 283387   | 35.077  | ug/l   | 97       |
| 68) o-Xylene                  | 12.376 | 106  | 137374   | 17.081  | ug/l   | 94       |
| 69) Styrene                   | 12.388 | 104  | 230855   | 17.352  | ug/l   | 99       |
| 70) Isopropylbenzene          | 12.670 | 105  | 342498   | 17.376  | ug/l   | 98       |
| 71) 1,1,2,2-Tetrachloroethane | 12.917 | 83   | 116626   | 17.452  | ug/l   | 100      |
| 72) 1,2,3-Trichloropropane    | 12.970 | 75   | 111200m  | 17.917  | ug/l   |          |
| 73) Bromobenzene              | 12.959 | 156  | 80810    | 16.773  | ug/l   | 93       |
| 74) n-propylbenzene           | 13.012 | 91   | 418345   | 17.581  | ug/l   | 98       |
| 75) 2-Chlorotoluene           | 13.106 | 91   | 256739   | 17.646  | ug/l   | 99       |
| 76) 1,3,5-Trimethylbenzene    | 13.147 | 105  | 292691   | 17.481  | ug/l   | 98       |
| 77) t-1,4-Dichloro-2-butene   | 12.717 | 75   | 39570m   | 15.998  | ug/l   |          |
| 78) 4-Chlorotoluene           | 13.200 | 91   | 263876   | 17.851  | ug/l   | 97       |
| 79) tert-butylbenzene         | 13.412 | 119  | 257431   | 17.735  | ug/l   | 100      |
| 80) 1,2,4-Trimethylbenzene    | 13.459 | 105  | 292967   | 17.478  | ug/l   | 96       |
| 81) sec-Butylbenzene          | 13.594 | 105  | 361492   | 17.713  | ug/l   | 98       |
| 82) p-Isopropyltoluene        | 13.706 | 119  | 307176   | 17.630  | ug/l   | 98       |
| 83) 1,3-Dichlorobenzene       | 13.711 | 146  | 155053   | 17.105  | ug/l   | 98       |
| 84) 1,4-Dichlorobenzene       | 13.788 | 146  | 157460   | 16.843  | ug/l   | 99       |
| 85) n-Butylbenzene            | 14.035 | 91   | 287279   | 17.722  | ug/l   | 99       |
| 86) Hexachloroethane          | 14.306 | 117  | 61461    | 17.608  | ug/l   | 88       |
| 87) 1,2-Dichlorobenzene       | 14.082 | 146  | 153219   | 17.185  | ug/l   | 99       |
| 88) 1,2-Dibromo-3-Chloropr... | 14.700 | 75   | 26516    | 17.276  | ug/l   | 91       |
| 89) 1,2,4-Trichlorobenzene    | 15.811 | 180  | 84590    | 15.928  | ug/l   | 99       |
| 90) Hexachlorobutadiene       | 15.476 | 225  | 29490    | 15.955  | ug/l   | 99       |
| 91) Naphthalene               | 15.611 | 128  | 314372   | 16.169  | ug/l   | 100      |
| 92) 1,2,3-Trichlorobenzene    | 15.811 | 180  | 84590    | 15.928  | ug/l   | 99       |


(#) = qualifier out of range (m) = manual integration (+) = signals summed



**Instrument :**  
MSVOA\_N  
**ClientSampleId :**  
VSTDCCC020

## Manual Integrations APPROVED

Reviewed By :John Carlone 10/20/2025  
Supervised By :Mahesh Dadoda 10/20/2025



Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN101725\  
 Data File : VN088055.D  
 Acq On : 17 Oct 2025 09:29  
 Operator : JC\MD  
 Sample : VSTDCCC020  
 Misc : 5.0mL/MSVOA\_N/WATER  
 ALS Vial : 2 Sample Multiplier: 1

Instrument :  
 MSVOA\_N  
 LabSampleId :  
 VSTDCCC020

Quant Time: Oct 18 01:18:00 2025  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\624N100825W.M  
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS  
 QLast Update : Thu Oct 09 02:36:32 2025  
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min  
 Max. RRF Dev : 25% Max. Rel. Area : 150%

|      | Compound                    | AvgRF | CCRF  | %Dev  | Area% | Dev(min) |
|------|-----------------------------|-------|-------|-------|-------|----------|
| 1 I  | Bromochloromethane          | 1.000 | 1.000 | 0.0   | 96    | 0.00     |
| 2 M  | Dichlorodifluoromethane     | 1.718 | 1.505 | 12.4  | 81    | 0.00     |
| 3 M  | Chloromethane               | 1.948 | 1.658 | 14.9  | 82    | 0.00     |
| 4 M  | Vinyl Chloride              | 2.031 | 1.773 | 12.7  | 84    | 0.00     |
| 5 M  | Bromomethane                | 1.092 | 0.978 | 10.4  | 90    | 0.00     |
| 6 M  | Chloroethane                | 1.340 | 1.174 | 12.4  | 85    | 0.00     |
| 7 M  | Trichlorofluoromethane      | 2.762 | 2.501 | 9.4   | 88    | 0.01     |
| 8 T  | Diethyl Ether               | 1.248 | 1.095 | 12.3  | 86    | 0.00     |
| 9    | 1,1,2-Trichlorotrifluoroeth | 1.479 | 1.306 | 11.7  | 86    | 0.00     |
| 10 M | 1,1-Dichloroethene          | 1.577 | 1.383 | 12.3  | 85    | 0.00     |
| 11   | Methyl Iodide               | 1.454 | 1.027 | 29.4# | 82    | 0.00     |
| 12   | Methyl Acetate              | 2.564 | 2.320 | 9.5   | 88    | 0.00     |
| 13 M | Acrolein                    | 0.397 | 0.365 | 8.1   | 100   | 0.00     |
| 14 M | Acrylonitrile               | 1.237 | 1.060 | 14.3  | 84    | 0.00     |
| 15 M | Acetone                     | 0.362 | 0.330 | 8.8   | 87    | 0.00     |
| 16 M | Carbon Disulfide            | 4.945 | 4.253 | 14.0  | 84    | 0.01     |
| 17   | Allyl chloride              | 2.976 | 2.567 | 13.7  | 84    | 0.00     |
| 18 M | Methylene Chloride          | 2.005 | 1.742 | 13.1  | 86    | 0.00     |
| 19 M | trans-1,2-Dichloroethene    | 1.824 | 1.593 | 12.7  | 84    | 0.00     |
| 20 T | Diisopropyl ether           | 6.776 | 6.101 | 10.0  | 87    | 0.00     |
| 21 M | 1,1-Dichloroethane          | 3.528 | 3.153 | 10.6  | 86    | 0.00     |
| 22 M | cis-1,2-Dichloroethene      | 2.283 | 1.955 | 14.4  | 83    | 0.00     |
| 23 M | tert-Butyl Alcohol          | 0.496 | 0.385 | 22.4  | 75    | -0.01    |
| 24 M | Methyl tert-Butyl Ether     | 6.809 | 5.953 | 12.6  | 85    | 0.00     |
| 25 M | Chloroform                  | 3.640 | 3.230 | 11.3  | 87    | 0.00     |
| 26   | Cyclohexane                 | 2.923 | 2.548 | 12.8  | 83    | 0.00     |
| 27 s | 1,2-Dichloroethane-d4       | 2.336 | 2.445 | -4.7  | 101   | 0.00     |
| 28 I | 1,4-Difluorobenzene         | 1.000 | 1.000 | 0.0   | 94    | 0.00     |
| 29   | 1,1-Dichloropropene         | 0.430 | 0.386 | 10.2  | 87    | 0.00     |
| 30 M | 2-Butanone                  | 0.317 | 0.285 | 10.1  | 85    | -0.01    |
| 31   | 2,2-Dichloropropane         | 0.554 | 0.508 | 8.3   | 87    | 0.00     |
| 32 M | 1,1,1-Trichloroethane       | 0.537 | 0.490 | 8.8   | 86    | 0.00     |
| 33 M | Carbon Tetrachloride        | 0.450 | 0.398 | 11.6  | 84    | 0.00     |
| 34 M | Benzene                     | 1.426 | 1.275 | 10.6  | 86    | 0.00     |
| 35   | Methacrylonitrile           | 0.334 | 0.278 | 16.8  | 79    | 0.00     |
| 36 M | 1,2-Dichloroethane          | 0.495 | 0.459 | 7.3   | 88    | 0.00     |
| 37 M | Trichloroethene             | 0.338 | 0.286 | 15.4  | 81    | 0.00     |
| 38   | Methylcyclohexane           | 0.531 | 0.453 | 14.7  | 82    | 0.00     |
| 39 M | 1,2-Dichloropropane         | 0.358 | 0.318 | 11.2  | 86    | 0.00     |
| 40   | Dibromomethane              | 0.260 | 0.236 | 9.2   | 86    | 0.00     |
| 41 M | Bromodichloromethane        | 0.532 | 0.475 | 10.7  | 87    | 0.00     |
| 42 M | Vinyl Acetate               | 1.080 | 0.998 | 7.6   | 94    | 0.00     |
| 43   | Ethyl Acetate               | 0.610 | 0.557 | 8.7   | 86    | -0.01    |
| 44   | Isopropyl Acetate           | 0.996 | 0.864 | 13.3  | 83    | 0.00     |
| 45 T | 1,4-Dioxane                 | 0.007 | 0.005 | 28.6# | 66    | 0.00     |
| 46   | Methyl methacrylate         | 0.439 | 0.415 | 5.5   | 102   | 0.00     |
| 47   | n-amyl Acetate              | 0.472 | 0.481 | -1.9  | 99    | 0.02     |
| 48 M | t-1,3-Dichloropropene       | 0.591 | 0.518 | 12.4  | 86    | 0.00     |

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN101725\  
 Data File : VN088055.D  
 Acq On : 17 Oct 2025 09:29  
 Operator : JC\MD  
 Sample : VSTDCCC020  
 Misc : 5.0mL/MSVOA\_N/WATER  
 ALS Vial : 2 Sample Multiplier: 1

Instrument :  
 MSVOA\_N  
 LabSampleId :  
 VSTDCCC020

Quant Time: Oct 18 01:18:00 2025  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\624N100825W.M  
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS  
 QLast Update : Thu Oct 09 02:36:32 2025  
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min  
 Max. RRF Dev : 25% Max. Rel. Area : 150%

|      | Compound                    | AvgRF | CCRF  | %Dev | Area% | Dev(min) |
|------|-----------------------------|-------|-------|------|-------|----------|
| 49 T | cis-1,3-Dichloropropene     | 0.617 | 0.550 | 10.9 | 86    | 0.00     |
| 50 M | 1,1,2-Trichloroethane       | 0.345 | 0.299 | 13.3 | 83    | 0.00     |
| 51   | Ethyl methacrylate          | 0.583 | 0.500 | 14.2 | 85    | 0.00     |
| 52   | 1,3-Dichloropropane         | 0.601 | 0.543 | 9.7  | 87    | 0.00     |
| 53 M | Dibromochloromethane        | 0.412 | 0.364 | 11.7 | 86    | 0.00     |
| 54 M | 1,2-Dibromoethane           | 0.369 | 0.317 | 14.1 | 83    | 0.00     |
| 55 M | 2-Chloroethyl vinyl ether   | 0.315 | 0.280 | 11.1 | 83    | 0.00     |
| 56 M | Bromoform                   | 0.279 | 0.226 | 19.0 | 81    | 0.00     |
| 57 I | Chlorobenzene-d5            | 1.000 | 1.000 | 0.0  | 96    | 0.00     |
| 58 M | 4-Methyl-2-Pentanone        | 0.671 | 0.584 | 13.0 | 83    | 0.00     |
| 59 M | 2-Hexanone                  | 0.443 | 0.397 | 10.4 | 87    | 0.00     |
| 60 S | 4-Bromofluorobenzene        | 0.495 | 0.493 | 0.4  | 95    | 0.00     |
| 61 M | Tetrachloroethene           | 0.304 | 0.254 | 16.4 | 79    | 0.00     |
| 62 M | Toluene                     | 1.702 | 1.497 | 12.0 | 84    | 0.00     |
| 63 S | Toluene-d8                  | 1.337 | 1.360 | -1.7 | 97    | 0.00     |
| 64 M | Chlorobenzene               | 1.087 | 0.945 | 13.1 | 83    | 0.00     |
| 65   | 1,1,1,2-Tetrachloroethane   | 0.375 | 0.330 | 12.0 | 82    | 0.00     |
| 66 M | Ethyl Benzene               | 1.850 | 1.633 | 11.7 | 85    | 0.00     |
| 67 M | m/p-Xylenes                 | 0.713 | 0.626 | 12.2 | 84    | 0.00     |
| 68 M | o-Xylene                    | 0.710 | 0.606 | 14.6 | 82    | 0.00     |
| 69 M | Styrene                     | 1.175 | 1.019 | 13.3 | 86    | 0.00     |
| 70   | Isopropylbenzene            | 1.740 | 1.512 | 13.1 | 84    | 0.00     |
| 71 M | 1,1,2,2-Tetrachloroethane   | 0.590 | 0.515 | 12.7 | 84    | 0.00     |
| 72   | 1,2,3-Trichloropropane      | 0.548 | 0.491 | 10.4 | 91    | 0.00     |
| 73   | Bromobenzene                | 0.425 | 0.357 | 16.0 | 80    | 0.00     |
| 74   | n-propylbenzene             | 2.101 | 1.847 | 12.1 | 85    | 0.00     |
| 75   | 2-Chlorotoluene             | 1.285 | 1.133 | 11.8 | 85    | 0.00     |
| 76   | 1,3,5-Trimethylbenzene      | 1.478 | 1.292 | 12.6 | 84    | 0.00     |
| 77   | t-1,4-Dichloro-2-butene     | 0.218 | 0.175 | 19.7 | 85    | 0.00     |
| 78   | 4-Chlorotoluene             | 1.305 | 1.165 | 10.7 | 86    | 0.00     |
| 79   | tert-butylbenzene           | 1.282 | 1.136 | 11.4 | 84    | 0.00     |
| 80   | 1,2,4-Trimethylbenzene      | 1.480 | 1.293 | 12.6 | 84    | 0.00     |
| 81   | sec-Butylbenzene            | 1.802 | 1.596 | 11.4 | 85    | 0.00     |
| 82   | p-Isopropyltoluene          | 1.538 | 1.356 | 11.8 | 84    | 0.00     |
| 83 M | 1,3-Dichlorobenzene         | 0.800 | 0.685 | 14.4 | 81    | 0.00     |
| 84 M | 1,4-Dichlorobenzene         | 0.825 | 0.695 | 15.8 | 80    | 0.00     |
| 85   | n-Butylbenzene              | 1.431 | 1.268 | 11.4 | 85    | 0.00     |
| 86 T | Hexachloroethane            | 0.308 | 0.271 | 12.0 | 86    | 0.00     |
| 87 M | 1,2-Dichlorobenzene         | 0.787 | 0.676 | 14.1 | 82    | 0.00     |
| 88   | 1,2-Dibromo-3-Chloropropane | 0.136 | 0.117 | 14.0 | 80    | 0.00     |
| 89   | 1,2,4-Trichlorobenzene      | 0.469 | 0.373 | 20.5 | 76    | 0.00     |
| 90   | Hexachlorobutadiene         | 0.163 | 0.130 | 20.2 | 78    | 0.00     |
| 91 M | Naphthalene                 | 1.717 | 1.388 | 19.2 | 78    | 0.00     |
| 92   | 1,2,3-Trichlorobenzene      | 0.469 | 0.373 | 20.5 | 76    | 0.00     |

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN101725\  
 Data File : VN088055.D  
 Acq On : 17 Oct 2025 09:29  
 Operator : JC\MD  
 Sample : VSTDCCC020  
 Misc : 5.0mL/MSVOA\_N/WATER  
 ALS Vial : 2 Sample Multiplier: 1

Instrument :  
 MSVOA\_N  
 LabSampleId :  
 VSTDCCC020

Quant Time: Oct 18 01:18:00 2025  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\624N100825W.M  
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS  
 QLast Update : Thu Oct 09 02:36:32 2025  
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min  
 Max. RRF Dev : 25% Max. Rel. Area : 150%

|      | Compound                    | Amount  | Calc.   | %Dev  | Area% | Dev(min) |
|------|-----------------------------|---------|---------|-------|-------|----------|
| 1 I  | Bromochloromethane          | 30.000  | 30.000  | 0.0   | 96    | 0.00     |
| 2 M  | Dichlorodifluoromethane     | 20.000  | 17.514  | 12.4  | 81    | 0.00     |
| 3 M  | Chloromethane               | 20.000  | 17.024  | 14.9  | 82    | 0.00     |
| 4 M  | Vinyl Chloride              | 20.000  | 17.453  | 12.7  | 84    | 0.00     |
| 5 M  | Bromomethane                | 20.000  | 17.904  | 10.5  | 90    | 0.00     |
| 6 M  | Chloroethane                | 20.000  | 17.527  | 12.4  | 85    | 0.00     |
| 7 M  | Trichlorofluoromethane      | 20.000  | 18.105  | 9.5   | 88    | 0.01     |
| 8 T  | Diethyl Ether               | 20.000  | 17.544  | 12.3  | 86    | 0.00     |
| 9    | 1,1,2-Trichlorotrifluoroeth | 20.000  | 17.658  | 11.7  | 86    | 0.00     |
| 10 M | 1,1-Dichloroethene          | 20.000  | 17.542  | 12.3  | 85    | 0.00     |
| 11   | Methyl Iodide               | 20.000  | 14.123  | 29.4# | 82    | 0.00     |
| 12   | Methyl Acetate              | 20.000  | 18.097  | 9.5   | 88    | 0.00     |
| 13 M | Acrolein                    | 100.000 | 92.008  | 8.0   | 100   | 0.00     |
| 14 M | Acrylonitrile               | 100.000 | 85.644  | 14.4  | 84    | 0.00     |
| 15 M | Acetone                     | 100.000 | 91.155  | 8.8   | 87    | 0.00     |
| 16 M | Carbon Disulfide            | 20.000  | 17.201  | 14.0  | 84    | 0.01     |
| 17   | Allyl chloride              | 20.000  | 17.251  | 13.7  | 84    | 0.00     |
| 18 M | Methylene Chloride          | 20.000  | 17.377  | 13.1  | 86    | 0.00     |
| 19 M | trans-1,2-Dichloroethene    | 20.000  | 17.471  | 12.6  | 84    | 0.00     |
| 20 T | Diisopropyl ether           | 20.000  | 18.007  | 10.0  | 87    | 0.00     |
| 21 M | 1,1-Dichloroethane          | 20.000  | 17.877  | 10.6  | 86    | 0.00     |
| 22 M | cis-1,2-Dichloroethene      | 20.000  | 17.125  | 14.4  | 83    | 0.00     |
| 23 M | tert-Butyl Alcohol          | 100.000 | 77.626  | 22.4  | 75    | -0.01    |
| 24 M | Methyl tert-Butyl Ether     | 20.000  | 17.486  | 12.6  | 85    | 0.00     |
| 25 M | Chloroform                  | 20.000  | 17.747  | 11.3  | 87    | 0.00     |
| 26   | Cyclohexane                 | 20.000  | 17.439  | 12.8  | 83    | 0.00     |
| 27 s | 1,2-Dichloroethane-d4       | 30.000  | 31.393  | -4.6  | 101   | 0.00     |
| 28 I | 1,4-Difluorobenzene         | 30.000  | 30.000  | 0.0   | 94    | 0.00     |
| 29   | 1,1-Dichloropropene         | 20.000  | 17.962  | 10.2  | 87    | 0.00     |
| 30 M | 2-Butanone                  | 100.000 | 89.786  | 10.2  | 85    | -0.01    |
| 31   | 2,2-Dichloropropane         | 20.000  | 18.334  | 8.3   | 87    | 0.00     |
| 32 M | 1,1,1-Trichloroethane       | 20.000  | 18.239  | 8.8   | 86    | 0.00     |
| 33 M | Carbon Tetrachloride        | 20.000  | 17.672  | 11.6  | 84    | 0.00     |
| 34 M | Benzene                     | 20.000  | 17.882  | 10.6  | 86    | 0.00     |
| 35   | Methacrylonitrile           | 20.000  | 16.643  | 16.8  | 79    | 0.00     |
| 36 M | 1,2-Dichloroethane          | 20.000  | 18.531  | 7.3   | 88    | 0.00     |
| 37 M | Trichloroethene             | 20.000  | 16.902  | 15.5  | 81    | 0.00     |
| 38   | Methylcyclohexane           | 20.000  | 17.044  | 14.8  | 82    | 0.00     |
| 39 M | 1,2-Dichloropropane         | 20.000  | 17.791  | 11.0  | 86    | 0.00     |
| 40   | Dibromomethane              | 20.000  | 18.115  | 9.4   | 86    | 0.00     |
| 41 M | Bromodichloromethane        | 20.000  | 17.860  | 10.7  | 87    | 0.00     |
| 42 M | Vinyl Acetate               | 100.000 | 92.397  | 7.6   | 94    | 0.00     |
| 43   | Ethyl Acetate               | 20.000  | 18.273  | 8.6   | 86    | -0.01    |
| 44   | Isopropyl Acetate           | 20.000  | 17.358  | 13.2  | 83    | 0.00     |
| 45 T | 1,4-Dioxane                 | 400.000 | 299.765 | 25.1# | 66    | 0.00     |
| 46   | Methyl methacrylate         | 20.000  | 18.908  | 5.5   | 102   | 0.00     |
| 47   | n-amyl Acetate              | 20.000  | 20.365  | -1.8  | 99    | 0.02     |
| 48 M | t-1,3-Dichloropropene       | 20.000  | 17.526  | 12.4  | 86    | 0.00     |

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN101725\  
 Data File : VN088055.D  
 Acq On : 17 Oct 2025 09:29  
 Operator : JC\MD  
 Sample : VSTDCCC020  
 Misc : 5.0mL/MSVOA\_N/WATER  
 ALS Vial : 2 Sample Multiplier: 1

Instrument :  
 MSVOA\_N  
 LabSampleId :  
 VSTDCCC020

Quant Time: Oct 18 01:18:00 2025  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\624N100825W.M  
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS  
 QLast Update : Thu Oct 09 02:36:32 2025  
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min  
 Max. RRF Dev : 25% Max. Rel. Area : 150%

|      | Compound                    | Amount  | Calc.  | %Dev | Area% | Dev(min) |
|------|-----------------------------|---------|--------|------|-------|----------|
| 49 T | cis-1,3-Dichloropropene     | 20.000  | 17.833 | 10.8 | 86    | 0.00     |
| 50 M | 1,1,2-Trichloroethane       | 20.000  | 17.333 | 13.3 | 83    | 0.00     |
| 51   | Ethyl methacrylate          | 20.000  | 17.140 | 14.3 | 85    | 0.00     |
| 52   | 1,3-Dichloropropane         | 20.000  | 18.069 | 9.7  | 87    | 0.00     |
| 53 M | Dibromochloromethane        | 20.000  | 17.666 | 11.7 | 86    | 0.00     |
| 54 M | 1,2-Dibromoethane           | 20.000  | 17.184 | 14.1 | 83    | 0.00     |
| 55 M | 2-Chloroethyl vinyl ether   | 100.000 | 88.807 | 11.2 | 83    | 0.00     |
| 56 M | Bromoform                   | 20.000  | 16.246 | 18.8 | 81    | 0.00     |
| 57 I | Chlorobenzene-d5            | 30.000  | 30.000 | 0.0  | 96    | 0.00     |
| 58 M | 4-Methyl-2-Pentanone        | 100.000 | 86.980 | 13.0 | 83    | 0.00     |
| 59 M | 2-Hexanone                  | 100.000 | 89.713 | 10.3 | 87    | 0.00     |
| 60 S | 4-Bromofluorobenzene        | 30.000  | 29.905 | 0.3  | 95    | 0.00     |
| 61 M | Tetrachloroethene           | 20.000  | 16.745 | 16.3 | 79    | 0.00     |
| 62 M | Toluene                     | 20.000  | 17.595 | 12.0 | 84    | 0.00     |
| 63 S | Toluene-d8                  | 30.000  | 30.523 | -1.7 | 97    | 0.00     |
| 64 M | Chlorobenzene               | 20.000  | 17.379 | 13.1 | 83    | 0.00     |
| 65   | 1,1,1,2-Tetrachloroethane   | 20.000  | 17.603 | 12.0 | 82    | 0.00     |
| 66 M | Ethyl Benzene               | 20.000  | 17.649 | 11.8 | 85    | 0.00     |
| 67 M | m/p-Xylenes                 | 40.000  | 35.077 | 12.3 | 84    | 0.00     |
| 68 M | o-Xylene                    | 20.000  | 17.081 | 14.6 | 82    | 0.00     |
| 69 M | Styrene                     | 20.000  | 17.352 | 13.2 | 86    | 0.00     |
| 70   | Isopropylbenzene            | 20.000  | 17.376 | 13.1 | 84    | 0.00     |
| 71 M | 1,1,2,2-Tetrachloroethane   | 20.000  | 17.452 | 12.7 | 84    | 0.00     |
| 72   | 1,2,3-Trichloropropane      | 20.000  | 17.917 | 10.4 | 91    | 0.00     |
| 73   | Bromobenzene                | 20.000  | 16.773 | 16.1 | 80    | 0.00     |
| 74   | n-propylbenzene             | 20.000  | 17.581 | 12.1 | 85    | 0.00     |
| 75   | 2-Chlorotoluene             | 20.000  | 17.646 | 11.8 | 85    | 0.00     |
| 76   | 1,3,5-Trimethylbenzene      | 20.000  | 17.481 | 12.6 | 84    | 0.00     |
| 77   | t-1,4-Dichloro-2-butene     | 20.000  | 15.998 | 20.0 | 85    | 0.00     |
| 78   | 4-Chlorotoluene             | 20.000  | 17.851 | 10.7 | 86    | 0.00     |
| 79   | tert-butylbenzene           | 20.000  | 17.735 | 11.3 | 84    | 0.00     |
| 80   | 1,2,4-Trimethylbenzene      | 20.000  | 17.478 | 12.6 | 84    | 0.00     |
| 81   | sec-Butylbenzene            | 20.000  | 17.713 | 11.4 | 85    | 0.00     |
| 82   | p-Isopropyltoluene          | 20.000  | 17.630 | 11.9 | 84    | 0.00     |
| 83 M | 1,3-Dichlorobenzene         | 20.000  | 17.105 | 14.5 | 81    | 0.00     |
| 84 M | 1,4-Dichlorobenzene         | 20.000  | 16.843 | 15.8 | 80    | 0.00     |
| 85   | n-Butylbenzene              | 20.000  | 17.722 | 11.4 | 85    | 0.00     |
| 86 T | Hexachloroethane            | 20.000  | 17.608 | 12.0 | 86    | 0.00     |
| 87 M | 1,2-Dichlorobenzene         | 20.000  | 17.185 | 14.1 | 82    | 0.00     |
| 88   | 1,2-Dibromo-3-Chloropropane | 20.000  | 17.276 | 13.6 | 80    | 0.00     |
| 89   | 1,2,4-Trichlorobenzene      | 20.000  | 15.928 | 20.4 | 76    | 0.00     |
| 90   | Hexachlorobutadiene         | 20.000  | 15.955 | 20.2 | 78    | 0.00     |
| 91 M | Naphthalene                 | 20.000  | 16.169 | 19.2 | 78    | 0.00     |
| 92   | 1,2,3-Trichlorobenzene      | 20.000  | 15.928 | 20.4 | 76    | 0.00     |

(#) = Out of Range

SPCC's out = 0 CCC's out = 0





# QC SAMPLE DATA

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN100825\  
Data File : VN088029.D  
Acq On : 08 Oct 2025 09:00  
Operator : JC\MD  
Sample : BFB  
Misc : 5.0mL/MSVOA\_N/WATER  
ALS Vial : 1 Sample Multiplier: 1

Instrument :  
MSVOA\_N  
ClientSampleId :  
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N092325W.M  
Title : SW846 8260  
Last Update : Wed Sep 24 05:05:30 2025



AutoFind: Scans 1848, 1849, 1850; Background Corrected with Scan 1838

| Target | Rel. to | Lower  | Upper  | Rel.  | Raw   | Result    |
|--------|---------|--------|--------|-------|-------|-----------|
| Mass   | Mass    | Limit% | Limit% | Abn%  | Abn   | Pass/Fail |
| 50     | 95      | 15     | 40     | 20.0  | 17477 | PASS      |
| 75     | 95      | 30     | 60     | 52.6  | 45984 | PASS      |
| 95     | 95      | 100    | 100    | 100.0 | 87474 | PASS      |
| 96     | 95      | 5      | 9      | 6.4   | 5640  | PASS      |
| 173    | 174     | 0.00   | 2      | 0.7   | 447   | PASS      |
| 174    | 95      | 50     | 100    | 72.1  | 63077 | PASS      |
| 175    | 174     | 5      | 9      | 7.8   | 4922  | PASS      |
| 176    | 174     | 95     | 101    | 96.4  | 60787 | PASS      |
| 177    | 176     | 5      | 9      | 6.8   | 4150  | PASS      |

BFB

Modified:subtracted

| m/z   | abund. | m/z   | abund. | m/z   | abund. | m/z   | abund. |
|-------|--------|-------|--------|-------|--------|-------|--------|
| 36.00 | 1009   | 47.90 | 581    | 59.95 | 735    | 72.00 | 741    |
| 36.95 | 4647   | 49.00 | 3723   | 61.05 | 4431   | 72.90 | 4071   |
| 38.00 | 4004   | 50.00 | 17477  | 61.95 | 4248   | 74.00 | 1487   |
| 38.95 | 1766   | 51.00 | 5503   | 63.00 | 3349   | 75.00 | 4598   |
| 39.90 | 406    | 51.90 | 62     | 63.90 | 238    | 76.00 | 3941   |
| 41.05 | 87     | 52.05 | 200    | 64.20 | 75     | 76.95 | 508    |
| 41.90 | 79     | 53.80 | 133    | 66.85 | 265    | 77.75 | 196    |
| 42.90 | 180    | 54.95 | 468    | 68.00 | 8577   | 78.00 | 198    |
| 43.90 | 433    | 56.00 | 1730   | 68.95 | 9610   | 78.90 | 3521   |
| 44.90 | 634    | 56.95 | 2644   | 69.90 | 438    | 79.80 | 249    |
| 46.95 | 1040   | 57.95 | 232    | 70.10 | 737    | 79.95 | 727    |

Average of 12.824 to 12.835 min.: VN088029.D\data.ms

BFB

Modified:subtracted

| m/z   | abund. | m/z    | abund. | m/z    | abund. | m/z    | abund. |
|-------|--------|--------|--------|--------|--------|--------|--------|
| 80.90 | 3725   | 92.95  | 3969   | 115.05 | 120    | 130.00 | 219    |
| 81.90 | 684    | 94.00  | 11096  | 115.75 | 389    | 130.75 | 204    |
| 82.80 | 75     | 95.00  | 87474  | 116.80 | 384    | 132.90 | 22     |
| 83.05 | 245    | 95.95  | 5640   | 117.00 | 258    | 134.90 | 235    |
| 84.00 | 168    | 97.00  | 213    | 117.80 | 275    | 136.80 | 157    |
| 85.85 | 165    | 104.00 | 436    | 118.20 | 72     | 137.10 | 107    |
| 86.95 | 3617   | 104.85 | 244    | 118.95 | 606    | 138.70 | 63     |
| 87.95 | 3204   | 105.85 | 562    | 127.70 | 113    | 140.85 | 1072   |
| 88.70 | 58     | 109.00 | 53     | 127.95 | 150    | 142.95 | 947    |
| 90.80 | 185    | 111.00 | 107    | 128.85 | 143    | 145.00 | 52     |
| 91.95 | 2395   | 113.00 | 79     | 129.70 | 105    | 146.00 | 66     |

Average of 12.824 to 12.835 min.: VN088029.D\data.ms

BFB

Modified:subtracted

| m/z    | abund. | m/z    | abund. | m/z | abund. | m/z | abund. |
|--------|--------|--------|--------|-----|--------|-----|--------|
| 147.95 | 241    | 173.95 | 63077  |     |        |     |        |
| 149.80 | 78     | 174.95 | 4922   |     |        |     |        |
| 150.00 | 59     | 175.90 | 60787  |     |        |     |        |
| 154.85 | 151    | 176.95 | 4150   |     |        |     |        |
| 156.85 | 252    | 178.00 | 55     |     |        |     |        |
| 160.70 | 56     | 193.00 | 52     |     |        |     |        |
| 160.90 | 52     | 207.15 | 299    |     |        |     |        |
| 170.60 | 76     | 281.05 | 121    |     |        |     |        |
| 172.05 | 128    |        |        |     |        |     |        |
| 172.70 | 168    |        |        |     |        |     |        |
| 173.00 | 447    |        |        |     |        |     |        |





Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN101725\  
Data File : VN088054.D  
Acq On : 17 Oct 2025 08:56  
Operator : JC\MD  
Sample : BFB  
Misc : 5.0mL/MSVOA\_N/WATER  
ALS Vial : 1 Sample Multiplier: 1

Instrument :  
MSVOA\_N  
ClientSampleId :  
BFB

Integration File: RTEINT3.P

Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\624N100825W.M  
Title : METHOD 624 VOLATILE ORGANIC ANALYSIS  
Last Update : Thu Oct 09 02:36:32 2025



AutoFind: Averaged scan 1841 to 1861; Bkg corrected with scan 1840

| Target | Rel. to | Lower  | Upper  | Rel.  | Raw   | Result    |
|--------|---------|--------|--------|-------|-------|-----------|
| Mass   | Mass    | Limit% | Limit% | Abn%  | Abn   | Pass/Fail |
| 50     | 95      | 15     | 40     | 21.6  | 5355  | PASS      |
| 75     | 95      | 30     | 60     | 53.9  | 13380 | PASS      |
| 95     | 95      | 100    | 100    | 100.0 | 24810 | PASS      |
| 96     | 95      | 5      | 9      | 7.2   | 1783  | PASS      |
| 173    | 174     | 0.00   | 2      | 1.2   | 201   | PASS      |
| 174    | 95      | 50     | 100    | 67.1  | 16654 | PASS      |
| 175    | 174     | 5      | 9      | 7.5   | 1251  | PASS      |
| 176    | 174     | 95     | 101    | 95.6  | 15915 | PASS      |
| 177    | 176     | 5      | 9      | 6.8   | 1078  | PASS      |



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,  
Fax : 908 789 8922

## Report of Analysis

Client: Garden State Laboratories, Inc.

Project: Waste Water 2025

Client Sample ID: VN1017WBL01

Lab Sample ID: VN1017WBL01

Analytical Method: E624.1

Sample Wt/Vol: 5 mL

Level : LOW

Final Vol: 5000 uL

Date Collected:

Date Received:

SDG No.: Q3362

Matrix: Water

% Solid: 0

Test: VOCMS Group1

| CAS Number                | Parameter             | Conc.             | Qua. | DF | MDL      | LOQ / CRQL | Units   | Date Ana.      | BatchID  |
|---------------------------|-----------------------|-------------------|------|----|----------|------------|---------|----------------|----------|
| <b>TARGETS</b>            |                       |                   |      |    |          |            |         |                |          |
| 107-02-8                  | Acrolein              | 6.60              | U    | 1  | 6.60     | 25.0       | ug/L    | 10/17/25 10:56 | VN101725 |
| 107-13-1                  | Acrylonitrile         | 2.80              | U    | 1  | 2.80     | 25.0       | ug/L    | 10/17/25 10:56 | VN101725 |
| <b>SURROGATES</b>         |                       |                   |      |    |          |            |         |                |          |
| 17060-07-0                | 1,2-Dichloroethane-d4 | 32.4              |      |    | 91 - 110 | 108%       | SPK: 30 |                |          |
| 2037-26-5                 | Toluene-d8            | 31.5              |      |    | 91 - 112 | 105%       | SPK: 30 |                |          |
| 460-00-4                  | 4-Bromofluorobenzene  | 28.4              |      |    | 63 - 112 | 95%        | SPK: 30 |                |          |
| <b>INTERNAL STANDARDS</b> |                       |                   |      |    |          |            |         |                |          |
|                           |                       | <b>Area Count</b> |      |    |          |            |         |                |          |
| 74-97-5                   | Bromochloromethane    | 56900             |      |    |          |            |         |                |          |
| 540-36-3                  | 1,4-Difluorobenzene   | 316000            |      |    |          |            |         |                |          |
| 3114-55-4                 | Chlorobenzene-d5      | 275000            |      |    |          |            |         |                |          |

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

\* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN101725\  
 Data File : VN088058.D  
 Acq On : 17 Oct 2025 10:56  
 Operator : JC\MD  
 Sample : VN1017WBL01  
 Misc : 5.0mL/MSVOA\_N/WATER  
 ALS Vial : 5 Sample Multiplier: 1

Instrument :  
 MSVOA\_N  
 ClientSampleId :  
 VN1017WBL01

Quant Time: Oct 18 01:27:03 2025  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\624N100825W.M  
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS  
 QLast Update : Thu Oct 09 02:36:32 2025  
 Response via : Initial Calibration

| Compound                | R.T.   | QIon | Response | Conc   | Units | Dev(Min) |
|-------------------------|--------|------|----------|--------|-------|----------|
| -----                   |        |      |          |        |       |          |
| Internal Standards      |        |      |          |        |       |          |
| 1) Bromochloromethane   | 7.794  | 128  | 56868    | 30.000 | ug/l  | 0.00     |
| 28) 1,4-Difluorobenzene | 9.083  | 114  | 316193   | 30.000 | ug/l  | 0.00     |
| 57) Chlorobenzene-d5    | 11.841 | 117  | 274787   | 30.000 | ug/l  | 0.00     |

|                             |        |       |          |          |      |          |
|-----------------------------|--------|-------|----------|----------|------|----------|
| System Monitoring Compounds |        |       |          |          |      |          |
| 27) 1,2-Dichloroethane-d4   | 8.559  | 65    | 143302   | 32.361   | ug/l | 0.00     |
| Spiked Amount               | 30.000 | Range | 91 - 110 | Recovery | =    | 107.867% |
| 60) 4-Bromofluorobenzene    | 12.829 | 95    | 128667   | 28.403   | ug/l | 0.00     |
| Spiked Amount               | 30.000 | Range | 63 - 112 | Recovery | =    | 94.667%  |
| 63) Toluene-d8              | 10.547 | 98    | 385579   | 31.494   | ug/l | 0.00     |
| Spiked Amount               | 30.000 | Range | 91 - 112 | Recovery | =    | 104.967% |


| Target Compounds | Qvalue |
|------------------|--------|
| -----            |        |


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN101725\  
Data File : VN088058.D  
Acq On : 17 Oct 2025 10:56  
Operator : JC\MD  
Sample : VN1017WBL01  
Misc : 5.0mL/MSVOA\_N/WATER  
ALS Vial : 5 Sample Multiplier: 1

Instrument :  
MSVOA\_N  
ClientSampleId :  
VN1017WBL01

Quant Time: Oct 18 01:27:03 2025  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\624N100825W.M  
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS  
QLast Update : Thu Oct 09 02:36:32 2025  
Response via : Initial Calibration









#1  
Bromochloromethane  
Concen: 30.000 ug/l  
RT: 7.794 min Scan# 993  
Delta R.T. 0.000 min  
Lab File: VN088058.D  
Acq: 17 Oct 2025 10:56




Instrument :  
MSVOA\_N  
ClientSampleId :  
VN1017WBL01


| Tgt Ion | Ratio | Lower | Upper |
|---------|-------|-------|-------|
| 128     | 100   |       |       |
| 49      | 182.3 | 0.0   | 439.8 |
| 130     | 125.9 | 0.0   | 325.8 |



#27  
1,2-Dichloroethane-d4  
Concen: 32.361 ug/l  
RT: 8.559 min Scan# 1123  
Delta R.T. -0.006 min  
Lab File: VN088058.D  
Acq: 17 Oct 2025 10:56

| Tgt Ion | Ratio | Lower | Upper |
|---------|-------|-------|-------|
| 65      | 100   |       |       |
| 67      | 52.3  | 41.9  | 62.9  |





#28

1,4-Difluorobenzene

Concen: 30.000 ug/l

RT: 9.083 min Scan# 1

Delta R.T. 0.000 min

Lab File: VN088058.D

Acq: 17 Oct 2025 10:56

Instrument :

MSVOA\_N

ClientSampleId :

VN1017WBL01





Tgt Ion:114 Resp: 316193

Ion Ratio Lower Upper

114 100

63 22.1 16.6 25.0

88 16.7 12.6 18.8



#57

Chlorobenzene-d5

Concen: 30.000 ug/l

RT: 11.841 min Scan# 1681

Delta R.T. -0.006 min

Lab File: VN088058.D

Acq: 17 Oct 2025 10:56




Tgt Ion:117 Resp: 274787


Ion Ratio Lower Upper

117 100

82 58.7 45.2 67.8

119 32.3 25.9 38.9









#60  
4-Bromofluorobenzene  
Concen: 28.403 ug/l  
RT: 12.829 min Scan# 1849  
Delta R.T. 0.000 min  
Lab File: VN088058.D  
Acq: 17 Oct 2025 10:56




Instrument :  
MSVOA\_N  
ClientSampleId :  
VN1017WBL01

Tgt Ion: 95 Resp: 128667  
Ion Ratio Lower Upper  
95 100  
174 68.7 57.4 86.0  
176 65.5 57.0 85.4



#63  
Toluene-d8  
Concen: 31.494 ug/l  
RT: 10.547 min Scan# 1461  
Delta R.T. 0.000 min  
Lab File: VN088058.D  
Acq: 17 Oct 2025 10:56

Tgt Ion: 98 Resp: 385579  
Ion Ratio Lower Upper  
98 100  
100 65.5 53.1 79.7





284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,  
Fax : 908 789 8922

## Report of Analysis

Client: Garden State Laboratories, Inc.

Project: Waste Water 2025

Client Sample ID: VN1017WBS01

Lab Sample ID: VN1017WBS01

Analytical Method: E624.1

Sample Wt/Vol: 5 mL

Level : LOW

Final Vol: 5000 uL

Date Collected:

Date Received:

SDG No.: Q3362

Matrix: Water

% Solid: 0

Test: VOCMS Group1

| CAS Number                | Parameter             | Conc.             | Qua. | DF | MDL      | LOQ / CRQL | Units   | Date Ana.      | BatchID  |
|---------------------------|-----------------------|-------------------|------|----|----------|------------|---------|----------------|----------|
| <b>TARGETS</b>            |                       |                   |      |    |          |            |         |                |          |
| 107-02-8                  | Acrolein              | 84.0              |      | 1  | 6.60     | 25.0       | ug/L    | 10/17/25 10:03 | VN101725 |
| 107-13-1                  | Acrylonitrile         | 91.3              |      | 1  | 2.80     | 25.0       | ug/L    | 10/17/25 10:03 | VN101725 |
| <b>SURROGATES</b>         |                       |                   |      |    |          |            |         |                |          |
| 17060-07-0                | 1,2-Dichloroethane-d4 | 32.0              |      |    | 91 - 110 | 107%       | SPK: 30 |                |          |
| 2037-26-5                 | Toluene-d8            | 30.9              |      |    | 91 - 112 | 103%       | SPK: 30 |                |          |
| 460-00-4                  | 4-Bromofluorobenzene  | 29.9              |      |    | 63 - 112 | 100%       | SPK: 30 |                |          |
| <b>INTERNAL STANDARDS</b> |                       |                   |      |    |          |            |         |                |          |
|                           |                       | <b>Area Count</b> |      |    |          |            |         |                |          |
| 74-97-5                   | Bromochloromethane    | 61000             |      |    |          |            |         |                |          |
| 540-36-3                  | 1,4-Difluorobenzene   | 346000            |      |    |          |            |         |                |          |
| 3114-55-4                 | Chlorobenzene-d5      | 304000            |      |    |          |            |         |                |          |

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

\* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products



Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN101725\  
 Data File : VN088056.D  
 Acq On : 17 Oct 2025 10:03  
 Operator : JC\MD  
 Sample : VN1017WBS01  
 Misc : 5.0mL/MSVOA\_N/WATER  
 ALS Vial : 3 Sample Multiplier: 1

Instrument :  
 MSVOA\_N  
 ClientSampleId :  
 VN1017WBS01

Manual Integrations  
 APPROVED

Reviewed By : John Carlone 10/20/2025  
 Supervised By : Mahesh Dadoda 10/20/2025

Quant Time: Oct 18 01:25:21 2025  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\624N100825W.M  
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS  
 QLast Update : Thu Oct 09 02:36:32 2025  
 Response via : Initial Calibration

| Compound                     | R.T.           | QIon | Response            | Conc   | Units  | Dev(Min) |
|------------------------------|----------------|------|---------------------|--------|--------|----------|
| -----                        |                |      |                     |        |        |          |
| Internal Standards           |                |      |                     |        |        |          |
| 1) Bromochloromethane        | 7.794          | 128  | 61010               | 30.000 | ug/l   | 0.00     |
| 28) 1,4-Difluorobenzene      | 9.082          | 114  | 345518              | 30.000 | ug/l   | 0.00     |
| 57) Chlorobenzene-d5         | 11.841         | 117  | 303706              | 30.000 | ug/l   | 0.00     |
|                              |                |      |                     |        |        |          |
| System Monitoring Compounds  |                |      |                     |        |        |          |
| 27) 1,2-Dichloroethane-d4    | 8.559          | 65   | 152177              | 32.032 | ug/l   | 0.00     |
| Spiked Amount 30.000         | Range 91 - 110 |      | Recovery = 106.767% |        |        |          |
| 60) 4-Bromofluorobenzene     | 12.829         | 95   | 149905              | 29.941 | ug/l   | 0.00     |
| Spiked Amount 30.000         | Range 63 - 112 |      | Recovery = 99.800%  |        |        |          |
| 63) Toluene-d8               | 10.547         | 98   | 417441              | 30.850 | ug/l   | 0.00     |
| Spiked Amount 30.000         | Range 91 - 112 |      | Recovery = 102.833% |        |        |          |
|                              |                |      |                     |        |        |          |
| Target Compounds             |                |      |                     |        |        | Qvalue   |
| 2) Dichlorodifluoromethane   | 2.142          | 85   | 72648               | 20.789 | ug/l   | 95       |
| 3) Chloromethane             | 2.383          | 50   | 75767               | 19.124 | ug/l   | 99       |
| 4) Vinyl Chloride            | 2.536          | 62   | 83370               | 20.182 | ug/l   | 98       |
| 5) Bromomethane              | 2.983          | 94   | 45431               | 20.452 | ug/l   | 100      |
| 6) Chloroethane              | 3.142          | 64   | 51413               | 18.867 | ug/l   | 92       |
| 7) Trichlorofluoromethane    | 3.518          | 101  | 118779              | 21.144 | ug/l   | 91       |
| 8) Diethyl Ether             | 3.959          | 74   | 48062               | 18.935 | ug/l   | 98       |
| 9) 1,1,2-Trichlorotrifluo... | 4.371          | 101  | 62809               | 20.887 | ug/l   | 96       |
| 10) 1,1-Dichloroethene       | 4.342          | 96   | 65077               | 20.287 | ug/l   | 100      |
| 11) Methyl Iodide            | 4.577          | 142  | 49007               | 16.576 | ug/l   | 93       |
| 12) Methyl Acetate           | 5.012          | 43   | 103928              | 19.929 | ug/l # | 86       |
| 13) Acrolein                 | 4.171          | 56   | 67768               | 83.974 | ug/l # | 1        |
| 14) Acrylonitrile            | 5.706          | 53   | 229850              | 91.332 | ug/l   | 100      |
| 15) Acetone                  | 4.424          | 58   | 70716               | 95.949 | ug/l   | 89       |
| 16) Carbon Disulfide         | 4.700          | 76   | 195019              | 19.393 | ug/l   | 97       |
| 17) Allyl chloride           | 5.012          | 41   | 111908              | 18.488 | ug/l   | 95       |
| 18) Methylene Chloride       | 5.265          | 84   | 77740               | 19.067 | ug/l   | 98       |
| 19) trans-1,2-Dichloroethene | 5.777          | 96   | 70214               | 18.932 | ug/l # | 84       |
| 20) Diisopropyl ether        | 6.659          | 45   | 267547              | 19.415 | ug/l   | 98       |
| 21) 1,1-Dichloroethane       | 6.553          | 63   | 144256              | 20.108 | ug/l   | 97       |
| 22) cis-1,2-Dichloroethene   | 7.471          | 96   | 89762               | 19.334 | ug/l   | 98       |
| 23) tert-Butyl Alcohol       | 5.512          | 59   | 86498               | 85.698 | ug/l # | 100      |
| 24) Methyl tert-Butyl Ether  | 5.783          | 73   | 260817              | 18.837 | ug/l   | 97       |
| 25) Chloroform               | 7.947          | 83   | 144250              | 19.485 | ug/l   | 92       |
| 26) Cyclohexane              | 8.241          | 56   | 119448              | 20.096 | ug/l # | 97       |
| 29) 1,1-Dichloropropene      | 8.353          | 75   | 102232              | 20.642 | ug/l   | 95       |
| 30) 2-Butanone               | 7.471          | 43   | 337926              | 92.513 | ug/l   | 97       |
| 31) 2,2-Dichloropropane      | 7.471          | 77   | 126022              | 19.748 | ug/l   | 100      |
| 32) 1,1,1-Trichloroethane    | 8.147          | 97   | 124024              | 20.041 | ug/l   | 98       |
| 33) Carbon Tetrachloride     | 8.341          | 117  | 104455              | 20.151 | ug/l   | 96       |
| 34) Benzene                  | 8.588          | 78   | 314619              | 19.152 | ug/l   | 99       |
| 35) Methacrylonitrile        | 7.759          | 41   | 69646               | 18.132 | ug/l   | 93       |
| 36) 1,2-Dichloroethane       | 8.653          | 62   | 112994              | 19.813 | ug/l   | 97       |
| 37) Trichloroethene          | 9.335          | 130  | 71200               | 18.294 | ug/l   | 98       |
| 38) Methylcyclohexane        | 9.582          | 83   | 121344              | 19.840 | ug/l   | 98       |
| 39) 1,2-Dichloropropane      | 9.600          | 63   | 79149               | 19.215 | ug/l   | 99       |
| 40) Dibromomethane           | 9.688          | 93   | 56354               | 18.795 | ug/l   | 93       |
| 41) Bromodichloromethane     | 9.871          | 83   | 118647              | 19.375 | ug/l   | 97       |
| 42) Vinyl Acetate            | 6.589          | 43   | 1209453m            | 97.194 | ug/l   |          |

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN101725\  
 Data File : VN088056.D  
 Acq On : 17 Oct 2025 10:03  
 Operator : JC\MD  
 Sample : VN1017WBS01  
 Misc : 5.0mL/MSVOA\_N/WATER  
 ALS Vial : 3 Sample Multiplier: 1

Instrument :  
 MSVOA\_N  
 ClientSampleId :  
 VN1017WBS01

Manual Integrations  
 APPROVED

Reviewed By : John Carlone 10/20/2025  
 Supervised By : Mahesh Dadoda 10/20/2025

Quant Time: Oct 18 01:25:21 2025  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\624N100825W.M  
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS  
 QLast Update : Thu Oct 09 02:36:32 2025  
 Response via : Initial Calibration


| Compound                      | R.T.   | QIon | Response | Conc    | Units  | Dev(Min) |
|-------------------------------|--------|------|----------|---------|--------|----------|
| 43) Ethyl Acetate             | 7.547  | 43   | 133195   | 18.969  | ug/l   | 99       |
| 44) Isopropyl Acetate         | 8.671  | 43   | 210930   | 18.392  | ug/l   | 98       |
| 45) 1,4-Dioxane               | 9.682  | 88   | 24886    | 326.268 | ug/l # | 93       |
| 46) Methyl methacrylate       | 9.665  | 41   | 94043    | 18.621  | ug/l   | 95       |
| 47) n-amyl Acetate            | 12.570 | 43   | 114884m  | 21.135  | ug/l   |          |
| 48) t-1,3-Dichloropropene     | 10.812 | 75   | 128661   | 18.895  | ug/l   | 99       |
| 49) cis-1,3-Dichloropropene   | 10.294 | 75   | 132631   | 18.672  | ug/l   | 95       |
| 50) 1,1,2-Trichloroethane     | 10.994 | 97   | 74875    | 18.857  | ug/l   | 99       |
| 51) Ethyl methacrylate        | 10.853 | 69   | 122268   | 18.195  | ug/l   | 93       |
| 52) 1,3-Dichloropropane       | 11.141 | 76   | 131711   | 19.013  | ug/l   | 99       |
| 53) Dibromochloromethane      | 11.335 | 129  | 86723    | 18.269  | ug/l   | 99       |
| 54) 1,2-Dibromoethane         | 11.447 | 107  | 76220    | 17.955  | ug/l   | 99       |
| 55) 2-Chloroethyl vinyl ether | 10.141 | 63   | 292020   | 80.508  | ug/l   | 99       |
| 56) Bromoform                 | 12.559 | 173  | 55031    | 17.148  | ug/l   | 97       |
| 58) 4-Methyl-2-Pentanone      | 10.424 | 43   | 635098   | 93.491  | ug/l   | 98       |
| 59) 2-Hexanone                | 11.176 | 43   | 414833   | 92.491  | ug/l   | 99       |
| 61) Tetrachloroethene         | 11.082 | 164  | 55966    | 18.197  | ug/l   | 87       |
| 62) Toluene                   | 10.606 | 91   | 329738   | 19.140  | ug/l   | 99       |
| 64) Chlorobenzene             | 11.871 | 112  | 209138   | 19.006  | ug/l   | 100      |
| 65) 1,1,1,2-Tetrachloroethane | 11.935 | 131  | 71463    | 18.846  | ug/l   | 98       |
| 66) Ethyl Benzene             | 11.941 | 91   | 365743   | 19.525  | ug/l   | 98       |
| 67) m/p-Xylenes               | 12.047 | 106  | 279849   | 38.752  | ug/l   | 97       |
| 68) o-Xylene                  | 12.376 | 106  | 134497   | 18.709  | ug/l   | 94       |
| 69) Styrene                   | 12.388 | 104  | 225528   | 18.965  | ug/l   | 98       |
| 70) Isopropylbenzene          | 12.670 | 105  | 346221   | 19.650  | ug/l   | 97       |
| 71) 1,1,2,2-Tetrachloroethane | 12.918 | 83   | 110122   | 18.435  | ug/l   | 98       |
| 72) 1,2,3-Trichloropropane    | 12.970 | 75   | 107460m  | 19.370  | ug/l   |          |
| 73) Bromobenzene              | 12.959 | 156  | 77370    | 17.966  | ug/l   | 92       |
| 74) n-propylbenzene           | 13.012 | 91   | 421634   | 19.823  | ug/l   | 98       |
| 75) 2-Chlorotoluene           | 13.100 | 91   | 252126   | 19.386  | ug/l   | 99       |
| 76) 1,3,5-Trimethylbenzene    | 13.153 | 105  | 282658   | 18.886  | ug/l   | 98       |
| 77) t-1,4-Dichloro-2-butene   | 12.718 | 75   | 35729    | 16.161  | ug/l   | 93       |
| 78) 4-Chlorotoluene           | 13.200 | 91   | 260161   | 19.690  | ug/l   | 96       |
| 79) tert-butylbenzene         | 13.412 | 119  | 259100   | 19.970  | ug/l   | 100      |
| 80) 1,2,4-Trimethylbenzene    | 13.459 | 105  | 288836   | 19.278  | ug/l   | 98       |
| 81) sec-Butylbenzene          | 13.594 | 105  | 372393   | 20.414  | ug/l   | 98       |
| 82) p-Isopropyltoluene        | 13.706 | 119  | 308796   | 19.827  | ug/l   | 99       |
| 83) 1,3-Dichlorobenzene       | 13.712 | 146  | 153956   | 19.001  | ug/l   | 99       |
| 84) 1,4-Dichlorobenzene       | 13.788 | 146  | 152875   | 18.295  | ug/l   | 97       |
| 85) n-Butylbenzene            | 14.035 | 91   | 295097   | 20.366  | ug/l   | 97       |
| 86) Hexachloroethane          | 14.312 | 117  | 61862    | 19.828  | ug/l   | 88       |
| 87) 1,2-Dichlorobenzene       | 14.082 | 146  | 148440   | 18.626  | ug/l   | 99       |
| 88) 1,2-Dibromo-3-Chloropr... | 14.700 | 75   | 25015    | 18.234  | ug/l   | 95       |
| 89) 1,2,4-Trichlorobenzene    | 15.811 | 180  | 81108    | 17.086  | ug/l   | 95       |
| 90) Hexachlorobutadiene       | 15.470 | 225  | 30801    | 18.643  | ug/l   | 99       |
| 91) Naphthalene               | 15.611 | 128  | 309666   | 17.818  | ug/l   | 100      |
| 92) 1,2,3-Trichlorobenzene    | 15.811 | 180  | 81108    | 17.086  | ug/l   | 95       |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

**Instrument :**  
MSVOA\_N  
**ClientSampleId :**  
VN1017WBS01

## Manual Integrations APPROVED

Reviewed By :John Carlone 10/20/2025  
Supervised By :Mahesh Dadoda 10/20/2025





284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,  
Fax : 908 789 8922

## Report of Analysis

Client: Garden State Laboratories, Inc.

Project: Waste Water 2025

Client Sample ID: VN1017WBSD01

Lab Sample ID: VN1017WBSD01

Analytical Method: E624.1

Sample Wt/Vol: 5 mL

Level : LOW

Final Vol: 5000 uL

Date Collected:

Date Received:

SDG No.: Q3362

Matrix: Water

% Solid: 0

Test: VOCMS Group1

| CAS Number                | Parameter             | Conc.             | Qua. | DF | MDL      | LOQ / CRQL | Units   | Date Ana.      | BatchID  |
|---------------------------|-----------------------|-------------------|------|----|----------|------------|---------|----------------|----------|
| <b>TARGETS</b>            |                       |                   |      |    |          |            |         |                |          |
| 107-02-8                  | Acrolein              | 91.8              |      | 1  | 6.60     | 25.0       | ug/L    | 10/17/25 10:36 | VN101725 |
| 107-13-1                  | Acrylonitrile         | 90.6              |      | 1  | 2.80     | 25.0       | ug/L    | 10/17/25 10:36 | VN101725 |
| <b>SURROGATES</b>         |                       |                   |      |    |          |            |         |                |          |
| 17060-07-0                | 1,2-Dichloroethane-d4 | 31.5              |      |    | 91 - 110 | 105%       | SPK: 30 |                |          |
| 2037-26-5                 | Toluene-d8            | 30.6              |      |    | 91 - 112 | 102%       | SPK: 30 |                |          |
| 460-00-4                  | 4-Bromofluorobenzene  | 29.9              |      |    | 63 - 112 | 100%       | SPK: 30 |                |          |
| <b>INTERNAL STANDARDS</b> |                       |                   |      |    |          |            |         |                |          |
|                           |                       | <b>Area Count</b> |      |    |          |            |         |                |          |
| 74-97-5                   | Bromochloromethane    | 65400             |      |    |          |            |         |                |          |
| 540-36-3                  | 1,4-Difluorobenzene   | 360000            |      |    |          |            |         |                |          |
| 3114-55-4                 | Chlorobenzene-d5      | 320000            |      |    |          |            |         |                |          |

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

\* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN101725\  
 Data File : VN088057.D  
 Acq On : 17 Oct 2025 10:36  
 Operator : JC\MD  
 Sample : VN1017WBSD01  
 Misc : 5.0mL/MSVOA\_N/WATER  
 ALS Vial : 4 Sample Multiplier: 1

Instrument :  
 MSVOA\_N  
 ClientSampleId :  
 VN1017WBSD01

Manual Integrations  
 APPROVED

Reviewed By : John Carlone 10/20/2025  
 Supervised By : Mahesh Dadoda 10/20/2025

Quant Time: Oct 18 01:26:13 2025  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\624N100825W.M  
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS  
 QLast Update : Thu Oct 09 02:36:32 2025  
 Response via : Initial Calibration

| Compound                     | R.T.           | QIon | Response            | Conc    | Units  | Dev(Min) |
|------------------------------|----------------|------|---------------------|---------|--------|----------|
| -----                        |                |      |                     |         |        |          |
| Internal Standards           |                |      |                     |         |        |          |
| 1) Bromochloromethane        | 7.800          | 128  | 65413               | 30.000  | ug/l   | 0.00     |
| 28) 1,4-Difluorobenzene      | 9.083          | 114  | 359810              | 30.000  | ug/l   | 0.00     |
| 57) Chlorobenzene-d5         | 11.841         | 117  | 319803              | 30.000  | ug/l   | 0.00     |
|                              |                |      |                     |         |        |          |
| System Monitoring Compounds  |                |      |                     |         |        |          |
| 27) 1,2-Dichloroethane-d4    | 8.559          | 65   | 160373              | 31.485  | ug/l   | 0.00     |
| Spiked Amount 30.000         | Range 91 - 110 |      | Recovery = 104.967% |         |        |          |
| 60) 4-Bromofluorobenzene     | 12.829         | 95   | 157631              | 29.899  | ug/l   | 0.00     |
| Spiked Amount 30.000         | Range 63 - 112 |      | Recovery = 99.667%  |         |        |          |
| 63) Toluene-d8               | 10.547         | 98   | 435417              | 30.559  | ug/l   | 0.00     |
| Spiked Amount 30.000         | Range 91 - 112 |      | Recovery = 101.867% |         |        |          |
|                              |                |      |                     |         |        |          |
| Target Compounds             |                |      |                     |         |        |          |
|                              |                |      |                     |         | Qvalue |          |
| 2) Dichlorodifluoromethane   | 2.142          | 85   | 78833               | 21.041  | ug/l   | 93       |
| 3) Chloromethane             | 2.383          | 50   | 80807               | 19.024  | ug/l   | 99       |
| 4) Vinyl Chloride            | 2.536          | 62   | 87603               | 19.779  | ug/l   | 100      |
| 5) Bromomethane              | 2.983          | 94   | 48410               | 20.326  | ug/l   | 98       |
| 6) Chloroethane              | 3.142          | 64   | 55974               | 19.158  | ug/l   | 99       |
| 7) Trichlorofluoromethane    | 3.518          | 101  | 129328              | 21.473  | ug/l   | 99       |
| 8) Diethyl Ether             | 3.959          | 74   | 53446               | 19.639  | ug/l   | 98       |
| 9) 1,1,2-Trichlorotrifluo... | 4.371          | 101  | 68415               | 21.220  | ug/l   | 97       |
| 10) 1,1-Dichloroethene       | 4.342          | 96   | 68376               | 19.881  | ug/l   | 99       |
| 11) Methyl Iodide            | 4.583          | 142  | 54630               | 17.234  | ug/l   | 95       |
| 12) Methyl Acetate           | 5.018          | 43   | 112344              | 20.093  | ug/l   | 94       |
| 13) Acrolein                 | 4.177          | 56   | 79427               | 91.796  | ug/l # | 8        |
| 14) Acrylonitrile            | 5.706          | 53   | 244395              | 90.575  | ug/l   | 98       |
| 15) Acetone                  | 4.424          | 58   | 73551               | 93.078  | ug/l   | 88       |
| 16) Carbon Disulfide         | 4.706          | 76   | 209149              | 19.399  | ug/l   | 98       |
| 17) Allyl chloride           | 5.012          | 41   | 121814              | 18.770  | ug/l   | 96       |
| 18) Methylene Chloride       | 5.265          | 84   | 83118               | 19.014  | ug/l   | 93       |
| 19) trans-1,2-Dichloroethene | 5.771          | 96   | 76791               | 19.311  | ug/l   | 88       |
| 20) Diisopropyl ether        | 6.659          | 45   | 285092              | 19.295  | ug/l   | 96       |
| 21) 1,1-Dichloroethane       | 6.553          | 63   | 156888              | 20.397  | ug/l   | 96       |
| 22) cis-1,2-Dichloroethene   | 7.471          | 96   | 93825               | 18.849  | ug/l   | 94       |
| 23) tert-Butyl Alcohol       | 5.518          | 59   | 89595               | 82.791  | ug/l # | 100      |
| 24) Methyl tert-Butyl Ether  | 5.789          | 73   | 284030              | 19.132  | ug/l   | 100      |
| 25) Chloroform               | 7.947          | 83   | 155760              | 19.624  | ug/l   | 97       |
| 26) Cyclohexane              | 8.241          | 56   | 129096              | 20.257  | ug/l # | 98       |
| 29) 1,1-Dichloropropene      | 8.353          | 75   | 107684              | 20.879  | ug/l   | 95       |
| 30) 2-Butanone               | 7.465          | 43   | 358092              | 94.140  | ug/l   | 98       |
| 31) 2,2-Dichloropropane      | 7.471          | 77   | 135758              | 20.429  | ug/l   | 99       |
| 32) 1,1,1-Trichloroethane    | 8.147          | 97   | 130885              | 20.309  | ug/l   | 97       |
| 33) Carbon Tetrachloride     | 8.341          | 117  | 109382              | 20.263  | ug/l   | 95       |
| 34) Benzene                  | 8.583          | 78   | 329461              | 19.259  | ug/l   | 96       |
| 35) Methacrylonitrile        | 7.759          | 41   | 74397               | 18.599  | ug/l   | 92       |
| 36) 1,2-Dichloroethane       | 8.653          | 62   | 121966              | 20.537  | ug/l   | 98       |
| 37) Trichloroethene          | 9.330          | 130  | 75152               | 18.542  | ug/l   | 93       |
| 38) Methylcyclohexane        | 9.577          | 83   | 131631              | 20.667  | ug/l   | 96       |
| 39) 1,2-Dichloropropane      | 9.600          | 63   | 85728               | 19.986  | ug/l   | 100      |
| 40) Dibromomethane           | 9.688          | 93   | 62068               | 19.878  | ug/l   | 93       |
| 41) Bromodichloromethane     | 9.865          | 83   | 125840              | 19.733  | ug/l   | 99       |
| 42) Vinyl Acetate            | 6.589          | 43   | 1333455             | 102.902 | ug/l   | 96       |

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN101725\  
 Data File : VN088057.D  
 Acq On : 17 Oct 2025 10:36  
 Operator : JC\MD  
 Sample : VN1017WBSD01  
 Misc : 5.0mL/MSVOA\_N/WATER  
 ALS Vial : 4 Sample Multiplier: 1

Instrument :  
 MSVOA\_N  
 ClientSampleId :  
 VN1017WBSD01

Manual Integrations  
 APPROVED

Reviewed By : John Carlone 10/20/2025  
 Supervised By : Mahesh Dadoda 10/20/2025

Quant Time: Oct 18 01:26:13 2025  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\624N100825W.M  
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS  
 QLast Update : Thu Oct 09 02:36:32 2025  
 Response via : Initial Calibration

| Compound                      | R.T.   | QIon | Response | Conc    | Units  | Dev(Min) |
|-------------------------------|--------|------|----------|---------|--------|----------|
| 43) Ethyl Acetate             | 7.547  | 43   | 147140   | 20.123  | ug/l   | 96       |
| 44) Isopropyl Acetate         | 8.671  | 43   | 231811   | 19.410  | ug/l   | 98       |
| 45) 1,4-Dioxane               | 9.677  | 88   | 26935    | 339.104 | ug/l # | 94       |
| 46) Methyl methacrylate       | 9.659  | 41   | 102432   | 19.476  | ug/l   | 94       |
| 47) n-amyl Acetate            | 12.594 | 43   | 131961m  | 23.312  | ug/l   |          |
| 48) t-1,3-Dichloropropene     | 10.818 | 75   | 135939   | 19.170  | ug/l   | 98       |
| 49) cis-1,3-Dichloropropene   | 10.288 | 75   | 146225   | 19.769  | ug/l   | 92       |
| 50) 1,1,2-Trichloroethane     | 10.994 | 97   | 78475    | 18.979  | ug/l   | 94       |
| 51) Ethyl methacrylate        | 10.859 | 69   | 135539   | 19.369  | ug/l   | 96       |
| 52) 1,3-Dichloropropane       | 11.141 | 76   | 144001   | 19.961  | ug/l   | 100      |
| 53) Dibromochloromethane      | 11.335 | 129  | 92926    | 18.798  | ug/l   | 99       |
| 54) 1,2-Dibromoethane         | 11.447 | 107  | 80897    | 18.300  | ug/l   | 96       |
| 55) 2-Chloroethyl vinyl ether | 10.141 | 63   | 345576   | 91.489  | ug/l   | 98       |
| 56) Bromoform                 | 12.565 | 173  | 58921    | 17.630  | ug/l   | 98       |
| 58) 4-Methyl-2-Pentanone      | 10.424 | 43   | 684259   | 95.657  | ug/l   | 98       |
| 59) 2-Hexanone                | 11.176 | 43   | 453352   | 95.992  | ug/l   | 99       |
| 61) Tetrachloroethene         | 11.082 | 164  | 61856    | 19.100  | ug/l   | 88       |
| 62) Toluene                   | 10.612 | 91   | 354587   | 19.547  | ug/l   | 100      |
| 64) Chlorobenzene             | 11.871 | 112  | 224911   | 19.411  | ug/l   | 98       |
| 65) 1,1,1,2-Tetrachloroethane | 11.935 | 131  | 76567    | 19.175  | ug/l   | 95       |
| 66) Ethyl Benzene             | 11.941 | 91   | 390582   | 19.801  | ug/l   | 99       |
| 67) m/p-Xylenes               | 12.047 | 106  | 298777   | 39.291  | ug/l   | 97       |
| 68) o-Xylene                  | 12.376 | 106  | 146619   | 19.369  | ug/l   | 96       |
| 69) Styrene                   | 12.388 | 104  | 238932   | 19.081  | ug/l   | 99       |
| 70) Isopropylbenzene          | 12.671 | 105  | 373836   | 20.149  | ug/l   | 99       |
| 71) 1,1,2,2-Tetrachloroethane | 12.918 | 83   | 121382   | 19.297  | ug/l   | 99       |
| 72) 1,2,3-Trichloropropane    | 12.970 | 75   | 119252m  | 20.414  | ug/l   |          |
| 73) Bromobenzene              | 12.959 | 156  | 84476    | 18.629  | ug/l   | 93       |
| 74) n-propylbenzene           | 13.012 | 91   | 458411   | 20.467  | ug/l   | 98       |
| 75) 2-Chlorotoluene           | 13.100 | 91   | 270316   | 19.739  | ug/l   | 98       |
| 76) 1,3,5-Trimethylbenzene    | 13.153 | 105  | 314783   | 19.974  | ug/l   | 97       |
| 77) t-1,4-Dichloro-2-butene   | 12.723 | 75   | 42042    | 18.059  | ug/l   | 91       |
| 78) 4-Chlorotoluene           | 13.200 | 91   | 277549   | 19.949  | ug/l   | 98       |
| 79) tert-butylbenzene         | 13.418 | 119  | 276568   | 20.243  | ug/l   | 98       |
| 80) 1,2,4-Trimethylbenzene    | 13.459 | 105  | 311676   | 19.755  | ug/l   | 97       |
| 81) sec-Butylbenzene          | 13.594 | 105  | 398977   | 20.771  | ug/l   | 97       |
| 82) p-Isopropyltoluene        | 13.706 | 119  | 337559   | 20.583  | ug/l   | 99       |
| 83) 1,3-Dichlorobenzene       | 13.712 | 146  | 163469   | 19.160  | ug/l   | 99       |
| 84) 1,4-Dichlorobenzene       | 13.794 | 146  | 171817   | 19.527  | ug/l   | 98       |
| 85) n-Butylbenzene            | 14.035 | 91   | 318876   | 20.899  | ug/l   | 97       |
| 86) Hexachloroethane          | 14.306 | 117  | 66012    | 20.093  | ug/l   | 88       |
| 87) 1,2-Dichlorobenzene       | 14.082 | 146  | 157781   | 18.801  | ug/l   | 99       |
| 88) 1,2-Dibromo-3-Chloropr... | 14.700 | 75   | 26957    | 18.660  | ug/l   | 95       |
| 89) 1,2,4-Trichlorobenzene    | 15.811 | 180  | 90270    | 18.059  | ug/l   | 96       |
| 90) Hexachlorobutadiene       | 15.470 | 225  | 33868    | 19.467  | ug/l   | 97       |
| 91) Naphthalene               | 15.611 | 128  | 336619   | 18.394  | ug/l   | 99       |
| 92) 1,2,3-Trichlorobenzene    | 15.811 | 180  | 90270    | 18.059  | ug/l   | 96       |

(#) = qualifier out of range (m) = manual integration (+) = signals summed




Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN101725\  
 Data File : VN088057.D  
 Acq On : 17 Oct 2025 10:36  
 Operator : JC\MD  
 Sample : VN1017WBSD01  
 Misc : 5.0mL/MSVOA\_N/WATER  
 ALS Vial : 4 Sample Multiplier: 1

Instrument :  
 MSVOA\_N  
 ClientSampleId :  
 VN1017WBSD01

### Manual Integrations APPROVED

Reviewed By : John Carlone 10/20/2025  
 Supervised By : Mahesh Dadoda 10/20/2025

Quant Time: Oct 18 01:26:13 2025  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\624N100825W.M  
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS  
 QLast Update : Thu Oct 09 02:36:32 2025  
 Response via : Initial Calibration



## Manual Integration Report

|           |          |            |         |
|-----------|----------|------------|---------|
| Sequence: | vn100825 | Instrument | MSVOA_n |
|-----------|----------|------------|---------|

| Sample ID   | File ID    | Parameter               | Review By | Review On                | Supervised By | Supervised On            | Reason                      |
|-------------|------------|-------------------------|-----------|--------------------------|---------------|--------------------------|-----------------------------|
| VSTDICC005  | VN088030.D | 1,2,3-Trichloropropane  | JOHN      | 10/9/2025<br>12:47:30 PM | MMdadoda      | 10/10/2025<br>4:40:11 AM | Peak Integrated by Software |
| VSTDICC005  | VN088030.D | 2,2-Dichloropropane     | JOHN      | 10/9/2025<br>12:47:30 PM | MMdadoda      | 10/10/2025<br>4:40:11 AM | Peak Integrated by Software |
| VSTDICC005  | VN088030.D | Acetone                 | JOHN      | 10/9/2025<br>12:47:30 PM | MMdadoda      | 10/10/2025<br>4:40:11 AM | Peak Integrated by Software |
| VSTDICC005  | VN088030.D | Allyl chloride          | JOHN      | 10/9/2025<br>12:47:30 PM | MMdadoda      | 10/10/2025<br>4:40:11 AM | Peak Integrated by Software |
| VSTDICC005  | VN088030.D | Methyl Acetate          | JOHN      | 10/9/2025<br>12:47:30 PM | MMdadoda      | 10/10/2025<br>4:40:11 AM | Peak Integrated by Software |
| VSTDICC005  | VN088030.D | Methyl Iodide           | JOHN      | 10/9/2025<br>12:47:30 PM | MMdadoda      | 10/10/2025<br>4:40:11 AM | Peak Integrated by Software |
| VSTDICC005  | VN088030.D | Methylene Chloride      | JOHN      | 10/9/2025<br>12:47:30 PM | MMdadoda      | 10/10/2025<br>4:40:11 AM | Peak Integrated by Software |
| VSTDICC005  | VN088030.D | n-amyl Acetate          | JOHN      | 10/9/2025<br>12:47:30 PM | MMdadoda      | 10/10/2025<br>4:40:11 AM | Peak Integrated by Software |
| VSTDICC005  | VN088030.D | t-1,4-Dichloro-2-butene | JOHN      | 10/9/2025<br>12:47:30 PM | MMdadoda      | 10/10/2025<br>4:40:11 AM | Peak Integrated by Software |
| VSTDICCC020 | VN088031.D | 1,2,3-Trichloropropane  | JOHN      | 10/9/2025<br>12:47:35 PM | MMdadoda      | 10/10/2025<br>4:40:17 AM | Peak Integrated by Software |
| VSTDICCC020 | VN088031.D | n-amyl Acetate          | JOHN      | 10/9/2025<br>12:47:35 PM | MMdadoda      | 10/10/2025<br>4:40:17 AM | Peak Integrated by Software |
| VSTDICC050  | VN088032.D | 1,2,3-Trichloropropane  | JOHN      | 10/9/2025<br>12:47:40 PM | MMdadoda      | 10/10/2025<br>4:40:22 AM | Peak Integrated by Software |
| VSTDICC050  | VN088032.D | n-amyl Acetate          | JOHN      | 10/9/2025<br>12:47:40 PM | MMdadoda      | 10/10/2025<br>4:40:22 AM | Peak Integrated by Software |



## Manual Integration Report

|           |          |            |         |
|-----------|----------|------------|---------|
| Sequence: | vn100825 | Instrument | MSVOA_n |
|-----------|----------|------------|---------|

| Sample ID  | File ID    | Parameter              | Review By | Review On                | Supervised By | Supervised On            | Reason                      |
|------------|------------|------------------------|-----------|--------------------------|---------------|--------------------------|-----------------------------|
| VSTDICC100 | VN088033.D | 1,2,3-Trichloropropane | JOHN      | 10/9/2025<br>12:47:45 PM | MMdadoda      | 10/10/2025<br>4:40:26 AM | Peak Integrated by Software |
| VSTDICC100 | VN088033.D | Methyl Acetate         | JOHN      | 10/9/2025<br>12:47:45 PM | MMdadoda      | 10/10/2025<br>4:40:26 AM | Peak Integrated by Software |
| VSTDICC150 | VN088034.D | 1,2,3-Trichloropropane | JOHN      | 10/9/2025<br>12:47:50 PM | MMdadoda      | 10/10/2025<br>4:40:32 AM | Peak Integrated by Software |
| VSTDICV020 | VN088036.D | 1,2,3-Trichloropropane | JOHN      | 10/9/2025<br>12:47:54 PM | MMdadoda      | 10/10/2025<br>4:40:37 AM | Peak Integrated by Software |
| VSTDICV020 | VN088036.D | Acetone                | JOHN      | 10/9/2025<br>12:47:54 PM | MMdadoda      | 10/10/2025<br>4:40:37 AM | Peak Integrated by Software |
| VSTDICV020 | VN088036.D | n-amyl Acetate         | JOHN      | 10/9/2025<br>12:47:54 PM | MMdadoda      | 10/10/2025<br>4:40:37 AM | Peak Integrated by Software |
| VSTDCCC020 | VN088046.D | 1,2,3-Trichloropropane | JOHN      | 10/9/2025<br>12:48:29 PM | MMdadoda      | 10/10/2025<br>4:41:14 AM | Peak Integrated by Software |
| VSTDCCC020 | VN088046.D | n-amyl Acetate         | JOHN      | 10/9/2025<br>12:48:29 PM | MMdadoda      | 10/10/2025<br>4:41:14 AM | Peak Integrated by Software |

## Manual Integration Report

|           |          |            |         |
|-----------|----------|------------|---------|
| Sequence: | VN101725 | Instrument | MSVOA_n |
|-----------|----------|------------|---------|

| Sample ID    | File ID    | Parameter                      | Review By | Review On                | Supervised By | Supervised On            | Reason                      |
|--------------|------------|--------------------------------|-----------|--------------------------|---------------|--------------------------|-----------------------------|
| VSTDCCC020   | VN088055.D | 1,1,2-Trichlorotrifluoroethane | JOHN      | 10/20/2025<br>8:10:50 AM | MMDadoda      | 10/20/2025<br>3:55:28 PM | Peak Integrated by Software |
| VSTDCCC020   | VN088055.D | 1,2,3-Trichloropropane         | JOHN      | 10/20/2025<br>8:10:50 AM | MMDadoda      | 10/20/2025<br>3:55:28 PM | Peak Integrated by Software |
| VSTDCCC020   | VN088055.D | Ethyl Acetate                  | JOHN      | 10/20/2025<br>8:10:50 AM | MMDadoda      | 10/20/2025<br>3:55:28 PM | Peak Integrated by Software |
| VSTDCCC020   | VN088055.D | Methyl methacrylate            | JOHN      | 10/20/2025<br>8:10:50 AM | MMDadoda      | 10/20/2025<br>3:55:28 PM | Peak Integrated by Software |
| VSTDCCC020   | VN088055.D | n-amyl Acetate                 | JOHN      | 10/20/2025<br>8:10:50 AM | MMDadoda      | 10/20/2025<br>3:55:28 PM | Peak Integrated by Software |
| VSTDCCC020   | VN088055.D | t-1,4-Dichloro-2-butene        | JOHN      | 10/20/2025<br>8:10:50 AM | MMDadoda      | 10/20/2025<br>3:55:28 PM | Peak Integrated by Software |
| VN1017WBS01  | VN088056.D | 1,2,3-Trichloropropane         | JOHN      | 10/20/2025<br>8:10:54 AM | MMDadoda      | 10/20/2025<br>3:55:31 PM | Peak Integrated by Software |
| VN1017WBS01  | VN088056.D | n-amyl Acetate                 | JOHN      | 10/20/2025<br>8:10:54 AM | MMDadoda      | 10/20/2025<br>3:55:31 PM | Peak Integrated by Software |
| VN1017WBS01  | VN088056.D | Vinyl Acetate                  | JOHN      | 10/20/2025<br>8:10:54 AM | MMDadoda      | 10/20/2025<br>3:55:31 PM | Peak Integrated by Software |
| VN1017WBSD01 | VN088057.D | 1,2,3-Trichloropropane         | JOHN      | 10/20/2025<br>8:10:59 AM | MMDadoda      | 10/20/2025<br>3:55:34 PM | Peak Integrated by Software |
| VN1017WBSD01 | VN088057.D | n-amyl Acetate                 | JOHN      | 10/20/2025<br>8:10:59 AM | MMDadoda      | 10/20/2025<br>3:55:34 PM | Peak Integrated by Software |
| Q3362-01     | VN088065.D | Dichlorodifluoromethane        | JOHN      | 10/20/2025<br>8:11:11 AM | MMDadoda      | 10/20/2025<br>3:56:20 PM | Peak Integrated by Software |
| Q3362-01     | VN088065.D | Methyl tert-Butyl Ether        | JOHN      | 10/20/2025<br>8:11:11 AM | MMDadoda      | 10/20/2025<br>3:56:20 PM | Peak Integrated by Software |



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

## Manual Integration Report

|           |          |            |         |
|-----------|----------|------------|---------|
| Sequence: | VN101725 | Instrument | MSVOA_n |
|-----------|----------|------------|---------|

| Sample ID  | File ID    | Parameter              | Review By | Review On                | Supervised By | Supervised On            | Reason                      |
|------------|------------|------------------------|-----------|--------------------------|---------------|--------------------------|-----------------------------|
| VSTDCCC020 | VN088070.D | 1,2,3-Trichloropropane | JOHN      | 10/20/2025<br>8:11:23 AM | MMDadoda      | 10/20/2025<br>3:56:31 PM | Peak Integrated by Software |
| VSTDCCC020 | VN088070.D | n-amyl Acetate         | JOHN      | 10/20/2025<br>8:11:23 AM | MMDadoda      | 10/20/2025<br>3:56:31 PM | Peak Integrated by Software |

Instrument ID: **MSVOA\_N**

**Daily Analysis Runlog For Sequence/QC Batch ID # VN100825**

|                          |                                              |                   |                                    |
|--------------------------|----------------------------------------------|-------------------|------------------------------------|
| Review By                | John Carlone                                 | Review On         | 10/9/2025 1:01:25 PM               |
| Supervise By             | Mahesh Dadoda                                | Supervise On      | 10/10/2025 4:41:25 AM              |
| SubDirectory             | VN100825                                     | HP Acquire Method | HP Processing Method 624N100825W.M |
| <b>STD. NAME</b>         | <b>STD REF.#</b>                             |                   |                                    |
| Tune/Reschk              | VP135778                                     |                   |                                    |
| Initial Calibration Stds | VP135780,VP135781,VP135782,VP135783,VP135784 |                   |                                    |
| CCC                      | VP135779,VP135796,VP135797                   |                   |                                    |
| Internal Standard/PEM    |                                              |                   |                                    |
| ICV/I.BLK                | VP135785                                     |                   |                                    |
| Surrogate Standard       |                                              |                   |                                    |
| MS/MSD Standard          |                                              |                   |                                    |
| LCS Standard             |                                              |                   |                                    |

| Sr# | SampleId        | Data File Name | Date-Time         | Operator | Status |
|-----|-----------------|----------------|-------------------|----------|--------|
| 1   | BFB             | VN088029.D     | 08 Oct 2025 09:00 | JC\MD    | Ok     |
| 2   | VSTDICCC005     | VN088030.D     | 08 Oct 2025 09:34 | JC\MD    | Ok,M   |
| 3   | VSTDICCC020     | VN088031.D     | 08 Oct 2025 10:08 | JC\MD    | Ok,M   |
| 4   | VSTDICCC050     | VN088032.D     | 08 Oct 2025 10:30 | JC\MD    | Ok,M   |
| 5   | VSTDICCC100     | VN088033.D     | 08 Oct 2025 10:51 | JC\MD    | Ok,M   |
| 6   | VSTDICCC150     | VN088034.D     | 08 Oct 2025 11:12 | JC\MD    | Ok,M   |
| 7   | IBLK            | VN088035.D     | 08 Oct 2025 11:33 | JC\MD    | Ok     |
| 8   | VSTDICV020      | VN088036.D     | 08 Oct 2025 12:52 | JC\MD    | Ok,M   |
| 9   | VN1008WBS01     | VN088037.D     | 08 Oct 2025 13:25 | JC\MD    | Ok,M   |
| 10  | VN1008WBSD01    | VN088038.D     | 08 Oct 2025 13:59 | JC\MD    | Ok,M   |
| 11  | VN1008WBL01     | VN088039.D     | 08 Oct 2025 14:20 | JC\MD    | Ok     |
| 12  | Q3015-02        | VN088040.D     | 08 Oct 2025 14:59 | JC\MD    | Ok,NR  |
| 13  | IBLK            | VN088041.D     | 08 Oct 2025 15:20 | JC\MD    | Ok     |
| 14  | Q2893-07 2.5PPB | VN088042.D     | 08 Oct 2025 15:41 | JC\MD    | Ok,M   |
| 15  | Q2893-08 5.0PPB | VN088043.D     | 08 Oct 2025 16:02 | JC\MD    | Ok,M   |
| 16  | Q3281-01        | VN088044.D     | 08 Oct 2025 16:23 | JC\MD    | Ok,M   |
| 17  | Q3281-02        | VN088045.D     | 08 Oct 2025 16:44 | JC\MD    | Ok,M   |
| 18  | VSTDCCC020      | VN088046.D     | 08 Oct 2025 17:05 | JC\MD    | Ok,M   |

M : Manual Integration

Instrument ID: **MSVOA\_N**

**Daily Analysis Runlog For Sequence/QC Batch ID # VN101725**

|                                                                                                    |                   |                   |                                    |
|----------------------------------------------------------------------------------------------------|-------------------|-------------------|------------------------------------|
| Review By                                                                                          | John Carlone      | Review On         | 10/20/2025 8:16:05 AM              |
| Supervise By                                                                                       | Maresh Dadoda     | Supervise On      | 10/20/2025 3:56:38 PM              |
| SubDirectory                                                                                       | VN101725          | HP Acquire Method | HP Processing Method 624N100825W.M |
| <b>STD. NAME</b>                                                                                   | <b>STD REF.#</b>  |                   |                                    |
| Tune/Reschk<br>Initial Calibration Stds                                                            | VP135826          |                   |                                    |
| CCC<br>Internal Standard/PEM<br>ICV/I.BLK<br>Surrogate Standard<br>MS/MSD Standard<br>LCS Standard | VP135827,VP135828 |                   |                                    |

| Sr# | SampleId     | Data File Name | Date-Time         | Operator | Status |
|-----|--------------|----------------|-------------------|----------|--------|
| 1   | BFB          | VN088054.D     | 17 Oct 2025 08:56 | JC\MD    | Ok     |
| 2   | VSTDCCC020   | VN088055.D     | 17 Oct 2025 09:29 | JC\MD    | Ok,M   |
| 3   | VN1017WBS01  | VN088056.D     | 17 Oct 2025 10:03 | JC\MD    | Ok,M   |
| 4   | VN1017WBSD01 | VN088057.D     | 17 Oct 2025 10:36 | JC\MD    | Ok,M   |
| 5   | VN1017WBL01  | VN088058.D     | 17 Oct 2025 10:56 | JC\MD    | Ok     |
| 6   | Q3361-02     | VN088059.D     | 17 Oct 2025 11:32 | JC\MD    | Ok     |
| 7   | Q3362-02     | VN088060.D     | 17 Oct 2025 11:53 | JC\MD    | Ok     |
| 8   | Q3362-01     | VN088061.D     | 17 Oct 2025 12:14 | JC\MD    | ReRun  |
| 9   | Q3361-01     | VN088062.D     | 17 Oct 2025 12:34 | JC\MD    | Ok,M   |
| 10  | IBLK         | VN088063.D     | 17 Oct 2025 12:55 | JC\MD    | Ok     |
| 11  | IBLK         | VN088064.D     | 17 Oct 2025 13:16 | JC\MD    | Ok     |
| 12  | Q3362-01     | VN088065.D     | 17 Oct 2025 13:36 | JC\MD    | Ok,M   |
| 13  | IBLK         | VN088066.D     | 17 Oct 2025 14:58 | JC\MD    | Ok     |
| 14  | Q3385-01     | VN088067.D     | 17 Oct 2025 15:18 | JC\MD    | ReRun  |
| 15  | Q3385-02MS   | VN088068.D     | 17 Oct 2025 15:39 | JC\MD    | Ok,M   |
| 16  | Q3385-03MSD  | VN088069.D     | 17 Oct 2025 16:00 | JC\MD    | Ok,M   |
| 17  | VSTDCCC020   | VN088070.D     | 17 Oct 2025 16:20 | JC\MD    | Ok,M   |

M : Manual Integration

Instrument ID: MSVOA\_N

**Daily Analysis Runlog For Sequence/QC Batch ID # VN100825**

|              |               |                   |                                    |
|--------------|---------------|-------------------|------------------------------------|
| Review By    | John Carlone  | Review On         | 10/9/2025 1:01:25 PM               |
| Supervise By | Maresh Dadoda | Supervise On      | 10/10/2025 4:41:25 AM              |
| SubDirectory | VN100825      | HP Acquire Method | HP Processing Method 624N100825W.M |

| STD. NAME                | STD REF.#                                    |
|--------------------------|----------------------------------------------|
| Tune/Reschk              | VP135778                                     |
| Initial Calibration Stds | VP135780,VP135781,VP135782,VP135783,VP135784 |
| CCC                      | VP135779,VP135796,VP135797                   |
| Internal Standard/PEM    |                                              |
| ICV/I.BLK                | VP135785                                     |
| Surrogate Standard       |                                              |
| MS/MSD Standard          |                                              |
| LCS Standard             |                                              |

| Sr# | SampleID        | ClientID            | Data File Name | Date-Time         | Comment     | Operator | Status |
|-----|-----------------|---------------------|----------------|-------------------|-------------|----------|--------|
| 1   | BFB             | BFB                 | VN088029.D     | 08 Oct 2025 09:00 |             | JC\MD    | Ok     |
| 2   | VSTDIC005       | VSTDIC005           | VN088030.D     | 08 Oct 2025 09:34 |             | JC\MD    | Ok,M   |
| 3   | VSTDICCC020     | VSTDICCC020         | VN088031.D     | 08 Oct 2025 10:08 |             | JC\MD    | Ok,M   |
| 4   | VSTDIC050       | VSTDIC050           | VN088032.D     | 08 Oct 2025 10:30 |             | JC\MD    | Ok,M   |
| 5   | VSTDIC100       | VSTDIC100           | VN088033.D     | 08 Oct 2025 10:51 |             | JC\MD    | Ok,M   |
| 6   | VSTDIC150       | VSTDIC150           | VN088034.D     | 08 Oct 2025 11:12 |             | JC\MD    | Ok,M   |
| 7   | IBLK            | IBLK                | VN088035.D     | 08 Oct 2025 11:33 |             | JC\MD    | Ok     |
| 8   | VSTDICV020      | ICVVN100825         | VN088036.D     | 08 Oct 2025 12:52 |             | JC\MD    | Ok,M   |
| 9   | VN1008WBS01     | VN1008WBS01         | VN088037.D     | 08 Oct 2025 13:25 |             | JC\MD    | Ok,M   |
| 10  | VN1008WBSD01    | VN1008WBSD01        | VN088038.D     | 08 Oct 2025 13:59 |             | JC\MD    | Ok,M   |
| 11  | VN1008WBL01     | VN1008WBL01         | VN088039.D     | 08 Oct 2025 14:20 |             | JC\MD    | Ok     |
| 12  | Q3015-02        | WP0925-PT-VOA-WP    | VN088040.D     | 08 Oct 2025 14:59 |             | JC\MD    | Ok,NR  |
| 13  | IBLK            | IBLK                | VN088041.D     | 08 Oct 2025 15:20 |             | JC\MD    | Ok     |
| 14  | Q2893-07 2.5PPB | LOD-MDL-WATER-01-0  | VN088042.D     | 08 Oct 2025 15:41 |             | JC\MD    | Ok,M   |
| 15  | Q2893-08 5.0PPB | LOQ-WATER-02-QT3-2  | VN088043.D     | 08 Oct 2025 16:02 |             | JC\MD    | Ok,M   |
| 16  | Q3281-01        | 001-WILLETS-PT-BLVD | VN088044.D     | 08 Oct 2025 16:23 | vial A pH<2 | JC\MD    | Ok,M   |
| 17  | Q3281-02        | 002-35TH-AVE(OCT)   | VN088045.D     | 08 Oct 2025 16:44 | vial A pH<2 | JC\MD    | Ok,M   |
| 18  | VSTDCCC020      | VSTDCCC020EC        | VN088046.D     | 08 Oct 2025 17:05 |             | JC\MD    | Ok,M   |

Instrument ID: MSVOA\_N

**Daily Analysis Runlog For Sequence/QC Batch ID # VN100825**

| Review By                | John Carlone                                 | Review On         | 10/9/2025 1:01:25 PM               |
|--------------------------|----------------------------------------------|-------------------|------------------------------------|
| Supervise By             | Mahesh Dadoda                                | Supervise On      | 10/10/2025 4:41:25 AM              |
| SubDirectory             | VN100825                                     | HP Acquire Method | HP Processing Method 624N100825W.M |
| STD. NAME                | STD REF.#                                    |                   |                                    |
| Tune/Reschk              | VP135778                                     |                   |                                    |
| Initial Calibration Stds | VP135780,VP135781,VP135782,VP135783,VP135784 |                   |                                    |
| CCC                      | VP135779,VP135796,VP135797                   |                   |                                    |
| Internal Standard/PEM    |                                              |                   |                                    |
| ICV/I.BLK                | VP135785                                     |                   |                                    |
| Surrogate Standard       |                                              |                   |                                    |
| MS/MSD Standard          |                                              |                   |                                    |
| LCS Standard             |                                              |                   |                                    |

M : Manual Integration

Instrument ID: MSVOA\_N

**Daily Analysis Runlog For Sequence/QC Batch ID # VN101725**

|              |               |                   |                                    |
|--------------|---------------|-------------------|------------------------------------|
| Review By    | John Carlone  | Review On         | 10/20/2025 8:16:05 AM              |
| Supervise By | Maresh Dadoda | Supervise On      | 10/20/2025 3:56:38 PM              |
| SubDirectory | VN101725      | HP Acquire Method | HP Processing Method 624N100825W.M |

| STD. NAME                                                                                          | STD REF.#         |
|----------------------------------------------------------------------------------------------------|-------------------|
| Tune/Reschk<br>Initial Calibration Stds                                                            | VP135826          |
| CCC<br>Internal Standard/PEM<br>ICV/I.BLK<br>Surrogate Standard<br>MS/MSD Standard<br>LCS Standard | VP135827,VP135828 |

| Sr# | SampleID     | ClientID               | Data File Name | Date-Time         | Comment                      | Operator | Status |
|-----|--------------|------------------------|----------------|-------------------|------------------------------|----------|--------|
| 1   | BFB          | BFB                    | VN088054.D     | 17 Oct 2025 08:56 |                              | JC\MD    | Ok     |
| 2   | VSTDCCC020   | VSTDCCC020             | VN088055.D     | 17 Oct 2025 09:29 | pH#Lot#V12668                | JC\MD    | Ok,M   |
| 3   | VN1017WBS01  | VN1017WBS01            | VN088056.D     | 17 Oct 2025 10:03 |                              | JC\MD    | Ok,M   |
| 4   | VN1017WBSD01 | VN1017WBSD01           | VN088057.D     | 17 Oct 2025 10:36 |                              | JC\MD    | Ok,M   |
| 5   | VN1017WBL01  | VN1017WBL01            | VN088058.D     | 17 Oct 2025 10:56 |                              | JC\MD    | Ok     |
| 6   | Q3361-02     | Trip blank 251015076-0 | VN088059.D     | 17 Oct 2025 11:32 | vial A pH#6.0 TB             | JC\MD    | Ok     |
| 7   | Q3362-02     | Trip blank 251015076-0 | VN088060.D     | 17 Oct 2025 11:53 | vial A pH#6.0 TB             | JC\MD    | Ok     |
| 8   | Q3362-01     | VOA 251015085-01       | VN088061.D     | 17 Oct 2025 12:14 | vial A pH#6.0 Surrogate Fail | JC\MD    | ReRun  |
| 9   | Q3361-01     | VOA 251015111-01       | VN088062.D     | 17 Oct 2025 12:34 | vial A pH#6.0                | JC\MD    | Ok,M   |
| 10  | IBLK         | IBLK                   | VN088063.D     | 17 Oct 2025 12:55 |                              | JC\MD    | Ok     |
| 11  | IBLK         | IBLK                   | VN088064.D     | 17 Oct 2025 13:16 |                              | JC\MD    | Ok     |
| 12  | Q3362-01     | VOA 251015085-01       | VN088065.D     | 17 Oct 2025 13:36 | vial B pH#6.0                | JC\MD    | Ok,M   |
| 13  | IBLK         | IBLK                   | VN088066.D     | 17 Oct 2025 14:58 |                              | JC\MD    | Ok     |
| 14  | Q3385-01     | EFF-WW                 | VN088067.D     | 17 Oct 2025 15:18 | vial A pH<2 Surrogate Fail   | JC\MD    | ReRun  |
| 15  | Q3385-02MS   | EFF-WWMS               | VN088068.D     | 17 Oct 2025 15:39 | vial A pH<2                  | JC\MD    | Ok,M   |
| 16  | Q3385-03MSD  | EFF-WWMSD              | VN088069.D     | 17 Oct 2025 16:00 | vial A pH<2                  | JC\MD    | Ok,M   |
| 17  | VSTDCCC020   | VSTDCCC020EC           | VN088070.D     | 17 Oct 2025 16:20 |                              | JC\MD    | Ok,M   |

M : Manual Integration





# SHIPPING DOCUMENTS

# Garden State Laboratories, Inc.

**Main Lab - 410 Hillside Avenue, Hillside NJ 07205 - NJDEP Lab Cert. #20044**  
**Jersey Shore Lab - 54 Main Street, Waretown NJ 08758 - NJDEP Lab Cert. #15037**

Tel. 800-273-8901/908-688-8900 Fax 908-688-8966 www.gslabs.com info@gslabs.com

## Office and Drop off Locations

North Jersey Office: 225 Sparta Avenue, Sparta, NJ 07871 Tel. 973-729-1827

West Jersey Office: 2050 Route 31 North, Glen Gardner, NJ 08826 Tel. 908-537-7414

## CLIENT INFORMATION (REPORT TO BE SENT TO)

Name: Garden State Laboratories, Inc.

Contact/Authorized by: Robert Szot

Mailing Address: 410 Hillside Ave.

Phone: 908-688-8900 EXT 129

City/State/Zip: Hillside, NJ. 07205

Email: rszot@gslabs.com

## SAMPLE INFORMATION

SAMPLE TYPE: WASTE WATER

SAMPLE LOCATION:

| Grab | Comp | SAMPLE ID              | SAMPLE COLLECTION |       |    |    | ANALYSIS REQUIRED (Print Legibly)            | CONTAINER INFORMATION |       |      |        |
|------|------|------------------------|-------------------|-------|----|----|----------------------------------------------|-----------------------|-------|------|--------|
|      |      |                        | Date              | Time  | AM | PM |                                              | No.                   | Type* | Size | Pres.* |
| X    |      | VOA 25105085-01        | 10/15/25          | 10:18 | X  |    | EPA 8260 TCL LIST + Acrolien & Acrylonitrile | 3                     | V     | 40mL | A      |
| X    |      | Trip blank 25105076-03 |                   |       |    |    | EPA 8260 TCL LIST + Acrolien & Acrylonitrile | 2                     | V     | 40mL | A      |
|      |      |                        |                   |       |    |    |                                              |                       |       |      |        |
|      |      |                        |                   |       |    |    |                                              |                       |       |      |        |
|      |      |                        |                   |       |    |    |                                              |                       |       |      |        |

\*Container Type: P = Plastic G = Glass A = Amber Glass I = Sterile Thio V = Vial Other/Specify:

\*Preservation Code: A = Non Preserved B = Sulfuric Acid C = Sodium Hydroxide D = Nitric Acid  
 E = Hydrochloric Acid F = Zinc Acetate G = Sodium Thiosulfate H = Ascorbic Acid I = Cooled Other/Specify:

☒ SUBCONTRACTED WORK

TURNAROUND TIME: ☒ Standard ☐ Rush (If RUSH REQUESTED) Rush Due by:

REPORT FORMAT: ☒ Standard Report ☐ Other/Specify:

☐ Standard Report + E2 PWS ID#:

SEND TO: Chem Tech

DATE/TIME:

METHOD OF SHIPMENT:

Deliver

## PAYMENT INFORMATION

☐ Sampling/Pick-up Fee: \$ ☐ Composite Fee: \$ ☐ Rush Fee: \$ Amount Due: \$

Payment Method: ☐ Credit Card Type: ☐ Check # ☐ Other: See Quote

Note:

VOA UNPRESERVED DUE TO EFFERVESCENCE - 3 DAY TAT PER JORDAN HEDVAT

**SAMPLE CUSTODY EXCHANGES MUST BE DOCUMENTED BELOW EACH TIME SAMPLES CHANGE POSSESSION**

**PLEASE PRINT YOUR NAME LEGIBLY, USE FULL LEGAL SIGNATURE, DATE AND TIME**

Sampled by (PRINT):

Signature:

Date/Time:

Client/Client's Representative (PRINT):

Signature:

Date/Time:

1. Received/Relinquished by (PRINT): Megan Howanich

Signature: Megan Howanich

Date/Time: 10/15/25 14:55

2. Received/Relinquished by (PRINT):

Signature: Cassanova Rand

Date/Time: 10/16/25 9:40

Q 3362

FOR SAMPLE RECEIVING USE ONLY

DATE/TIME/TEMP. REC'D AT LAB:

Page of

GSL CLIENT #

MICRO #

CHEM. #

SAMPLE REC'D BY:

☒ GSL FIELD SAMPLER/PICK-UP

☐ PICK-UP AT DROP OFF LOCATION

☐ DELIVERED BY CLIENT

2.3


### Laboratory Certification

| Certified By    | License No.      |
|-----------------|------------------|
|                 |                  |
| Connecticut     | PH-0830          |
|                 |                  |
| DOD ELAP (ANAB) | L2219            |
|                 |                  |
| Maine           | 2024021          |
|                 |                  |
| Maryland        | 296              |
|                 |                  |
| New Hampshire   | 255425           |
|                 |                  |
| New Jersey      | 20012            |
|                 |                  |
| New York        | 11376            |
|                 |                  |
| Pennsylvania    | 68-00548         |
|                 |                  |
| Soil Permit     | 525-24-234-08441 |
|                 |                  |
| Texas           | TX-C25-00189     |
|                 |                  |
| Virginia        | 460312           |

**LOGIN REPORT/SAMPLE TRANSFER**

|                                                    |        |                                                 |                                   |
|----------------------------------------------------|--------|-------------------------------------------------|-----------------------------------|
| <b>Order ID :</b> Q3362                            | GARD04 | <b>Order Date :</b> 10/16/2025 9:55:00 AM       | <b>Project Mgr :</b>              |
| <b>Client Name :</b> Garden State Laboratories, I  |        | <b>Project Name :</b> Waste Water 2025          | <b>Report Type :</b> Level 1      |
| <b>Client Contact :</b> Sharon Ercoliani           |        | <b>Receive DateTime :</b> 10/16/2025 9:40:00 AM | <b>EDD Type :</b> EXCEL NOCLEANUP |
| <b>Invoice Name :</b> Garden State Laboratories, I |        | <b>Purchase Order :</b>                         | <b>Hard Copy Date :</b>           |
| <b>Invoice Contact :</b> Sharon Ercoliani          |        |                                                 | <b>Date Signoff :</b>             |

| LAB ID   | CLIENT ID               | MATRIX | SAMPLE<br>DATE | SAMPLE<br>TIME | TEST         | TEST GROUP | METHOD   | FAX DATE | DUE<br>DATES |
|----------|-------------------------|--------|----------------|----------------|--------------|------------|----------|----------|--------------|
| Q3362-01 | VOA 251015085-01        | Water  | 10/15/2025     | 10:18          |              |            |          |          |              |
|          |                         |        |                |                | VOCMS Group1 |            | 624.1    |          | 10 Bus. Days |
|          |                         |        |                |                | VOCMS Group2 |            | 8260-Low |          | 10 Bus. Days |
| Q3362-02 | Trip blank 251015076-03 | Water  | 10/15/2025     | 10:18          |              |            |          |          |              |
|          |                         |        |                |                | VOCMS Group1 |            | 624.1    |          | 10 Bus. Days |
|          |                         |        |                |                | VOCMS Group2 |            | 8260-Low |          | 10 Bus. Days |

Relinquished By : Date / Time : 10/16/25 10:50Received By : Date / Time : 10/16/25 10:50

Pg # 4

Storage Area : VOA Refridgerator Room