

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: GARD04  
 Lab Code: ACE SDG NO.: Q3362  
 Lab File ID: VX048201.D Date Analyzed: 10/17/2025  
 Instrument ID: MSVOA\_X Time Analyzed: 09:55  
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

|                         | IS4<br>AREA # | RT #   |  |  |  |  |
|-------------------------|---------------|--------|--|--|--|--|
| 12 HOUR STD             | 103673        | 12.00€ |  |  |  |  |
| UPPER LIMIT             | 207346        | 12.50€ |  |  |  |  |
| LOWER LIMIT             | 51836.5       | 11.50€ |  |  |  |  |
| EPA SAMPLE NO.          |               |        |  |  |  |  |
| VOA 251015085-01        | 112369        | 12.01  |  |  |  |  |
| Trip blank 251015076-03 | 114293        | 12.01  |  |  |  |  |
| VX1017WBL01             | 138959        | 12.01  |  |  |  |  |
| VX1017WBS01             | 112861        | 12.01  |  |  |  |  |
| VX1017WBSD01            | 106057        | 12.00  |  |  |  |  |

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area  
 AREA LOWER LIMIT = -50% of internal standard area  
 RT UPPER LIMIT = +0.50 minutes of internal standard RT  
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

# Column used to flag values outside QC limits with an asterisk.  
 \* Values outside of QC limits.