

ALLIANCE 284 Sheffield Street, Mountainside New Jersey 07092

NEW JERSEY LAB ID#: 20012; NEW YORK LAB ID#: 11376

METALS CONFORMANCE/NON-CONFORMANCE SUMMARY

ORDER ID: Q3363

MATRIX: Solid

METHOD: 6020B,7471B

| | NA | NO | YES |
|---|----|----|-----|
| 1. Calibration Summary met criteria. | | | ✓ |
| 2. ICP Interference Check Sample Results Summary Submitted. | | | ✓ |
| 3. Serial Dilution Summary (if applicable) Submitted. | | | ✓ |
| The Serial Dilution met criteria for all compounds. | | | |
| 4. Laboratory Control Sample Summary (if applicable) Submitted. | | | ✓ |
| 5. Blank Contamination - If yes, list compounds and concentrations in each blank: | | ✓ | |
| 6. Matrix Spike/Matrix Spike Duplicate Recoveries Met Criteria | | ✓ | |
| If not met, list those compounds and their recoveries which fall outside the acceptable range. | | | |
| The Matrix Spike (COMP01-20251015MS) analysis met criteria for all compounds except for Arsenic, Calcium, Chromium, Copper, Magnesium, Manganese, Nickel, Potassium, Selenium, Silver and Zinc due to Chemical Interference during Digestion process. The Matrix Spike Duplicate (COMP01-20251015MSD) analysis met criteria for all compounds except for Arsenic, Chromium, Manganese, Silver and Zinc due to Chemical Interference during Digestion process. | | | |
| 7. Sample Duplicate Analysis Met QC Criteria | | | ✓ |
| If not met, list those compounds and their recoveries which fall outside the acceptable range. | | | |
| 8. Digestion Holding Time Met | | | ✓ |
| If not met, list number of days exceeded for each sample: | | | |
| 9. Analysis Holding Time Met | | | ✓ |
| If not met, list those compounds and their recoveries which fall outside the acceptable range. | | | |

ALLIANCE 284 Sheffield Street, Mountainside New Jersey 07092

NEW JERSEY LAB ID#: 20012: NEW YORK LAB ID#: 11376

METALS CONFORMANCE/NON-CONFORMANCE SUMMARY (CONTINUED)

NA NO YES

ADDITIONAL COMMENTS:

The Post Digest Spike (COMP01-20251015A) analysis met criteria for all compounds except for Arsenic, Calcium, Chromium, Copper, Magnesium, Manganese, Nickel, Potassium, Silver and Zinc due to unknown chemical interference of matrix with the addition of spike amount after digestion and before analysis; matrix has suppression effect during addition of spike.

Collision cell is being used to remove potential interferences. The analytes Na, Mg, Al, K, V, Cr, Mn, Fe, Co, Ni, Cu, Zn, As are being analyzed with collision cell and analytes Be, B, Ca, Ti, Se, Sr, Zr, Mo, Ag, Cd, Sn, Sb, Ba, Tl, Pb, U are being analyzed with Non-Collision Cell. Helium gas is used for the Collision Cell analysis.

Internal standard 89Y(1) was out Side qc limit for samples Q3363-04 in Original so for this sample affected parameters are reported from 5X Dilution.

Internal standard 89Y(1 and 2) was out Side qc limit for samples Q3363-05+Qcs in Original so for these samples affected parameters are reported from 5X Dilution.

QA REVIEW

Date