

Initial and Continuing Calibration Verification

Client: AECOM

SDG No.: Q3363

Project: AE1-CTY 3.2.Z-PR-DOCK-Des-ODC - 60693795-1719206

RunNo.: LB137582

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV1 TOC	mg/L	968	1000	97	90-110	07/22/2025
Sample ID: CCV1 TOC	mg/L	1030	1000	103	90-110	10/20/2025
Sample ID: CCV2 TOC	mg/L	1050	1000	105	90-110	10/20/2025
Sample ID: CCV3 TOC	mg/L	1010	1000	101	90-110	10/20/2025

Initial and Continuing Calibration Verification

Client: AECOM

SDG No.: Q3363

Project: AE1-CTY 3.2.Z-PR-DOCK-Des-ODC - 60693795-1719206

RunNo.: LB137587

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV Hexavalent Chromium	mg/L	0.494	0.5	99	90-110	10/20/2025
Sample ID: CCV1 Hexavalent Chromium	mg/L	0.495	0.5	99	90-110	10/20/2025
Sample ID: CCV2 Hexavalent Chromium	mg/L	0.494	0.5	99	90-110	10/20/2025
Sample ID: CCV3 Hexavalent Chromium	mg/L	0.490	0.5	98	90-110	10/20/2025
Sample ID: CCV4 Hexavalent Chromium	mg/L	0.492	0.5	98	90-110	10/20/2025