

Instrument ID: CV1

Daily Analysis Runlog For Sequence/QC Batch ID # LB137570

Review By	jaswal	Review On	10/17/2025 5:13:27 PM
Supervise By	Sagar	Supervise On	10/17/2025 5:17:19 PM

STD. NAME	STD REF.#
ICAL Standard	MP87618,MP87619,MP87620,MP87621,MP87622,MP87623
ICV Standard	MP87624
CCV Standard	MP87626
ICSA Standard	
CRI Standard	MP87628
LCS Standard	
Chk Standard	MP87625,MP87627,MP87629,MP87631

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	S0	S0	CAL1	10/17/25 09:54		Sagar	OK
2	S0.2	S0.2	CAL2	10/17/25 09:56		Sagar	OK
3	S2.5	S2.5	CAL3	10/17/25 09:59		Sagar	OK
4	S5	S5	CAL4	10/17/25 10:01		Sagar	OK
5	S7.5	S7.5	CAL5	10/17/25 10:04		Sagar	OK
6	S10	S10	CAL6	10/17/25 10:06		Sagar	OK
7	ICV77	ICV77	ICV	10/17/25 10:10		Sagar	OK
8	ICB77	ICB77	ICB	10/17/25 10:12		Sagar	OK
9	CCV50	CCV50	CCV	10/17/25 10:15		Sagar	OK
10	CCB50	CCB50	CCB	10/17/25 10:17		Sagar	OK
11	CRA	CRA	CRDL	10/17/25 10:20		Sagar	OK
12	HighStd	HighStd	HIGH STD	10/17/25 10:22		Sagar	OK
13	ChkStd	ChkStd	SAM	10/17/25 10:24		Sagar	OK
14	PB170143BL	PB170143BL	MB	10/17/25 10:27		Sagar	OK
15	PB170143BS	PB170143BS	LCS	10/17/25 10:30		Sagar	OK
16	Q3352-01	TP-3	SAM	10/17/25 10:32		Sagar	OK
17	Q3352-01DUP	TP-3DUP	DUP	10/17/25 10:34		Sagar	OK
18	Q3352-01MS	TP-3MS	MS	10/17/25 10:37		Sagar	OK

Instrument ID: CV1

Daily Analysis Runlog For Sequence/QC Batch ID # LB137570

Review By	jaswal	Review On	10/17/2025 5:13:27 PM
Supervise By	Sagar	Supervise On	10/17/2025 5:17:19 PM
STD. NAME	STD REF.#		
ICAL Standard	MP87618,MP87619,MP87620,MP87621,MP87622,MP87623		
ICV Standard	MP87624		
CCV Standard	MP87626		
ICSA Standard			
CRI Standard	MP87628		
LCS Standard			
Chk Standard	MP87625,MP87627,MP87629,MP87631		

19	Q3352-01MSD	TP-3MSD	MSD	10/17/25 10:39		Sagar	OK
20	Q3354-01	HT1403	SAM	10/17/25 10:42		Sagar	OK
21	CCV51	CCV51	CCV	10/17/25 10:44		Sagar	OK
22	CCB51	CCB51	CCB	10/17/25 10:47		Sagar	OK
23	Q3355-01	EO-03-101525	SAM	10/17/25 10:49		Sagar	OK
24	Q3363-01	PR132-SO1-085011-2	SAM	10/17/25 10:52	Hg High	Sagar	Dilution
25	Q3363-02	PR132-SO1-039085-2	SAM	10/17/25 10:54	Hg High	Sagar	Dilution
26	Q3363-03	PR132-SO3-028010-2	SAM	10/17/25 10:56	Hg High	Sagar	Dilution
27	Q3363-04	PR132-SO3-010013-2	SAM	10/17/25 10:59	Hg High	Sagar	Dilution
28	Q3363-05	COMP01-20251015	SAM	10/17/25 11:03	Hg High	Sagar	Dilution
29	Q3366-01	TR-06-101625	SAM	10/17/25 11:06		Sagar	OK
30	Q3366-03	TR-1-101625	SAM	10/17/25 11:09		Sagar	OK
31	Q3367-01	BU-03-101625	SAM	10/17/25 11:11		Sagar	OK
32	Q3369-01	MH-22	SAM	10/17/25 11:13		Sagar	OK
33	CCV52	CCV52	CCV	10/17/25 11:16		Sagar	OK
34	CCB52	CCB52	CCB	10/17/25 11:18		Sagar	OK
35	Q3369-05	MH-21	SAM	10/17/25 11:21		Sagar	OK
36	Q3352-01L	TP-3L	SD	10/17/25 11:23		Sagar	OK
37	Q3352-01A	TP-3A	PS	10/17/25 11:25		Sagar	OK
38	Q3363-01DL	PR132-SO1-085011-2	SAM	10/17/25 11:27	10X For Hg	Sagar	Confirms

Instrument ID: CV1

Daily Analysis Runlog For Sequence/QC Batch ID # LB137570

Review By	jaswal	Review On	10/17/2025 5:13:27 PM
Supervise By	Sagar	Supervise On	10/17/2025 5:17:19 PM
STD. NAME	STD REF.#		
ICAL Standard	MP87618,MP87619,MP87620,MP87621,MP87622,MP87623		
ICV Standard	MP87624		
CCV Standard	MP87626		
ICSA Standard			
CRI Standard	MP87628		
LCS Standard			
Chk Standard	MP87625,MP87627,MP87629,MP87631		

39	Q3363-02DL	PR132-SO1-039085-2	SAM	10/17/25 11:30	NOT USE	Sagar	Not Ok
40	Q3363-03DL	PR132-SO3-028010-2	SAM	10/17/25 11:33	10X For Hg	Sagar	Confirms
41	Q3363-04DL	PR132-SO3-010013-2	SAM	10/17/25 11:35	10X For Hg	Sagar	Confirms
42	Q3363-05DL	COMP01-20251015D	SAM	10/17/25 11:38	10X For Hg	Sagar	Confirms
43	Q3363-02DL2	PR132-SO1-039085-2	SAM	10/17/25 11:41	50X For Hg	Sagar	Confirms
44	CCV53	CCV53	CCV	10/17/25 11:43		Sagar	OK
45	CCB53	CCB53	CCB	10/17/25 11:46		Sagar	OK

Instrument ID: P7

Daily Analysis Runlog For Sequence/QC Batch ID # LB137619

Review By	Janvi	Review On	10/24/2025 4:51:23 PM
Supervise By	jaswal	Supervise On	10/24/2025 6:19:47 PM
STD. NAME	STD REF.#		
ICAL Standard	MP87545,MP87554,MP87553,MP87551,MP87550,MP87549,MP87548,MP87547,MP87546		
ICV Standard	MP87566,MP87553		
CCV Standard	MP87555		
ICSA Standard	MP87556,MP87557		
CRI Standard	MP87553		
LCS Standard			
Chk Standard	MP87559,MP87560		

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	TUNE	TUNE	TUNE	10/22/25 10:45		Jaswal	OK
2	S0	S0	CAL1	10/22/25 11:54		Jaswal	OK
3	S2	S2	CAL3	10/22/25 11:57		Jaswal	OK
4	S3	S3	CAL4	10/22/25 12:04		Jaswal	OK
5	S4	S4	CAL5	10/22/25 12:07		Jaswal	OK
6	S5	S5	CAL6	10/22/25 12:10		Jaswal	OK
7	S6	S6	CAL7	10/22/25 12:13		Jaswal	OK
8	S7	S7	CAL8	10/22/25 12:16		Jaswal	OK
9	S8	S8	CAL9	10/22/25 12:19		Jaswal	OK
10	ICV01	ICV01	ICV	10/22/25 12:38		Jaswal	OK
11	LLICV01	LLICV01	LLICV	10/22/25 13:12		Jaswal	OK
12	ICB01	ICB01	ICB	10/22/25 13:16		Jaswal	OK
13	ICSA01	ICSA01	ICSA	10/22/25 13:20		Jaswal	OK
14	ICSAB01	ICSAB01	ICSAB	10/22/25 13:23		Jaswal	OK
15	CCV01	CCV01	CCV	10/22/25 13:27		Jaswal	OK
16	CCB01	CCB01	CCB	10/22/25 13:30		Jaswal	OK
17	CRI01	CRI01	CRDL	10/22/25 13:33		Jaswal	OK
18	PB170191BS	PB170191BS	LCS	10/22/25 13:40		Jaswal	OK

Instrument ID: P7

Daily Analysis Runlog For Sequence/QC Batch ID # LB137619

Review By	Janvi	Review On	10/24/2025 4:51:23 PM
Supervise By	jaswal	Supervise On	10/24/2025 6:19:47 PM
STD. NAME	STD REF.#		
ICAL Standard	MP87545,MP87554,MP87553,MP87551,MP87550,MP87549,MP87548,MP87547,MP87546		
ICV Standard	MP87566,MP87553		
CCV Standard	MP87555		
ICSA Standard	MP87556,MP87557		
CRI Standard	MP87553		
LCS Standard			
Chk Standard	MP87559,MP87560		

19	Q3386-03	3443	SAM	10/22/25 13:43		Jaswal	OK
20	Q3386-04	3444	SAM	10/22/25 13:46		Jaswal	OK
21	Q3386-04DUP	3444DUP	DUP	10/22/25 13:49		Jaswal	OK
22	Q3386-04L	3444L	SD	10/22/25 13:53		Jaswal	OK
23	Q3386-04MS	3444MS	MS	10/22/25 13:56		Jaswal	OK
24	Q3386-04MSD	3444MSD	MSD	10/22/25 13:59		Jaswal	OK
25	Q3386-04A	3444A	PS	10/22/25 14:02	0.1ml mp87570,0.2ml each mp87575,mp87578,mp875 81-10ml sample	Jaswal	OK
26	PB170191BL	PB170191BL	MB	10/22/25 14:08		Jaswal	OK
27	CCV02	CCV02	CCV	10/22/25 14:11		Jaswal	OK
28	CCB02	CCB02	CCB	10/22/25 14:18		Jaswal	OK
29	Q3363-01	PR132-SO1-085011-2	SAM	10/22/25 14:21		Jaswal	OK
30	Q3363-02	PR132-SO1-039085-2	SAM	10/22/25 14:24	Cr High	Jaswal	Dilution
31	Q3363-03	PR132-SO3-028010-2	SAM	10/22/25 14:27	Cr High	Jaswal	Dilution
32	Q3363-04	PR132-SO3-010013-2	SAM	10/22/25 14:30	INT_STD 89Y(1) Fail	Jaswal	Dilution
33	Q3363-05	COMP01-20251015	SAM	10/22/25 14:34	Al,Fe High, INT_STD 89Y(1&2) Fail	Jaswal	Dilution
34	Q3363-05DUP	COMP01-20251015D	DUP	10/22/25 14:37	Al,Fe High, INT_STD 89Y(1&2) Fail	Jaswal	Dilution
35	Q3363-05L	COMP01-20251015L	SD	10/22/25 14:40	Al,Fe High, INT_STD 89Y(1&2) Fail	Jaswal	Dilution
36	Q3363-05MS	COMP01-20251015M	MS	10/22/25 14:43	Al,Fe,Pb High, INT_STD 89Y(1&2) Fail	Jaswal	Dilution

Instrument ID: P7

Daily Analysis Runlog For Sequence/QC Batch ID # LB137619

Review By	Janvi	Review On	10/24/2025 4:51:23 PM
Supervise By	jaswal	Supervise On	10/24/2025 6:19:47 PM

STD. NAME	STD REF.#
ICAL Standard	MP87545,MP87554,MP87553,MP87551,MP87550,MP87549,MP87548,MP87547,MP87546
ICV Standard	MP87566,MP87553
CCV Standard	MP87555
ICSA Standard	MP87556,MP87557
CRI Standard	MP87553
LCS Standard	
Chk Standard	MP87559,MP87560

37	Q3363-05MSD	COMP01-20251015M	MSD	10/22/25 14:46	Al,Fe,Pb High, INT_STD 89Y(1&2) Fail	Jaswal	Dilution
38	Q3363-05A	COMP01-20251015A	PS	10/22/25 14:49	INT_STD 89Y(1&2) Fail(0.1ml mp87570,0.2ml each mp87575,mp87578,mp875 81-10ml sample)	Jaswal	Dilution
39	CCV03	CCV03	CCV	10/22/25 14:54		Jaswal	OK
40	CCB03	CCB03	CCB	10/22/25 15:05		Jaswal	OK
41	PB170139BL	PB170139BL	MB	10/22/25 15:14		Jaswal	OK
42	PB170139BS	PB170139BS	LCS	10/22/25 15:17		Jaswal	OK
43	PB170158BL	PB170158BL	MB	10/22/25 15:20		Jaswal	OK
44	PB170158BS	PB170158BS	LCS	10/22/25 15:24		Jaswal	OK
45	PB170186BL	PB170186BL	MB	10/22/25 15:27		Jaswal	OK
46	PB170186BS	PB170186BS	LCS	10/22/25 15:30		Jaswal	OK
47	Q3375-01	PR132-S02-074099-2	SAM	10/22/25 15:33		Jaswal	OK
48	Q3375-02	PR132-S04-000013-2	SAM	10/22/25 15:36		Jaswal	OK
49	Q3375-03	PR132-S05-000057-2	SAM	10/22/25 15:39	Cr High	Jaswal	Dilution
50	Q3375-04	PR132-S02-010074-2	SAM	10/22/25 15:42		Jaswal	OK
51	CCV04	CCV04	CCV	10/22/25 15:45		Jaswal	OK
52	CCB04	CCB04	CCB	10/22/25 16:00		Jaswal	OK
53	Q3375-05	PR132-DP02-202510	SAM	10/22/25 16:03	Cr,Fe,Pb High	Jaswal	Dilution
54	Q3376-01	PR132-S05-057082-2	SAM	10/22/25 16:06	Cr High	Jaswal	Dilution

Instrument ID: P7

Daily Analysis Runlog For Sequence/QC Batch ID # LB137619

Review By	Janvi	Review On	10/24/2025 4:51:23 PM
Supervise By	jaswal	Supervise On	10/24/2025 6:19:47 PM

STD. NAME	STD REF.#
ICAL Standard	MP87545,MP87554,MP87553,MP87551,MP87550,MP87549,MP87548,MP87547,MP87546
ICV Standard	MP87566,MP87553
CCV Standard	MP87555
ICSA Standard	MP87556,MP87557
CRI Standard	MP87553
LCS Standard	
Chk Standard	MP87559,MP87560

55	Q3376-02	PR132-DP01-202510	SAM	10/22/25 16:09	Cr High	Jaswal	Dilution
56	Q3376-03	PR132-COMP2-2025	SAM	10/22/25 16:13	Cr High	Jaswal	Dilution
57	Q3376-04	PR132-S12-047072-2	SAM	10/22/25 16:16	Cr High	Jaswal	Dilution
58	Q3376-05	PR132-S12-020047-2	SAM	10/22/25 16:19		Jaswal	OK
59	Q3376-05L	PR132-S12-020047-2	SD	10/22/25 16:25		Jaswal	OK
60	Q3376-05MSD	PR132-S12-020047-2	MSD	10/22/25 16:31		Jaswal	OK
61	Q3376-05DUP	PR132-S12-020047-2	DUP	10/22/25 16:40		Jaswal	OK
62	Q3376-05MS	PR132-S12-020047-2	MS	10/22/25 16:43		Jaswal	OK
63	CCV05	CCV05	CCV	10/22/25 16:46		Jaswal	OK
64	CCB05	CCB05	CCB	10/22/25 16:55		Jaswal	OK
65	Q3376-05A	PR132-S12-020047-2	PS	10/22/25 17:00	0.1ml mp87570,0.2ml each mp87575,mp87578,mp87581-10ml sample	Jaswal	OK
66	Q3393-01	PR132-S10-000022-2	SAM	10/22/25 17:15	Cr High	Jaswal	Dilution
67	Q3393-02	PR132-S10-022047-2	SAM	10/22/25 17:18	Cr High	Jaswal	Dilution
68	Q3393-03	PR132-S08-013038-2	SAM	10/22/25 17:22	Al,Fe High, INT_STD 89Y(1&2) Fail	Jaswal	Dilution
69	Q3393-04	PR132-S06-000021-2	SAM	10/22/25 17:25	Cr High	Jaswal	Dilution
70	Q3393-05	PR132-S08-000013-2	SAM	10/22/25 17:28	Cr,Fe High, INT_STD 89Y(2) Fail	Jaswal	Dilution
71	Q3395-01	PR132-S09-000051-2	SAM	10/22/25 17:31	Cr High	Jaswal	Dilution
72	Q3395-02	PR132-S07-000102-2	SAM	10/22/25 17:34	Cr High	Jaswal	Dilution

Instrument ID: P7

Daily Analysis Runlog For Sequence/QC Batch ID # LB137619

Review By	Janvi	Review On	10/24/2025 4:51:23 PM
Supervise By	jaswal	Supervise On	10/24/2025 6:19:47 PM
STD. NAME	STD REF.#		
ICAL Standard	MP87545,MP87554,MP87553,MP87551,MP87550,MP87549,MP87548,MP87547,MP87546		
ICV Standard	MP87566,MP87553		
CCV Standard	MP87555		
ICSA Standard	MP87556,MP87557		
CRI Standard	MP87553		
LCS Standard			
Chk Standard	MP87559,MP87560		

73	Q3395-02DUP	PR132-S07-000102-2	DUP	10/22/25 17:41	Cr High	Jaswal	Dilution
74	Q3395-02L	PR132-S07-000102-2	SD	10/22/25 17:44		Jaswal	OK
75	CCV06	CCV06	CCV	10/22/25 17:59		Jaswal	OK
76	CCB06	CCB06	CCB	10/22/25 18:10		Jaswal	OK
77	Q3395-03	PR132-S07-000102-2	MS	10/22/25 18:15	Cr,Fe,Pb,Ni High	Jaswal	Dilution
78	Q3395-04	PR132-S07-000102-2	MSD	10/22/25 18:18	Cr,Fe,Pb,Ni High	Jaswal	Dilution
79	Q3395-02A	PR132-S07-000102-2	PS	10/22/25 18:21	Pb High(0.1ml mp87570,0.2ml each mp87575,mp87578,mp875 81-10ml sample	Jaswal	Dilution
80	Q3395-05	PR132-S07-102116-2	SAM	10/22/25 18:23	Al,Cr,Fe High	Jaswal	Dilution
81	Q3395-06	PR132-COMP03-2025	SAM	10/22/25 18:26		Jaswal	OK
82	Q3395-07	PR132-S08-038048-2	SAM	10/22/25 18:29	Fe High	Jaswal	Dilution
83	CCV07	CCV07	CCV	10/22/25 18:41		Jaswal	OK
84	CCB07	CCB07	CCB	10/22/25 18:49		Jaswal	OK
85	Q3363-02DL	PR132-SO1-039085-2	SAM	10/22/25 19:12	5X For Cr	Jaswal	Confirms
86	Q3363-03DL	PR132-SO3-028010-2	SAM	10/22/25 19:15	5X For Cr	Jaswal	Confirms
87	Q3363-04DL	PR132-SO3-010013-2	SAM	10/22/25 19:18	5X For INT_STD 89Y(1)	Jaswal	Confirms
88	Q3363-05DL	COMP01-20251015D	SAM	10/22/25 19:22	5X For INT_STD 89Y(1&2),5X For Al,Fe	Jaswal	Confirms
89	Q3363-05DUPDL	COMP01-20251015D	DUP	10/22/25 19:25	5X For INT_STD 89Y(1&2),5X For Al,Fe	Jaswal	Confirms
90	Q3363-05LDL	COMP01-20251015L	SD	10/22/25 19:28	25X For INT_STD 89Y(1&2),5X For Al,Fe	Jaswal	Confirms

Instrument ID: P7

Daily Analysis Runlog For Sequence/QC Batch ID # LB137619

Review By	Janvi	Review On	10/24/2025 4:51:23 PM
Supervise By	jaswal	Supervise On	10/24/2025 6:19:47 PM
STD. NAME	STD REF.#		
ICAL Standard	MP87545,MP87554,MP87553,MP87551,MP87550,MP87549,MP87548,MP87547,MP87546		
ICV Standard	MP87566,MP87553		
CCV Standard	MP87555		
ICSA Standard	MP87556,MP87557		
CRI Standard	MP87553		
LCS Standard			
Chk Standard	MP87559,MP87560		

91	Q3363-05MSDL	COMP01-20251015M	MS	10/22/25 19:31	5X For INT_STD 89Y(1&2),5X For Al,Fe,Pb	Jaswal	Confirms
92	Q3363-05MSDDL	COMP01-20251015M	MSD	10/22/25 19:34	5X For INT_STD 89Y(1&2),5X For Al,Fe,Pb	Jaswal	Confirms
93	Q3363-05ADL	COMP01-20251015A	PS	10/22/25 19:37	5X For INT_STD 89Y(1&2)	Jaswal	Confirms
94	Q3375-03DL	PR132-S05-000057-2	SAM	10/22/25 19:41	5X For Cr	Jaswal	Confirms
95	CCV08	CCV08	CCV	10/22/25 19:44		Jaswal	OK
96	CCB08	CCB08	CCB	10/22/25 19:52		Jaswal	OK
97	Q3375-05DL	PR132-DP02-202510	SAM	10/22/25 19:57	5X For Cr,Fe,Pb	Jaswal	Confirms
98	Q3376-01DL	PR132-S05-057082-2	SAM	10/22/25 20:00	5X For Cr	Jaswal	Confirms
99	Q3376-02DL	PR132-DP01-202510	SAM	10/22/25 20:03	5X For Cr	Jaswal	Confirms
100	Q3376-03DL	PR132-COMP2-2025	SAM	10/22/25 20:06	5X For Cr	Jaswal	Confirms
101	Q3376-04DL	PR132-S12-047072-2	SAM	10/22/25 20:09	5X For Cr	Jaswal	Confirms
102	Q3393-01DL	PR132-S10-000022-2	SAM	10/22/25 20:12	5X For Cr	Jaswal	Confirms
103	Q3393-02DL	PR132-S10-022047-2	SAM	10/22/25 20:16	5X For Cr	Jaswal	Confirms
104	Q3393-03DL	PR132-S08-013038-2	SAM	10/22/25 20:19	5X For Al,Fe, 5X For INT_STD 89Y(1&2)	Jaswal	Confirms
105	Q3393-04DL	PR132-S06-000021-2	SAM	10/22/25 20:22	5X For Cr	Jaswal	Confirms
106	Q3393-05DL	PR132-S08-000013-2	SAM	10/22/25 20:25	5X For Cr,Fe , 5X For INT_STD 89Y(2)	Jaswal	Confirms
107	CCV09	CCV09	CCV	10/22/25 20:28		Jaswal	OK
108	CCB09	CCB09	CCB	10/22/25 20:31		Jaswal	OK
109	Q3395-01DL	PR132-S09-000051-2	SAM	10/22/25 20:35	5X For Cr	Jaswal	Confirms

Instrument ID: P7

Daily Analysis Runlog For Sequence/QC Batch ID # LB137619

Review By	Janvi	Review On	10/24/2025 4:51:23 PM
Supervise By	jaswal	Supervise On	10/24/2025 6:19:47 PM
STD. NAME	STD REF.#		
ICAL Standard	MP87545,MP87554,MP87553,MP87551,MP87550,MP87549,MP87548,MP87547,MP87546		
ICV Standard	MP87566,MP87553		
CCV Standard	MP87555		
ICSA Standard	MP87556,MP87557		
CRI Standard	MP87553		
LCS Standard			
Chk Standard	MP87559,MP87560		

110	Q3395-02DL	PR132-S07-000102-2	SAM	10/22/25 20:38	5X For Cr	Jaswal	Confirms
111	Q3395-02DUPDL	PR132-S07-000102-2	DUP	10/22/25 20:41	5X For Cr	Jaswal	Confirms
112	Q3395-02LDL	PR132-S07-000102-2	SD	10/22/25 20:44	Not Required	Jaswal	Not Ok
113	Q3395-03DL	PR132-S07-000102-2	MS	10/22/25 20:48	5X For Cr,Pb,Fe,Ni	Jaswal	Confirms
114	Q3395-04DL	PR132-S07-000102-2	MSD	10/22/25 20:51	5X For Cr,Pb,Fe,Ni	Jaswal	Confirms
115	Q3395-02ADL	PR132-S07-000102-2	PS	10/22/25 20:54	5X For Pb,Ni	Jaswal	Confirms
116	Q3395-05DL	PR132-S07-102116-2	SAM	10/22/25 20:57	5X For Al,Cr,Fe	Jaswal	Confirms
117	Q3395-07DL	PR132-S08-038048-2	SAM	10/22/25 21:00	5X For Fe	Jaswal	Confirms
118	CCV10	CCV10	CCV	10/22/25 21:03		Jaswal	OK
119	CCB10	CCB10	CCB	10/22/25 21:06		Jaswal	OK