



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

IDENTIFICATION SUMMARY
FOR MULTICOMPONENT ANALYTES

SAMPLE NO.

PB170141BS

Lab Name: Alliance

Contract: AECO02

Lab Code: ACE

SDG NO.: Q3363

Lab Sample ID: PB170141BS

Date(s) Analyzed: 10/17/2025 10/17/2025

Instrument ID (1): ECD_O

Instrument ID (2): ECD_O

GC Column: (1): ZB-MR1 ID: 0.32 (mm)

GC Column: (2): ZB-MR2 ID: 0.32 (mm)

Data file P0114454.D

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	MEAN CONCENTRATION	%RPD
			FROM	TO			
Aroclor-1016	1	5.493	5.443	5.543	167	165	1.2
COLUMN 1	2	5.515	5.465	5.565	164		
	3	5.576	5.526	5.626	160		
	4	5.672	5.622	5.722	171		
	5	5.964	5.914	6.014	165		
	1	4.739	4.689	4.789	164		
COLUMN 2	2	4.758	4.708	4.808	164	167	
	3	4.933	4.883	4.983	171		
	4	4.975	4.925	5.025	169		
	5	5.188	5.138	5.238	164		
Aroclor-1260	1	7.077	7.027	7.127	170	153	9.94
COLUMN 1	2	7.331	7.281	7.381	166		
	3	7.688	7.638	7.738	141		
	4	7.91	7.86	7.96	148		
	5	8.217	8.167	8.267	141		
	1	6.221	6.171	6.271	181		
COLUMN 2	2	6.408	6.358	6.458	186	169	
	3	6.561	6.511	6.611	168		
	4	7.033	6.983	7.083	159		
	5	7.273	7.223	7.323	153		

IDENTIFICATION SUMMARY
FOR MULTICOMPONENT ANALYTES

SAMPLE NO.

PR132-SO3-028010-202

Lab Name: Alliance

Contract: AECO02

Lab Code: ACE

SDG NO.: Q3363

Lab Sample ID: Q3363-03

Date(s) Analyzed: 10/17/2025 10/17/2025

Instrument ID (1): ECD_O

Instrument ID (2): ECD_O

GC Column: (1): ZB-MR1 ID: 0.32 (mm)

GC Column: (2): ZB-MR2 ID: 0.32 (mm)

Data file PO114457.D

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	MEAN CONCENTRATION	%RPD
			FROM	TO			
Aroclor-1242	1	5.494	5.444	5.544	46.8	66.4	21.33
COLUMN 1	2	5.516	5.466	5.566	55.2		
	3	5.578	5.528	5.628	46.0		
	4	5.671	5.621	5.721	105		
	5	6.399	6.349	6.449	78.4		
	1	4.739	4.689	4.789	34.4		
COLUMN 2	2	4.759	4.709	4.809	39.8	53.6	
	3	4.935	4.885	4.985	39.2		
	4	5.018	4.968	5.068	52.5		
	5	5.542	5.492	5.592	102		
Aroclor-1254	1	6.335	6.285	6.385	35.1	30.3	13.01
COLUMN 1	2	6.552	6.502	6.602	31.9		
	3	6.918	6.868	6.968	46.6		
	4	7.194	7.144	7.244	18.2		
	5	7.608	7.558	7.658	19.7		
	1	5.542	5.492	5.592	45.9		
COLUMN 2	2	5.686	5.636	5.736	33.1	26.6	
	3	6.09	6.04	6.14	18.3		
	4	6.317	6.267	6.367	17.3		
	5	6.734	6.684	6.784	18.5		

IDENTIFICATION SUMMARY
FOR MULTICOMPONENT ANALYTES

SAMPLE NO.

MH-22MS

Lab Name: Alliance

Contract: AECO02

Lab Code: ACE

SDG NO.: Q3363

Lab Sample ID: Q3369-01MS

Date(s) Analyzed: 10/17/2025 10/17/2025

Instrument ID (1): ECD_O

Instrument ID (2): ECD_O

GC Column: (1): ZB-MR1 ID: 0.32 (mm)

GC Column: (2): ZB-MR2 ID: 0.32 (mm)

Data file PO114459.D

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	MEAN CONCENTRATION	%RPD
			FROM	TO			
Aroclor-1016	1	5.494	5.444	5.544	148	165	7.02
COLUMN 1	2	5.515	5.465	5.565	144		
	3	5.578	5.528	5.628	148		
	4	5.671	5.621	5.721	210		
	5	5.965	5.915	6.015	177		
	1	4.74	4.69	4.79	154		
COLUMN 2	2	4.76	4.71	4.81	221	177	
	3	4.934	4.884	4.984	173		
	4	4.977	4.927	5.027	161		
	5	5.189	5.139	5.239	177		
Aroclor-1260	1	7.078	7.028	7.128	188	160	1.89
COLUMN 1	2	7.331	7.281	7.381	197		
	3	7.689	7.639	7.739	134		
	4	7.91	7.86	7.96	147		
	5	8.218	8.168	8.268	132		
	1	6.222	6.172	6.272	197	157	
COLUMN 2	2	6.409	6.359	6.459	168		
	3	6.562	6.512	6.612	156		
	4	7.035	6.985	7.085	135		
	5	7.275	7.225	7.325	127		

IDENTIFICATION SUMMARY
FOR MULTICOMPONENT ANALYTES

SAMPLE NO.

MH-22MSD

Lab Name: Alliance

Contract: AECO02

Lab Code: ACE

SDG NO.: Q3363

Lab Sample ID: Q3369-01MSD

Date(s) Analyzed: 10/17/2025 10/17/2025

Instrument ID (1): ECD_O

Instrument ID (2): ECD_O

GC Column: (1): ZB-MR1 ID: 0.32 (mm)

GC Column: (2): ZB-MR2 ID: 0.32 (mm)

Data file PO114460.D

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	MEAN CONCENTRATION	%RPD
			FROM	TO			
Aroclor-1016	1	5.494	5.444	5.544	149	153	11.11
COLUMN 1	2	5.515	5.465	5.565	137		
	3	5.578	5.528	5.628	142		
	4	5.672	5.622	5.722	183		
	5	5.964	5.914	6.014	155		
	1	4.74	4.69	4.79	149		
COLUMN 2	2	4.761	4.711	4.811	213	171	
	3	4.934	4.884	4.984	167		
	4	4.977	4.927	5.027	153		
	5	5.189	5.139	5.239	172		
Aroclor-1260	1	7.078	7.028	7.128	184	156	4.59
COLUMN 1	2	7.331	7.281	7.381	192		
	3	7.689	7.639	7.739	130		
	4	7.91	7.86	7.96	147		
	5	8.218	8.168	8.268	129		
	1	6.222	6.172	6.272	185		
COLUMN 2	2	6.409	6.359	6.459	159	149	
	3	6.562	6.512	6.612	149		
	4	7.034	6.984	7.084	130		
	5	7.275	7.225	7.325	123		

IDENTIFICATION SUMMARY
FOR MULTICOMPONENT ANALYTES

SAMPLE NO.

PR132-SO1-039085-202

Lab Name: Alliance

Contract: AECO02

Lab Code: ACE

SDG NO.: Q3363

Lab Sample ID: Q3363-02

Date(s) Analyzed: 10/17/2025 10/17/2025

Instrument ID (1): ECD_P

Instrument ID (2): ECD_P

GC Column: (1): ZB-MR1 ID: 0.32 (mm)

GC Column: (2): ZB-MR2 ID: 0.32 (mm)

Data file PP075861.D

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	MEAN CONCENTRATION	%RPD
			FROM	TO			
Aroclor-1242	1	5.805	5.755	5.855	122	161	19.8
COLUMN 1	2	5.826	5.776	5.876	147		
	3	5.888	5.838	5.938	135		
	4	5.983	5.933	6.033	189		
	5	6.716	6.666	6.766	211		
COLUMN 2	1	4.887	4.837	4.937	90.4	132	
	2	4.941	4.891	4.991	50.7		
	3	5.067	5.017	5.117	91.0		
	4	5.144	5.094	5.194	187		
	5	5.669	5.619	5.719	238		
Aroclor-1254	1	6.652	6.602	6.702	115	109	
COLUMN 1	2	6.868	6.818	6.918	144		
	3	7.235	7.185	7.285	132		
	4	7.512	7.462	7.562	72.8		
	5	7.928	7.878	7.978	80.1		
COLUMN 2	1	5.669	5.619	5.719	93.1	84.5	
	2	5.814	5.764	5.864	142		
	3	6.229	6.179	6.279	106		
	4	6.443	6.393	6.493	45.9		
	5	6.857	6.807	6.907	35.6		