

### CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: AECO02

Lab Code: ACE

SDG NO.: Q3363

Continuing Calib Date: 10/17/2025

Initial Calibration Date(s): 10/09/2025

10/09/2025

Continuing Calib Time: 15:50

Initial Calibration Time(s): 14:14

22:29

GC Column: ZB-MR1 ID: 0.32 (mm)

| COMPOUND             |     | CCAL<br>RT | AVG<br>RT | RT WINDOW |       | DIFF<br>RT |
|----------------------|-----|------------|-----------|-----------|-------|------------|
|                      |     |            |           | FROM      | TO    |            |
| Aroclor-1016-1       | (1) | 5.50       | 5.49      | 5.39      | 5.59  | 0.00       |
| Aroclor-1016-2       | (2) | 5.52       | 5.51      | 5.41      | 5.61  | -0.01      |
| Aroclor-1016-3       | (3) | 5.58       | 5.58      | 5.48      | 5.68  | 0.00       |
| Aroclor-1016-4       | (4) | 5.67       | 5.67      | 5.57      | 5.77  | 0.00       |
| Aroclor-1016-5       | (5) | 5.97       | 5.96      | 5.86      | 6.06  | -0.01      |
| Aroclor-1260-1       | (1) | 7.08       | 7.08      | 6.98      | 7.18  | 0.00       |
| Aroclor-1260-2       | (2) | 7.33       | 7.33      | 7.23      | 7.43  | 0.00       |
| Aroclor-1260-3       | (3) | 7.69       | 7.69      | 7.59      | 7.79  | 0.00       |
| Aroclor-1260-4       | (4) | 7.91       | 7.91      | 7.81      | 8.01  | 0.00       |
| Aroclor-1260-5       | (5) | 8.22       | 8.21      | 8.11      | 8.31  | -0.01      |
| Tetrachloro-m-xylene |     | 4.36       | 4.36      | 4.26      | 4.46  | 0.00       |
| Decachlorobiphenyl   |     | 9.94       | 9.93      | 9.83      | 10.03 | -0.01      |

### CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: AECO02

Lab Code: ACE

SDG NO.: Q3363

Continuing Calib Date: 10/17/2025

Initial Calibration Date(s): 10/09/2025

10/09/2025

Continuing Calib Time: 15:50

Initial Calibration Time(s): 14:14

22:29

GC Column: ZB-MR2 ID: 0.32 (mm)

| COMPOUND             |     | CCAL<br>RT | AVG<br>RT | RT WINDOW |      | DIFF<br>RT |
|----------------------|-----|------------|-----------|-----------|------|------------|
|                      |     |            |           | FROM      | TO   |            |
| Aroclor-1016-1       | (1) | 4.74       | 4.74      | 4.64      | 4.84 | 0.00       |
| Aroclor-1016-2       | (2) | 4.76       | 4.76      | 4.66      | 4.86 | 0.00       |
| Aroclor-1016-3       | (3) | 4.93       | 4.94      | 4.84      | 5.04 | 0.01       |
| Aroclor-1016-4       | (4) | 4.98       | 4.98      | 4.88      | 5.08 | 0.00       |
| Aroclor-1016-5       | (5) | 5.19       | 5.19      | 5.09      | 5.29 | 0.00       |
| Aroclor-1260-1       | (1) | 6.22       | 6.23      | 6.13      | 6.33 | 0.01       |
| Aroclor-1260-2       | (2) | 6.41       | 6.41      | 6.31      | 6.51 | 0.00       |
| Aroclor-1260-3       | (3) | 6.56       | 6.57      | 6.47      | 6.67 | 0.01       |
| Aroclor-1260-4       | (4) | 7.04       | 7.04      | 6.94      | 7.14 | 0.01       |
| Aroclor-1260-5       | (5) | 7.28       | 7.28      | 7.18      | 7.38 | 0.00       |
| Tetrachloro-m-xylene |     | 3.66       | 3.66      | 3.56      | 3.76 | 0.00       |
| Decachlorobiphenyl   |     | 8.67       | 8.67      | 8.57      | 8.77 | 0.01       |