

### CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: AECO02

Lab Code: ACE

SDG NO.: Q3363

Continuing Calib Date: 10/17/2025

Initial Calibration Date(s): 09/23/2025

09/23/2025

Continuing Calib Time: 16:18

Initial Calibration Time(s): 08:45

17:14

GC Column: ZB-MR1 ID: 0.32 (mm)

| COMPOUND             |     | CCAL<br>RT | AVG<br>RT | RT WINDOW |       | DIFF<br>RT |
|----------------------|-----|------------|-----------|-----------|-------|------------|
|                      |     |            |           | FROM      | TO    |            |
| Aroclor-1016-1       | (1) | 5.80       | 5.81      | 5.71      | 5.91  | 0.01       |
| Aroclor-1016-2       | (2) | 5.83       | 5.83      | 5.73      | 5.93  | 0.00       |
| Aroclor-1016-3       | (3) | 5.89       | 5.89      | 5.79      | 5.99  | 0.00       |
| Aroclor-1016-4       | (4) | 5.99       | 5.99      | 5.89      | 6.09  | 0.00       |
| Aroclor-1016-5       | (5) | 6.28       | 6.28      | 6.18      | 6.38  | 0.00       |
| Aroclor-1260-1       | (1) | 7.40       | 7.40      | 7.30      | 7.50  | 0.00       |
| Aroclor-1260-2       | (2) | 7.65       | 7.65      | 7.55      | 7.75  | 0.00       |
| Aroclor-1260-3       | (3) | 8.01       | 8.01      | 7.91      | 8.11  | 0.00       |
| Aroclor-1260-4       | (4) | 8.23       | 8.24      | 8.14      | 8.34  | 0.01       |
| Aroclor-1260-5       | (5) | 8.56       | 8.56      | 8.46      | 8.66  | 0.00       |
| Tetrachloro-m-xylene |     | 4.65       | 4.66      | 4.56      | 4.76  | 0.01       |
| Decachlorobiphenyl   |     | 10.43      | 10.43     | 10.33     | 10.53 | 0.00       |

### CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: AECO02

Lab Code: ACE

SDG NO.: Q3363

Continuing Calib Date: 10/17/2025

Initial Calibration Date(s): 09/23/2025

09/23/2025

Continuing Calib Time: 16:18

Initial Calibration Time(s): 08:45

17:14

GC Column: ZB-MR2 ID: 0.32 (mm)

| COMPOUND             |     | CCAL<br>RT | AVG<br>RT | RT WINDOW<br>FROM   TO |      | DIFF<br>RT |
|----------------------|-----|------------|-----------|------------------------|------|------------|
| Aroclor-1016-1       | (1) | 4.89       | 4.89      | 4.79                   | 4.99 | 0.00       |
| Aroclor-1016-2       | (2) | 4.94       | 4.95      | 4.85                   | 5.05 | 0.01       |
| Aroclor-1016-3       | (3) | 5.06       | 5.07      | 4.97                   | 5.17 | 0.01       |
| Aroclor-1016-4       | (4) | 5.10       | 5.11      | 5.01                   | 5.21 | 0.01       |
| Aroclor-1016-5       | (5) | 5.32       | 5.33      | 5.23                   | 5.43 | 0.01       |
| Aroclor-1260-1       | (1) | 6.35       | 6.36      | 6.26                   | 6.46 | 0.01       |
| Aroclor-1260-2       | (2) | 6.53       | 6.54      | 6.44                   | 6.64 | 0.01       |
| Aroclor-1260-3       | (3) | 6.69       | 6.70      | 6.60                   | 6.80 | 0.01       |
| Aroclor-1260-4       | (4) | 7.16       | 7.17      | 7.07                   | 7.27 | 0.01       |
| Aroclor-1260-5       | (5) | 7.40       | 7.40      | 7.30                   | 7.50 | 0.01       |
| Tetrachloro-m-xylene |     | 3.79       | 3.80      | 3.70                   | 3.90 | 0.01       |
| Decachlorobiphenyl   |     | 8.79       | 8.80      | 8.70                   | 8.90 | 0.01       |