

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** AECO02
Lab Code: ACE **SDG NO.:** Q3363
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/09/2025 10/09/2025

Client Sample No.: CCAL01 **Date Analyzed:** 10/17/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO114448.D **Time Analyzed:** 09:54

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.493	5.392	5.592	531.760	500.000	6.4
Aroclor-1016-2	5.515	5.414	5.614	514.370	500.000	2.9
Aroclor-1016-3	5.577	5.476	5.676	511.540	500.000	2.3
Aroclor-1016-4	5.672	5.571	5.771	539.820	500.000	8.0
Aroclor-1016-5	5.964	5.863	6.063	539.430	500.000	7.9
Aroclor-1260-1	7.077	6.975	7.175	515.840	500.000	3.2
Aroclor-1260-2	7.332	7.229	7.429	505.990	500.000	1.2
Aroclor-1260-3	7.688	7.585	7.785	521.640	500.000	4.3
Aroclor-1260-4	7.910	7.807	8.007	510.820	500.000	2.2
Aroclor-1260-5	8.217	8.113	8.313	499.190	500.000	-0.2
Decachlorobiphenyl	9.941	9.834	10.034	50.690	50.000	1.4
Tetrachloro-m-xylene	4.355	4.255	4.455	56.020	50.000	12.0

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** AECO02
Lab Code: ACE **SDG NO.:** Q3363
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/09/2025 10/09/2025

Client Sample No.: CCAL01 **Date Analyzed:** 10/17/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO114448.D **Time Analyzed:** 09:54

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.740	4.643	4.843	536.390	500.000	7.3
Aroclor-1016-2	4.759	4.662	4.862	525.960	500.000	5.2
Aroclor-1016-3	4.934	4.836	5.036	555.790	500.000	11.2
Aroclor-1016-4	4.976	4.879	5.079	564.050	500.000	12.8
Aroclor-1016-5	5.189	5.092	5.292	540.980	500.000	8.2
Aroclor-1260-1	6.222	6.125	6.325	556.010	500.000	11.2
Aroclor-1260-2	6.409	6.313	6.513	569.040	500.000	13.8
Aroclor-1260-3	6.562	6.466	6.666	519.230	500.000	3.8
Aroclor-1260-4	7.034	6.939	7.139	561.290	500.000	12.3
Aroclor-1260-5	7.274	7.179	7.379	532.690	500.000	6.5
Decachlorobiphenyl	8.664	8.571	8.771	47.610	50.000	-4.8
Tetrachloro-m-xylene	3.658	3.560	3.760	53.810	50.000	7.6