

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** AECO02
Lab Code: ACE **SDG NO.:** Q3363
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/09/2025 10/09/2025

Client Sample No.: CCAL02 **Date Analyzed:** 10/17/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO114461.D **Time Analyzed:** 15:50

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.495	5.392	5.592	509.020	500.000	1.8
Aroclor-1016-2	5.516	5.414	5.614	497.030	500.000	-0.6
Aroclor-1016-3	5.578	5.476	5.676	490.250	500.000	-2.0
Aroclor-1016-4	5.674	5.571	5.771	520.520	500.000	4.1
Aroclor-1016-5	5.965	5.863	6.063	516.110	500.000	3.2
Aroclor-1260-1	7.078	6.975	7.175	500.590	500.000	0.1
Aroclor-1260-2	7.331	7.229	7.429	488.920	500.000	-2.2
Aroclor-1260-3	7.689	7.585	7.785	502.790	500.000	0.6
Aroclor-1260-4	7.911	7.807	8.007	484.610	500.000	-3.1
Aroclor-1260-5	8.217	8.113	8.313	483.900	500.000	-3.2
Decachlorobiphenyl	9.943	9.834	10.034	49.500	50.000	-1.0
Tetrachloro-m-xylene	4.356	4.255	4.455	54.530	50.000	9.1

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** AECO02
Lab Code: ACE **SDG NO.:** Q3363
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/09/2025 10/09/2025

Client Sample No.: CCAL02 **Date Analyzed:** 10/17/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO114461.D **Time Analyzed:** 15:50

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.740	4.643	4.843	523.670	500.000	4.7
Aroclor-1016-2	4.760	4.662	4.862	523.380	500.000	4.7
Aroclor-1016-3	4.934	4.836	5.036	546.600	500.000	9.3
Aroclor-1016-4	4.976	4.879	5.079	546.950	500.000	9.4
Aroclor-1016-5	5.189	5.092	5.292	534.400	500.000	6.9
Aroclor-1260-1	6.222	6.125	6.325	534.580	500.000	6.9
Aroclor-1260-2	6.409	6.313	6.513	547.830	500.000	9.6
Aroclor-1260-3	6.562	6.466	6.666	492.090	500.000	-1.6
Aroclor-1260-4	7.035	6.939	7.139	522.110	500.000	4.4
Aroclor-1260-5	7.275	7.179	7.379	502.940	500.000	0.6
Decachlorobiphenyl	8.665	8.571	8.771	45.140	50.000	-9.7
Tetrachloro-m-xylene	3.659	3.560	3.760	53.380	50.000	6.8