

### CALIBRATION VERIFICATION SUMMARY

**Lab Name:** Alliance **Contract:** AECO02  
**Lab Code:** ACE **SDG NO.:** Q3363  
**GC Column:** ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 09/23/2025 09/23/2025

**Client Sample No.:** CCAL03 **Date Analyzed:** 10/17/2025  
**Lab Sample No.:** AR1660CCC500 **Data File :** PP075835.D **Time Analyzed:** 09:48

| COMPOUND             | RT     | RT WINDOW |        | CALC<br>AMOUNT(ng) | NOM<br>AMOUNT(ng) | %D   |
|----------------------|--------|-----------|--------|--------------------|-------------------|------|
|                      |        | FROM      | TO     |                    |                   |      |
| Aroclor-1016-1       | 5.801  | 5.708     | 5.908  | 560.820            | 500.000           | 12.2 |
| Aroclor-1016-2       | 5.822  | 5.729     | 5.929  | 574.390            | 500.000           | 14.9 |
| Aroclor-1016-3       | 5.885  | 5.792     | 5.992  | 576.750            | 500.000           | 15.4 |
| Aroclor-1016-4       | 5.982  | 5.889     | 6.089  | 604.370            | 500.000           | 20.9 |
| Aroclor-1016-5       | 6.274  | 6.181     | 6.381  | 592.100            | 500.000           | 18.4 |
| Aroclor-1260-1       | 7.392  | 7.299     | 7.499  | 542.950            | 500.000           | 8.6  |
| Aroclor-1260-2       | 7.644  | 7.551     | 7.751  | 530.010            | 500.000           | 6.0  |
| Aroclor-1260-3       | 8.002  | 7.909     | 8.109  | 578.720            | 500.000           | 15.7 |
| Aroclor-1260-4       | 8.230  | 8.137     | 8.337  | 587.280            | 500.000           | 17.5 |
| Aroclor-1260-5       | 8.556  | 8.464     | 8.664  | 582.040            | 500.000           | 16.4 |
| Decachlorobiphenyl   | 10.426 | 10.334    | 10.534 | 46.550             | 50.000            | -6.9 |
| Tetrachloro-m-xylene | 4.648  | 4.555     | 4.755  | 52.660             | 50.000            | 5.3  |

### CALIBRATION VERIFICATION SUMMARY

**Lab Name:** Alliance **Contract:** AECO02  
**Lab Code:** ACE **SDG NO.:** Q3363  
**GC Column:** ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 09/23/2025 09/23/2025

**Client Sample No.:** CCAL03 **Date Analyzed:** 10/17/2025  
**Lab Sample No.:** AR1660CCC500 **Data File :** PP075835.D **Time Analyzed:** 09:48

| COMPOUND             | RT    | RT WINDOW |       | CALC<br>AMOUNT(ng) | NOM<br>AMOUNT(ng) | %D   |
|----------------------|-------|-----------|-------|--------------------|-------------------|------|
|                      |       | FROM      | TO    |                    |                   |      |
| Aroclor-1016-1       | 4.883 | 4.793     | 4.993 | 558.610            | 500.000           | 11.7 |
| Aroclor-1016-2       | 4.940 | 4.851     | 5.051 | 586.980            | 500.000           | 17.4 |
| Aroclor-1016-3       | 5.060 | 4.971     | 5.171 | 525.120            | 500.000           | 5.0  |
| Aroclor-1016-4       | 5.100 | 5.011     | 5.211 | 574.590            | 500.000           | 14.9 |
| Aroclor-1016-5       | 5.314 | 5.225     | 5.425 | 549.910            | 500.000           | 10.0 |
| Aroclor-1260-1       | 6.342 | 6.255     | 6.455 | 540.530            | 500.000           | 8.1  |
| Aroclor-1260-2       | 6.529 | 6.441     | 6.641 | 545.220            | 500.000           | 9.0  |
| Aroclor-1260-3       | 6.682 | 6.595     | 6.795 | 520.850            | 500.000           | 4.2  |
| Aroclor-1260-4       | 7.151 | 7.065     | 7.265 | 490.770            | 500.000           | -1.8 |
| Aroclor-1260-5       | 7.391 | 7.304     | 7.504 | 514.870            | 500.000           | 3.0  |
| Decachlorobiphenyl   | 8.787 | 8.703     | 8.903 | 51.610             | 50.000            | 3.2  |
| Tetrachloro-m-xylene | 3.788 | 3.696     | 3.896 | 51.430             | 50.000            | 2.9  |