

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** AECO02
Lab Code: ACE **SDG NO.:** Q3363
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 09/23/2025 09/23/2025

Client Sample No.: CCAL04 **Date Analyzed:** 10/17/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP075850.D **Time Analyzed:** 16:18

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.804	5.708	5.908	560.110	500.000	12.0
Aroclor-1016-2	5.826	5.729	5.929	554.310	500.000	10.9
Aroclor-1016-3	5.888	5.792	5.992	539.710	500.000	7.9
Aroclor-1016-4	5.985	5.889	6.089	574.030	500.000	14.8
Aroclor-1016-5	6.277	6.181	6.381	557.720	500.000	11.5
Aroclor-1260-1	7.396	7.299	7.499	506.900	500.000	1.4
Aroclor-1260-2	7.648	7.551	7.751	492.860	500.000	-1.4
Aroclor-1260-3	8.006	7.909	8.109	530.570	500.000	6.1
Aroclor-1260-4	8.234	8.137	8.337	552.390	500.000	10.5
Aroclor-1260-5	8.560	8.464	8.664	558.370	500.000	11.7
Decachlorobiphenyl	10.431	10.334	10.534	44.890	50.000	-10.2
Tetrachloro-m-xylene	4.652	4.555	4.755	51.250	50.000	2.5

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** AECO02
Lab Code: ACE **SDG NO.:** Q3363
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 09/23/2025 09/23/2025

Client Sample No.: CCAL04 **Date Analyzed:** 10/17/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP075850.D **Time Analyzed:** 16:18

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.885	4.793	4.993	585.760	500.000	17.2
Aroclor-1016-2	4.943	4.851	5.051	574.070	500.000	14.8
Aroclor-1016-3	5.062	4.971	5.171	569.370	500.000	13.9
Aroclor-1016-4	5.103	5.011	5.211	571.670	500.000	14.3
Aroclor-1016-5	5.317	5.225	5.425	570.940	500.000	14.2
Aroclor-1260-1	6.346	6.255	6.455	503.720	500.000	0.7
Aroclor-1260-2	6.533	6.441	6.641	498.150	500.000	-0.4
Aroclor-1260-3	6.686	6.595	6.795	476.240	500.000	-4.8
Aroclor-1260-4	7.155	7.065	7.265	420.700	500.000	-15.9
Aroclor-1260-5	7.395	7.304	7.504	432.680	500.000	-13.5
Decachlorobiphenyl	8.792	8.703	8.903	46.660	50.000	-6.7
Tetrachloro-m-xylene	3.791	3.696	3.896	49.640	50.000	-0.7